



MAXXAM ANALYTICS
#1 2080 39 Ave. NE, Calgary, AB
T2E 6P7

maxxam.ca
Toll Free 800-386-7247
Fax 403-219-3673

AMBIENT AIR MONITORING MONTHLY DATA REPORT
PEACE RIVER AREA MONITORING PROGRAM COMMITTEE
THREE CREEKS 842B STATION

JOB #: 8449-2017-09-80-C

September 2017

Prepared for:

PEACE RIVER AREA MONITORING PROGRAM COMMITTEE

Attention: LILY LIN

DATE: **October 27, 2017**

Prepared by:

Wunmi Adekanmbi, M.Sc., EPt.

Project Manager, Customer Service, Air Services

Reviewed by:

Cheri Sinclair, B.Sc.

Supervisor, Customer Service, Air Services

SUMMARY

In September 2017, Maxxam Analytics was contracted to manage the ambient air quality monitoring and maintenance activities at the Three Creeks 842b Station, near Peace River Oil Sands Area 2, Alberta. The monitoring station provides continuous meteorological measurements and air quality data for compliance parameters, as requested by the PRAMP Committee.

All data collected this month was compliant with the requirements outlined in the Air Monitoring Directive (Alberta Environment and Parks, 2016).

The operational time for all continuous ambient air analyzers, meteorological systems and data acquisition systems, with the exception of THC/CH₄/NMHC, were above the 90% requirement.

THC/CH₄/NMHC: 220 hours of downtime were recorded this month. Operational time was 69.4%.

- The analyzer malfunctioned on September 15, troubleshooting was completed, calibrations were attempted and the analyzer was subsequently replaced on September 18. Eighty-three hours of downtime were recorded.
- Fifteen hours of downtime were recorded on September 20 due to maintenance activities and additional quality checks performed to address an unstable span response.
- In response to a "low concentration" alarm on Sept 26, a shut-down calibration was attempted on Sept 27, but failed. The analyzer was replaced with the newly-repaired analyzer, originally removed on Sept 15. Data was invalidated to the point when a decline in concentrations was identified. 122 hours of downtime were incurred.
- The 90% operational time requirement was not met this month. AEP reference number: 329861.

The summary of results is presented on the following pages.

Any deviations or modifications made to the sampling or analytical methods are outlined in Section 1.0, Discussion. On this basis, Maxxam Analytics is issuing this completed report to Peace River Area Monitoring Program Committee, Three Creeks 842b Station.

Should you have any questions concerning the results or if we can be of further assistance, please contact us at 403-219-3661 or toll-free at 1-800-386-7247.

Monthly Continuous Data Summary

Peace River Area Monitoring Program Committee Three Creeks 842b Station					MAXIMUM VALUES							OPERATIONAL TIME (%)	
									1-HOUR				
					MONTHLY AVERAGE	READING	DAY	HOUR	WIND SPEED (kph)	WIND DIRECTION (sector)	READING		
PARAMETER	OBJECTIVES		EXCEEDANCES			1-hr	24-hr	1-hr	24-hr	READING	DAY		
SO ₂ (ppb)	172	48	0	0	0	1	7	9	14.2	SSW	0	1	100.0
TRS (ppb)	-	-	-	-	0.18	0.73	8	7	6.5	WNW	0.32	8	100.0
THC (ppm)	-	-	-	-	1.99	2.27	29	23	3	E	2.06	4	69.4
CH ₄ (ppm)	-	-	-	-	1.99	2.27	29	23	3	E	2.06	4	69.4
NMHC (ppm)	-	-	-	-	0.00	0.00	1	0	8.9	S	0.00	1	69.4
RELATIVE HUMIDITY (%)	-	-	-	-	70	96	2	22	5	SSE	86	19	100.0
BAROMETRIC PRESSURE (millibar)	-	-	-	-	942	958	4	5	2	E	956	4	100.0
AMBIENT TEMPERATURE (°C)	-	-	-	-	11.1	29.0	6	16	6	SSW	18.8	7	100.0
STATION TEMPERATURE (°C)	-	-	-	-	21.9	25.8	27	16	9.4	WSW	22.8	26	100.0
VECTOR WS (kph)	-	-	-	-	3.4	31.6	11	16	-	WSW	17.4	10	100.0
VECTOR WD (sec)	-	-	-	-	205 (SSW)	-	-	-	-	-	-	-	100.0

SOUR GAS PROCESSING INDUSTRY
MONTHLY REPORT SUMMARY

Three Creeks 842b Station

Peace River Area Monitoring Program Committee

Plant Name / Location

Company

Licence Number	Report Date	
	YEAR	MONTH
N/A	2017	September

CONTINUOUS AMBIENT MONITORING					
PARAMETER	% TIME OPERATIONAL	ONE - HOUR AVERAGE		24 - HOUR AVERAGE	
		MAXIMUM VALUES	NO. READINGS > REGULATION	MAXIMUM VALUES	NO. READINGS > REGULATION
SO ₂	100.0	0.001 ppm	0	0.000 ppm	0
TRS	100.0	0.001 ppm	-	0.000 ppm	-
THC	69.4	2.27 ppm	-	2.06 ppm	-
CH ₄	69.4	2.27 ppm	-	2.06 ppm	-
NMHC	69.4	0.00 ppm	-	0.00 ppm	-
RH	100.0	96 %	-	86 %	-
BP	100.0	958 mb	-	956 mb	-
Ambient TPX	100.0	29.0 °C	-	18.8 °C	-
Station TPX	100.0	25.8 °C	-	22.8 °C	-
Wind Speed	100.0	31.6 kph	-	17.4 kph	-
Wind Direction	100.0	-	-	-	-

SIGNATURE OF COMPANY REPRESENTATIVE

FOR ALBERTA ENVIRONMENT USE ONLY

Exceedance Summary Report

SO₂ 1-Hour Exceedances

Measured concentrations of sulphur dioxide were below the 1-hour AAAQO of 172 ppb.

SO₂ 24-Hour Exceedances

Measured concentrations of sulphur dioxide were below the 24-hour AAAQO of 48.0 ppb.

In accordance with EPEA and the Substance Release Regulation.

In accordance with A Guide to Release Reporting and the Alberta Ambient Air Quality Objectives and Guidelines Summary.

TABLE OF CONTENTS

<u>Title</u>		<u>Page</u>
SUMMARY		1
MONTHLY CONTINUOUS DATA SUMMARY REPORT		2
SOUR GAS SUMMARY REPORT		3
EXCEEDANCE SUMMARY REPORT		4
TABLE OF CONTENTS		5
1.0 Discussion		7
2.0 Project Personnel		9
3.0 Plant Monthly Required AMD Summary		9
4.0 Calculations and Results		9
5.0 Methods and Procedures		10
 Appendix I	Continuous Monitoring Data Results	14
	Sulphur Dioxide	15
	Total Reduced Sulphur	23
	Total Hydrocarbon	31
	Methane	39
	Non-Methane Hydrocarbon	47
	Wind Speed	55
	Wind Direction	62
	Relative Humidity	65
	Barometric Pressure	68
	Ambient Temperature	71
	Station Temperature	74
 Appendix II	Equipment Calibration Results	77
	Sulphur Dioxide	78
	Total Reduced Sulphur	81
	Total Hydrocarbon	84
	Wind System	94
	Calibrators	96

	Calibration Gases	100
Appendix III	Report Certification Form	104
Appendix IV	Data Validation Certification Form	106

1.0 Discussion

This monthly report consists of continuous monitoring results for the following parameters: Sulphur Dioxide (SO₂), Total Reduced Sulphur (TRS), Total Hydrocarbon (THC), Methane (CH₄), Non-Methane Hydrocarbon (NMHC), Relative Humidity (RH), Barometric Pressure (BP), Ambient Temperature (AmbTPX), Station Temperature (StnTPX), Wind Speed (WS) and Wind Direction (WD).

Sample filters for all continuous air monitors are changed before the calibration begins. The sample manifold is cleaned during the site visit each month.

Control checks, consisting of a zero and span, are conducted daily on all continuous air monitors. In place of the air sample, zero air (from scrubbed air or gas cylinders) is used for zero checks, and a known concentration of the pollutant being analyzed is used for span checks. These checks are controlled by automatic timers and valves. The total zero span cycle is completed within an hour, the commencement of the zero span cycle is at the beginning of the hour.

Multipoint calibrations are done a minimum of once a month for each continuous air monitor. An additional calibration is required under the following conditions: 1) within three days after the initial start-up and stabilization of a newly installed instrument, 2) prior to shut-down or moving of an instrument which has been working to specification, and 3) when major repair has been done on the instrument.

Time during the first multi-point calibration is not considered downtime (Data is flagged as C). If more than one calibration is performed during the month, the time during the additional calibration is considered as downtime (Data is flagged as C1).

Only one zero/span check is run per day. Time during the zero/span check is not considered as downtime (Data is flagged as S). If an extra zero/span check is performed, the time during the additional check is considered as downtime (Data is flagged as S1).

The AMD requires each instrument and accompanying data recording system to be operational 90% of the time, at a minimum, for each monthly monitoring period.

All sampling, analysis, and QA/QC for this project was performed by Maxxam Analytics and complies with the Alberta Air Monitoring Directive.

Data contained in this monthly report has undergone the verification and validation based on the requirements of the AMD Chapter 6: [Ambient Data Quality \(December, 2016\)](#). The descriptions of the data verification and validation process can be found in Section 5 of this report. Instantaneous data, where applicable, is provided for reference purposes and has not undergone zero correction. The minimum and maximum statistics are highlighted in the data table and are for reference only. The highlighted cells are based on the software's interpretation of the exact position of the minimum or maximum value. The visual presentation of these statistics may not be the obvious choice in a data range due to rounding, truncating or analyzer specifications.

Hourly/minute data have been reviewed based on daily zero/span results and multi-point calibration results. Data may be considered invalid if a zero-corrected span check in excess of +/- 10% of the span concentration (established by the previous multi-point calibration) is encountered and/or significant differences in the calibration factor occurs (greater than 10%).

SULPHUR DIOXIDE (SO₂)

- Operational time, for the monitoring period was 100%.
- The routine monthly calibration was performed on September 15.
- One instance of maximum instantaneous data was discarded on September 30 at 17:00 as it was considered an anomalous spike.

TOTAL REDUCED SULPHUR (TRS)

- Operational time, for the monitoring period was 100%.
- The routine monthly calibration was performed on September 15.
- Maximum instantaneous data collected after the calibration on September 15 appear to trend lower than hourly data collected at the same period. This is because for instantaneous data, negative readings were corrected to zero but baseline zero correction was not applied.

TOTAL HYDROCARBONS (THC), METHANE (CH₄) and NON-METHANE HYDROCARBONS (NMHC)

- Operational time, for the monitoring period was 69.4%, equivalent to 220 hours of downtime.
- Upon arrival at the station on September 15 for a scheduled monthly calibration, the analyzer (Thermo 55i, s/n: 1236656188) was found recording anomalous concentrations. Troubleshooting was completed and it was determined that the internal switching valve had failed. Data review revealed that this malfunction occurred after the daily zero/span check completed earlier that day. As these are very delicate components that should not be repaired in the field, arrangements were made to mobilize an alternate analyzer to the station. A replacement analyzer (Thermo 55i, s/n: 1505664392) was installed on September 16 and column conditioning was performed overnight. A post-repair calibration was attempted on September 17, however, the Methane component of the calibration did not stabilize at zero. The problem was traced to a slightly unstable baseline signal and changes were made to the processing to address this. A successful post-repair calibration was completed on September 18. The expected span value was updated after the daily zero/span check on September 19. Data was invalidated back to the last valid zero/span check on September 15. Eighty-three hours of downtime were incurred due to this event.
- Instability was noted in Methane's span response as evidenced by a sudden drift in the results of the scheduled and additional span checks of September 20. A shut-down calibration attempt later that day proved abortive, as a 15-minute stabilization was not achieved at as-found high point, due to intermittent anomalous injections. The actuator operation and alignment were checked, gas supply pressures were verified and adjusted and the zero chromatograph was reset. A successful post-repair calibration was subsequently completed. Data was invalidated to the point when anomalous injections were identified, determined to be hour 01:00 of September 20. Fifteen hours of downtime were recorded due to this event.
- In response to a "low-concentration" alarm triggered on September 26 at hour 23:00, a technician was dispatched to the site on September 27. A shut-down calibration attempt proved unsuccessful. The analyzer was therefore removed and the original analyzer (removed for maintenance on September 15 and now repaired) was re-installed. Column conditioning was performed overnight and a successful installation calibration was completed on September 28. Data was invalidated to the point when a decline in concentrations was identified, determined to be hour 07:00 of September 23. 122 hours of downtime were recorded due to this event.
- The 90% operational time requirement was not met this month. AEP reference number: 329861.

WIND SPEED (WS) and WIND DIRECTION (WD)

- Operational time, for the monitoring period was 100%.
- Wind data is reported as vector wind speed and vector wind direction. Wind direction is defined as the direction from which the wind is blowing from and is measured in degrees from true north.

RELATIVE HUMIDITY (RH)

- Operational time, for the monitoring period was 100%.

BAROMETRIC PRESSURE (BP)

- Operational time, for the monitoring period was 100%.

AMBIENT TEMPERATURE (AmbTPX)

- Operational time, for the monitoring period was 100%.

STATION TEMPERATURE (StnTPX)

- Operational time, for the monitoring period was 100%.

2.0 Project Personnel

Karla Reesor was the contact for Peace River Area Monitoring Program Committee and the Maxxam field technicians were Raja Ashraf and Christopher Wesson.

3.0 Plant Monthly Required AMD Summary

All data collected this month was compliant with the requirements outlined in the Air Monitoring Directive (Alberta Environment and Parks, 2016).

The operational time for all continuous ambient air analyzers, meteorological systems and data acquisition systems, with the exception of THC/CH₄/NMHC, were above the 90% requirement.

4.0 Calculations and Results

All calculations and reporting of results follow the methods described in the AMD, 2016.

5.0 Methods and Procedures

The following methods and procedures were used to complete the monitoring program:

Maxxam AIR SOP-00001: Methane, Non-Methane Hydrocarbon Analyzer Monitoring

Maxxam AIR SOP-00208: RM Young Wind Monitor Calibration

Maxxam AIR SOP-00209: Ambient Sulphur Monitoring

There were no deviations from the prescribed methods.

The following instruments were used to perform the test program:

Sulphur Dioxide - API 100A UV Fluorescent Analyzer

Total Reduced Sulphur - Thermo 43i-TLE UV Fluorescent Analyzer

Methane, Non-Methane Hydrocarbon - Thermo 55i FID Analyzer

Wind System - RM Young Unit

Relative Humidity - RM Young Unit

Barometric Pressure - Met One Unit

Ambient Temperature - RM Young Unit

Datalogger - ESC 8832

The following steps were used to complete the data verification and validation process:

Level 0 Preliminary Verification

Level 0 data are raw data obtained directly from the data acquisition system (DAS). Under the step of Level 0, these data undergo a certain amount of manual or automated screening and flagging. It included a) identification of periods of missing data; b) verification of time stamps against reference time; c) verification that instrument diagnostics/datalogger flags indicate normal operation; d) comparison of data to upper and lower limits; e) rate of change flagging indicating that data changed too rapidly or not at all; and f) verification that zero, span and multipoint performance checks are within specifications. This level of verification is performed on a daily basis.

Level 1 Primary Validation

Validation actions under the step of Level 1 include a) review of all screening flags assigned during preliminary verification; b) review of all supporting site information and documentation; c) review of operational acceptance limits for each parameter/analyser; d) review of daily zero/span and monthly calibration results for all gaseous parameters; and e) application of any necessary adjustments to data (e.g. baseline adjustments, below zero adjustments). This level of validation is performed on a monthly basis.

Level 2 Final Validation


The purpose of Level 2 validation is to verify that there are no inconsistencies among related data, or among regional data measured at nearby sites.

Level 3 Independent Data Review

Level 3 validation is the last step of data review, and it is completed by an individual that is independent of both field operations and primary data validation. A final independent QA review and endorsement is performed during this step before data is submitted to Alberta Environment.

Post-Final Validation

The Post-Final Validation step serves to re-evaluate the data that errors or omissions are discovered and/or suspected after the initial submittal of data. Any data issues or patterns which were not clear on a monthly basis are highlighted during this step. This validation is performed on an annual basis.



Source: Air Monitoring Directive (December 2016), Chapter 6, Ambient Data Quality; Figure 1 Data Collection and Management Process Flow Chart

APPENDIX I
CONTINUOUS MONITORING DATA RESULTS

SULPHUR DIOXIDE

SULPHUR DIOXIDE Hourly Averages (SO₂ ppb)

	HR START (MST)	0:00	1:00	2:00	3:00	4:00	5:00	6:00	7:00	8:00	9:00	10:00	11:00	12:00	13:00	14:00	15:00	16:00	17:00	18:00	19:00	20:00	21:00	22:00	23:00	DAILY MIN.	DAILY MAX.	24-HR AVG.	RDGS.	
	HR END (MST)	0:59	1:59	2:59	3:59	4:59	5:59	6:59	7:59	8:59	9:59	10:59	11:59	12:59	13:59	14:59	15:59	16:59	17:59	18:59	19:59	20:59	21:59	22:59	23:59					
DAY																														
1		0	0	0	0	0	S	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	24
2		0	0	0	0	0	S	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	24
3		0	0	0	0	0	S	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	24
4		0	0	0	0	0	S	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	24
5		0	0	0	0	0	S	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	24
6		0	0	0	0	0	S	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	24
7		0	0	0	0	0	S	0	0	0	0	1	1	1	1	1	1	0	0	0	0	0	0	0	0	0	1	0	24	
8		0	0	0	0	0	S	1	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	1	0	24
9		0	0	0	0	0	S	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	24
10		0	0	0	0	0	S	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	24
11		0	0	0	0	0	S	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	1	0	0	1	0	24	
12		0	0	0	0	0	S	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	24	
13		0	0	0	0	0	S	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	24	
14		0	0	0	0	0	S	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	24	
15		0	0	0	0	0	S	0	0	0	C	C	C	C	C	C	0	0	0	0	0	0	0	0	0	0	0	0	0	24
16		0	1	1	0	0	S	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	1	0	0	1	0	24		
17		1	1	0	0	0	S	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	1	0	24	
18		0	0	0	0	0	S	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	24	
19		0	0	0	0	0	S	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	24	
20		0	0	0	0	0	S	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	24	
21		0	0	0	0	0	S	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	24	
22		0	0	0	0	0	S	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	24	
23		0	0	0	0	0	S	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	24	
24		0	0	0	0	0	S	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	24	
25		0	0	0	0	0	S	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	24	
26		0	0	0	0	0	S	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	24	
27		0	0	0	0	0	S	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	1	0	0	1	0	24		
28		0	0	0	0	0	S	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	1	0	24		
29		1	0	0	0	0	S	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	1	0	24		
30		0	0	0	0	0	S	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	24		
HOURLY MAX		1	1	1	0	0	NA	1	0	0	1	1	1	1	1	1	1	0	1	0	0	1	0	1	1	0				
HOURLY AVG		0	0	0	0	0	NA	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	24			

STATUS FLAG CODES


C	- MONTHLY CALIBRATION	Q	- QUALITY ASSURANCE
C1	- REPEAT CALIBRATION	R	- RECOVERY
Y	- MAINTENANCE	X	- MACHINE MALFUNCTION
S	- DAILY ZERO/SPAN CHECK	G	- OUT FOR REPAIR
S1	- REPEAT ZERO/SPAN CHECK	P	- POWER FAILURE

OBJECTIVE LIMIT:

ALBERTA ENVIRONMENT: 1-HR 172 ppb 24-HR 48 ppb

MONTHLY SUMMARY

NUMBER OF 1-HR EXCEEDANCES:	0
NUMBER OF 24-HR EXCEEDANCES:	0
NUMBER OF NON-ZERO READINGS:	18
MINIMUM 1-HR AVERAGE	0 ppb @ HOUR
MAXIMUM 1-HR AVERAGE:	1 ppb @ HOUR
MAXIMUM 24-HR AVERAGE:	0 ppb
I2S CALIBRATION TIME:	30 hrs
MONTHLY CALIBRATION TIME:	6 hrs
AMD OPERATION UPTIME:	100.0 %
STANDARD DEVIATION:	0
MONTHLY AVERAGE:	0 ppb






PEACE RIVER AREA MONITORING PROGRAM COMMITTEE

Three Creeks 842b Station - September 2017

SULPHUR DIOXIDE Hourly Averages (SO₂ ppb)





PEACE RIVER AREA MONITORING PROGRAM COMMITTEE

Three Creeks 842b Station - September 2017

SULPHUR DIOXIDE Instantaneous Maximum (SO₂ ppb)

	HR START (MST)	0:00	1:00	2:00	3:00	4:00	5:00	6:00	7:00	8:00	9:00	10:00	11:00	12:00	13:00	14:00	15:00	16:00	17:00	18:00	19:00	20:00	21:00	22:00	23:00	DAILY MIN.	DAILY MAX.	24-HR AVG.	RDGS.	
	HR END (MST)	0:59	1:59	2:59	3:59	4:59	5:59	6:59	7:59	8:59	9:59	10:59	11:59	12:59	13:59	14:59	15:59	16:59	17:59	18:59	19:59	20:59	21:59	22:59	23:59					
DAY																														
1		1	1	1	1	1	S	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	24
2		1	1	1	1	1	S	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	24
3		1	1	1	1	1	S	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	24
4		1	1	1	1	1	S	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	24
5		1	1	1	1	1	S	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	24
6		1	1	1	1	1	S	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	24
7		1	1	1	1	1	S	1	1	2	2	2	2	2	2	2	1	1	1	1	1	1	1	1	2	1	1	1	24	
8		1	1	1	1	1	S	2	2	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	24
9		1	1	1	1	1	S	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	24
10		1	1	1	1	1	S	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	24
11		1	1	1	1	1	S	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	2	1	1	1	24	
12		1	1	1	1	1	S	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	24	
13		1	1	1	1	1	S	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	24	
14		1	1	1	1	1	S	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	24	
15		1	1	1	1	1	S	1	1	1	C	C	C	C	C	C	1	1	0	1	1	1	1	0	1	1	1	24		
16		1	1	1	1	1	S	1	1	1	1	1	1	0	1	1	1	1	1	1	1	1	1	1	0	1	1	1	24	
17		1	1	1	1	1	S	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	24	
18		1	1	1	1	1	S	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	24	
19		1	1	1	1	1	S	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	24	
20		1	1	1	1	1	S	1	1	1	1	1	1	1	1	1	1	1	0	1	1	1	1	1	1	0	1	1	24	
21		0	1	0	0	1	S	0	0	0	1	0	0	0	1	0	1	0	0	1	0	0	0	1	0	1	0	24		
22		0	0	0	1	0	S	1	1	1	0	0	0	0	1	1	1	1	1	1	1	0	1	1	0	1	1	24		
23		1	1	1	1	1	S	0	1	1	0	0	0	0	1	0	1	1	1	1	1	1	1	1	0	1	1	24		
24		1	1	1	1	0	S	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	0	1	1	24		
25		1	1	1	1	1	S	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	24		
26		1	1	0	0	0	S	1	1	1	1	1	0	0	1	1	1	1	1	1	1	1	1	1	0	1	1	24		
27		1	1	0	1	0	S	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	0	1	1	24		
28		1	1	1	1	1	S	1	1	1	1	0	0	0	0	0	0	1	1	1	1	0	1	1	0	1	1	24		
29		1	1	0	0	0	S	0	0	0	1	0	1	1	1	1	1	1	1	0	0	0	1	1	0	1	1	24		
30		1	0	1	1	0	S	0	0	0	0	0	0	0	0	1	1	X	0	0	0	0	0	0	0	1	0	23		
HOURLY MAX		1	1	1	1	1	NA	2	2	2	2	2	2	2	2	2	2	1	1	1	1	1	1	2	1					
HOURLY AVG		1	1	1	1	1	NA	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1			

STATUS FLAG CODES

C	- MONTHLY CALIBRATION	Q	- QUALITY ASSURANCE
C1	- REPEAT CALIBRATION	R	- RECOVERY
Y	- MAINTENANCE	X	- MACHINE MALFUNCTION
S	- DAILY ZERO/SPAN CHECK	G	- OUT FOR REPAIR
S1	- REPEAT ZERO/SPAN CHECK	P	- POWER FAILURE

MONTHLY SUMMARY


NUMBER OF NON-ZERO READINGS:	610
MAXIMUM INSTANTANEOUS VALUE:	2 ppb
@ HOUR	8
ON DAY	7
I2S CALIBRATION TIME:	30 hrs
MONTHLY CALIBRATION TIME:	6 hrs
STANDARD DEVIATION:	0
OPERATIONAL TIME:	719 hrs



PEACE RIVER AREA MONITORING PROGRAM COMMITTEE

Three Creeks 842b Station - September 2017

SULPHUR DIOXIDE Instantaneous Maximum (SO₂ ppb)



Wind: PRAMP_842
Poll.: PRAMP_842-SO2[ppb]
Monthly: 17/09
Type: PollutionRose
Direction: Blowing From (Wind Frequency)
Based On 1 Hr.

Calm:


5.12%

Calm Avg: 0.01 [ppb]


Direction	0-3	3-10	10-85	85-170	>170.0	Total
N	0.4	0.0	0.0	0.0	0.0	0.4
NE	3.7	0.0	0.0	0.0	0.0	3.7
E	9.8	0.0	0.0	0.0	0.0	9.8
SE	23.1	0.0	0.0	0.0	0.0	23.1
S	18.6	0.0	0.0	0.0	0.0	18.6
SW	23.3	0.0	0.0	0.0	0.0	23.3
W	10.8	0.0	0.0	0.0	0.0	10.8
NW	5.3	0.0	0.0	0.0	0.0	5.3
Summary	94.9	0.0	0.0	0.0	0.0	94.9

%	Icon	Classes (ppb)	95	0-3	0	3-10	0	10-85	0	85-170	0	>170.0
---	------	---------------	----	-----	---	------	---	-------	---	--------	---	--------

PRAMP_842 Poll.: PRAMP_842-SO2[ppb] 2017/09/01 00:00 - 2017/09/30 23:00 Calm: 5.12% Calm Poll Avg: 0.01[ppb]



SO₂[ppb] Calibration: PRAMP_842 Monthly: 17/09 Type: Span



TOTAL REDUCED SULPHUR




PEACE RIVER AREA MONITORING PROGRAM COMMITTEE

Three Creeks 842b Station - September 2017

TOTAL REDUCED SULPHUR Hourly Averages (TRS ppb)

	HR START (MST)	0:00	1:00	2:00	3:00	4:00	5:00	6:00	7:00	8:00	9:00	10:00	11:00	12:00	13:00	14:00	15:00	16:00	17:00	18:00	19:00	20:00	21:00	22:00	23:00	DAILY MIN.	DAILY MAX.	24-HR AVG.	RDGS.	
	HR END (MST)	0:59	1:59	2:59	3:59	4:59	5:59	6:59	7:59	8:59	9:59	10:59	11:59	12:59	13:59	14:59	15:59	16:59	17:59	18:59	19:59	20:59	21:59	22:59	23:59					
DAY																														
1	0.20	0.28	0.20	0.17	0.19	S	0.21	0.18	0.15	0.13	0.12	0.11	0.11	0.12	0.10	0.11	0.12	0.13	0.14	0.14	0.18	0.16	0.18	0.10	0.28	0.15	24			
2	0.22	0.20	0.23	0.23	0.25	S	0.31	0.35	0.32	0.25	0.29	0.28	0.25	0.22	0.22	0.23	0.18	0.18	0.18	0.19	0.23	0.24	0.18	0.35	0.24	24				
3	0.27	0.24	0.25	0.23	0.26	S	0.38	0.34	0.27	0.23	0.16	0.12	0.12	0.12	0.11	0.13	0.12	0.09	0.11	0.11	0.15	0.16	0.14	0.14	0.09	0.38	0.18	24		
4	0.17	0.15	0.16	0.13	0.17	S	0.22	0.19	0.22	0.19	0.12	0.14	0.18	0.17	0.13	0.14	0.12	0.13	0.15	0.18	0.21	0.17	0.14	0.14	0.12	0.22	0.16	24		
5	0.14	0.14	0.14	0.14	0.15	S	0.23	0.25	0.23	0.19	0.20	0.28	0.23	0.17	0.16	0.14	0.14	0.17	0.20	0.19	0.21	0.22	0.20	0.14	0.28	0.19	24			
6	0.21	0.24	0.27	0.19	0.16	S	0.18	0.18	0.26	0.22	0.20	0.22	0.23	0.25	0.22	0.21	0.15	0.15	0.17	0.24	0.22	0.18	0.20	0.15	0.15	0.27	0.20	24		
7	0.15	0.15	0.16	0.16	0.16	S	0.19	0.19	0.20	0.23	0.30	0.26	0.25	0.24	0.24	0.26	0.29	0.32	0.30	0.31	0.32	0.32	0.35	0.15	0.35	0.24	24			
8	0.35	0.40	0.38	0.39	0.42	S	0.45	0.73	0.43	0.31	0.39	0.34	0.29	0.25	0.23	0.21	0.19	0.18	0.22	0.26	0.27	0.22	0.19	0.18	0.18	0.73	0.32	24		
9	0.21	0.18	0.17	0.20	0.17	S	0.22	0.18	0.17	0.18	0.18	0.19	0.23	0.22	0.19	0.18	0.20	0.19	0.19	0.19	0.20	0.19	0.21	0.17	0.23	0.19	24			
10	0.20	0.23	0.21	0.19	0.22	S	0.29	0.33	0.32	0.29	0.27	0.25	0.24	0.23	0.22	0.21	0.20	0.21	0.23	0.22	0.24	0.23	0.25	0.19	0.33	0.24	24			
11	0.24	0.23	0.25	0.31	0.26	S	0.21	0.19	0.19	0.22	0.24	0.24	0.16	0.14	0.15	0.15	0.12	0.12	0.12	0.14	0.15	0.16	0.39	0.19	0.12	0.39	0.20	24		
12	0.20	0.28	0.18	0.18	0.22	S	0.26	0.26	0.28	0.28	0.18	0.18	0.17	0.18	0.15	0.15	0.13	0.14	0.14	0.16	0.18	0.19	0.17	0.13	0.28	0.19	24			
13	0.18	0.16	0.15	0.20	0.16	S	0.22	0.21	0.20	0.21	0.20	0.17	0.16	0.15	0.12	0.13	0.12	0.13	0.14	0.35	0.22	0.19	0.26	0.22	0.12	0.35	0.18	24		
14	0.20	0.26	0.37	0.36	0.26	S	0.25	0.23	0.29	0.31	0.22	0.16	0.12	0.20	0.20	0.15	0.12	0.20	0.12	0.11	0.12	0.09	0.12	0.10	0.09	0.37	0.20	24		
15	0.09	0.08	0.11	0.10	S	0.27	0.24	0.20	0.17	0.18	C	C	C	C	C	0.10	0.06	0.06	0.07	0.07	0.08	0.11	0.11	0.06	0.27	0.12	24			
16	0.10	0.11	0.12	0.13	0.14	S	0.18	0.15	0.15	0.12	0.13	0.13	0.12	0.13	0.13	0.14	0.11	0.11	0.14	0.13	0.13	0.13	0.15	0.13	0.10	0.18	0.13	24		
17	0.13	0.14	0.13	0.13	0.14	S	0.16	0.15	0.15	0.14	0.13	0.12	0.12	0.10	0.11	0.09	0.10	0.12	0.12	0.13	0.14	0.13	0.11	0.11	0.09	0.16	0.13	24		
18	0.13	0.13	0.13	0.13	0.14	S	0.19	0.16	0.16	0.15	0.16	0.17	0.16	0.19	0.18	0.19	0.19	0.20	0.19	0.15	0.17	0.17	0.20	0.13	0.20	0.17	24			
19	0.20	0.21	0.20	0.20	0.21	S	0.21	0.21	0.19	0.14	0.13	0.13	0.12	0.13	0.13	0.18	0.24	0.21	0.21	0.20	0.23	0.22	0.21	0.17	0.12	0.24	0.19	24		
20	0.16	0.16	0.14	0.15	0.15	S	0.19	0.15	0.17	0.14	0.14	0.13	0.13	0.12	0.13	0.12	0.11	0.12	0.14	0.15	0.15	0.14	0.13	0.11	0.19	0.14	24			
21	0.14	0.15	0.17	0.15	0.16	S	0.18	0.17	0.16	0.16	0.16	0.15	0.15	0.14	0.14	0.13	0.13	0.14	0.13	0.13	0.12	0.13	0.14	0.12	0.18	0.15	24			
22	0.15	0.16	0.15	0.14	0.15	S	0.20	0.14	0.15	0.15	0.15	0.15	0.14	0.14	0.14	0.14	0.12	0.14	0.14	0.13	0.12	0.13	0.12	0.12	0.20	0.14	24			
23	0.12	0.13	0.12	0.11	0.11	S	0.14	0.12	0.13	0.14	0.15	0.13	0.14	0.13	0.14	0.12	0.14	0.13	0.14	0.14	0.13	0.15	0.15	0.14	0.11	0.15	0.13	24		
24	0.12	0.12	0.10	0.10	0.13	S	0.20	0.20	0.19	0.17	0.16	0.19	0.18	0.17	0.14	0.13	0.15	0.15	0.18	0.24	0.24	0.26	0.26	0.30	0.10	0.30	0.18	24		
25	0.30	0.34	0.32	0.31	0.30	S	0.35	0.29	0.25	0.24	0.21	0.19	0.17	0.15	0.16	0.13	0.12	0.13	0.17	0.23	0.20	0.24	0.21	0.19	0.12	0.35	0.23	24		
26	0.20	0.23	0.27	0.27	0.21	S	0.21	0.19	0.23	0.24	0.20	0.16	0.15	0.15	0.12	0.13	0.13	0.17	0.18	0.19	0.27	0.29	0.30	0.12	0.30	0.20	24			
27	0.32	0.35	0.26	0.24	0.29	S	0.32	0.28	0.30	0.30	0.26	0.20	0.18	0.15	0.14	0.14	0.14	0.16	0.16	0.16	0.16	0.16	0.15	0.15	0.13	0.35	0.21	24		
28	0.16	0.13	0.13	0.14	0.15	S	0.16	0.14	0.15	0.16	0.14	0.14	0.12	0.11	0.09	0.10	0.11	0.12	0.13	0.14	0.12	0.14	0.14	0.14	0.09	0.16	0.13	24		
29	0.14	0.13	0.17	0.18	0.16	S	0.21	0.17	0.21	0.23	0.23	0.22	0.21	0.24	0.15	0.11	0.13	0.13	0.16	0.15	0.14	0.15	0.20	0.22	0.11	0.24	0.18	24		
30	0.20	0.19	0.23	0.22	0.22	S	0.22	0.19	0.20	0.17	0.14	0.14	0.13	0.12	0.11	0.12	0.12	0.11	0.11	0.13	0.12	0.11	0.10	0.10	0.23	0.15	24			
HOURLY MAX		0.35	0.40	0.38	0.39	0.42	NA	0.45	0.73	0.43	0.31	0.39	0.34	0.29	0.26	0.24	0.24	0.26	0.29	0.32	0.35	0.31	0.32	0.39	0.35					
HOURLY AVG		0.19	0.20	0.20	0.19	0.19	NA	0.23	0.23	0.22	0.20	0.19	0.18	0.17	0.17	0.15	0.15	0.15	0.16	0.18	0.17	0.18	0.19	0.18						


24 HR AVERAGES September 2017



MONTHLY SUMMARY

NUMBER OF NON-ZERO READINGS:	685		
MINIMUM 1-HR AVERAGE	0.06	ppb	@ HOUR
MAXIMUM 1-HR AVERAGE:	0.73	ppb	@ HOUR
MAXIMUM 24-HR AVERAGE:	0.32	ppb	ON DAY
I2S CALIBRATION TIME:	30	hrs	
MONTHLY CALIBRATION TIME:	5	hrs	AMD OPERATION UPTIME:
STANDARD DEVIATION:	0.07		100.0 %
		MONTLY AVERAGE:	0.18 ppb

TOTAL REDUCED SULPHUR Hourly Averages (TRS ppb)





PEACE RIVER AREA MONITORING PROGRAM COMMITTEE

Three Creeks 842b Station - September 2017

TOTAL REDUCED SULPHUR Instantaneous Maximum (TRS ppb)

	HR START (MST) HR END (MST)	0:00 0:59	1:00 1:59	2:00 2:59	3:00 3:59	4:00 4:59	5:00 5:59	6:00 6:59	7:00 7:59	8:00 8:59	9:00 9:59	10:00 10:59	11:00 11:59	12:00 12:59	13:00 13:59	14:00 14:59	15:00 15:59	16:00 16:59	17:00 17:59	18:00 18:59	19:00 19:59	20:00 20:59	21:00 21:59	22:00 22:59	23:00 23:59	DAILY MIN.	DAILY MAX.	24-HR AVG.	RDGS.	
DAY																														
1		0.41	0.46	0.36	0.33	0.33	S	0.36	0.30	0.30	0.33	0.28	0.22	0.22	0.25	0.22	0.25	0.25	0.25	0.23	0.28	0.25	0.28	0.30	0.28	0.33	0.22	0.46	0.30	24
2		0.43	0.33	0.33	0.33	0.41	S	0.54	0.54	0.43	0.41	0.41	0.41	0.41	0.41	0.38	0.36	0.36	0.38	0.33	0.28	0.30	0.30	0.33	0.36	0.38	0.28	0.54	0.38	24
3		0.38	0.38	0.38	0.38	0.43	S	0.49	0.49	0.46	0.38	0.30	0.25	0.26	0.25	0.23	0.28	0.25	0.23	0.25	0.28	0.30	0.30	0.30	0.28	0.23	0.49	0.33	24	
4		0.30	0.30	0.33	0.30	0.36	S	0.44	0.33	0.38	0.33	0.28	0.28	0.36	0.38	0.28	0.28	0.25	0.28	0.28	0.30	0.30	0.33	0.30	0.28	0.28	0.25	0.44	0.31	24
5		0.28	0.28	0.28	0.28	0.28	S	0.38	0.41	0.36	0.36	0.30	0.41	0.38	0.30	0.28	0.25	0.28	0.25	0.30	0.33	0.33	0.30	0.25	0.41	0.32	24			
6		0.36	0.36	0.38	0.33	0.30	S	0.28	0.30	0.38	0.36	0.30	0.38	0.38	0.36	0.36	0.33	0.28	0.28	0.33	0.36	0.36	0.30	0.33	0.28	0.38	0.33	24		
7		0.25	0.28	0.30	0.30	0.28	S	0.33	0.30	0.33	0.36	0.43	0.46	0.41	0.38	0.36	0.36	0.36	0.36	0.38	0.41	0.41	0.43	0.41	0.43	0.25	0.46	0.36	24	
8		0.43	0.46	0.49	0.47	0.51	S	0.80	0.98	0.59	0.41	0.46	0.44	0.38	0.36	0.33	0.28	0.25	0.25	0.33	0.33	0.30	0.28	0.25	0.98	0.43	24			
9		0.25	0.28	0.25	0.30	0.25	S	0.30	0.25	0.25	0.25	0.25	0.28	0.30	0.28	0.25	0.25	0.28	0.25	0.28	0.28	0.30	0.25	0.30	0.27	0.30	0.27	24		
10		0.30	0.30	0.33	0.22	0.28	S	0.36	0.41	0.41	0.36	0.33	0.30	0.33	0.30	0.30	0.30	0.28	0.30	0.36	0.33	0.36	0.43	0.22	0.43	0.33	24			
11		0.36	0.36	0.39	0.46	0.38	S	0.36	0.30	0.33	0.33	0.39	0.36	0.33	0.28	0.28	0.30	0.28	0.25	0.28	0.30	0.36	0.46	0.25	0.70	0.35	24			
12		0.46	0.64	0.33	0.33	0.38	S	0.41	0.46	0.43	0.49	0.30	0.33	0.33	0.36	0.30	0.30	0.28	0.28	0.30	0.33	0.36	0.36	0.30	0.28	0.64	0.36	24		
13		0.30	0.38	0.33	0.33	0.28	S	0.38	0.36	0.31	0.30	0.30	0.28	0.30	0.30	0.26	0.25	0.23	0.26	0.36	0.47	0.36	0.30	0.51	0.38	0.23	0.51	0.33	24	
14		0.36	0.38	0.54	0.49	0.41	S	0.39	0.39	0.46	0.46	0.39	0.33	0.28	0.56	0.41	0.30	0.30	0.30	0.30	0.28	0.30	0.30	0.30	0.28	0.56	0.37	24		
15		0.41	0.30	0.33	0.33	0.30	S	0.49	0.39	0.33	0.30	0.30	C	C	C	C	0.17	0.05	0.05	0.07	0.05	0.05	0.05	0.07	0.05	0.49	0.22	24		
16		0.05	0.05	0.05	0.07	0.09	S	0.09	0.05	0.07	0.02	0.05	0.05	0.05	0.05	0.05	0.07	0.02	0.02	0.05	0.07	0.05	0.07	0.07	0.02	0.09	0.05	24		
17		0.05	0.05	0.07	0.05	0.05	S	0.09	0.09	0.07	0.05	0.09	0.07	0.05	0.05	0.07	0.07	0.07	0.07	0.05	0.09	0.05	0.05	0.02	0.09	0.06	24			
18		0.05	0.05	0.05	0.02	0.05	S	0.09	0.09	0.05	0.07	0.09	0.07	0.07	0.12	0.09	0.09	0.12	0.09	0.09	0.12	0.05	0.07	0.07	0.09	0.02	0.12	0.08	24	
19		0.09	0.07	0.09	0.09	0.09	S	0.12	0.09	0.09	0.02	0.02	0.02	0.02	0.00	0.00	0.09	0.12	0.12	0.07	0.07	0.09	0.07	0.02	0.00	0.12	0.07	24		
20		0.05	0.07	0.02	0.02	0.02	S	0.07	0.02	0.05	0.02	0.05	0.02	0.00	0.00	0.00	0.02	0.00	0.02	0.07	0.02	0.02	0.02	0.00	0.07	0.03	24			
21		0.02	0.07	0.09	0.05	0.07	S	0.07	0.07	0.05	0.07	0.09	0.05	0.07	0.07	0.05	0.05	0.05	0.05	0.07	0.09	0.07	0.05	0.09	0.06	24				
22		0.07	0.09	0.07	0.07	0.09	S	0.23	0.07	0.09	0.08	0.09	0.09	0.07	0.05	0.07	0.07	0.07	0.07	0.07	0.05	0.05	0.07	0.05	0.23	0.08	24			
23		0.07	0.07	0.07	0.07	0.05	S	0.07	0.05	0.07	0.09	0.07	0.07	0.07	0.07	0.07	0.07	0.07	0.09	0.09	0.07	0.05	0.07	0.05	0.09	0.07	24			
24		0.07	0.09	0.07	0.05	0.07	S	0.15	0.15	0.15	0.09	0.09	0.12	0.15	0.12	0.07	0.05	0.09	0.05	0.09	0.15	0.17	0.17	0.22	0.05	0.22	0.11	24		
25		0.22	0.25	0.20	0.22	0.17	S	0.28	0.20	0.15	0.17	0.09	0.12	0.09	0.07	0.07	0.07	0.05	0.07	0.15	0.18	0.09	0.20	0.17	0.09	0.05	0.28	0.15	24	
26		0.15	0.17	0.22	0.22	0.17	S	0.17	0.17	0.22	0.17	0.12	0.09	0.09	0.09	0.05	0.07	0.07	0.12	0.12	0.09	0.12	0.28	0.22	0.23	0.05	0.28	0.15	24	
27		0.30	0.28	0.33	0.15	0.22	S	0.30	0.20	0.23	0.25	0.18	0.17	0.09	0.07	0.07	0.07	0.07	0.07	0.07	0.09	0.09	0.07	0.07	0.09	0.07	0.33	0.15	24	
28		0.09	0.07	0.07	0.05	0.09	S	0.07	0.05	0.07	0.07	0.07	0.05	0.02	0.02	0.02	0.02	0.02	0.02	0.05	0.05	0.05	0.05	0.02	0.09	0.05	24			
29		0.05	0.05	0.09	0.09	0.05	S	0.12	0.09	0.12	0.12	0.15	0.15	0.12	0.17	0.07	0.02	0.05	0.05	0.07	0.05	0.07	0.09	0.17	0.02	0.17	0.09	24		
30		0.12	0.12	0.15	0.15	0.12	S	0.12	0.12	0.12	0.12	0.05	0.07	0.05	0.02	0.05	0.02	0.07	0.05	0.02	0.05	0.15	0.07	0.05	0.02	0.15	0.08	24		
HOURLY MAX		0.46	0.64	0.54	0.49	0.51	NA	0.80	0.98	0.59	0.49	0.46	0.46	0.41	0.56	0.41	0.36	0.38	0.38	0.41	0.47	0.41	0.43	0.70	0.46					
HOURLY AVG		0.22	0.23	0.23	0.22	0.22	NA	0.28	0.26	0.25	0.23	0.21	0.20	0.20	0.20	0.17	0.17	0.17	0.16	0.18	0.20	0.19	0.20	0.21	0.20					


STATUS FLAG CODES

C	- MONTHLY CALIBRATION	Q	- QUALITY ASSURANCE
C1	- REPEAT CALIBRATION	R	- RECOVERY
Y	- MAINTENANCE	X	- MACHINE MALFUNCTION
S	- DAILY ZERO/SPAN CHECK	G	- OUT FOR REPAIR
S1	- REPEAT ZERO/SPAN CHECK	P	- POWER FAILURE

MONTHLY SUMMARY

NUMBER OF NON-ZERO READINGS:	679
MAXIMUM INSTANTANEOUS VALUE:	0.98 ppb @ HOUR 7 ON DAY 8
I2S CALIBRATION TIME:	30 hrs
MONTHLY CALIBRATION TIME:	5 hrs
STANDARD DEVIATION:	0.15

TOTAL REDUCED SULPHUR Instantaneous Maximum (TRS ppb)



Wind: PRAMP_842
Poll.: PRAMP_842-TRS[ppb]
Monthly: 17/09
Type: PollutionRose
Direction: Blowing From (Wind Frequency)
Based On 1 Hr.


Calm: 5.11%

Calm Avg: 0.17 [ppb]


Direction	0-1	1-3	3-10	>10.0	Total
N	0.4	0.0	0.0	0.0	0.4
NE	3.7	0.0	0.0	0.0	3.7
E	9.8	0.0	0.0	0.0	9.8
SE	23.1	0.0	0.0	0.0	23.1
S	18.7	0.0	0.0	0.0	18.7
SW	23.2	0.0	0.0	0.0	23.2
W	10.8	0.0	0.0	0.0	10.8
NW	5.3	0.0	0.0	0.0	5.3
Summary	94.9	0.0	0.0	0.0	94.9

% Icon Classes (ppb)	95	0-1	0	1-3	0	3-10	0	>10.0
----------------------	----	-----	---	-----	---	------	---	-------

PRAMP_842 Poll.: PRAMP_842-TRS[ppb] 2017/09/01 00:00 - 2017/09/30 23:00 Calm: 5.11% Calm Poll Avg: 0.17[ppb]




TRS[ppb] Calibration: PRAMP_842 Monthly: 17/09 Type: Span



TOTAL HYDROCARBON


TOTAL HYDROCARBONS Hourly Averages (THC ppm)

	HR START (MST)	0:00	1:00	2:00	3:00	4:00	5:00	6:00	7:00	8:00	9:00	10:00	11:00	12:00	13:00	14:00	15:00	16:00	17:00	18:00	19:00	20:00	21:00	22:00	23:00	DAILY MIN.	DAILY MAX.	24-HR AVG.	RDGS.
	HR END (MST)	0:59	1:59	2:59	3:59	4:59	5:59	6:59	7:59	8:59	9:59	10:59	11:59	12:59	13:59	14:59	15:59	16:59	17:59	18:59	19:59	20:59	21:59	22:59	23:59				
DAY																													
1	1.99	2.00	1.99	1.99	1.99	1.99	S	2.00	1.99	2.00	1.99	1.98	1.99	1.99	2.00	2.00	1.96	1.99	1.99	1.98	1.99	1.99	2.01	1.96	2.01	1.99	24		
2	2.02	2.00	2.00	2.01	2.02	S	2.02	2.02	2.02	1.99	2.00	1.99	1.99	2.00	2.00	1.98	1.96	1.96	1.95	1.93	1.93	1.95	1.96	1.93	2.02	1.99	24		
3	1.97	1.97	1.97	1.97	1.97	S	1.99	1.99	1.98	1.97	1.96	1.96	1.96	1.96	1.97	1.97	1.97	1.97	1.97	1.99	2.00	2.00	1.96	2.00	1.97	24			
4	2.05	2.06	2.10	2.13	2.24	S	2.24	2.24	2.15	2.03	2.00	2.00	2.00	2.00	2.00	2.00	1.99	1.99	2.00	2.02	2.04	2.05	2.02	2.04	1.99	2.24	2.06	24	
5	2.02	2.03	2.02	2.03	2.04	S	2.02	2.05	2.05	2.04	2.04	2.04	2.04	2.03	2.02	2.02	2.00	2.01	2.00	2.00	2.01	2.01	2.02	2.06	2.12	2.00	2.12	2.03	24
6	2.11	2.19	2.24	2.15	2.08	S	2.06	2.05	2.07	2.06	2.04	2.04	2.04	2.03	2.02	2.01	2.01	2.01	2.02	2.03	2.06	2.07	2.05	2.01	2.24	2.06	24		
7	2.05	2.04	2.03	2.03	2.04	S	2.06	2.05	2.05	2.05	2.07	2.07	2.07	2.06	2.04	2.03	2.03	2.04	2.05	2.04	2.04	2.06	2.07	2.07	2.03	2.05	24		
8	2.08	2.08	2.06	2.05	2.07	S	1.99	1.98	1.98	1.99	1.99	1.93	1.96	1.93	1.95	1.95	1.96	1.96	1.95	1.95	1.96	1.96	1.96	1.95	1.93	2.08	1.98	24	
9	1.96	1.98	1.98	1.99	2.00	S	2.03	2.03	2.00	1.98	1.99	2.00	2.00	2.01	2.00	1.99	2.00	2.00	1.99	2.00	2.00	2.00	1.99	1.97	2.03	1.99	24		
10	1.98	1.98	1.97	1.99	1.99	S	2.00	2.00	2.00	2.00	2.00	2.00	2.00	1.99	1.99	2.00	2.00	2.00	1.99	2.00	2.00	2.00	1.99	1.97	2.00	1.99	24		
11	1.98	1.98	1.99	1.99	2.01	S	2.00	1.98	1.99	1.98	2.01	2.00	1.96	1.97	1.98	1.98	1.96	1.97	1.98	1.98	1.98	1.98	1.98	1.96	2.01	1.98	24		
12	1.98	1.98	1.98	2.02	2.01	S	2.01	2.01	2.01	2.00	2.00	1.99	1.98	1.98	1.98	1.98	1.98	1.99	1.99	2.00	2.00	1.99	2.02	1.98	2.02	1.99	24		
13	2.03	2.02	2.01	2.01	2.02	S	2.01	2.00	2.00	2.00	1.99	1.99	2.00	2.00	2.01	2.00	2.01	2.00	2.00	2.03	2.01	2.01	2.02	2.01	1.99	2.03	2.01	24	
14	2.03	2.02	2.03	2.02	2.03	S	2.05	2.07	2.05	2.03	2.03	2.02	2.01	2.01	2.01	2.02	2.01	2.01	2.08	2.04	2.01	2.02	2.01	2.08	2.03	24			
15	2.02	2.02	2.04	2.06	S	X	X	X	X	X	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	2.02	2.06	2.04	6	
16	Y	Y	Y	Y	Y	S	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	-	-	-	-	
17	Y	Y	Y	Y	Y	S	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	-	-	-	-	
18	Y	Y	Y	Y	Y	S	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	C	C	C	2.02	2.04	2.04	1.98	1.98	2.04	2.02	7
19	1.96	1.94	1.94	1.94	1.95	S	1.99	1.98	1.97	1.98	1.98	1.97	1.97	1.97	1.96	1.96	1.96	1.96	1.97	1.97	1.96	1.96	1.96	1.95	1.94	1.99	1.96	24	
20	1.95	1.95	X	X	X	S	1.93	1.93	1.93	1.93	1.93	1.93	1.93	1.93	1.92	1.93	1.92	1.91	1.91	1.91	1.91	1.91	1.90	1.93	1.92	1.87	1.95	2.04	9
21	1.92	1.93	1.92	1.93	1.93	S	1.93	1.93	1.93	1.93	1.93	1.93	1.93	1.92	1.93	1.92	1.91	1.91	1.91	1.91	1.91	1.91	1.91	1.91	1.91	2.00	1.93	24	
22	1.97	1.96	1.97	1.98	2.01	S	2.18	2.12	2.08	1.94	1.91	1.91	1.91	1.92	1.91	1.91	1.91	1.92	1.95	1.97	1.96	1.95	1.91	2.18	1.96	24			
23	1.96	1.96	1.97	1.98	1.98	S	1.98	2.02	X	X	X	X	X	X	X	X	X	X	X	X	X	X	X	X	1.96	2.02	1.98	8	
24	X	X	X	X	X	S	X	X	X	X	X	X	X	X	X	X	X	X	X	X	X	X	X	X	-	-	-	-	
25	X	X	X	X	X	S	X	X	X	X	X	X	X	X	X	X	X	X	X	X	X	X	X	X	-	-	-	-	
26	X	X	X	X	X	S	X	X	X	X	X	X	X	X	X	X	X	X	X	X	X	X	X	X	-	-	-	-	
27	X	X	X	X	X	S	X	X	X	X	X	X	X	X	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	-	-	-	-	
28	Y	Y	Y	Y	Y	S	Y	Y	C1	C1	C1	1.95	1.94	1.93	1.92	1.92	1.92	1.92	1.93	1.95	1.97	2.00	2.01	1.92	2.01	1.95	14		
29	2.01	2.00	2.05	2.13	2.06	S	2.15	2.08	2.01	2.00	2.00	1.99	1.99	1.99	1.96	1.93	1.94	1.94	1.94	1.94	1.95	1.95	1.93	2.27	2.02	24			
30	2.09	1.99	1.97	1.95	1.93	S	1.94	1.95	1.94	1.94	1.93	1.94	1.93	1.93	1.93	1.93	1.91	1.93	1.94	1.95	1.95	1.91	2.09	1.95	24				
	HOURLY MAX	2.11	2.19	2.24	2.15	2.24	NA	2.24	2.24	2.15	2.06	2.07	2.07	2.06	2.04	2.03	2.03	2.04	2.05	2.04	2.08	2.06	2.13	2.27					
	HOURLY AVG	2.01	2.00	2.01	2.02	2.02	NA	2.03	2.03	2.01	1.99	1.99	1.98	1.98	1.98	1.97	1.97	1.97	1.97	1.97	1.98	1.98	1.99	2.00	2.01	1.95	24		



NUMBER OF NON-ZERO READINGS:	476
MINIMUM 1-HR AVERAGE	1.87 ppm @ HOUR
MAXIMUM 1-HR AVERAGE:	2.27 ppm @ HOUR
MAXIMUM 24-HR AVERAGE:	2.06 ppm
ON DAY	20
ON DAY	29
ON DAY	4
I2S CALIBRATION TIME:	21 hrs
OPERATIONAL TIME:	500 hrs
MONTHLY CALIBRATION TIME:	3 hrs
AMD OPERATION UPTIME:	69.4 %
STANDARD DEVIATION:	0.05
MONTHLY AVERAGE:	1.99 ppm

TOTAL HYDROCARBONS Hourly Averages (THC ppm)





PEACE RIVER AREA MONITORING PROGRAM COMMITTEE

Three Creeks 842b Station - September 2017

TOTAL HYDROCARBONS Instantaneous Maximum (THC ppm)

HR START (MST)	0:00	1:00	2:00	3:00	4:00	5:00	6:00	7:00	8:00	9:00	10:00	11:00	12:00	13:00	14:00	15:00	16:00	17:00	18:00	19:00	20:00	21:00	22:00	23:00	DAILY MIN.	DAILY MAX.	24-HR AVG.	RDGS.		
HR END (MST)	0:59	1:59	2:59	3:59	4:59	5:59	6:59	7:59	8:59	9:59	10:59	11:59	12:59	13:59	14:59	15:59	16:59	17:59	18:59	19:59	20:59	21:59	22:59	23:59						
DAY																														
1	2.02	2.02	2.04	2.02	2.02	S	2.02	2.01	2.04	2.01	2.01	2.01	2.01	2.01	2.01	2.01	2.03	2.01	2.01	2.01	2.01	2.01	2.01	2.05	2.05	2.01	2.05	2.02	24	
2	2.07	2.02	2.02	2.02	2.05	S	2.06	2.04	2.04	2.03	2.03	2.03	2.04	2.06	2.03	2.01	2.01	2.01	2.01	2.01	2.00	1.99	2.02	2.02	1.99	2.07	2.03	24		
3	2.04	2.04	2.04	2.05	2.07	S	2.06	2.06	2.04	2.03	2.01	2.02	2.02	2.01	2.01	2.03	2.01	2.01	2.02	2.11	2.14	2.08	2.06	2.07	2.01	2.14	2.04	24		
4	2.09	2.14	2.19	2.23	2.35	S	2.29	2.32	2.21	2.09	2.02	2.22	2.03	2.03	2.02	2.01	2.03	2.01	2.08	2.09	2.07	2.08	2.06	2.07	2.01	2.35	2.12	24		
5	2.04	2.08	2.04	2.07	2.08	S	2.05	2.08	2.07	2.06	2.06	2.05	2.05	2.04	2.03	2.02	2.02	2.04	2.05	2.05	2.05	2.07	2.14	2.18	2.02	2.18	2.06	24		
6	2.21	2.27	2.36	2.28	2.18	S	2.12	2.08	2.10	2.08	2.07	2.06	2.06	2.04	2.11	2.04	2.03	2.03	2.05	2.08	2.09	2.17	2.08	2.03	2.36	2.11	24			
7	2.10	2.06	2.05	2.06	2.05	S	2.14	2.07	2.07	2.08	2.12	2.09	2.09	2.08	2.07	2.04	2.05	2.06	2.06	2.05	2.06	2.09	2.12	2.10	2.04	2.14	2.08	24		
8	2.11	2.09	2.08	2.07	2.11	S	2.03	1.99	2.00	2.01	2.01	1.99	1.98	1.97	1.97	1.97	1.97	1.97	1.98	1.97	1.98	1.98	1.98	1.98	1.97	1.97	2.11	2.01	24	
9	1.99	2.01	2.03	2.02	2.04	S	2.06	2.05	2.03	2.01	2.01	2.03	2.03	2.04	2.02	2.01	2.01	2.01	2.01	2.01	2.00	1.99	1.99	1.99	1.99	2.06	2.02	24		
10	1.99	2.00	2.01	2.02	2.01	S	2.03	2.01	2.04	2.01	2.01	2.01	2.01	2.01	2.01	2.01	2.01	2.01	2.03	2.03	2.05	2.04	1.99	2.18	2.02	24				
11	2.00	2.01	2.02	2.03	2.03	S	2.02	2.01	2.00	2.03	2.03	2.02	2.02	1.98	2.00	2.01	1.98	1.99	1.99	1.99	1.99	2.00	2.01	1.98	2.03	2.01	24			
12	2.01	2.01	2.01	2.17	2.04	S	2.10	2.10	2.03	2.02	2.02	2.09	2.01	2.01	2.01	2.01	2.00	2.00	2.03	2.02	2.03	2.06	2.11	2.06	2.00	2.17	2.04	24		
13	2.08	2.04	2.07	2.03	2.05	S	2.05	2.04	2.02	2.02	2.01	2.01	2.01	2.02	2.27	2.02	2.02	2.05	2.07	2.05	2.04	2.04	2.03	2.01	2.27	2.05	24			
14	2.06	2.04	2.06	2.03	2.06	S	2.11	2.12	2.11	2.07	2.04	2.23	2.03	2.02	2.06	2.03	2.03	2.06	2.14	2.09	2.03	2.03	2.02	2.23	2.07	24				
15	2.04	2.04	2.05	2.05	2.11	S	X	X	X	X	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	2.04	2.11	2.06	6	
16	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	-	-	-	-	
17	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	-	-	-	-	
18	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	C	C	C	2.06	2.08	2.08	2.05	2.05	2.08	2.07	7		
19	2.03	1.99	1.97	1.97	1.99	S	2.03	2.01	2.00	2.01	2.01	2.00	2.00	2.00	2.00	2.00	2.00	2.00	2.00	2.00	2.00	2.00	2.00	1.99	1.99	1.97	2.03	2.00	24	
20	1.98	1.98	X	X	X	X	X	X	Y	Y	Y	Y	C1	C1	C1	C1	C1	1.90	1.91	1.93	1.93	1.93	1.97	1.95	1.90	1.98	1.94	9		
21	1.94	1.95	1.96	1.95	1.96	S	1.95	2.00	1.97	1.96	1.97	1.97	1.96	1.97	1.96	1.95	1.94	1.93	1.95	1.95	1.97	1.97	2.04	2.05	1.93	2.05	1.97	24		
22	2.03	2.05	2.02	2.02	2.06	S	2.31	2.21	2.22	2.00	1.95	1.94	1.94	1.94	1.94	1.94	1.93	1.93	1.97	1.99	2.01	1.99	1.93	1.93	2.31	2.02	24			
23	1.99	1.99	2.00	2.03	2.04	S	2.02	2.06	X	X	X	X	X	X	X	X	X	X	X	X	X	X	X	X	1.99	2.06	2.02	8		
24	X	X	X	X	X	X	X	X	X	X	X	X	X	X	X	X	X	X	X	X	X	X	X	X	-	-	-	-		
25	X	X	X	X	X	X	X	X	X	X	X	X	X	X	X	X	X	X	X	X	X	X	X	X	-	-	-	-		
26	X	X	X	X	X	X	X	X	X	X	X	X	X	X	X	X	X	X	X	X	X	X	X	X	-	-	-	-		
27	X	X	X	X	X	X	X	X	X	X	X	X	X	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	-	-	-	-		
28	Y	Y	Y	Y	Y	Y	Y	C1	C1	C1	1.97	1.96	1.95	1.95	1.94	1.93	1.94	1.93	1.96	1.97	1.96	2.01	2.02	2.02	1.93	2.02	1.97	14		
29	2.05	2.02	2.20	2.22	2.10	S	2.23	2.17	2.02	2.02	2.03	2.01	2.01	1.99	1.95	1.96	2.03	1.96	1.97	1.99	2.05	2.31	2.35	1.95	2.35	2.07	24			
30	2.30	2.02	2.05	1.98	1.96	S	1.98	2.02	1.99	1.97	1.96	1.96	1.95	1.94	1.93	1.96	1.95	1.92	1.95	1.97	1.97	1.96	1.92	2.30	1.98	24				
HOURLY MAX	2.30	2.27	2.36	2.28	2.35	NA	2.31	2.32	2.22	2.09	2.12	2.22	2.23	2.08	2.27	2.04	2.06	2.18	2.08	2.11	2.14	2.09	2.31	2.35						
HOURLY AVG	2.05	2.04	2.06	2.06	2.06	NA	2.08	2.07	2.05	2.03	2.02	2.03	2.02	2.01	2.02	2.00	2.00	2.00	2.01	2.02	2.03	2.05	2.05							


STATUS FLAG CODES

C	- MONTHLY CALIBRATION	Q	- QUALITY ASSURANCE
C1	- REPEAT CALIBRATION	R	- RECOVERY
Y	- MAINTENANCE	X	- MACHINE MALFUNCTION
S	- DAILY ZERO/SPAN CHECK	G	- OUT FOR REPAIR
S1	- REPEAT ZERO/SPAN CHECK	P	- POWER FAILURE

MONTHLY SUMMARY

NUMBER OF NON-ZERO READINGS:	476
MAXIMUM INSTANTANEOUS VALUE:	2.36 ppm @ HOUR 2 ON DAY 6
I2S CALIBRATION TIME:	21 hrs
MONTHLY CALIBRATION TIME:	3 hrs
STANDARD DEVIATION:	0.07
OPERATIONAL TIME:	500 hrs

TOTAL HYDROCARBONS Instantaneous Maximum (THC ppm)



Wind: PRAMP_842
 Poll.: PRAMP_842-THC55[ppm]
 Monthly: 17/09
 Type: PollutionRose
 Direction: Blowing From (Wind Frequency)
 Based On 1 Hr.


Calm: 6.72%

Calm Avg: 1.99 [ppm]

Direction	0-2	2-3	3-5	5-10	>10.0	Total
N	0.6	0.0	0.0	0.0	0.0	0.6
NE	3.2	0.4	0.0	0.0	0.0	3.6
E	4.0	4.4	0.0	0.0	0.0	8.4
SE	6.1	12.4	0.0	0.0	0.0	18.5
S	9.9	6.7	0.0	0.0	0.0	16.6
SW	15.1	8.6	0.0	0.0	0.0	23.7
W	11.6	2.9	0.0	0.0	0.0	14.5
NW	5.9	1.5	0.0	0.0	0.0	7.4
Summary	56.3	37.0	0.0	0.0	0.0	93.3

% Icon	Classes (ppm)	56	0-2	37	2-3	0	3-5	0	5-10	0	>10.0
--------	---------------	----	-----	----	-----	---	-----	---	------	---	-------

PRAMP_842 Poll.: PRAMP_842-THC55[ppm] 2017/09/01 00:00 - 2017/09/30 23:00 Calm: 6.72% Calm Poll Avg: 1.99[ppm]




THC55[ppm] Calibration: PRAMP_842 Monthly: 17/09 Type: Span



METHANE

METHANE Hourly Averages (CH₄ ppm)

	HR START (MST)	0:00	1:00	2:00	3:00	4:00	5:00	6:00	7:00	8:00	9:00	10:00	11:00	12:00	13:00	14:00	15:00	16:00	17:00	18:00	19:00	20:00	21:00	22:00	23:00	DAILY MIN.	DAILY MAX.	24-HR AVG.	RDGS.	
	HR END (MST)	0:59	1:59	2:59	3:59	4:59	5:59	6:59	7:59	8:59	9:59	10:59	11:59	12:59	13:59	14:59	15:59	16:59	17:59	18:59	19:59	20:59	21:59	22:59	23:59					
DAY																														
1	1.99	2.00	1.99	1.99	1.99	1.99	S	2.00	1.99	2.00	1.99	1.98	1.99	1.99	2.00	2.00	1.96	1.99	1.99	1.98	1.99	1.99	2.01	1.96	2.01	1.99	24			
2	2.02	2.00	2.00	2.01	2.02	S	2.02	2.02	2.02	1.99	2.00	1.99	1.99	2.00	2.00	1.98	1.96	1.96	1.95	1.93	1.93	1.95	1.96	1.93	2.02	1.99	24			
3	1.97	1.97	1.97	1.97	1.97	S	1.98	1.99	1.98	1.97	1.96	1.96	1.96	1.96	1.97	1.97	1.97	1.97	1.97	1.99	2.00	1.99	2.00	1.96	2.00	1.97	24			
4	2.05	2.06	2.10	2.13	2.24	S	2.24	2.24	2.15	2.03	2.00	2.00	2.00	2.00	2.00	2.00	1.99	1.99	1.99	1.99	2.00	2.02	2.04	2.05	2.02	2.03	1.99	2.24	2.06	24
5	2.02	2.03	2.02	2.03	2.03	S	2.02	2.05	2.05	2.03	2.04	2.03	2.03	2.02	2.02	2.00	2.00	2.00	2.00	2.01	2.01	2.02	2.06	2.12	2.00	2.12	2.03	24		
6	2.11	2.19	2.24	2.15	2.08	S	2.06	2.05	2.07	2.05	2.04	2.04	2.03	2.02	2.01	2.01	2.00	2.00	2.02	2.03	2.06	2.07	2.05	2.00	2.24	2.06	24			
7	2.05	2.04	2.03	2.03	2.04	S	2.06	2.05	2.05	2.05	2.07	2.07	2.07	2.06	2.04	2.03	2.03	2.04	2.05	2.04	2.04	2.06	2.07	2.07	2.03	2.05	24			
8	2.08	2.08	2.05	2.07	S	1.99	1.98	1.98	1.99	1.99	1.93	1.95	1.92	1.95	1.95	1.96	1.96	1.95	1.95	1.96	1.96	1.95	1.95	1.95	1.92	2.08	1.98	24		
9	1.96	1.98	1.98	1.99	2.00	S	2.03	2.03	2.00	1.98	1.99	2.00	2.00	2.01	2.00	1.99	2.00	2.00	1.99	2.00	2.00	1.99	1.97	2.00	1.99	1.99	2.03	1.99	24	
10	1.98	1.98	1.97	1.99	1.99	S	2.00	2.00	2.00	2.00	2.00	1.99	1.99	1.99	1.99	2.00	2.00	2.00	1.99	2.00	2.00	2.00	1.99	1.97	2.00	1.99	24			
11	1.98	1.98	1.99	1.99	2.01	S	2.00	1.98	1.99	1.98	2.00	2.00	1.96	1.97	1.98	1.98	1.96	1.97	1.98	1.98	1.98	1.98	1.98	1.96	2.01	1.98	24			
12	1.98	1.98	1.98	2.01	2.01	S	2.01	2.01	2.01	2.00	2.00	1.99	1.98	1.98	1.98	1.98	1.98	1.99	1.99	2.00	2.00	1.99	2.02	1.98	2.02	1.99	24			
13	2.03	2.02	2.01	2.01	2.02	S	2.01	2.00	2.00	2.00	1.99	1.99	2.00	2.00	2.01	2.00	2.01	2.00	2.00	2.03	2.01	2.01	2.02	2.01	1.99	2.03	2.01	24		
14	2.03	2.02	2.03	2.02	2.03	S	2.05	2.07	2.05	2.03	2.03	2.02	2.01	2.01	2.01	2.02	2.01	2.01	2.08	2.04	2.01	2.02	2.01	2.08	2.03	24				
15	2.02	2.02	2.04	2.06	S	X	X	X	X	X	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	2.02	2.06	2.04	6		
16	Y	Y	Y	Y	Y	S	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	-	-	-	-	
17	Y	Y	Y	Y	Y	S	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	-	-	-	-	
18	Y	Y	Y	Y	Y	S	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	C	C	C	2.02	2.04	2.04	1.98	1.98	2.04	2.02	7	
19	1.96	1.94	1.94	1.94	1.95	S	1.99	1.98	1.97	1.98	1.98	1.97	1.97	1.97	1.97	1.96	1.96	1.96	1.97	1.97	1.96	1.96	1.96	1.95	1.94	1.99	1.96	24		
20	1.95	1.95	X	X	X	S	1.93	1.93	1.93	1.93	1.93	1.93	1.93	1.93	1.92	1.93	1.92	1.91	1.91	1.91	1.91	1.91	1.90	1.93	1.92	1.87	1.95	1.91	9	
21	1.92	1.93	1.92	1.93	1.93	S	1.93	1.93	1.93	1.93	1.93	1.93	1.93	1.92	1.93	1.92	1.91	1.91	1.91	1.91	1.91	1.91	1.91	1.91	1.91	2.00	1.93	24		
22	1.97	1.96	1.97	1.98	2.01	S	2.18	2.12	2.08	1.94	1.91	1.91	1.91	1.92	1.91	1.91	1.92	1.92	1.95	1.97	1.96	1.95	1.91	2.18	1.96	24				
23	1.96	1.96	1.97	1.98	1.98	S	1.98	2.02	X	X	X	X	X	X	X	X	X	X	X	X	X	X	X	X	1.96	2.02	1.98	8		
24	X	X	X	X	X	S	X	X	X	X	X	X	X	X	X	X	X	X	X	X	X	X	X	X	-	-	-	-		
25	X	X	X	X	X	S	X	X	X	X	X	X	X	X	X	X	X	X	X	X	X	X	X	X	-	-	-	-		
26	X	X	X	X	X	S	X	X	X	X	X	X	X	X	X	X	X	X	X	X	X	X	X	X	-	-	-	-		
27	X	X	X	X	X	S	X	X	X	X	X	X	X	X	X	Y	Y	Y	Y	Y	Y	Y	Y	Y	-	-	-	-		
28	Y	Y	Y	Y	Y	S	Y	Y	C1	C1	C1	1.95	1.94	1.93	1.92	1.92	1.92	1.92	1.93	1.95	1.97	2.00	2.01	1.92	2.01	1.95	14			
29	2.01	2.00	2.05	2.13	2.06	S	2.15	2.08	2.01	2.00	2.00	1.99	1.99	1.99	1.96	1.93	1.94	1.94	1.94	1.94	1.95	1.95	1.99	2.13	1.93	2.27	2.02	24		
30	2.09	1.99	1.97	1.95	1.93	S	1.94	1.94	1.94	1.94	1.93	1.94	1.93	1.93	1.93	1.93	1.93	1.91	1.93	1.94	1.95	1.95	1.91	2.09	1.95	24				
HOURLY MAX		2.11	2.19	2.24	2.15	2.24	NA	2.24	2.24	2.15	2.05	2.07	2.07	2.06	2.04	2.03	2.03	2.04	2.05	2.04	2.08	2.06	2.13	2.27						
HOURLY AVG		2.01	2.00	2.01	2.02	2.02	NA	2.03	2.03	2.01	1.99	1.99	1.98	1.98	1.98	1.97	1.97	1.97	1.97	1.97	1.98	1.99	2.00	2.01						

24 HR AVERAGES September 2017

MONTHLY SUMMARY


NUMBER OF NON-ZERO READINGS:	476				
MINIMUM 1-HR AVERAGE	1.87	ppm	@ HOUR	17	ON DAY
MAXIMUM 1-HR AVERAGE:	2.27	ppm	@ HOUR	23	ON DAY
MAXIMUM 24-HR AVERAGE:	2.06	ppm			ON DAY
I2S CALIBRATION TIME:	21	hrs			500 hrs
MONTHLY CALIBRATION TIME:	3	hrs			69.4 %
STANDARD DEVIATION:	0.05				MONTHLY AVERAGE:
					1.99 ppm



PEACE RIVER AREA MONITORING PROGRAM COMMITTEE

Three Creeks 842b Station - September 2017

METHANE Hourly Averages (CH₄ ppm)





PEACE RIVER AREA MONITORING PROGRAM COMMITTEE

Three Creeks 842b Station - September 2017


METHANE MAX Instantaneous Maximum (CH₄ ppm)

	HR START (MST)	0:00	1:00	2:00	3:00	4:00	5:00	6:00	7:00	8:00	9:00	10:00	11:00	12:00	13:00	14:00	15:00	16:00	17:00	18:00	19:00	20:00	21:00	22:00	23:00	DAILY MIN.	DAILY MAX.	24-HR AVG.	RDGS.	
	HR END (MST)	0:59	1:59	2:59	3:59	4:59	5:59	6:59	7:59	8:59	9:59	10:59	11:59	12:59	13:59	14:59	15:59	16:59	17:59	18:59	19:59	20:59	21:59	22:59	23:59					
DAY																														
1		2.02	2.01	2.01	2.02	2.01	S	2.02	2.00	2.03	2.01	2.01	2.00	2.00	2.00	2.01	2.00	2.00	2.00	2.00	1.99	2.00	2.01	2.04	2.04	1.99	2.04	2.01	24	
2		2.06	2.01	2.02	2.01	2.04	S	2.05	2.04	2.04	2.02	2.02	2.02	2.03	2.03	2.06	2.03	2.00	2.00	2.01	1.99	2.01	2.02	2.02	1.99	2.06	2.02	24		
3		2.04	2.03	2.04	2.04	2.07	S	2.06	2.05	2.04	2.02	2.01	2.01	2.00	2.01	2.00	2.03	2.00	2.00	2.01	2.11	2.14	2.07	2.05	2.06	2.00	2.14	2.04	24	
4		2.08	2.14	2.17	2.21	2.35	S	2.29	2.32	2.20	2.08	2.01	2.02	2.01	2.02	2.02	2.00	2.02	2.00	2.07	2.08	2.07	2.08	2.06	2.07	2.00	2.35	2.10	24	
5		2.04	2.07	2.03	2.06	2.07	S	2.04	2.07	2.06	2.05	2.05	2.05	2.04	2.03	2.02	2.01	2.03	2.04	2.05	2.06	2.13	2.18	2.01	2.18	2.05	24			
6		2.20	2.26	2.36	2.28	2.17	S	2.12	2.07	2.09	2.07	2.06	2.05	2.06	2.04	2.03	2.02	2.02	2.02	2.04	2.07	2.08	2.16	2.07	2.02	2.36	2.10	24		
7		2.07	2.06	2.04	2.05	2.05	S	2.14	2.06	2.07	2.07	2.12	2.08	2.08	2.07	2.06	2.04	2.05	2.05	2.05	2.05	2.05	2.09	2.12	2.09	2.04	2.14	2.07	24	
8		2.11	2.09	2.08	2.06	2.11	S	2.02	1.99	2.00	2.01	2.01	1.98	1.97	1.96	1.96	1.96	1.97	1.97	1.96	1.97	1.97	1.98	1.97	1.96	2.11	2.00	24		
9		1.99	2.00	2.02	2.02	2.03	S	2.05	2.05	2.02	2.00	2.02	2.03	2.01	2.00	2.01	2.01	2.01	2.01	2.00	1.99	1.98	1.98	1.98	1.98	1.98	2.05	2.01	24	
10		1.98	1.99	2.01	2.01	2.01	S	2.03	2.01	2.01	2.01	2.01	2.01	2.01	2.01	2.01	2.17	2.01	2.03	2.02	2.03	2.04	2.03	1.98	2.17	2.02	24			
11		1.99	2.00	2.01	2.02	2.02	S	2.02	2.01	2.00	2.02	2.03	2.01	2.01	1.98	2.00	2.00	1.98	1.98	1.98	1.99	2.00	2.00	1.98	2.03	2.00	24			
12		2.00	2.01	2.00	2.04	2.04	S	2.09	2.03	2.02	2.01	2.01	2.01	2.01	2.00	2.00	2.00	1.99	2.00	2.01	2.02	2.02	2.05	2.11	2.05	1.99	2.11	2.02	24	
13		2.07	2.03	2.06	2.02	2.05	S	2.05	2.04	2.02	2.02	2.01	2.00	2.01	2.27	2.01	2.01	2.04	2.06	2.04	2.04	2.04	2.02	2.27	2.04	2.04	24			
14		2.06	2.04	2.05	2.03	2.06	S	2.11	2.12	2.08	2.04	2.03	2.03	2.22	2.02	2.02	2.02	2.02	2.02	2.05	2.14	2.09	2.03	2.02	2.22	2.06	24			
15		2.04	2.03	2.04	2.04	2.11	S	X	X	X	X	X	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	2.03	2.11	2.05	6
16		Y	Y	Y	Y	Y	S	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	-	-	-	-
17		Y	Y	Y	Y	Y	S	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	-	-	-	-	
18		Y	Y	Y	Y	Y	S	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	C	C	C	2.05	2.06	2.06	2.03	2.03	2.06	2.05	7
19		2.01	1.97	1.96	1.96	1.97	S	2.01	2.00	1.99	1.99	2.00	1.99	1.99	1.99	1.98	1.98	1.98	1.98	1.99	1.99	1.98	1.99	1.98	1.97	1.96	2.01	1.98	24	
20		1.97	1.97	X	X	X	S	X	X	Y	Y	Y	Y	Y	C1	C1	C1	C1	1.89	1.90	1.92	1.91	1.92	1.97	1.94	1.89	1.97	1.93	9	
21		1.94	1.94	1.94	1.94	1.94	S	1.95	1.95	1.95	1.94	1.95	1.95	1.94	1.94	1.94	1.94	1.93	1.92	1.95	1.94	1.96	1.95	2.03	1.92	2.03	1.95	24		
22		2.02	2.04	2.01	2.01	2.05	S	2.30	2.20	2.20	1.98	1.93	1.92	1.93	1.93	1.92	1.92	1.92	1.96	1.98	1.99	1.98	1.98	1.92	2.30	2.00	24			
23		1.98	1.98	1.99	2.01	2.02	S	2.00	2.05	X	X	X	X	X	X	X	X	X	X	X	X	X	X	X	1.98	2.05	2.00	8		
24		X	X	X	X	X	S	X	X	X	X	X	X	X	X	X	X	X	X	X	X	X	X	X	-	-	-	-		
25		X	X	X	X	X	S	X	X	X	X	X	X	X	X	X	X	X	X	X	X	X	X	X	-	-	-	-		
26		X	X	X	X	X	S	X	X	X	X	X	X	X	X	X	X	X	X	X	X	X	X	X	-	-	-	-		
27		X	X	X	X	X	S	X	X	X	X	X	X	X	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	-	-	-	-		
28		Y	Y	Y	Y	Y	S	C1	C1	C1	1.96	1.95	1.95	1.94	1.93	1.93	1.93	1.93	1.96	1.97	1.96	2.00	2.01	2.02	1.93	2.02	1.96	14		
29		2.04	2.02	2.19	2.21	2.10	S	2.22	2.15	2.02	2.01	2.03	2.00	2.01	1.99	1.95	1.95	1.95	2.02	1.95	1.96	1.98	2.05	2.31	2.35	2.07	24			
30		2.29	2.02	2.04	1.97	1.95	S	1.97	2.01	1.98	1.96	1.94	1.95	1.95	1.94	1.94	1.94	1.92	1.94	1.95	1.96	1.96	1.96	1.92	2.29	1.98	24			
HOURLY MAX		2.29	2.26	2.36	2.28	2.35	NA	2.30	2.32	2.20	2.08	2.12	2.08	2.22	2.07	2.27	2.04	2.05	2.17	2.07	2.11	2.14	2.09	2.31	2.35					
HOURLY AVG		2.05	2.03	2.05	2.05	2.06	NA	2.08	2.06	2.04	2.02	2.01	2.00	2.01	2.00	2.01	1.99	1.99	1.99	2.01	2.01	2.02	2.05	2.04						

STATUS FLAG CODES	
C	- MONTHLY CALIBRATION
C1	- REPEAT CALIBRATION
Y	- MAINTENANCE
S	- DAILY ZERO/SPAN CHECK
S1	- REPEAT ZERO/SPAN CHECK
Q	- QUALITY ASSURANCE
R	- RECOVERY
X	- MACHINE MALFUNCTION
G	- OUT FOR REPAIR
P	- POWER FAILURE

MONTHLY SUMMARY	
NUMBER OF NON-ZERO READINGS:	
476	
MAXIMUM INSTANTANEOUS VALUE:	
2.36	ppm
@ HOUR	2
ON DAY	6
IZS CALIBRATION TIME:	
21	hrs
MONTHLY CALIBRATION TIME:	3 hrs
STANDARD DEVIATION:	0.07
OPERATIONAL TIME:	
500	hrs

METHANE MAX Instantaneous Maximum (CH_4 ppm)



Wind: PRAMP_842
 Poll.: PRAMP_842-CH4[ppm]
 Monthly: 17/09
 Type: PollutionRose
 Direction: Blowing From (Wind Frequency)
 Based On 1 Hr.

Calm:

6.72%


Calm Avg:

1.99 [ppm]


Direction	0-2	2-3	3-5	5-10	>10.0	Total
N	0.6	0.0	0.0	0.0	0.0	0.6
NE	3.2	0.4	0.0	0.0	0.0	3.6
E	4.0	4.4	0.0	0.0	0.0	8.4
SE	6.1	12.4	0.0	0.0	0.0	18.5
S	10.3	6.3	0.0	0.0	0.0	16.6
SW	15.3	8.4	0.0	0.0	0.0	23.7
W	11.8	2.7	0.0	0.0	0.0	14.5
NW	5.9	1.5	0.0	0.0	0.0	7.4
Summary	57.1	36.1	0.0	0.0	0.0	93.3



PRAMP_842 Poll.: PRAMP_842-CH4[ppm] 2017/09/01 00:00 - 2017/09/30 23:00 Calm: 6.72% Calm Poll Avg: 1.99[ppm]



CH4[ppm] Calibration: PRAMP_842 Monthly: 17/09 Type: Span



NON-METHANE HYDROCARBON




PEACE RIVER AREA MONITORING PROGRAM COMMITTEE

Three Creeks 842b Station - September 2017

NON-METHANE HYDROCARBONS Hourly Averages (NMHC ppm)

	HR START (MST)	0:00	1:00	2:00	3:00	4:00	5:00	6:00	7:00	8:00	9:00	10:00	11:00	12:00	13:00	14:00	15:00	16:00	17:00	18:00	19:00	20:00	21:00	22:00	23:00	DAILY MIN.	DAILY MAX.	24-HR AVG.	RDGS.			
	HR END (MST)	0:59	1:59	2:59	3:59	4:59	5:59	6:59	7:59	8:59	9:59	10:59	11:59	12:59	13:59	14:59	15:59	16:59	17:59	18:59	19:59	20:59	21:59	22:59	23:59							
DAY																																
1		0.00	0.00	0.00	0.00	0.00	S	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	24				
2		0.00	0.00	0.00	0.00	0.00	S	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	24				
3		0.00	0.00	0.00	0.00	0.00	S	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	24				
4		0.00	0.00	0.00	0.00	0.00	S	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	24				
5		0.00	0.00	0.00	0.00	0.00	S	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	24				
6		0.00	0.00	0.00	0.00	0.00	S	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	24				
7		0.00	0.00	0.00	0.00	0.00	S	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	24				
8		0.00	0.00	0.00	0.00	0.00	S	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	24				
9		0.00	0.00	0.00	0.00	0.00	S	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	24				
10		0.00	0.00	0.00	0.00	0.00	S	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	24				
11		0.00	0.00	0.00	0.00	0.00	S	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	24				
12		0.00	0.00	0.00	0.00	0.00	S	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	24				
13		0.00	0.00	0.00	0.00	0.00	S	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	24				
14		0.00	0.00	0.00	0.00	0.00	S	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	24				
15		0.00	0.00	0.00	0.00	0.00	S	X	X	X	X	X	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	0.00	0.00	0.00	6		
16		Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	-	-	-	-	
17		Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	-	-	-	-	
18		Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	C	C	C	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	7		
19		0.00	0.00	0.00	0.00	0.00	S	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	24		
20		0.00	0.00	X	X	X	X	X	Y	Y	Y	Y	C1	C1	C1	C1	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	9
21		0.00	0.00	0.00	0.00	0.00	S	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	24			
22		0.00	0.00	0.00	0.00	0.00	S	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	24			
23		0.00	0.00	0.00	0.00	0.00	S	0.00	0.00	X	X	X	X	X	X	X	X	X	X	X	X	X	X	X	X	0.00	0.00	0.00	8			
24		X	X	X	X	X	X	X	X	X	X	X	X	X	X	X	X	X	X	X	X	X	X	X	X	-	-	-	-			
25		X	X	X	X	X	X	X	X	X	X	X	X	X	X	X	X	X	X	X	X	X	X	X	-	-	-	-				
26		X	X	X	X	X	X	X	X	X	X	X	X	X	X	X	X	X	X	X	X	X	X	X	-	-	-	-				
27		X	X	X	X	X	X	X	X	X	X	X	X	X	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	-	-	-	-			
28		Y	Y	Y	Y	Y	Y	Y	C1	C1	C1	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	14				
29		0.00	0.00	0.00	0.00	0.00	S	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	24				
30		0.00	0.00	0.00	0.00	0.00	S	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	24				
	HOURLY MAX	0.00	0.00	0.00	0.00	0.00	NA	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00					
	HOURLY AVG	0.00	0.00	0.00	0.00	0.00	NA	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00					




NUMBER OF NON-ZERO READINGS:	0				
MINIMUM 1-HR AVERAGE	0.00	ppm @ HOUR	0	ON DAY	1
MAXIMUM 1-HR AVERAGE:	0.00	ppm @ HOUR	0	ON DAY	1
MAXIMUM 24-HR AVERAGE:	0.00	ppm		ON DAY	1
I2S CALIBRATION TIME:	21	hrs		OPERATIONAL TIME:	
MONTHLY CALIBRATION TIME:	3	hrs		AMD OPERATION UPTIME:	69.4 %
STANDARD DEVIATION:	0.00			MONTHLY AVERAGE:	0.00 ppm



PEACE RIVER AREA MONITORING PROGRAM COMMITTEE

Three Creeks 842b Station - September 2017

NON-METHANE HYDROCARBONS Hourly Averages (NMHC ppm)





PEACE RIVER AREA MONITORING PROGRAM COMMITTEE

Three Creeks 842b Station - September 2017


NON-METHANE HYDROCARBONS Instantaneous Maximum (NMHC ppm)

	HR START (MST)	0:00	1:00	2:00	3:00	4:00	5:00	6:00	7:00	8:00	9:00	10:00	11:00	12:00	13:00	14:00	15:00	16:00	17:00	18:00	19:00	20:00	21:00	22:00	23:00	DAILY MIN.	DAILY MAX.	24-HR AVG.	RDGS.	
	HR END (MST)	0:59	1:59	2:59	3:59	4:59	5:59	6:59	7:59	8:59	9:59	10:59	11:59	12:59	13:59	14:59	15:59	16:59	17:59	18:59	19:59	20:59	21:59	22:59	23:59					
DAY																														
1		0.00	0.00	0.05	0.00	0.00	S	0.01	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.03	0.00	0.00	0.01	0.02	0.00	0.00	0.01	0.00	0.05	0.01	24			
2		0.00	0.00	0.01	0.01	0.00	S	0.00	0.00	0.01	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.01	0.00	0.00	0.00	0.00	0.01	0.00	24			
3		0.00	0.00	0.01	0.00	0.00	S	0.00	0.00	0.00	0.00	0.01	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.01	0.00	0.00	0.00	0.01	0.00	0.00	24			
4		0.00	0.00	0.01	0.01	0.00	S	0.00	0.00	0.00	0.00	0.00	0.21	0.02	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.01	0.00	24			
5		0.01	0.01	0.00	0.00	0.01	S	0.00	0.00	0.00	0.02	0.00	0.00	0.01	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.02	0.00	24			
6		0.00	0.00	0.01	0.00	0.00	S	0.00	0.00	0.00	0.00	0.00	0.00	0.01	0.09	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.09	0.00	24		
7		0.04	0.00	0.00	0.01	0.01	S	0.00	0.00	0.00	0.00	0.01	0.01	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.04	0.00	24		
8		0.00	0.00	0.00	0.00	0.00	S	0.00	0.00	0.00	0.00	0.00	0.02	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.02	0.00	24		
9		0.00	0.00	0.00	0.00	0.00	S	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.01	0.00	0.00	0.01	0.00	0.00	0.00	0.00	0.00	0.01	0.00	24			
10		0.00	0.00	0.00	0.00	0.00	S	0.00	0.00	0.04	0.00	0.00	0.00	0.00	0.00	0.00	0.01	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.04	0.00	24			
11		0.00	0.00	0.00	0.00	0.00	S	0.00	0.00	0.00	0.00	0.00	0.01	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.01	0.00	24			
12		0.00	0.00	0.00	0.14	0.00	S	0.01	0.06	0.00	0.00	0.01	0.09	0.00	0.01	0.00	0.00	0.00	0.03	0.00	0.01	0.00	0.00	0.00	0.00	0.14	0.02	24		
13		0.00	0.00	0.05	0.00	0.02	S	0.01	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.01	0.00	0.05	0.00	24		
14		0.00	0.00	0.01	0.01	0.01	S	0.00	0.00	0.06	0.03	0.01	0.00	0.11	0.00	0.00	0.00	0.04	0.00	0.01	0.00	0.00	0.00	0.00	0.00	0.11	0.01	24		
15		0.00	0.00	0.01	0.00	0.01	S	X	X	X	X	X	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	0.00	0.01	0.00	6
16		Y	Y	Y	Y	Y	S	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	-	-	-	-	
17		Y	Y	Y	Y	Y	S	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	-	-	-	-	
18		Y	Y	Y	Y	Y	S	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	C	C	C	0.00	0.00	0.00	0.03	0.00	0.03	0.01	7	
19		0.00	0.00	0.02	0.00	0.00	S	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.02	0.00	24		
20		0.00	0.00	X	X	X	S	X	X	Y	Y	Y	Y	C1	C1	C1	C1	0.01	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.01	0.00	9		
21		0.00	0.00	0.00	0.01	0.00	S	0.00	0.06	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.06	0.00	24			
22		0.00	0.00	0.00	0.00	0.00	S	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	24			
23		0.00	0.00	0.00	0.00	0.00	S	0.00	0.00	X	X	X	X	X	X	X	X	X	X	X	X	X	X	X	0.00	0.00	0.00	8		
24		X	X	X	X	X	S	X	X	X	X	X	X	X	X	X	X	X	X	X	X	X	X	X	-	-	-	-		
25		X	X	X	X	X	S	X	X	X	X	X	X	X	X	X	X	X	X	X	X	X	X	X	-	-	-	-		
26		X	X	X	X	X	S	X	X	X	X	X	X	X	X	X	X	X	X	X	X	X	X	X	-	-	-	-		
27		X	X	X	X	X	S	X	X	X	X	X	X	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	-	-	-	-		
28		Y	Y	Y	Y	Y	S	C1	C1	C1	0.02	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.02	0.00	14			
29		0.00	0.00	0.00	0.00	0.00	S	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.03	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.03	0.00	24				
30		0.00	0.00	0.00	0.00	0.00	S	0.01	0.00	0.00	0.00	0.00	0.00	0.01	0.00	0.02	0.00	0.00	0.00	0.00	0.00	0.01	0.00	0.02	0.00	24				
HOURLY MAX		0.04	0.01	0.05	0.14	0.02	NA	0.01	0.06	0.06	0.03	0.02	0.21	0.11	0.01	0.09	0.03	0.04	0.03	0.03	0.02	0.01	0.00	0.01	0.03					
HOURLY AVG		0.00	0.00	0.01	0.01	0.00	NA	0.00	0.01	0.01	0.00	0.00	0.02	0.01	0.00	0.01	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.02	0.00				

STATUS FLAG CODES	
C	- MONTHLY CALIBRATION
C1	- REPEAT CALIBRATION
Y	- MAINTENANCE
S	- DAILY ZERO/SPAN CHECK
S1	- REPEAT ZERO/SPAN CHECK
Q	- QUALITY ASSURANCE
R	- RECOVERY
X	- MACHINE MALFUNCTION
G	- OUT FOR REPAIR
P	- POWER FAILURE

MONTHLY SUMMARY	
NUMBER OF NON-ZERO READINGS:	
71	
MAXIMUM INSTANTANEOUS VALUE:	
0.21 ppm @ HOUR 11 ON DAY 4	
IZS CALIBRATION TIME: 21 hrs	
MONTHLY CALIBRATION TIME: 3 hrs	
STANDARD DEVIATION: 0.02	

NON-METHANE HYDROCARBONS Instantaneous Maximum (NMHC ppm)



Wind: PRAMP_842
Poll.: PRAMP_842-NMHC[ppm]
Monthly: 17/09
Type: PollutionRose
Direction: Blowing From (Wind Frequency)
Based On 1 Hr.


Calm: 6.72%

Calm Avg: 0.00 [ppm]


Direction	0-0.1	0.1-0.3	0.3-1	1-2	>2.0	Total
N	0.6	0.0	0.0	0.0	0.0	0.6
NE	3.6	0.0	0.0	0.0	0.0	3.6
E	8.4	0.0	0.0	0.0	0.0	8.4
SE	18.5	0.0	0.0	0.0	0.0	18.5
S	16.6	0.0	0.0	0.0	0.0	16.6
SW	23.7	0.0	0.0	0.0	0.0	23.7
W	14.5	0.0	0.0	0.0	0.0	14.5
NW	7.4	0.0	0.0	0.0	0.0	7.4
Summary	93.3	0.0	0.0	0.0	0.0	93.3

% Icon Classes (ppm)	93	0-0.1	0	0.1-0.3	0	0.3-1	0	1-2	0	>2.0
----------------------	----	-------	---	---------	---	-------	---	-----	---	------

PRAMP_842 Poll.: PRAMP_842-NMHC[ppm] 2017/09/01 00:00 - 2017/09/30 23:00 Calm: 6.72% Calm Poll Avg: 0.00[ppm]



NMHC[ppm] Calibration: PRAMP_842 Monthly: 17/09 Type: Span




WIND SPEED

WIND SPEED Hourly Averages (WS kph)

	HR START (MST)	0:00	1:00	2:00	3:00	4:00	5:00	6:00	7:00	8:00	9:00	10:00	11:00	12:00	13:00	14:00	15:00	16:00	17:00	18:00	19:00	20:00	21:00	22:00	23:00	DAILY MIN.	DAILY MAX.	24-HR AVG.	RDGS.	
	HR END (MST)	0:59	1:59	2:59	3:59	4:59	5:59	6:59	7:59	8:59	9:59	10:59	11:59	12:59	13:59	14:59	15:59	16:59	17:59	18:59	19:59	20:59	21:59	22:59	23:59					
DAY																														
1		8.9	8.9	8.0	10.1	11.7	10.2	6.5	10.2	15.2	19.1	21.0	22.3	21.4	22.1	21.7	25.6	24.1	23.0	15.3	15.7	16.0	14.2	10.8	8.1	6.5	25.6	14.0	24	
2		7.5	8.0	7.8	4.6	3.7	4.3	5.9	5.4	5.2	7.9	5.9	8.8	8.0	8.5	7.7	5.7	8.0	8.6	9.7	7.0	5.1	4.7	5.0	7.1	3.7	9.7	5.6	24	
3		6.7	4.7	6.6	3.2	3.7	3.0	1.6	3.1	2.9	1.7	4.4	4.8	7.2	7.5	10.2	11.9	11.3	10.0	6.7	1.6	1.8	0.6	0.9	1.5	0.6	11.9	2.7	24	
4		1.8	1.7	1.8	1.9	2.7	2.0	2.3	2.9	2.5	5.6	9.4	10.9	9.6	9.8	10.2	9.4	10.4	11.0	5.3	3.7	4.9	5.5	7.6	8.6	1.7	11.0	4.4	24	
5		8.4	7.9	6.5	1.9	5.3	7.7	8.5	8.9	8.1	9.4	9.4	7.8	10.4	10.3	10.5	9.9	8.3	7.7	4.2	1.5	2.3	3.5	4.2	3.5	1.5	10.5	5.2	24	
6		3.3	4.9	4.3	1.8	3.2	4.2	5.7	5.6	4.8	5.5	4.7	7.1	9.5	8.0	8.3	6.9	5.9	6.6	4.9	1.1	5.7	5.9	4.8	7.4	1.1	9.5	4.1	24	
7		9.5	10.8	11.1	8.6	7.8	3.0	4.6	6.1	9.4	14.2	12.1	12.0	12.0	10.8	10.1	9.9	10.1	5.9	5.3	6.3	3.7	6.5	4.0	2.5	2.5	14.2	7.2	24	
8		2.6	5.4	4.2	4.4	5.7	7.9	11.0	6.5	4.0	4.5	4.4	7.8	9.0	12.6	11.7	12.5	10.1	9.8	8.6	6.5	7.7	7.9	8.8	2.6	12.6	5.2	24		
9		6.8	6.8	4.6	2.9	2.6	3.8	3.5	3.5	6.3	9.1	11.5	14.1	14.8	10.8	15.6	16.6	14.8	12.3	9.6	9.7	9.8	9.9	9.3	8.5	2.6	16.6	6.9	24	
10		10.6	12.3	11.7	19.3	19.2	16.5	16.8	17.1	19.4	19.0	22.3	22.6	23.5	28.0	24.8	26.5	23.4	20.5	18.1	14.8	13.8	13.7	12.1	6.4	28.0	17.4	24		
11		10.0	9.6	8.2	6.8	5.2	8.5	10.7	7.9	8.4	8.3	8.4	11.6	17.9	26.0	18.9	19.1	31.6	26.5	25.0	19.5	15.9	12.2	8.7	3.8	3.8	31.6	9.3	24	
12		5.3	6.3	5.9	3.9	4.1	3.4	2.0	4.4	5.6	8.2	9.4	8.5	8.2	7.6	9.6	8.1	6.8	8.4	3.7	2.6	1.5	1.6	1.9	4.1	1.5	9.6	3.0	24	
13		2.9	2.1	1.7	3.0	3.0	2.3	0.8	2.5	4.7	7.6	5.0	6.6	9.0	10.2	10.7	8.4	9.3	6.7	2.4	3.2	2.2	2.7	3.9	5.2	0.8	10.7	3.4	24	
14		6.3	7.0	3.9	1.6	2.8	2.6	3.0	1.9	3.3	4.0	3.9	3.9	4.6	7.6	6.8	6.4	5.7	1.4	0.9	2.2	4.3	9.3	10.5	6.7	0.9	10.5	1.5	24	
15		4.3	4.9	6.2	6.8	1.6	3.0	5.0	7.4	11.2	14.7	18.1	15.6	16.9	16.1	17.1	17.5	16.2	15.3	10.1	6.7	8.1	10.1	9.0	10.1	1.6	18.1	10.1	24	
16		12.8	12.8	11.8	10.3	9.2	9.3	10.0	10.4	12.7	17.3	20.4	18.8	17.7	17.7	16.2	16.1	15.9	12.5	6.7	6.3	8.9	10.8	12.0	10.2	6.3	20.4	12.4	24	
17		8.8	8.7	7.8	7.2	7.2	5.6	5.3	5.2	5.3	9.9	10.4	10.1	9.1	9.9	10.5	10.1	10.1	9.0	9.1	7.6	6.6	7.2	6.5	7.4	7.5	5.2	10.5	7.2	24
18		6.3	5.3	5.7	7.0	6.1	6.5	7.5	7.0	7.9	6.8	7.0	7.8	6.8	5.6	7.3	6.1	2.6	1.9	2.7	2.8	1.6	1.8	4.6	5.6	1.6	7.9	4.7	24	
19		6.0	4.1	3.8	2.1	2.0	3.5	3.2	4.4	5.9	5.2	7.2	8.9	9.2	9.0	8.3	11.2	13.0	12.1	11.8	10.8	11.4	11.4	13.5	13.7	2.0	13.7	6.4	24	
20		13.8	13.9	13.7	12.5	12.8	13.7	15.7	16.0	14.9	14.1	14.0	13.8	14.6	13.8	13.0	12.8	12.5	19.0	19.1	17.1	17.4	8.2	5.2	9.6	5.2	19.1	13.3	24	
21		8.4	7.2	10.2	10.6	11.2	9.9	9.8	7.5	9.4	10.1	9.2	10.2	11.0	12.1	12.8	11.3	10.3	8.0	1.7	0.5	0.7	1.1	2.0	1.6	0.5	12.8	7.0	24	
22		1.3	1.9	1.0	1.7	2.5	2.7	2.0	3.7	3.3	6.9	8.4	7.9	5.6	7.3	7.4	6.2	7.0	6.8	1.7	2.9	4.3	4.9	5.3	5.7	1.0	8.4	3.3	24	
23		4.9	4.7	5.7	6.0	4.7	5.2	4.3	4.4	5.6	11.4	11.7	11.5	13.5	12.7	10.9	10.1	7.9	4.5	3.8	1.8	3.4	4.7	6.4	7.6	1.8	13.5	5.1	24	
24		7.5	8.5	8.9	8.8	9.5	10.3	11.0	7.7	8.5	8.1	11.6	10.2	10.8	11.7	11.9	11.6	12.6	10.1	8.4	11.1	7.4	6.4	9.6	10.0	6.4	12.6	8.6	24	
25		8.7	8.1	9.0	9.5	8.3	7.5	9.3	10.7	10.8	13.9	15.2	14.7	15.2	15.8	14.2	14.5	15.5	14.8	14.8	7.9	6.2	5.7	5.2	5.1	5.4	5.1	15.8	9.0	24
26		6.3	4.9	4.7	2.5	1.4	1.8	2.6	2.2	2.8	7.0	7.3	8.4	8.9	10.7	13.7	15.3	14.1	5.3	2.2	3.9	2.6	4.1	5.6	3.5	1.4	15.3	4.5	24	
27		3.8	4.7	3.2	3.0	4.7	4.3	4.3	4.9	4.5	6.8	5.5	5.6	5.9	6.8	4.6	4.6	6.9	4.6	6.7	1.3	2.2	4.3	5.3	5.1	6.0	1.3	9.4	2.7	24
28		7.1	8.8	8.0	3.4	6.7	6.6	6.0	6.2	11.8	15.3	15.1	14.7	13.9	14.8	16.7	15.4	11.7	8.3	3.2	5.1	6.9	7.9	8.6	3.2	16.7	8.9	24		
29		5.9	2.4	4.8	4.7	2.0	2.0	2.3	5.7	5.2	6.0	4.1	5.8	6.9	6.6	5.2	6.9	2.8	1.4	2.6	2.9	2.9	4.2	3.1	1.4	6.9	1.9	24		
30		4.5	5.6	8.9	9.4	12.6	16.0	10.8	9.6	9.0	15.5	19.1	20.1	19.9	20.6	21.0	20.0	16.9	14.3	8.9	12.4	11.9	8.1	11.2	13.6	4.5	21.0	11.3	24	
HOURLY MAX		13.8	13.9	13.7	19.3	19.2	16.5	16.8	17.1	19.4	19.1	22.3	22.6	23.5	28.0	24.8	26.5	31.6	26.5	25.0	19.5	17.4	14.2	13.5	13.7					
HOURLY AVG		3.2	3.6	3.3	2.5	2.4	2.1	2.2	2.4	3.4	5.2	5.8	6.1	6.6	7.4	7.3	7.8	7.1	4.7	2.5	1.8	1.5	2.1	2.7	2.7					

24 HR AVERAGES September 2017



NUMBER OF NON-ZERO READINGS:

720

MINIMUM 1-HR AVERAGE:

0.5

kph @ HOUR

19

ON DAY

21

MAXIMUM 1-HR AVERAGE:

31.6

kph @ HOUR

16

ON DAY

11

MAXIMUM 24-HR AVERAGE:

17.4

kph

ON DAY

10

ON DAY

MONTHLY CALIBRATION TIME:

0

hrs

AMD OPERATION UPTIME:

100.0


%



PEACE RIVER AREA MONITORING PROGRAM COMMITTEE

Three Creeks 842b Station - September 2017

WIND SPEED Hourly Averages (WS kph)





PEACE RIVER AREA MONITORING PROGRAM COMMITTEE

Three Creeks 842b Station - September 2017

WIND SPEED Instantaneous Maximum (WS kph)

HR START (MST)	0:00 0:59	1:00 1:59	2:00 2:59	3:00 3:59	4:00 4:59	5:00 5:59	6:00 6:59	7:00 7:59	8:00 8:59	9:00 9:59	10:00 10:59	11:00 11:59	12:00 12:59	13:00 13:59	14:00 14:59	15:00 15:59	16:00 16:59	17:00 17:59	18:00 18:59	19:00 19:59	20:00 20:59	21:00 21:59	22:00 22:59	23:00 23:59	DAILY MIN.	DAILY MAX.	24-HR AVG.	RDGS.	
DAY																													
1	18.6	15.0	14.3	20.9	22.6	18.9	15.7	22.0	30.0	45.3	36.5	42.3	44.3	43.6	44.8	59.6	49.9	48.9	34.2	36.7	29.6	25.3	24.6	16.6	14.3	59.6	31.7	24	
2	17.1	15.1	14.4	9.8	7.2	10.8	11.3	10.3	11.0	14.8	11.3	17.3	13.8	14.5	13.8	9.6	15.7	16.4	17.0	12.5	9.3	10.3	11.1	13.1	7.2	17.3	12.8	24	
3	14.3	9.8	12.4	6.5	9.5	8.5	4.2	8.9	6.9	5.6	12.1	12.3	17.1	18.2	24.9	25.5	22.9	22.2	15.8	5.5	6.0	3.6	2.7	3.9	2.7	25.5	11.6	24	
4	5.5	3.9	3.9	4.1	4.9	4.4	4.7	5.5	6.2	11.5	18.2	21.6	23.2	22.2	23.8	21.3	20.1	9.8	8.0	9.9	11.5	14.2	17.4	3.9	23.8	12.4	24		
5	15.9	16.1	13.9	8.9	11.9	14.0	16.4	15.3	14.4	17.5	15.8	15.2	18.3	18.7	20.9	20.0	15.5	12.9	9.3	3.5	6.0	6.8	7.5	7.1	3.5	20.9	13.4	24	
6	7.4	7.7	8.4	6.0	7.2	7.2	11.4	11.7	10.4	10.2	14.2	14.6	16.6	16.8	13.4	14.1	10.5	12.9	12.1	4.2	13.4	13.4	11.4	15.6	4.2	16.8	11.3	24	
7	15.7	17.9	18.0	16.6	15.9	14.5	12.4	17.3	22.3	28.2	23.0	21.7	21.1	17.8	17.0	17.7	18.4	14.4	9.7	10.3	11.7	11.6	10.5	6.5	6.5	28.2	16.3	24	
8	6.8	9.2	7.8	7.9	12.9	13.8	22.6	20.9	9.4	11.2	13.2	17.2	17.0	22.4	21.5	21.8	22.9	20.4	17.4	11.2	13.2	12.6	14.4	15.1	6.8	22.9	15.1	24	
9	13.3	13.2	11.0	6.9	8.8	8.3	7.5	9.2	14.1	17.8	23.1	27.5	30.0	21.3	32.9	32.8	28.8	24.1	19.0	20.2	18.3	17.6	17.5	6.9	32.9	18.4	24		
10	20.6	21.7	31.9	39.7	40.5	32.5	33.4	34.7	40.7	37.3	44.3	47.5	45.8	55.4	47.4	51.9	50.9	39.9	38.4	31.3	27.3	25.3	21.4	12.0	55.4	36.3	24		
11	18.8	16.8	12.8	14.4	12.6	19.9	23.3	17.7	17.6	18.6	16.8	24.7	42.5	53.1	41.2	41.5	64.4	56.3	53.6	43.4	30.0	31.7	18.5	11.8	64.4	29.3	24		
12	10.7	11.7	12.2	11.0	10.5	7.4	7.7	10.5	13.0	18.9	19.8	19.4	21.8	19.3	49.3	30.2	24.0	28.1	10.5	5.0	4.5	5.3	5.2	9.9	4.5	49.3	15.2	24	
13	8.6	6.2	6.6	7.3	9.5	6.3	5.4	7.3	14.1	17.5	12.4	18.3	20.8	27.1	24.0	17.0	21.0	22.0	6.3	9.0	5.2	6.1	10.8	16.5	5.2	27.1	12.7	24	
14	17.9	16.3	9.9	3.7	5.7	5.9	8.5	6.6	7.5	9.1	9.0	11.7	14.0	20.3	16.0	21.4	12.3	7.2	3.3	8.8	12.9	16.2	17.0	14.8	3.3	21.4	11.5	24	
15	11.4	11.3	12.9	14.7	5.2	7.7	11.6	16.8	24.1	25.8	29.5	25.7	28.9	31.4	32.3	32.2	30.6	29.2	18.6	14.5	16.7	18.1	17.6	22.6	5.2	32.3	20.4	24	
16	23.0	23.9	21.1	20.1	18.9	19.5	20.5	20.7	27.4	32.0	36.8	34.0	33.0	34.7	30.9	34.8	30.3	26.2	19.9	15.9	18.2	22.2	21.9	21.8	15.9	36.8	25.3	24	
17	17.9	16.9	16.3	16.7	15.5	12.6	11.5	12.4	13.2	19.4	19.3	21.4	23.3	20.4	23.9	24.2	21.6	21.8	21.5	19.6	21.5	18.3	22.8	19.8	11.5	24.2	18.8	24	
18	14.9	15.5	16.8	18.1	16.3	19.6	18.1	17.9	18.6	20.7	17.7	18.3	20.3	14.4	20.3	15.6	15.6	8.9	7.1	5.9	10.0	8.6	7.0	9.9	14.1	5.9	20.7	14.8	24
19	10.6	7.2	9.4	6.0	7.8	7.7	6.4	10.5	12.2	11.1	19.9	20.5	27.6	21.7	17.9	31.6	30.9	29.0	26.7	26.3	30.5	34.3	35.0	6.0	35.0	19.7	24		
20	38.3	35.3	36.1	31.3	38.3	30.7	42.8	38.5	33.7	33.7	34.3	36.6	34.8	34.1	35.8	37.5	58.7	49.6	51.1	55.1	44.1	22.9	21.3	21.3	58.7	37.8	24		
21	23.5	19.2	23.9	26.4	25.9	24.0	21.4	23.5	24.1	27.0	22.6	27.4	25.7	27.5	25.8	26.3	23.1	19.5	5.1	3.2	2.6	3.9	4.6	3.9	2.6	27.5	19.2	24	
22	2.8	3.6	4.1	3.7	6.3	4.2	3.3	9.1	8.6	12.7	16.0	16.8	14.4	19.5	17.2	15.9	16.3	15.0	6.1	9.3	9.3	10.7	13.2	12.1	2.8	19.5	10.4	24	
23	10.5	11.6	12.2	12.7	11.9	11.1	8.9	8.5	16.0	20.9	24.0	25.5	28.0	24.5	21.8	21.5	15.9	9.8	7.2	5.6	8.0	10.5	13.5	14.2	5.6	28.0	14.8	24	
24	15.3	15.9	16.9	13.9	14.5	19.3	18.9	14.4	17.0	17.5	19.6	21.4	22.9	24.6	25.2	24.1	21.0	16.5	22.7	15.0	12.3	12.3	23.0	17.6	12.3	25.2	18.8	24	
25	16.4	13.2	14.2	15.4	22.2	14.8	16.2	19.9	21.6	25.0	26.5	26.6	29.5	29.0	29.4	29.5	28.2	33.6	16.8	11.6	12.7	9.1	10.3	9.6	9.1	33.6	20.1	24	
26	10.4	9.8	7.9	7.4	4.4	4.3	5.1	4.6	7.2	13.7	12.9	16.1	18.2	25.9	28.3	27.0	28.5	15.9	4.1	19.7	7.5	11.2	14.5	8.7	4.1	28.5	13.1	24	
27	8.4	8.4	6.3	6.9	8.3	8.1	8.8	8.4	9.8	11.6	12.1	13.3	17.2	16.8	15.2	16.3	21.4	14.2	6.6	4.5	8.4	10.1	12.5	10.6	4.5	21.4	11.0	24	
28	13.6	16.2	15.8	11.5	14.5	13.0	12.8	15.9	18.7	26.9	27.3	25.5	24.5	29.9	31.3	31.4	28.3	16.6	10.7	13.3	14.9	15.7	16.4	10.7	31.4	19.2	24		
29	14.9	7.3	8.1	11.1	7.0	5.0	6.0	9.9	10.4	11.4	10.3	15.5	14.5	13.0	12.0	14.2	13.7	7.9	3.4	5.6	5.5	7.6	7.0	9.2	3.4	15.5	9.6	24	
30	18.3	13.1	21.7	15.7	24.5	30.0	20.2	18.4	18.7	30.5	34.0	35.2	36.6	38.7	38.9	39.8	36.7	29.9	21.2	28.3	29.1	19.2	26.1	34.1	13.1	39.8	27.5	24	
HOURLY MAX	38.3	35.3	36.1	39.7	40.5	32.5	42.8	38.5	40.7	45.3	44.3	47.5	45.8	55.4	59.6	64.4	58.7	53.6	51.1	55.1	44.1	34.3	35.0						
HOURLY AVG	14.7	13.6	14.0	13.2	14.0	13.5	13.9	14.9	16.6	20.1	21.1	22.9	24.9	25.9	26.4	26.9	25.8	23.4	16.8	15.7	15.2	14.9	15.2	15.0					


STATUS FLAG CODES

C	- MONTHLY CALIBRATION	Q	- QUALITY ASSURANCE
C1	- REPEAT CALIBRATION	R	- RECOVERY
Y	- MAINTENANCE	X	- MACHINE MALFUNCTION
S	- DAILY ZERO/SPAN CHECK	G	- OUT FOR REPAIR
S1	- REPEAT ZERO/SPAN CHECK	P	- POWER FAILURE

MONTHLY SUMMARY

MAXIMUM INSTANTANEOUS VALUE:	64.4	kph	@ HOUR	16	ON DAY	11
OPERATIONAL TIME: 720 hrs						

WIND SPEED Instantaneous Maximum (WS kph)




Wind: PRAMP_842
 Monitor: WSP [kph]
 Monthly: 17/09
 Type: WindRose
 Direction: Blowing From (Wind Frequency)
 Based On 1 Hr.

Calm: 5.00%

Direction	1.8-6.0	6.0-12.0	12.0-20.0	20.0-29.0	29.0-39.0	>39.0	Total
N	0.0	0.1	0.4	0.0	0.0	0.0	0.6
NE	0.4	0.8	2.2	0.0	0.0	0.0	3.5
E	7.5	2.9	0.0	0.0	0.0	0.0	10.4
SE	10.3	12.2	0.8	0.0	0.0	0.0	23.3
S	5.1	9.6	4.2	0.1	0.0	0.0	19.0
SW	5.6	12.2	4.7	0.1	0.0	0.0	22.6
W	1.4	3.3	2.6	3.1	0.1	0.0	10.6
NW	1.0	3.1	1.0	0.0	0.0	0.0	5.0
Summary	31.3	44.3	16.0	3.3	0.1	0.0	95.0

%	Icon	Classes (kph)	31	1.8-6.0	44	6.0-12.0	16	12.0-20.0	3	20.0-29.0	0	29.0-39.0	0	>39.0
---	------	---------------	----	---------	----	----------	----	-----------	---	-----------	---	-----------	---	-------

PRAMP_842 2017/09/01 00:00 - 2017/09/30 23:00 Calm: 5.00% Calm Wind Avg Speed: 1.43(kph)



WIND DIRECTION



PEACE RIVER AREA MONITORING PROGRAM COMMITTEE

Three Creeks 842b Station - September 2017

WIND DIRECTION Hourly Averages (WD)

HR START (MST)	0:00	1:00	2:00	3:00	4:00	5:00	6:00	7:00	8:00	9:00	10:00	11:00	12:00	13:00	14:00	15:00	16:00	17:00	18:00	19:00	20:00	21:00	22:00	23:00	24-HOUR AVG QUADRANT	24-HR RDGS.	
HR END (MST)	0:59	1:59	2:59	3:59	4:59	5:59	6:59	7:59	8:59	9:59	10:59	11:59	12:59	13:59	14:59	15:59	16:59	17:59	18:59	19:59	20:59	21:59	22:59	23:59			
DAY																											
1	S	SSE	S	SW	SW	SW	SW	SW	WSW	WSW	W	W	WSW	W	W	W	W	W	WSW	24							
2	SW	SW	SW	SW	ESE	SE	SSE	SSE	SSW	S	SSW	SW	SW	WSW	SW	SW	SW	SW	SSW	SSE	S	SSW	SSW	SSW	SSW	24	
3	SW	S	S	SSE	SE	SSE	SSE	S	SW	W	W	NW	WSW	W	W	WSW	W	W	WNW	WSW	ENE	ESE	E	ESE	WSW	24	
4	ESE	E	E	ESE	E	E	ESE	SE	SSW	S	SSW	SSW	S	SSW	SSW	SSW	SSW	SSW	SSE	SE	SE	SE	SE	S	24		
5	SE	SE	SE	SE	SE	SE	SSE	SSE	S	SSW	SW	SW	SSW	S	SSW	SSW	SSW	SSW	SSE	ESE	ESE	ESE	ESE	S	24		
6	ESE	ESE	ESE	ESE	SE	SE	SE	SSE	SSE	SSE	SSW	SW	SSW	SSW	SSW	SSW	SSW	S	SE	ESE	ESE	ESE	ESE	SSE	24		
7	SE	SE	SE	SE	SSE	SSE	SSE	SSE	SSW	SSW	S	SSW	S	SSE	SSE	S	SSW	S	24								
8	SW	SW	SW	WSW	WSW	W	WNW	WNW	S	SSE	SW	SSW	SSW	SSW	SSW	SSW	SSW	S	S	SE	SSE	SSE	SSE	SSW	24		
9	SE	SE	E	E	E	E	ESE	SE	SSE	S	S	SSW	24														
10	SW	SSW	SW	SW	WSW	SW	WSW	WSW	WSW	WSW	WSW	W	W	W	W	WSW	24										
11	SW	SW	S	SE	SE	SE	SE	SE	SSE	S	S	SSW	SW	WSW	WSW	WSW	WSW	W	W	WSW	W	NW	NW	SW	24		
12	W	W	WSW	SW	SW	WSW	SW	WSW	W	WNW	W	WNW	W	WSW	W	WNW	W	NE	ENE	E	ENE	NE	NE	NW	24		
13	NE	SE	S	ESE	WNW	SSW	WNW	WNW	NW	S	SSW	WNW	24														
14	E	WSW	WSW	SW	SE	E	ESE	ESE	SW	SSW	SE	SE	S	W	WNW	W	W	W	S	E	SE	SSE	SSE	SSE	24		
15	SSE	SE	SE	SSE	SE	SSE	SSE	SSE	SSE	S	SSE	S	SSE	SE	SE	SSE	SSE	SSE	24								
16	SSE	SE	SSE	SSE	S	SSE	SE	SE	SSE	SSE	SSE	24															
17	SE	SE	SE	ESE	SE	ESE	ESE	ESE	SSE	SSE	S	SSE	ESE	ESE	ESE	ESE	ESE	SE	24								
18	ESE	E	E	E	E	E	E	E	E	E	E	ESE	E	E	ESE	ESE	S	ESE	24								
19	S	SSE	SSE	SE	ESE	ESE	ENE	ESE	NNW	NW	SW	WSW	WNW	WSW	WSW	ENE	NNE	24									
20	NNE	NNE	NE	NNE	NNE	NNE	NNE	NNE	NE	ESE	SE	SW	WSW	WNW	SW	ENE	NNE	24									
21	SW	NW	NW	SW	SW	SW	ESE	S	SSW	WNW	SW	E	NW	NW	NW	WNW	NW	W	S	SE	E	ESE	ESE	NNW	24		
22	E	ESE	E	E	E	E	ENE	ESE	SE	S	S	SSW	S	S	SSW	SE	SSE	S	SE	ESE	SE	SE	ESE	SSE	24		
23	E	ESE	E	SE	SSE	S	S	SSW	SSW	SW	SSW	S	SE	ESE	SE	SE	SE	SSE	24								
24	SE	SE	SE	SSE	SSE	SSE	SSE	SSE	SSE	SSE	SSE	S	SSW	SSW	SSW	SSW	S	24									
25	SSW	S	S	S	S	S	SSW	SSW	SSW	SW	WSW	24															
26	SSE	SSE	SE	SE	SSE	ESE	ESE	ESE	S	SW	SW	SW	SW	SW	SW	SW	SW	SW	ESE	SE	SSE	SW	SW	SSW	24		
27	SW	SW	SSE	SE	SE	SE	SSE	SSE	SSW	SSW	SW	WSW	S	SSW	WSW	SW	WSW	SW	SE	E	ENE	E	ESE	ESE	SSW	24	
28	ESE	SE	SSE	SSE	S	SSE	S	S	SSE	SSE	ESE	SE	SE	SE	SE	SE	SSE	24									
29	SE	E	E	ESE	ESE	E	ESE	SSE	S	SSW	SSW	SW	WSW	SW	WSW	WSW	WSW	WSW	ESE	ESE	ENE	E	ENE	E	24		
30	SW	SSE	WSW	WSW	SW	WSW	WSW	WSW	WSW	WSW	WSW	WSW	WSW	WSW	WSW	WSW	WSW	WSW	W	NWW	NW	NW	NW	W	24		


STATUS FLAG CODES

C	- MONTHLY CALIBRATION	Q	- QUALITY ASSURANCE
C1	- REPEAT CALIBRATION	R	- RECOVERY
Y	- MAINTENANCE	X	- MACHINE MALFUNCTION
S	- DAILY ZERO/SPAN CHECK	G	- OUT FOR REPAIR
S1	- REPEAT ZERO/SPAN CHECK	P	- POWER FAILURE

LAST CALIBRATION:	August 30, 2017
DECLINATION :	MAGNETIC DECLINATION 15 DEGREE EAST

MONTHLY CALIBRATION TIME:	0 hrs	OPERATIONAL TIME:	720 hrs
STANDARD DEVIATION:	67	AMD OPERATION UPTIME:	100.0 %
MONTHLY AVERAGE: 205 (SSW)			

WIND DIRECTION Hourly Averages (WD)



RELATIVE HUMIDITY


RELATIVE HUMIDITY Hourly Averages (RH %)

	HR START (MST)	0:00 0:59	1:00 1:59	2:00 2:59	3:00 3:59	4:00 4:59	5:00 5:59	6:00 6:59	7:00 7:59	8:00 8:59	9:00 9:59	10:00 10:59	11:00 11:59	12:00 12:59	13:00 13:59	14:00 14:59	15:00 15:59	16:00 16:59	17:00 17:59	18:00 18:59	19:00 19:59	20:00 20:59	21:00 21:59	22:00 22:59	23:00 23:59	DAILY MIN.	DAILY MAX.	24-HR AVG.	RDGS.
DAY																													
1	67	75	75	70	71	77	76	70	62	53	50	45	40	42	38	35	35	36	39	43	45	48	50	60	35	77	54	24	
2	64	70	70	70	76	80	80	79	75	73	82	82	86	90	91	91	92	90	90	92	94	95	96	96	64	96	84	24	
3	96	96	96	96	96	96	96	96	96	95	73	47	38	40	40	43	38	38	41	61	85	90	91	93	38	96	74	24	
4	93	94	95	95	95	95	96	96	87	65	45	33	31	30	30	30	28	28	30	45	65	73	65	63	65	28	96	62	24
5	65	64	66	79	83	80	81	76	67	59	55	49	42	37	34	30	29	31	42	66	83	88	89	89	29	89	62	24	
6	90	91	91	91	84	83	84	77	76	68	56	50	43	34	29	28	27	30	36	66	76	68	76	64	27	91	63	24	
7	58	58	62	66	67	78	80	69	61	56	55	53	51	48	44	42	47	55	63	66	71	77	79	82	42	82	62	24	
8	85	81	77	79	84	76	75	83	91	91	92	91	84	78	73	65	61	63	72	79	84	90	91	91	61	92	81	24	
9	92	91	91	92	91	93	94	94	90	85	77	77	81	78	64	57	55	56	58	67	71	71	76	77	55	94	78	24	
10	81	80	75	66	64	67	69	66	63	61	52	47	40	35	32	31	32	35	37	42	47	51	57	62	31	81	54	24	
11	63	63	70	85	88	86	82	79	77	75	70	64	52	46	50	49	34	36	38	44	48	60	67	72	34	88	62	24	
12	79	78	84	89	90	91	94	87	72	55	52	49	46	44	58	68	76	63	67	86	93	94	95	95	44	95	75	24	
13	96	96	96	95	95	95	95	96	95	93	89	92	79	72	57	46	48	39	40	61	82	85	86	84	39	96	79	24	
14	84	86	91	91	93	94	95	96	95	91	76	66	52	54	65	52	48	58	64	80	84	83	84	86	48	96	78	24	
15	90	92	92	92	93	96	96	93	86	74	58	45	39	35	31	30	33	36	43	50	57	61	67	70	30	96	65	24	
16	72	74	76	78	79	80	79	73	63	52	44	38	34	32	31	27	26	29	36	45	53	57	59	61	26	80	54	24	
17	66	69	74	76	78	79	82	77	67	58	50	45	39	34	32	31	31	35	46	55	63	64	65	69	31	82	58	24	
18	73	77	79	80	81	82	82	80	77	72	66	61	56	55	56	56	58	61	74	80	76	79	80	89	55	89	72	24	
19	91	92	95	95	95	95	95	96	95	90	79	75	72	73	82	72	67	68	72	86	90	92	93	92	67	96	86	24	
20	92	92	92	92	92	93	93	93	93	91	89	86	84	79	73	65	52	58	82	79	69	64	73	67	52	93	81	24	
21	71	75	80	83	85	88	89	87	78	77	76	71	70	68	66	60	55	61	75	90	94	95	96	96	55	96	79	24	
22	96	96	96	96	96	95	95	95	95	84	75	63	58	51	47	46	43	42	44	61	84	83	82	84	87	42	96	75	24
23	91	93	93	93	93	94	93	94	86	77	69	62	56	52	50	47	50	50	63	84	91	91	88	88	47	94	77	24	
24	88	87	88	91	93	94	92	91	86	79	70	67	65	61	56	54	54	54	60	65	69	74	79	78	54	94	76	24	
25	80	84	84	85	87	90	89	85	78	71	65	61	55	51	48	44	44	46	57	63	69	71	74	83	44	90	69	24	
26	86	87	91	92	94	95	96	96	84	78	66	57	50	45	43	48	47	54	71	82	87	82	75	79	43	96	74	24	
27	81	85	88	92	94	93	93	91	86	79	68	60	52	48	43	41	40	45	66	80	85	86	83	85	40	94	74	24	
28	89	81	81	88	86	88	88	84	78	64	55	47	41	36	30	30	31	32	40	47	54	57	61	66	30	89	61	24	
29	68	84	90	92	92	93	94	94	77	67	59	51	47	40	30	27	27	33	55	73	75	78	82	83	27	94	67	24	
30	77	75	73	77	77	67	73	77	82	71	58	52	47	43	42	41	42	46	56	60	61	60	66	67	65	41	82	63	24
HOURLY MAX	96	96	96	96	96	96	96	96	95	92	91	86	90	91	91	92	90	90	92	94	95	96	96	96					
HOURLY AVG	81	82	84	86	86	87	88	85	79	72	65	59	54	50	48	46	45	48	57	69	74	76	78	79					

STATUS FLAG CODES

C	- MONTHLY CALIBRATION	Q	- QUALITY ASSURANCE
C1	- REPEAT CALIBRATION	R	- RECOVERY
Y	- MAINTENANCE	X	- MACHINE MALFUNCTION
S	- DAILY ZERO/SPAN CHECK	G	- OUT FOR REPAIR
S1	- REPEAT ZERO/SPAN CHECK	P	- POWER FAILURE


24 HR AVERAGES September 2017



MONTHLY SUMMARY

MINIMUM 1-HR AVERAGE:	26	% @ HOUR	16	ON DAY	16
MAXIMUM 1-HR AVERAGE:	96	% @ HOUR	22	ON DAY	2
MAXIMUM 24-HR AVERAGE:	86	%		ON DAY	19
OPERATIONAL TIME:			720	hrs	
AMD OPERATION UPTIME:			100.0	%	
STANDARD DEVIATION:	20			MONTHLY AVERAGE:	70

RELATIVE HUMIDITY Hourly Averages (RH %)



BAROMETRIC PRESSURE


BAROMETRIC PRESSURE Hourly Averages (BP mbar)

	HR START (MST)	0:00	1:00	2:00	3:00	4:00	5:00	6:00	7:00	8:00	9:00	10:00	11:00	12:00	13:00	14:00	15:00	16:00	17:00	18:00	19:00	20:00	21:00	22:00	23:00	DAILY MIN.	DAILY MAX.	24-HR AVG.	RDGS.
	HR END (MST)	0:59	1:59	2:59	3:59	4:59	5:59	6:59	7:59	8:59	9:59	10:59	11:59	12:59	13:59	14:59	15:59	16:59	17:59	18:59	19:59	20:59	21:59	22:59	23:59				
DAY																													
1	941	941	941	941	941	941	941	941	941	940	940	940	940	940	940	940	940	940	940	940	941	941	942	942	940	942	941	24	
2	942	942	942	942	942	942	942	942	942	941	941	941	942	942	942	942	942	942	942	942	943	943	944	944	941	944	942	24	
3	944	944	945	946	946	946	947	948	949	949	950	950	950	951	951	951	952	952	953	954	955	956	944	956	950	944	956	24	
4	956	956	957	957	957	957	958	958	958	958	957	957	956	956	955	955	954	954	953	953	953	952	952	952	952	958	956	24	
5	952	952	951	951	951	950	950	950	950	949	949	949	949	948	948	947	947	946	947	947	946	946	946	946	946	952	949	24	
6	946	946	946	946	946	946	946	946	946	946	946	946	945	945	944	944	943	943	942	942	942	942	942	942	942	946	944	24	
7	942	941	941	940	940	940	939	939	939	939	938	938	938	937	937	936	937	937	937	937	937	937	937	936	942	938	24		
8	937	937	938	938	938	939	940	941	940	941	941	941	940	940	939	939	939	939	939	939	939	939	939	939	939	937	941	939	24
9	938	938	938	937	937	937	936	936	935	935	935	934	934	934	934	934	934	934	934	934	934	934	935	934	938	935	24		
10	935	936	936	936	937	937	938	939	939	940	940	940	941	941	941	941	942	942	942	941	941	935	942	939	942	939	24		
11	941	940	939	939	938	937	936	936	935	935	935	934	934	934	934	935	935	936	937	939	940	941	934	941	937	24			
12	941	942	943	943	943	944	944	944	945	945	945	945	944	944	944	945	945	945	945	945	946	946	946	946	946	946	944	24	
13	946	946	947	947	947	947	948	948	948	948	948	948	947	947	947	947	948	948	948	948	948	948	948	948	948	947	24		
14	947	948	948	948	948	949	949	949	949	949	949	949	949	949	949	950	950	950	951	951	951	951	947	951	949	24			
15	951	951	951	951	951	952	952	952	952	952	952	951	951	950	950	949	948	948	948	947	947	947	952	950	24				
16	947	946	946	946	945	945	945	945	944	944	944	943	943	942	941	941	940	940	940	939	939	939	947	943	24				
17	938	938	937	937	936	936	936	935	935	934	934	933	933	932	932	931	931	931	931	931	930	930	938	934	24				
18	930	930	929	929	929	929	929	929	929	930	930	930	930	930	930	931	931	931	931	931	931	931	929	931	930	24			
19	932	931	931	931	931	931	931	931	931	930	930	930	929	929	928	928	928	928	929	928	928	928	932	932	930	24			
20	928	928	929	930	930	931	931	932	932	932	933	933	933	933	934	934	936	937	937	938	938	939	939	933	933	24			
21	939	940	940	941	941	942	942	943	944	944	945	945	946	946	946	947	947	948	948	948	948	948	948	948	939	945	24		
22	948	948	948	948	948	948	948	948	948	948	947	947	946	946	945	945	945	945	945	944	944	944	948	947	24				
23	944	944	944	944	944	944	944	944	944	944	944	944	944	944	944	944	944	944	944	943	943	943	945	944	24				
24	943	943	942	942	942	942	942	941	941	941	941	941	940	940	940	939	939	939	939	938	938	938	943	940	24				
25	938	938	939	939	939	939	939	939	939	939	939	939	939	939	939	939	939	939	939	940	940	941	938	941	939	24			
26	941	942	942	942	943	943	944	944	945	945	945	945	945	945	945	945	945	945	945	945	945	945	945	945	944	24			
27	946	946	946	946	947	947	947	947	947	948	948	948	948	948	948	948	948	948	948	949	949	946	949	948	24				
28	949	948	948	948	947	947	947	947	947	946	945	945	944	943	943	942	941	941	940	940	940	939	939	949	944	24			
29	938	938	938	937	937	937	937	937	937	936	936	935	935	934	934	933	933	933	933	932	931	931	931	938	935	24			
30	931	931	931	932	932	933	933	933	933	933	933	933	933	933	933	934	934	935	936	937	938	931	938	933	24				
HOURLY MAX		956	956	957	957	958	958	958	958	958	957	957	956	956	955	955	954	954	953	953	954	955	956						
HOURLY AVG		942	942	942	942	942	942	942	942	942	942	942	942	942	941	941	941	941	941	942	942	942	942	942	942	938	933	24	

STATUS FLAG CODES

C	- MONTHLY CALIBRATION	Q	- QUALITY ASSURANCE
C1	- REPEAT CALIBRATION	R	- RECOVERY
Y	- MAINTENANCE	X	- MACHINE MALFUNCTION
S	- DAILY ZERO/SPAN CHECK	G	- OUT FOR REPAIR
S1	- REPEAT ZERO/SPAN CHECK	P	- POWER FAILURE


24 HR AVERAGES September 2017



MONTHLY SUMMARY

MINIMUM 1-HR AVERAGE:	928	mbar	@ HOUR	13	ON DAY	19
MAXIMUM 1-HR AVERAGE:	958	mbar	@ HOUR	5	ON DAY	4
MAXIMUM 24-HR AVERAGE:	956	mbar			ON DAY	4
OPERATIONAL TIME:				720	hrs	
AMD OPERATION UPTIME:				100.0	%	
STANDARD DEVIATION:	7				MONTHLY AVERAGE:	942 mbar


BAROMETRIC PRESSURE Hourly Averages (BP mbar)



AMBIENT TEMPERATURE


AMBIENT TEMPERATURE Hourly Averages (AmbTPX °C)

HR START (MST)	0:00	1:00	2:00	3:00	4:00	5:00	6:00	7:00	8:00	9:00	10:00	11:00	12:00	13:00	14:00	15:00	16:00	17:00	18:00	19:00	20:00	21:00	22:00	23:00	DAILY MIN.	DAILY MAX.	24-HR AVG.	RDGS.	
HR END (MST)	0:59	1:59	2:59	3:59	4:59	5:59	6:59	7:59	8:59	9:59	10:59	11:59	12:59	13:59	14:59	15:59	16:59	17:59	18:59	19:59	20:59	21:59	22:59	23:59					
DAY																													
1	10.0	8.0	8.1	9.5	9.3	8.1	8.4	10.4	12.9	15.0	15.5	17.2	18.4	17.9	19.0	19.6	19.4	19.2	18.8	16.9	15.7	14.6	13.6	11.9	8.0	19.6	14.1	24	
2	11.2	9.8	9.5	9.3	8.0	7.4	7.4	8.8	10.5	11.7	11.7	12.5	12.6	12.2	12.1	11.9	12.0	12.0	11.9	10.5	9.2	8.6	8.1	7.4	12.6	10.3	24		
3	9.0	8.5	8.0	7.7	7.1	7.0	6.2	6.8	8.5	11.0	14.7	18.2	19.5	18.7	18.4	17.7	18.5	18.1	17.9	13.5	8.3	6.3	4.7	3.5	19.5	11.6	24		
4	2.2	1.3	0.5	-0.2	-0.8	-1.1	-0.9	4.3	11.2	14.8	17.0	18.4	19.2	19.9	20.6	21.1	21.0	20.0	16.7	12.9	9.3	10.0	10.2	9.8	-1.1	21.1	10.7	24	
5	9.5	9.8	9.8	7.1	6.8	7.9	8.2	10.2	13.5	16.3	18.4	20.6	22.6	24.2	25.1	25.7	26.1	25.6	23.7	17.7	12.5	10.9	9.8	9.7	6.8	26.1	15.5	24	
6	8.5	7.9	8.4	8.6	9.7	9.5	9.6	12.5	14.1	16.7	20.5	22.6	24.5	26.4	27.7	28.4	29.0	28.2	26.2	19.0	14.7	16.5	13.8	15.4	7.9	29.0	17.4	24	
7	15.7	15.4	15.8	15.3	14.7	12.1	11.4	14.7	17.5	19.4	20.8	21.8	23.0	24.3	25.7	25.9	25.0	23.3	21.3	19.4	17.8	17.0	16.3	16.4	11.4	25.9	18.8	24	
8	15.4	16.2	16.3	15.5	14.3	15.9	16.2	14.9	13.9	13.7	14.3	14.6	15.6	16.4	17.2	19.0	19.5	18.8	17.1	15.5	14.4	13.0	12.4	12.4	12.4	19.5	15.5	24	
9	12.3	12.7	13.1	12.6	12.6	11.2	11.6	12.7	14.6	15.7	17.4	17.1	16.3	16.7	18.1	18.3	18.7	18.4	17.4	14.9	13.8	12.6	11.1	10.4	10.4	18.7	14.6	24	
10	9.6	9.6	9.5	10.6	10.2	9.3	8.9	9.7	10.5	11.6	13.5	14.3	16.3	16.8	17.6	17.8	17.2	16.1	15.4	14.1	13.2	12.8	11.7	10.7	8.9	17.8	12.8	24	
11	11.0	11.4	10.4	8.8	8.7	8.9	9.2	9.7	10.4	11.6	13.1	14.9	18.3	19.7	19.1	19.8	20.8	19.7	18.4	16.2	14.8	11.5	9.9	8.4	8.4	20.8	13.5	24	
12	7.0	7.4	5.6	4.3	3.9	3.5	2.0	5.6	9.8	13.7	14.5	15.6	16.1	16.3	13.0	11.6	10.2	13.0	11.3	6.4	4.2	3.8	3.5	2.8	2.0	16.3	8.5	24	
13	2.4	2.7	3.8	4.5	3.3	3.4	3.3	4.3	6.7	7.0	6.7	11.8	12.8	14.8	15.5	14.0	15.3	15.3	11.5	6.5	5.7	5.5	5.5	6.0	2.4	15.5	7.8	24	
14	6.0	5.8	4.9	4.3	3.3	2.0	0.8	0.7	3.6	5.1	8.6	10.4	13.1	13.5	11.1	13.4	13.7	11.6	11.2	6.2	5.2	5.3	4.6	3.6	0.7	13.7	7.0	24	
15	2.2	2.2	2.5	2.3	0.0	0.1	1.1	2.2	5.0	8.1	10.4	12.3	13.3	14.2	14.6	14.9	14.3	12.6	9.6	7.8	7.0	6.1	5.7	0.0	14.9	7.6	24		
16	6.0	5.8	5.3	4.7	4.2	3.6	3.6	5.2	8.0	10.7	12.5	14.2	15.5	16.3	17.1	17.4	17.4	16.6	14.2	11.5	10.2	9.4	8.8	8.1	3.6	17.4	10.3	24	
17	7.1	6.2	5.2	4.7	4.4	3.9	3.6	5.1	8.3	11.0	13.5	15.4	17.4	19.2	19.9	20.0	19.8	18.6	15.6	13.0	11.3	10.8	10.6	10.0	3.6	20.0	11.4	24	
18	8.9	7.7	7.1	6.6	6.2	6.0	5.8	5.9	6.9	8.6	10.5	12.5	13.2	13.3	13.0	13.3	13.4	11.2	9.3	8.9	8.2	8.0	7.6	5.8	13.4	9.4	24		
19	7.2	6.3	6.1	6.0	6.1	5.4	4.7	4.2	4.6	5.8	6.9	7.7	9.0	9.5	8.4	9.5	10.0	10.0	9.4	7.9	7.1	7.4	7.5	4.2	10.0	7.2	24		
20	7.6	7.6	7.4	7.0	6.8	6.4	6.4	6.6	6.8	6.7	7.1	7.2	7.2	7.3	8.0	8.8	9.0	8.0	5.5	5.9	6.6	6.3	5.5	5.5	9.0	7.1	24		
21	5.7	5.5	5.9	5.4	4.8	4.2	4.1	4.6	6.1	6.0	5.7	6.7	6.9	7.5	8.2	8.9	10.0	8.3	5.1	1.4	0.3	-0.5	-1.1	-1.1	10.0	5.0	24		
22	-1.9	-2.5	-2.6	-3.4	-3.9	-4.2	-4.3	-2.6	2.7	5.8	8.1	9.4	10.9	11.6	11.8	12.7	12.1	11.6	8.4	3.3	3.1	3.3	3.3	3.1	-4.3	12.7	4.0	24	
23	1.8	1.5	1.4	1.3	0.8	1.2	0.9	0.6	4.3	7.0	9.0	10.7	12.4	13.5	14.1	14.1	13.4	13.7	10.3	5.4	3.0	3.2	3.8	4.2	0.6	14.1	6.3	24	
24	4.4	4.5	4.1	3.4	3.2	3.2	4.0	4.8	6.8	8.7	10.8	12.2	13.3	14.7	16.3	16.7	16.8	15.4	14.3	13.9	13.1	12.6	12.5	11.8	3.2	16.8	10.1	24	
25	10.7	10.0	9.9	9.4	8.6	7.2	7.6	8.7	10.6	12.7	14.2	15.5	17.0	17.8	18.3	18.8	18.3	17.6	14.8	13.0	11.6	11.0	10.3	8.0	7.2	18.8	12.6	24	
26	7.1	7.0	5.5	4.3	2.5	2.0	1.2	2.6	8.8	10.7	13.0	15.0	16.6	17.7	18.1	17.4	17.4	16.0	12.5	10.7	11.6	12.4	12.6	10.5	1.2	18.1	10.6	24	
27	10.1	9.1	8.0	6.1	6.1	6.7	7.0	8.5	10.6	12.6	15.3	17.2	19.5	20.3	22.3	22.0	21.8	20.3	15.1	11.2	9.6	9.5	10.2	8.9	6.1	22.3	12.8	24	
28	8.4	10.4	10.2	7.1	8.1	7.9	7.7	9.1	10.7	15.0	17.6	19.9	22.0	23.3	24.2	24.4	24.3	23.5	20.5	18.1	15.9	15.1	14.2	13.2	7.1	24.4	15.5	24	
29	12.3	7.5	5.5	4.7	4.6	3.4	2.5	4.7	10.9	13.8	16.5	19.2	20.5	22.3	24.1	24.2	24.1	22.9	16.4	11.4	10.2	9.1	8.1	7.4	2.5	24.2	12.8	24	
30	8.8	9.8	10.6	11.1	11.8	12.3	11.1	10.4	10.0	11.1	11.9	13.0	14.7	15.7	16.3	16.8	16.4	14.6	13.2	12.4	10.9	8.9	8.3	7.3	16.8	12.0	24		
HOURLY MAX		15.7	16.2	16.3	15.5	14.7	15.9	16.2	14.9	17.5	19.4	20.8	22.6	24.5	26.4	27.7	28.4	29.0	28.2	26.2	19.4	17.8	17.0	16.3	16.4				
HOURLY AVG		7.9	7.5	7.2	6.6	6.2	5.8	5.6	6.9	9.3	11.3	13.0	14.6	15.9	16.7	17.2	17.5	17.5	16.8	14.6	11.6	10.0	9.4	8.8	8.3				

24 HR AVERAGES September 2017

MONTHLY SUMMARY

MINIMUM 1-HR AVERAGE:	-4.3	°C	@ HOUR	6	ON DAY	22
MAXIMUM 1-HR AVERAGE:	29.0	°C	@ HOUR	16	ON DAY	6
MAXIMUM 24-HR AVERAGE:	18.8	°C			ON DAY	7
OPERATIONAL TIME:				720	hrs	
AMD OPERATION UPTIME:				100.0	%	
STANDARD DEVIATION:	6.0			MONTLY AVERAGE:	11.1	°C

AMBIENT TEMPERATURE Hourly Averages (AmbTPX °C)




STATION TEMPERATURE

STATION TEMPERATURE Hourly Averages (StnTPX °C)

	0:00	1:00	2:00	3:00	4:00	5:00	6:00	7:00	8:00	9:00	10:00	11:00	12:00	13:00	14:00	15:00	16:00	17:00	18:00	19:00	20:00	21:00	22:00	23:00	DAILY MIN.	DAILY MAX.	24-HR AVG.	RDGS.		
HR START (MST)	0:00	1:00	2:00	3:00	4:00	5:00	6:00	7:00	8:00	9:00	10:00	11:00	12:00	13:00	14:00	15:00	16:00	17:00	18:00	19:00	20:00	21:00	22:00	23:00						
HR END (MST)	0:59	1:59	2:59	3:59	4:59	5:59	6:59	7:59	8:59	9:59	10:59	11:59	12:59	13:59	14:59	15:59	16:59	17:59	18:59	19:59	20:59	21:59	22:59	23:59						
DAY																														
1	22.5	22.6	22.7	22.9	22.6	22.7	22.7	22.2	22.0	21.9	22.0	21.9	22.0	22.1	22.1	22.2	22.3	22.3	22.4	22.4	21.9	22.9	22.3	24						
2	22.5	22.5	22.7	22.7	22.6	22.7	22.9	22.5	22.7	22.4	22.5	22.4	22.3	22.4	22.3	22.4	22.5	22.7	22.7	22.7	22.3	22.9	22.5	24						
3	22.5	22.6	22.7	22.7	22.9	22.7	22.6	22.7	22.2	22.1	21.9	22.0	21.8	22.0	22.1	22.1	22.2	22.3	22.7	23.0	21.8	23.0	22.4	24						
4	22.3	21.9	21.3	20.8	20.3	19.9	19.5	19.4	20.2	21.8	22.2	22.0	21.9	21.9	22.0	22.1	22.2	22.1	22.2	22.4	22.5	22.7	19.4	22.7	21.6	24				
5	22.7	22.7	22.7	22.6	22.9	22.7	22.9	22.6	22.3	22.1	21.9	21.7	21.7	21.6	21.8	21.8	22.0	21.9	22.0	22.2	22.5	22.5	21.6	22.9	22.2	24				
6	22.6	22.5	22.6	22.7	22.6	22.6	22.7	22.5	22.3	22.1	21.8	21.6	21.7	21.5	21.7	21.6	21.8	21.9	21.9	22.1	22.2	22.3	22.4	21.5	22.7	22.1	24			
7	22.4	22.4	22.4	22.3	22.4	22.4	22.6	22.3	22.1	22.0	21.9	21.8	21.8	21.7	21.8	21.8	22.0	22.1	22.1	22.2	22.2	22.3	21.7	22.6	22.1	24				
8	22.3	22.4	22.3	22.2	22.4	22.3	22.3	22.2	22.4	22.3	22.2	22.2	22.2	22.1	22.1	22.0	22.0	22.2	22.4	22.6	22.4	22.0	22.6	22.3	24					
9	22.3	22.4	22.5	22.4	22.5	22.6	22.5	22.3	22.3	22.2	22.1	22.2	22.2	22.0	22.0	22.1	22.2	22.3	22.3	22.5	22.5	22.0	22.6	22.3	24					
10	22.7	22.7	22.7	22.9	22.6	22.8	22.7	22.8	22.4	22.4	22.2	22.2	22.1	22.3	22.1	22.2	22.1	22.3	22.4	22.5	22.5	22.1	22.9	22.5	24					
11	22.6	22.5	22.4	22.6	22.7	22.5	22.8	22.5	22.6	22.4	22.1	22.0	22.1	22.1	22.1	22.1	22.2	22.3	22.4	22.6	22.5	22.0	22.8	22.4	24					
12	22.7	22.7	22.9	22.8	22.4	22.1	21.6	21.3	21.7	22.5	22.3	21.9	22.0	22.1	22.1	22.2	20.5	21.0	20.9	20.2	20.5	20.8	21.0	20.2	22.9	21.7	24			
13	21.0	20.9	20.9	21.0	20.3	20.7	20.9	21.0	20.5	20.7	20.7	20.3	20.3	20.1	20.6	21.0	21.9	22.5	22.1	20.4	20.6	20.7	20.9	20.5	20.1	22.5	20.9	24		
14	20.7	20.8	20.6	20.9	21.0	21.1	21.0	20.7	20.6	20.9	20.2	20.4	20.8	20.7	21.2	22.2	21.6	21.3	20.4	20.6	21.0	21.1	20.9	20.2	22.2	20.9	24			
15	20.5	20.4	20.2	20.1	19.9	19.7	19.4	19.2	20.1	22.4	23.1	22.8	22.3	22.1	22.4	22.2	22.3	22.6	23.0	22.2	20.8	21.1	20.6	19.2	23.1	21.3	24			
16	20.8	21.0	21.0	20.9	20.8	20.7	20.7	20.9	20.1	20.4	21.1	21.9	22.7	22.2	22.1	22.2	23.2	22.7	21.3	20.2	20.9	21.0	20.1	23.2	21.3	24				
17	20.6	20.7	21.0	21.3	20.6	21.0	21.1	21.1	20.6	22.1	22.4	21.9	21.5	21.5	21.7	21.4	21.6	22.5	23.4	22.6	21.8	20.8	20.9	20.6	23.4	21.5	24			
18	20.8	20.9	21.0	20.9	20.8	20.8	21.1	20.7	20.8	20.6	20.7	20.9	21.7	22.2	22.4	22.5	22.1	22.1	22.7	21.8	21.5	21.7	22.0	21.5	20.6	22.7	21.4	24		
19	21.8	21.7	22.1	21.4	22.0	21.5	21.8	21.9	21.9	21.4	21.8	21.5	21.5	21.9	21.6	21.8	21.5	21.5	21.9	21.4	21.8	21.6	21.4	22.1	21.7	24				
20	21.6	21.9	21.4	22.0	21.4	21.7	21.9	22.0	21.4	21.7	22.2	22.3	22.1	22.1	22.4	21.8	21.1	22.1	22.0	22.1	22.2	22.3	21.7	21.1	22.4	21.9	24			
21	22.0	22.0	22.1	22.2	22.1	22.1	22.0	22.2	22.0	21.8	22.2	21.7	22.1	21.9	22.1	21.9	22.3	21.6	21.4	20.9	20.5	20.0	20.0	22.3	21.8	24				
22	19.4	18.9	19.4	19.2	19.1	19.4	19.4	18.6	19.3	20.4	21.7	21.9	22.2	22.8	23.7	22.7	23.2	23.5	22.2	22.3	22.2	21.8	18.6	23.7	21.0	24				
23	21.3	21.0	20.7	20.4	20.1	20.0	19.9	19.7	19.8	20.6	21.7	22.0	22.1	22.7	23.0	22.5	23.4	22.5	23.2	22.1	22.2	22.2	21.9	19.7	23.4	21.5	24			
24	21.6	21.6	21.5	21.3	21.2	21.2	21.2	21.2	21.5	22.1	21.9	21.9	22.4	23.2	22.6	22.1	21.9	22.6	23.1	23.7	23.6	23.4	23.1	22.7	21.2	23.7	22.2	24		
25	22.2	21.8	22.3	22.0	22.0	22.3	21.9	22.3	21.6	21.7	22.5	23.5	22.3	22.7	23.9	23.0	23.5	23.7	24.8	24.4	23.5	22.6	22.0	22.2	21.6	24.8	22.7	24		
26	22.2	22.0	22.0	22.1	21.8	21.5	21.1	20.7	21.0	21.8	21.8	22.6	23.8	24.3	23.5	24.3	24.6	25.5	25.2	24.1	23.5	23.2	23.0	22.2	20.7	25.5	22.8	24		
27	22.1	22.0	22.4	21.8	22.0	22.3	22.1	21.9	22.3	22.8	24.2	21.9	23.9	24.8	25.4	25.8	21.1	22.5	23.3	22.6	21.8	21.7	21.7	21.1	25.8	22.7	24			
28	22.0	21.7	21.9	22.0	21.6	21.9	21.9	22.0	22.0	22.5	22.8	21.7	21.7	21.5	21.0	21.2	21.2	21.2	21.8	22.6	23.2	23.0	22.5	21.9	21.0	23.2	22.0	24		
29	22.0	21.7	22.1	22.2	22.1	22.0	21.7	21.4	21.8	21.9	22.5	22.3	21.7	21.8	21.2	21.0	21.3	21.7	22.2	22.7	21.8	22.1	21.9	21.8	21.0	22.7	21.9	24		
30	22.1	21.9	22.2	21.7	22.1	21.9	21.8	22.1	21.8	21.7	21.3	21.5	22.1	23.0	22.4	22.1	22.5	23.4	23.0	22.4	21.6	21.6	22.0	21.3	23.4	22.1	24			
HOURLY MAX		22.7	22.7	22.9	22.9	22.7	22.9	22.6	22.8	23.1	24.2	23.8	24.3	24.8	25.4	25.8	25.8	25.5	25.2	24.4	23.6	23.4	23.1	22.8						
HOURLY AVG		21.8	21.8	21.8	21.8	21.7	21.6	21.5	21.5	21.8	21.9	22.0	21.9	22.1	22.2	22.2	22.2	22.5	22.3	22.1	22.0	22.0	21.8							

24 HR AVERAGES September 2017



MONTHLY SUMMARY

MINIMUM 1-HR AVERAGE:	18.6	°C	@ HOUR	8	ON DAY	22


<tbl_r cells="7" ix="5" max



PEACE RIVER AREA MONITORING PROGRAM COMMITTEE

Three Creeks 842b Station - September 2017

STATION TEMPERATURE Hourly Averages (StnTPX °C)



APPENDIX II
EQUIPMENT CALIBRATION RESULTS

SULPHUR DIOXIDE



API 100A Sulphur Dioxide Analyzer Calibration


Date:	September 15, 2017	Barometer/B.P./units:	n/a	28.09	inHg
Company/Airshed:	PRAMP	Thermometer/Station Temp:	n/a	21.69	°C
Location/Station Name:	842b	Weather Conditions:	Sunny		
Parameter:	Sulphur Dioxide	Calibration Purpose:	routine monthly		
Start Time 24 hr. (mst):	9:44	Performed By/Reviewer:	Raja Abid		Tom Bourque
End Time 24 hr. (mst):	14:29	Cal Gas Expiry Date:	May 23, 2019		
Calibration Method:	Gas Dilution	Converter Model & s/n (if applicable):	n/a		

Analyzer:	
ID# or Serial Number:	838
Last Calibration Date:	August 10, 2017
Previous C.F.:	1.000
Range ppb:	500
As Found C.F.:	1.011
New C.F.:	1.000

Calibration Standards:	Standard Calibration Points for Ranges
Low Flow Meter ID/Expiry Date:	Definer Low 129069 expires February 5, 2018
High Flow Meter ID/Expiry Date:	Definer High 128686 expires February 5, 2018
Calibrator ID/Expiry Date:	API id# 690 expires March 17, 2018
Cal Gas Cylinder I.D. # :	LL 119513
Cal Gas Conc. (ppm):	50.6
Point	ppb
High	380
Mid	180
Low	90

ALL POINTS ARE 15 MINUTES OF STABILITY AS OF SEPTEMBER 23, 2015					
Calibrator Flow Rates (cc/min)			Calculated Concentration (ppb):	Indicated Concentration (ppb):	Correction Factors (C.F.):
Point	Diluent	Cal Gas	Total		
as found zero	5947	0.00	5947	0.0	0.2
as found high	5906	43.26	5949	367.9	364.0
adjusted zero	4900	0.00	4900	0.0	0.0
adjusted high	5906	43.26	5949	367.9	368.0
mid	5919	20.74	5940	176.7	175.5
low	5932	10.24	5942	87.2	86.0
calibrator zero	5947	0.00	5947	0.0	0.0
Average C.F.=					1.007

Linear Regression/Calibration Results:					
LIMITS					
Correlation Coefficient = 1.000 > or = 0.995					
Slope = 0.999 0.95-1.05					
b (Intercept as % of full scale)= 0.15% ± 3% F.S.					
% change in C.F. from last cal= -1.14% ± 10%					




As found:	As left:
Slope: 1.024	Slope: 1.034
Offset: 19.6	Offset: 19.9
Hyps: 685	Hyps: 686
Dcps: 2545	Dcps: 2544
Rcell Temp: 49.5	Rcell Temp: 49.5
Box Temp: 29.6	Box Temp: 31.0
Pmt Temp: 7.3	Pmt Temp: 7.2
Izs Temp: 60.1	Izs Temp: 60.2
Converter Temp: n/a	Converter Temp: n/a
Pres: 26.7	Pres: 26.6
Samp Fl: 646	Samp Fl: 643
Pmt: 53.6	Pmt: 51.5
Uv Lamp: 2091.1	Uv Lamp: 2053.1
Lamp Ratio: 84.2	Lamp Ratio: 82.7
Expected Value: 233.0	Expected Value: 237.1

Comments:	
The analyzer sample inlet filter was changed.	The analyzer cooling fan filter(s) were cleaned.
	The manifold blower was found to be working normally.

Station temperature recorded from data logger.
 Barometric Pressure recorded from Environment Canada site.
 Sample line disconnected by mistake after as-found high. Adjusted high restarted at 11:39.
 Flow measurements after mid-point

SO2[ppb] Station: PRAMP_842 Daily: 17/09/15 Type: AVG 1 Min. [1 Min.]



— SO2[ppb]

TOTAL REDUCED SULPHUR



Thermo 43I-TLE Total Reduced Sulphur Analyzer Calibration


Date:	September 15, 2017	Barometer/B.P./units:	n/a	28.09	inHg
Company/Airshed:	PRAMP	Thermometer/Station Temp:	n/a	21.69	°C
Location/Station Name:	842b	Weather Conditions:	Sunny		
Parameter:	Total Reduced Sulphur	Calibration Purpose:	routine monthly		
Start Time 24 hr. (mst):	11:07	Performed By/Reviewer:	Raja Abid		Tom Bourque
End Time 24 hr. (mst):	15:47	Cal Gas Expiry Date:	December 1, 2018		
Calibration Method:	Gas Dilution	Converter Model & s/n (if applicable):	CD NOVA-CDN 101s/n 534		

Analyzer:	
ID# or Serial Number:	1162460023
Last Calibration Date:	August 10, 2017
Previous C.F.:	1.001
Range ppb:	100
As Found C.F.:	0.984
New C.F.:	0.999

Calibration Standards:	Standard Calibration Points for Ranges	SO2 Scrubber Check (10 minutes):
Low Flow Meter ID/Expiry Date: Definer Low 129069 expires February 5, 2018	Point	Start/End Time 24 hr.: 11:51-12:07
High Flow Meter ID/Expiry Date: Definer High 128686 expires February 5, 2018	ppb	SO2 Analyzer Range: 500
Calibrator ID/Expiry Date: API id# 829 expires January 27, 2018	High	Target Concentration (ppb): 380
Cal Gas Cylinder I.D. #: BLM 001927	Mid	As Found Zero: 0.2
Cal Gas Conc. (ppm): 10.3	Low	Analyzer Response: (ppb): -0.2
		Zero Corrected Result (ppb): -0.4

ALL POINTS ARE 15 MINUTES OF STABILITY AS OF SEPTEMBER 23, 2015					
Calibrator Flow Rates (cc/min)			Calculated Concentration (ppb):	Indicated Concentration (ppb):	Correction Factors (C.F.):
Point	Diluent	Cal Gas	Total		
as found zero	7434	0.00	7434	0.0	n/a
as found high	7408	57.30	7466	79.1	0.984
adjusted zero	7434	0.00	7434	0.0	n/a
adjusted high	7409	57.30	7466	79.1	0.999
mid	7446	27.45	7473	37.8	1.006
low	7456	13.82	7470	19.1	1.025
calibrator zero	7434	0.00	7434	0.0	n/a
Average C.F.:					1.010


Linear Regression/Calibration Results:					
LIMITS					
Correlation Coefficient =	1.000	> or = 0.995			
Slope =	0.998	0.95-1.05			
b (Intercept as % of full scale)=	0.24%	± 3% F.S.			
% change in C.F. from last cal=	1.74%	± 10%			



As found:		As left:	
Bkg:	2.50	Bkg:	2.70
Coef:	0.872	Coef:	0.853
Pmt:	-725	Pmt:	-725.2
Flash:	978	Flash:	976
Internal:	30.3	Internal:	30.4
Chamber:	44.8	Chamber:	45.0
Perm Oven Gas:	45.00	Perm Oven Gas:	45.00
Perm Oven Heater:	44.11	Perm Oven Heater:	44.11
Pressure:	666.3	Pressure:	669.9
Sample Flow:	0.408	Sample Flow:	0.410
Lamp Intensity:	89	Lamp Intensity:	89
Converter:	850	Converter:	850
Converter Set:	850	Converter Set:	850
Averaging Time:	120	Averaging Time:	120
Expected Value:	59.0	Expected Value:	57.9

Comments:	The analyzer sample inlet filter was changed.	The analyzer cooling fan filter(s) were cleaned.	The manifold blower was found to be working normally.
Station temperature recorded from data logger. Barometric Pressure recorded from Environment Canada site. As-found restarted at 12:50 due to low gas pressure. Flow measurements after mid-point			

TRS[ppb] Station: PRAMP_842 Daily: 17/09/15 Type: AVG 1 Min. [1 Min.]




TOTAL HYDROCARBON




Thermo 55i Methane/Non-Methane Analyzer Calibration

<p>Date: September 18, 2017 Company/Airshed: PRAMP Location/Station Name: 842b Parameter: CH₄ / NMHC / THC Start/End Time 24 hr. (mst): 17:07 / 19:52 Calibration Method: Gas Dilution</p>	<p>Barometer/B.P./units: Brunton 05535 expires December 5, 2017 929 millibars Thermometer/Station Temp: F.S. 160459244 expires May 18, 2018 22 °C Weather Conditions: Mix of sun and clouds Calibration Purpose: installation Performed By/Reviewer: Chris Wesson Tom Bourque Cal Gas Expiry Date: November 25, 2023</p>																																																																																																																														
Analyzer: <table border="1" style="width: 100%; border-collapse: collapse;"> <tr> <td>ID# or Serial Number: 1505664392</td> <td colspan="3">Correction Factors:</td> </tr> <tr> <td>Measured Flow: 1.1 L/min</td> <td>CH₄</td> <td>As Found C.F.:</td> <td>New C.F.:</td> </tr> <tr> <td>Last Calibration Date: n/a</td> <td>n/a</td> <td>n/a</td> <td>0.999</td> </tr> <tr> <td>Range ppm: 20 CH₄/20 NMHC/40 THC</td> <td>NMHC</td> <td>n/a</td> <td>0.995</td> </tr> <tr> <td>CH₄ expressed as C₃H₈:</td> <td>THC</td> <td>n/a</td> <td>0.997</td> </tr> </table>		ID# or Serial Number: 1505664392	Correction Factors:			Measured Flow: 1.1 L/min	CH ₄	As Found C.F.:	New C.F.:	Last Calibration Date: n/a	n/a	n/a	0.999	Range ppm: 20 CH ₄ /20 NMHC/40 THC	NMHC	n/a	0.995	CH ₄ expressed as C ₃ H ₈ :	THC	n/a	0.997																																																																																																										
ID# or Serial Number: 1505664392	Correction Factors:																																																																																																																														
Measured Flow: 1.1 L/min	CH ₄	As Found C.F.:	New C.F.:																																																																																																																												
Last Calibration Date: n/a	n/a	n/a	0.999																																																																																																																												
Range ppm: 20 CH ₄ /20 NMHC/40 THC	NMHC	n/a	0.995																																																																																																																												
CH ₄ expressed as C ₃ H ₈ :	THC	n/a	0.997																																																																																																																												
Calibration Standards: <table border="1" style="width: 100%; border-collapse: collapse;"> <tr> <td>Low Flow Meter ID/Expiry Date: Defender Low 153358 expires January 19, 2018</td> <td colspan="3">Standard Calibration Points for Analyzer Range of 20/20/40 ppm</td> </tr> <tr> <td>High Flow Meter ID/Expiry Date: Defender High 152571 expires January 19, 2018</td> <td>Point</td> <td>CH₄</td> <td>NMHC</td> </tr> <tr> <td>Calibrator ID/Expiry Date: Sabio id# 17100415 expires May 16, 2018</td> <td>High</td> <td>13.00</td> <td>13.00</td> </tr> <tr> <td>Cal Gas Cylinder I.D. #: LL86139</td> <td>Mid</td> <td>7.00</td> <td>7.00</td> </tr> <tr> <td>CH₄ Cylinder Conc.= 599.0</td> <td>Low</td> <td>3.00</td> <td>3.00</td> </tr> <tr> <td>CH₄ expressed as C₃H₈= 580.3</td> <td>THC</td> <td>26.00</td> <td>14.00</td> </tr> </table>		Low Flow Meter ID/Expiry Date: Defender Low 153358 expires January 19, 2018	Standard Calibration Points for Analyzer Range of 20/20/40 ppm			High Flow Meter ID/Expiry Date: Defender High 152571 expires January 19, 2018	Point	CH ₄	NMHC	Calibrator ID/Expiry Date: Sabio id# 17100415 expires May 16, 2018	High	13.00	13.00	Cal Gas Cylinder I.D. #: LL86139	Mid	7.00	7.00	CH ₄ Cylinder Conc.= 599.0	Low	3.00	3.00	CH ₄ expressed as C ₃ H ₈ = 580.3	THC	26.00	14.00																																																																																																						
Low Flow Meter ID/Expiry Date: Defender Low 153358 expires January 19, 2018	Standard Calibration Points for Analyzer Range of 20/20/40 ppm																																																																																																																														
High Flow Meter ID/Expiry Date: Defender High 152571 expires January 19, 2018	Point	CH ₄	NMHC																																																																																																																												
Calibrator ID/Expiry Date: Sabio id# 17100415 expires May 16, 2018	High	13.00	13.00																																																																																																																												
Cal Gas Cylinder I.D. #: LL86139	Mid	7.00	7.00																																																																																																																												
CH ₄ Cylinder Conc.= 599.0	Low	3.00	3.00																																																																																																																												
CH ₄ expressed as C ₃ H ₈ = 580.3	THC	26.00	14.00																																																																																																																												
ALL POINTS ARE 15 MINUTES OF STABILITY AS OF SEPTEMBER 23, 2015																																																																																																																															
Calibrator Flow Rates (cc/min) <table border="1" style="width: 100%; border-collapse: collapse;"> <tr> <th>Point</th> <th>Diluent</th> <th>Cal Gas</th> <th>Total Flow</th> <th>Calculated CH₄ (ppm)</th> <th>Calculated NMHC (ppm)</th> <th>Calculated THC (ppm)</th> <th>Indicated CH₄ (ppm)</th> <th>Indicated NMHC (ppm)</th> <th>Indicated THC (ppm)</th> <th>Correction Factors:</th> </tr> <tr> <td>adjusted zero</td> <td>2506</td> <td>0.00</td> <td>2506</td> <td>0.00</td> <td>0.00</td> <td>0.00</td> <td>0.00</td> <td>0.00</td> <td>0.00</td> <td>CH₄</td> </tr> <tr> <td>adjusted high</td> <td>2445</td> <td>60.62</td> <td>2506</td> <td>14.49</td> <td>14.04</td> <td>28.53</td> <td>14.50</td> <td>14.10</td> <td>28.60</td> <td>NMHC</td> </tr> <tr> <td>mid</td> <td>2476</td> <td>30.62</td> <td>2507</td> <td>7.32</td> <td>7.09</td> <td>14.40</td> <td>7.29</td> <td>7.10</td> <td>14.39</td> <td>THC</td> </tr> <tr> <td>low</td> <td>2494</td> <td>15.85</td> <td>2510</td> <td>3.78</td> <td>3.66</td> <td>7.45</td> <td>3.74</td> <td>3.70</td> <td>7.44</td> <td></td> </tr> <tr> <td>calibrator zero</td> <td>2506</td> <td>0.00</td> <td>2506</td> <td>0.00</td> <td>0.00</td> <td>0.00</td> <td>0.00</td> <td>0.00</td> <td>0.00</td> <td>Average C.F.=</td> </tr> </table>				Point	Diluent	Cal Gas	Total Flow	Calculated CH ₄ (ppm)	Calculated NMHC (ppm)	Calculated THC (ppm)	Indicated CH ₄ (ppm)	Indicated NMHC (ppm)	Indicated THC (ppm)	Correction Factors:	adjusted zero	2506	0.00	2506	0.00	0.00	0.00	0.00	0.00	0.00	CH ₄	adjusted high	2445	60.62	2506	14.49	14.04	28.53	14.50	14.10	28.60	NMHC	mid	2476	30.62	2507	7.32	7.09	14.40	7.29	7.10	14.39	THC	low	2494	15.85	2510	3.78	3.66	7.45	3.74	3.70	7.44		calibrator zero	2506	0.00	2506	0.00	0.00	0.00	0.00	0.00	0.00	Average C.F.=																																																										
Point	Diluent	Cal Gas	Total Flow	Calculated CH ₄ (ppm)	Calculated NMHC (ppm)	Calculated THC (ppm)	Indicated CH ₄ (ppm)	Indicated NMHC (ppm)	Indicated THC (ppm)	Correction Factors:																																																																																																																					
adjusted zero	2506	0.00	2506	0.00	0.00	0.00	0.00	0.00	0.00	CH ₄																																																																																																																					
adjusted high	2445	60.62	2506	14.49	14.04	28.53	14.50	14.10	28.60	NMHC																																																																																																																					
mid	2476	30.62	2507	7.32	7.09	14.40	7.29	7.10	14.39	THC																																																																																																																					
low	2494	15.85	2510	3.78	3.66	7.45	3.74	3.70	7.44																																																																																																																						
calibrator zero	2506	0.00	2506	0.00	0.00	0.00	0.00	0.00	0.00	Average C.F.=																																																																																																																					
Linear Regression/Calibration Results:																																																																																																																															
<table border="1" style="width: 100%; border-collapse: collapse;"> <tr> <td>Correlation Coefficient = 1.000</td> <td>CH₄</td> <td>NMHC</td> <td>THC</td> <td>LIMITS</td> </tr> <tr> <td>Slope = 1.002</td> <td>1.000</td> <td>1.004</td> <td>1.003</td> <td>> or = 0.995</td> </tr> <tr> <td>b (Intercept as % of full scale)= -0.12%</td> <td>0.02%</td> <td>-0.05%</td> <td>-</td> <td>0.95-1.05</td> </tr> <tr> <td>% change in C.F. from last cal= n/a</td> <td>n/a</td> <td>n/a</td> <td>n/a</td> <td>± 3% F.S.</td> </tr> </table>				Correlation Coefficient = 1.000	CH ₄	NMHC	THC	LIMITS	Slope = 1.002	1.000	1.004	1.003	> or = 0.995	b (Intercept as % of full scale)= -0.12%	0.02%	-0.05%	-	0.95-1.05	% change in C.F. from last cal= n/a	n/a	n/a	n/a	± 3% F.S.																																																																																																								
Correlation Coefficient = 1.000	CH ₄	NMHC	THC	LIMITS																																																																																																																											
Slope = 1.002	1.000	1.004	1.003	> or = 0.995																																																																																																																											
b (Intercept as % of full scale)= -0.12%	0.02%	-0.05%	-	0.95-1.05																																																																																																																											
% change in C.F. from last cal= n/a	n/a	n/a	n/a	± 3% F.S.																																																																																																																											
As Left Instrument Diagnostics:																																																																																																																															
<table border="1" style="width: 100%; border-collapse: collapse;"> <tr> <td>Interface Board Voltages:</td> <td>Bias Supply: -293.9</td> <td>Calibration History cnt'd:</td> <td>NM Peak Area: n/a</td> </tr> <tr> <td>Temperatures:</td> <td>Detector Oven: 175.1</td> <td>Crucial Settings:</td> <td>Methane Start: 8.0</td> </tr> <tr> <td></td> <td>Filter: 175.0</td> <td></td> <td>Methane End: 16.0</td> </tr> <tr> <td></td> <td>Column Oven: 75.1</td> <td></td> <td>Backflush: 18.0</td> </tr> <tr> <td></td> <td>Internal: 33.4</td> <td></td> <td>NMHV Start: 25.0</td> </tr> <tr> <td>Cylinder Pressures/reg.:</td> <td>Carrier: 650 50</td> <td></td> <td>NMHV End: 54.0</td> </tr> <tr> <td></td> <td>Fuel: 900 50</td> <td></td> <td>Date: 18Sep2017</td> </tr> <tr> <td></td> <td>Span Gas: 2000 14</td> <td></td> <td>Time: 16:50</td> </tr> <tr> <td></td> <td>Zero Air Generator: 45</td> <td></td> <td>CH₄ PK HT: 0</td> </tr> <tr> <td>Internal Pressures:</td> <td>Carrier: 35.1</td> <td></td> <td>CH₄ RT: 13.4</td> </tr> <tr> <td></td> <td>Fuel: 46.4</td> <td></td> <td>CH₄ Baseline: 50</td> </tr> <tr> <td></td> <td>Air: 23.7</td> <td></td> <td>CH₄ LOD: 9</td> </tr> <tr> <td>FID Status:</td> <td>Status: LIT</td> <td></td> <td>CH₄ SD: 3</td> </tr> <tr> <td></td> <td>Counts: ~28000</td> <td></td> <td>CH₄ CONC: 0.00</td> </tr> <tr> <td></td> <td>Flame: 365.8</td> <td></td> <td>NM PK HT: 0</td> </tr> <tr> <td></td> <td>Det Base: 175.1</td> <td></td> <td>NM Peak Area: 0</td> </tr> <tr> <td>Flame and Power Stats:</td> <td>Last Power On: 18Sep2017@15:50</td> <td></td> <td>NM CONC: 0.00</td> </tr> <tr> <td></td> <td>Flameouts: 1</td> <td></td> <td>NM Base Start: 48</td> </tr> <tr> <td></td> <td>Det Oven at Start: 114.2</td> <td></td> <td>NM Base End: 61</td> </tr> <tr> <td></td> <td>Col Oven at Start: 61.2</td> <td></td> <td>NM LOD: 11</td> </tr> <tr> <td>Calibration History:</td> <td>Time: 01Jan1970@00:00</td> <td></td> <td>NM Start IDX: 52</td> </tr> <tr> <td></td> <td>Type: n/a</td> <td></td> <td>NM End IDX: 50</td> </tr> <tr> <td></td> <td>Status: n/a</td> <td></td> <td>NM Max Slope: 1.4e+00</td> </tr> <tr> <td></td> <td>Check/Adjust: n/a</td> <td></td> <td>NM Min Slope: -7.0e-01</td> </tr> <tr> <td></td> <td>CH₄ Span Conc: n/a</td> <td></td> <td>NM PT Count: 0</td> </tr> <tr> <td></td> <td>CH₄ SP Ratio: n/a</td> <td></td> <td>Previous CH4: 10.32</td> </tr> <tr> <td></td> <td>CH₄ RT: n/a</td> <td></td> <td>Previous NMHC: 11.1</td> </tr> <tr> <td></td> <td>CH₄ PK IDX: n/a</td> <td></td> <td>Previous THC: 21.44</td> </tr> <tr> <td></td> <td>CH₄ PK HT: n/a</td> <td></td> <td>New CH4: 10.32</td> </tr> <tr> <td></td> <td>NM Span Conc: n/a</td> <td></td> <td>New NMHC: 11.10</td> </tr> <tr> <td></td> <td>NM SP Ratio: n/a</td> <td></td> <td>New THC: 21.44</td> </tr> </table>				Interface Board Voltages:	Bias Supply: -293.9	Calibration History cnt'd:	NM Peak Area: n/a	Temperatures:	Detector Oven: 175.1	Crucial Settings:	Methane Start: 8.0		Filter: 175.0		Methane End: 16.0		Column Oven: 75.1		Backflush: 18.0		Internal: 33.4		NMHV Start: 25.0	Cylinder Pressures/reg.:	Carrier: 650 50		NMHV End: 54.0		Fuel: 900 50		Date: 18Sep2017		Span Gas: 2000 14		Time: 16:50		Zero Air Generator: 45		CH ₄ PK HT: 0	Internal Pressures:	Carrier: 35.1		CH ₄ RT: 13.4		Fuel: 46.4		CH ₄ Baseline: 50		Air: 23.7		CH ₄ LOD: 9	FID Status:	Status: LIT		CH ₄ SD: 3		Counts: ~28000		CH ₄ CONC: 0.00		Flame: 365.8		NM PK HT: 0		Det Base: 175.1		NM Peak Area: 0	Flame and Power Stats:	Last Power On: 18Sep2017@15:50		NM CONC: 0.00		Flameouts: 1		NM Base Start: 48		Det Oven at Start: 114.2		NM Base End: 61		Col Oven at Start: 61.2		NM LOD: 11	Calibration History:	Time: 01Jan1970@00:00		NM Start IDX: 52		Type: n/a		NM End IDX: 50		Status: n/a		NM Max Slope: 1.4e+00		Check/Adjust: n/a		NM Min Slope: -7.0e-01		CH ₄ Span Conc: n/a		NM PT Count: 0		CH ₄ SP Ratio: n/a		Previous CH4: 10.32		CH ₄ RT: n/a		Previous NMHC: 11.1		CH ₄ PK IDX: n/a		Previous THC: 21.44		CH ₄ PK HT: n/a		New CH4: 10.32		NM Span Conc: n/a		New NMHC: 11.10		NM SP Ratio: n/a		New THC: 21.44
Interface Board Voltages:	Bias Supply: -293.9	Calibration History cnt'd:	NM Peak Area: n/a																																																																																																																												
Temperatures:	Detector Oven: 175.1	Crucial Settings:	Methane Start: 8.0																																																																																																																												
	Filter: 175.0		Methane End: 16.0																																																																																																																												
	Column Oven: 75.1		Backflush: 18.0																																																																																																																												
	Internal: 33.4		NMHV Start: 25.0																																																																																																																												
Cylinder Pressures/reg.:	Carrier: 650 50		NMHV End: 54.0																																																																																																																												
	Fuel: 900 50		Date: 18Sep2017																																																																																																																												
	Span Gas: 2000 14		Time: 16:50																																																																																																																												
	Zero Air Generator: 45		CH ₄ PK HT: 0																																																																																																																												
Internal Pressures:	Carrier: 35.1		CH ₄ RT: 13.4																																																																																																																												
	Fuel: 46.4		CH ₄ Baseline: 50																																																																																																																												
	Air: 23.7		CH ₄ LOD: 9																																																																																																																												
FID Status:	Status: LIT		CH ₄ SD: 3																																																																																																																												
	Counts: ~28000		CH ₄ CONC: 0.00																																																																																																																												
	Flame: 365.8		NM PK HT: 0																																																																																																																												
	Det Base: 175.1		NM Peak Area: 0																																																																																																																												
Flame and Power Stats:	Last Power On: 18Sep2017@15:50		NM CONC: 0.00																																																																																																																												
	Flameouts: 1		NM Base Start: 48																																																																																																																												
	Det Oven at Start: 114.2		NM Base End: 61																																																																																																																												
	Col Oven at Start: 61.2		NM LOD: 11																																																																																																																												
Calibration History:	Time: 01Jan1970@00:00		NM Start IDX: 52																																																																																																																												
	Type: n/a		NM End IDX: 50																																																																																																																												
	Status: n/a		NM Max Slope: 1.4e+00																																																																																																																												
	Check/Adjust: n/a		NM Min Slope: -7.0e-01																																																																																																																												
	CH ₄ Span Conc: n/a		NM PT Count: 0																																																																																																																												
	CH ₄ SP Ratio: n/a		Previous CH4: 10.32																																																																																																																												
	CH ₄ RT: n/a		Previous NMHC: 11.1																																																																																																																												
	CH ₄ PK IDX: n/a		Previous THC: 21.44																																																																																																																												
	CH ₄ PK HT: n/a		New CH4: 10.32																																																																																																																												
	NM Span Conc: n/a		New NMHC: 11.10																																																																																																																												
	NM SP Ratio: n/a		New THC: 21.44																																																																																																																												
Comments: The analyzer sample inlet filter was changed. The manifold blower was found to be working normally.																																																																																																																															
Flow measurements after mid-point																																																																																																																															




Station: PRAMP_842 Daily: 17/09/18 Type: AVG 1 Min. [1 Min.]






Thermo 55i Methane/Non-Methane Analyzer Calibration

<p>Date: September 20, 2017 Company/Airshed: PRAMP Location/Station Name: 842b Parameter: CH₄ / NMHC / THC Start/End Time 24 hr. (mst): 13:24 / 16:29 Calibration Method: Gas Dilution</p>	<p>Barometer/B.P./units: Brunton 05535 expires December 5, 2017 932 millibars Thermometer/Station Temp: F.S. 160459244 expires May 18, 2018 22 °C Weather Conditions: Cloudy/Overcast Calibration Purpose: post repair Performed By/Reviewer: Chris Wesson Tom Bourque Cal Gas Expiry Date: November 25, 2023</p>																																																																																																																														
Analyzer: <table border="1" style="width: 100%; border-collapse: collapse;"> <tr> <td>ID# or Serial Number: 1505664392</td> <td colspan="3">Correction Factors:</td> </tr> <tr> <td>Measured Flow: 1.2 L/min</td> <td>Previous C.F.:</td> <td>As Found C.F.:</td> <td>New C.F.:</td> </tr> <tr> <td>Last Calibration Date: n/a</td> <td>CH₄ =</td> <td>n/a</td> <td>1.003</td> </tr> <tr> <td>Range ppm: 20 CH₄/20 NMHC/40 THC</td> <td>NMHC =</td> <td>n/a</td> <td>1.001</td> </tr> <tr> <td>CH₄ expressed as C₃H₈:</td> <td>THC =</td> <td>n/a</td> <td>1.002</td> </tr> </table>		ID# or Serial Number: 1505664392	Correction Factors:			Measured Flow: 1.2 L/min	Previous C.F.:	As Found C.F.:	New C.F.:	Last Calibration Date: n/a	CH ₄ =	n/a	1.003	Range ppm: 20 CH ₄ /20 NMHC/40 THC	NMHC =	n/a	1.001	CH ₄ expressed as C ₃ H ₈ :	THC =	n/a	1.002																																																																																																										
ID# or Serial Number: 1505664392	Correction Factors:																																																																																																																														
Measured Flow: 1.2 L/min	Previous C.F.:	As Found C.F.:	New C.F.:																																																																																																																												
Last Calibration Date: n/a	CH ₄ =	n/a	1.003																																																																																																																												
Range ppm: 20 CH ₄ /20 NMHC/40 THC	NMHC =	n/a	1.001																																																																																																																												
CH ₄ expressed as C ₃ H ₈ :	THC =	n/a	1.002																																																																																																																												
Calibration Standards: <table border="1" style="width: 100%; border-collapse: collapse;"> <tr> <td>Low Flow Meter ID/Expiry Date: Defender Low 153358 expires January 19, 2018</td> <td colspan="3">Standard Calibration Points for Analyzer Range of 20/20/40 ppm</td> </tr> <tr> <td>High Flow Meter ID/Expiry Date: Defender High 152571 expires January 19, 2018</td> <td>Point</td> <td>CH₄</td> <td>NMHC</td> </tr> <tr> <td>Calibrator ID/Expiry Date: Sabio id# 17100415 expires May 16, 2018</td> <td>High</td> <td>13.00</td> <td>13.00</td> </tr> <tr> <td>Cal Gas Cylinder I.D. #: LL86139</td> <td>Mid</td> <td>7.00</td> <td>7.00</td> </tr> <tr> <td>CH₄ Cylinder Conc.= 599.0</td> <td>Low</td> <td>3.00</td> <td>3.00</td> </tr> <tr> <td>CH₄ expressed as C₃H₈= 580.3</td> <td></td> <td></td> <td>6.00</td> </tr> </table>		Low Flow Meter ID/Expiry Date: Defender Low 153358 expires January 19, 2018	Standard Calibration Points for Analyzer Range of 20/20/40 ppm			High Flow Meter ID/Expiry Date: Defender High 152571 expires January 19, 2018	Point	CH ₄	NMHC	Calibrator ID/Expiry Date: Sabio id# 17100415 expires May 16, 2018	High	13.00	13.00	Cal Gas Cylinder I.D. #: LL86139	Mid	7.00	7.00	CH ₄ Cylinder Conc.= 599.0	Low	3.00	3.00	CH ₄ expressed as C ₃ H ₈ = 580.3			6.00																																																																																																						
Low Flow Meter ID/Expiry Date: Defender Low 153358 expires January 19, 2018	Standard Calibration Points for Analyzer Range of 20/20/40 ppm																																																																																																																														
High Flow Meter ID/Expiry Date: Defender High 152571 expires January 19, 2018	Point	CH ₄	NMHC																																																																																																																												
Calibrator ID/Expiry Date: Sabio id# 17100415 expires May 16, 2018	High	13.00	13.00																																																																																																																												
Cal Gas Cylinder I.D. #: LL86139	Mid	7.00	7.00																																																																																																																												
CH ₄ Cylinder Conc.= 599.0	Low	3.00	3.00																																																																																																																												
CH ₄ expressed as C ₃ H ₈ = 580.3			6.00																																																																																																																												
ALL POINTS ARE 15 MINUTES OF STABILITY AS OF SEPTEMBER 23, 2015																																																																																																																															
Calibrator Flow Rates (cc/min) <table border="1" style="width: 100%; border-collapse: collapse;"> <tr> <th>Point</th> <th>Diluent</th> <th>Cal Gas</th> <th>Total Flow</th> <th>Calculated CH₄ (ppm)</th> <th>Calculated NMHC (ppm)</th> <th>Calculated THC (ppm)</th> <th>Indicated CH₄ (ppm)</th> <th>Indicated NMHC (ppm)</th> <th>Indicated THC (ppm)</th> <th>Correction Factors:</th> </tr> <tr> <td>adjusted zero</td> <td>2493</td> <td>0.00</td> <td>2493</td> <td>0.00</td> <td>0.00</td> <td>0.00</td> <td>0.02</td> <td>0.00</td> <td>0.03</td> <td>CH₄</td> </tr> <tr> <td>adjusted high</td> <td>2430</td> <td>60.71</td> <td>2491</td> <td>14.60</td> <td>14.14</td> <td>28.74</td> <td>14.58</td> <td>14.12</td> <td>28.71</td> <td>NMHC</td> </tr> <tr> <td>mid</td> <td>2466</td> <td>30.54</td> <td>2497</td> <td>7.33</td> <td>7.10</td> <td>14.42</td> <td>7.28</td> <td>7.13</td> <td>14.42</td> <td>THC</td> </tr> <tr> <td>low</td> <td>2483</td> <td>15.82</td> <td>2499</td> <td>3.79</td> <td>3.67</td> <td>7.46</td> <td>3.74</td> <td>3.72</td> <td>7.47</td> <td></td> </tr> <tr> <td>calibrator zero</td> <td>2493</td> <td>0.00</td> <td>2493</td> <td>0.00</td> <td>0.00</td> <td>0.00</td> <td>0.01</td> <td>0.00</td> <td>0.02</td> <td>Average C.F.=</td> </tr> </table>				Point	Diluent	Cal Gas	Total Flow	Calculated CH ₄ (ppm)	Calculated NMHC (ppm)	Calculated THC (ppm)	Indicated CH ₄ (ppm)	Indicated NMHC (ppm)	Indicated THC (ppm)	Correction Factors:	adjusted zero	2493	0.00	2493	0.00	0.00	0.00	0.02	0.00	0.03	CH ₄	adjusted high	2430	60.71	2491	14.60	14.14	28.74	14.58	14.12	28.71	NMHC	mid	2466	30.54	2497	7.33	7.10	14.42	7.28	7.13	14.42	THC	low	2483	15.82	2499	3.79	3.67	7.46	3.74	3.72	7.47		calibrator zero	2493	0.00	2493	0.00	0.00	0.00	0.01	0.00	0.02	Average C.F.=																																																										
Point	Diluent	Cal Gas	Total Flow	Calculated CH ₄ (ppm)	Calculated NMHC (ppm)	Calculated THC (ppm)	Indicated CH ₄ (ppm)	Indicated NMHC (ppm)	Indicated THC (ppm)	Correction Factors:																																																																																																																					
adjusted zero	2493	0.00	2493	0.00	0.00	0.00	0.02	0.00	0.03	CH ₄																																																																																																																					
adjusted high	2430	60.71	2491	14.60	14.14	28.74	14.58	14.12	28.71	NMHC																																																																																																																					
mid	2466	30.54	2497	7.33	7.10	14.42	7.28	7.13	14.42	THC																																																																																																																					
low	2483	15.82	2499	3.79	3.67	7.46	3.74	3.72	7.47																																																																																																																						
calibrator zero	2493	0.00	2493	0.00	0.00	0.00	0.01	0.00	0.02	Average C.F.=																																																																																																																					
Linear Regression/Calibration Results: <table border="1" style="width: 100%; border-collapse: collapse;"> <tr> <td>Correlation Coefficient =</td> <td>CH₄</td> <td>NMHC</td> <td>THC</td> <td>LIMITS</td> </tr> <tr> <td>Slope =</td> <td>1.000</td> <td>1.000</td> <td>1.000</td> <td>> or = 0.995</td> </tr> <tr> <td>b (Intercept as % of full scale)=</td> <td>0.998</td> <td>0.998</td> <td>0.998</td> <td>0.95-1.05</td> </tr> <tr> <td>% change in C.F. from last cal=</td> <td>-0.07%</td> <td>0.15%</td> <td>0.06%</td> <td>± 3% F.S.</td> </tr> <tr> <td></td> <td>n/a</td> <td>n/a</td> <td>n/a</td> <td>n/a</td> </tr> </table>				Correlation Coefficient =	CH ₄	NMHC	THC	LIMITS	Slope =	1.000	1.000	1.000	> or = 0.995	b (Intercept as % of full scale)=	0.998	0.998	0.998	0.95-1.05	% change in C.F. from last cal=	-0.07%	0.15%	0.06%	± 3% F.S.		n/a	n/a	n/a	n/a																																																																																																			
Correlation Coefficient =	CH ₄	NMHC	THC	LIMITS																																																																																																																											
Slope =	1.000	1.000	1.000	> or = 0.995																																																																																																																											
b (Intercept as % of full scale)=	0.998	0.998	0.998	0.95-1.05																																																																																																																											
% change in C.F. from last cal=	-0.07%	0.15%	0.06%	± 3% F.S.																																																																																																																											
	n/a	n/a	n/a	n/a																																																																																																																											
As Left Instrument Diagnostics: <table border="1" style="width: 100%; border-collapse: collapse;"> <tr> <td>Interface Board Voltages:</td> <td>Bias Supply: -293.9</td> <td>Calibration History cnt'd:</td> <td>NM Peak Area: n/a</td> </tr> <tr> <td>Temperatures:</td> <td>Detector Oven: 175.0</td> <td>Crucial Settings:</td> <td>Methane Start: n/a</td> </tr> <tr> <td></td> <td>Filter: 175.0</td> <td></td> <td>Methane End: n/a</td> </tr> <tr> <td></td> <td>Column Oven: 75.1</td> <td></td> <td>Backflush: n/a</td> </tr> <tr> <td></td> <td>Internal: 32.5</td> <td></td> <td>NMHV Start: n/a</td> </tr> <tr> <td>Cylinder Pressures/reg.:</td> <td>Carrier: 600 50</td> <td></td> <td>NMHV End: n/a</td> </tr> <tr> <td></td> <td>Fuel: 800 50</td> <td></td> <td>Date: 20Sep2017</td> </tr> <tr> <td></td> <td>Span Gas: 2000 15</td> <td></td> <td>Time: 13:01</td> </tr> <tr> <td></td> <td>Zero Air Generator: 55</td> <td></td> <td>CH₄ PK HT: 0</td> </tr> <tr> <td>Internal Pressures:</td> <td>Carrier: 35.1</td> <td></td> <td>CH₄ RT: 13.6</td> </tr> <tr> <td></td> <td>Fuel: 46.4</td> <td></td> <td>CH₄ Baseline: -2</td> </tr> <tr> <td></td> <td>Air: 23.7</td> <td></td> <td>CH₄ LOD: 10</td> </tr> <tr> <td>FID Status:</td> <td>Status: LIT</td> <td></td> <td>CH₄ SD: 3</td> </tr> <tr> <td></td> <td>Counts: ~30000</td> <td></td> <td>CH₄ CONC: 0.00</td> </tr> <tr> <td></td> <td>Flame: 367.9</td> <td></td> <td>NM PK HT: 0</td> </tr> <tr> <td></td> <td>Det Base: 175.0</td> <td></td> <td>NM Peak Area: 0</td> </tr> <tr> <td>Flame and Power Stats:</td> <td>Last Power On: 20Sep2017@10:41</td> <td></td> <td>NM CONC: 0.00</td> </tr> <tr> <td></td> <td>Flameouts: 2</td> <td></td> <td>NM Base Start: -2</td> </tr> <tr> <td></td> <td>Det Oven at Start: 102.2</td> <td></td> <td>NM Base End: 8</td> </tr> <tr> <td></td> <td>Col Oven at Start: 56.0</td> <td></td> <td>NM LOD: 11</td> </tr> <tr> <td>Calibration History:</td> <td>Time: 01Jan1970@00:00</td> <td></td> <td>NM Start IDX: 4</td> </tr> <tr> <td></td> <td>Type: n/a</td> <td></td> <td>NM End IDX: 54</td> </tr> <tr> <td></td> <td>Status: n/a</td> <td></td> <td>NM Max Slope: 8.8e-01</td> </tr> <tr> <td></td> <td>Check/Adjust: n/a</td> <td></td> <td>NM Min Slope: -9.0e-01</td> </tr> <tr> <td></td> <td>CH₄ Span Conc: n/a</td> <td></td> <td>NM PT Count: 0</td> </tr> <tr> <td></td> <td>CH₄ SP Ratio: n/a</td> <td></td> <td>Previous CH4: 10.32</td> </tr> <tr> <td></td> <td>CH₄ RT: n/a</td> <td></td> <td>Previous NMHC: 11.1</td> </tr> <tr> <td></td> <td>CH₄ PK IDX: n/a</td> <td></td> <td>Previous THC: 21.44</td> </tr> <tr> <td></td> <td>CH₄ PK HT: n/a</td> <td></td> <td>New CH4: 8.76</td> </tr> <tr> <td></td> <td>NM Span Conc: n/a</td> <td></td> <td>New NMHC: 10.08</td> </tr> <tr> <td></td> <td>NM SP Ratio: n/a</td> <td></td> <td>New THC: 18.85</td> </tr> </table>				Interface Board Voltages:	Bias Supply: -293.9	Calibration History cnt'd:	NM Peak Area: n/a	Temperatures:	Detector Oven: 175.0	Crucial Settings:	Methane Start: n/a		Filter: 175.0		Methane End: n/a		Column Oven: 75.1		Backflush: n/a		Internal: 32.5		NMHV Start: n/a	Cylinder Pressures/reg.:	Carrier: 600 50		NMHV End: n/a		Fuel: 800 50		Date: 20Sep2017		Span Gas: 2000 15		Time: 13:01		Zero Air Generator: 55		CH ₄ PK HT: 0	Internal Pressures:	Carrier: 35.1		CH ₄ RT: 13.6		Fuel: 46.4		CH ₄ Baseline: -2		Air: 23.7		CH ₄ LOD: 10	FID Status:	Status: LIT		CH ₄ SD: 3		Counts: ~30000		CH ₄ CONC: 0.00		Flame: 367.9		NM PK HT: 0		Det Base: 175.0		NM Peak Area: 0	Flame and Power Stats:	Last Power On: 20Sep2017@10:41		NM CONC: 0.00		Flameouts: 2		NM Base Start: -2		Det Oven at Start: 102.2		NM Base End: 8		Col Oven at Start: 56.0		NM LOD: 11	Calibration History:	Time: 01Jan1970@00:00		NM Start IDX: 4		Type: n/a		NM End IDX: 54		Status: n/a		NM Max Slope: 8.8e-01		Check/Adjust: n/a		NM Min Slope: -9.0e-01		CH ₄ Span Conc: n/a		NM PT Count: 0		CH ₄ SP Ratio: n/a		Previous CH4: 10.32		CH ₄ RT: n/a		Previous NMHC: 11.1		CH ₄ PK IDX: n/a		Previous THC: 21.44		CH ₄ PK HT: n/a		New CH4: 8.76		NM Span Conc: n/a		New NMHC: 10.08		NM SP Ratio: n/a		New THC: 18.85
Interface Board Voltages:	Bias Supply: -293.9	Calibration History cnt'd:	NM Peak Area: n/a																																																																																																																												
Temperatures:	Detector Oven: 175.0	Crucial Settings:	Methane Start: n/a																																																																																																																												
	Filter: 175.0		Methane End: n/a																																																																																																																												
	Column Oven: 75.1		Backflush: n/a																																																																																																																												
	Internal: 32.5		NMHV Start: n/a																																																																																																																												
Cylinder Pressures/reg.:	Carrier: 600 50		NMHV End: n/a																																																																																																																												
	Fuel: 800 50		Date: 20Sep2017																																																																																																																												
	Span Gas: 2000 15		Time: 13:01																																																																																																																												
	Zero Air Generator: 55		CH ₄ PK HT: 0																																																																																																																												
Internal Pressures:	Carrier: 35.1		CH ₄ RT: 13.6																																																																																																																												
	Fuel: 46.4		CH ₄ Baseline: -2																																																																																																																												
	Air: 23.7		CH ₄ LOD: 10																																																																																																																												
FID Status:	Status: LIT		CH ₄ SD: 3																																																																																																																												
	Counts: ~30000		CH ₄ CONC: 0.00																																																																																																																												
	Flame: 367.9		NM PK HT: 0																																																																																																																												
	Det Base: 175.0		NM Peak Area: 0																																																																																																																												
Flame and Power Stats:	Last Power On: 20Sep2017@10:41		NM CONC: 0.00																																																																																																																												
	Flameouts: 2		NM Base Start: -2																																																																																																																												
	Det Oven at Start: 102.2		NM Base End: 8																																																																																																																												
	Col Oven at Start: 56.0		NM LOD: 11																																																																																																																												
Calibration History:	Time: 01Jan1970@00:00		NM Start IDX: 4																																																																																																																												
	Type: n/a		NM End IDX: 54																																																																																																																												
	Status: n/a		NM Max Slope: 8.8e-01																																																																																																																												
	Check/Adjust: n/a		NM Min Slope: -9.0e-01																																																																																																																												
	CH ₄ Span Conc: n/a		NM PT Count: 0																																																																																																																												
	CH ₄ SP Ratio: n/a		Previous CH4: 10.32																																																																																																																												
	CH ₄ RT: n/a		Previous NMHC: 11.1																																																																																																																												
	CH ₄ PK IDX: n/a		Previous THC: 21.44																																																																																																																												
	CH ₄ PK HT: n/a		New CH4: 8.76																																																																																																																												
	NM Span Conc: n/a		New NMHC: 10.08																																																																																																																												
	NM SP Ratio: n/a		New THC: 18.85																																																																																																																												
Comments: <p>The manifold blower was found to be working normally.</p>																																																																																																																															
Post-repair following verification of valve action and pressure adjustments Flow measurements after mid-point																																																																																																																															




Station: PRAMP_842 Periodically: 2017/09/20 11:00-2017/09/20 23:59 Type: AVG 1 Min. [1
Min.] [RAW]






Thermo 55i Methane/Non-Methane Analyzer Calibration

Date:	September 28, 2017	Barometer/B.P./units:	Brunton 05535 expires December 5, 2017	947	millibars							
Company/Airshed:	PRAMP	Thermometer/Station Temp:	F.S. 160459244 expires May 18, 2018	22	°C							
Location/Station Name:	842b	Weather Conditions:	Clear									
Parameter:	CH4 / NMHC / THC	Calibration Purpose:	Installation									
Start/End Time 24 hr. (mst):	07:01 / 09:57	Performed By/Reviewer:	Chris Wesson	Tom Bourque								
Calibration Method:	Gas Dilution	Cal Gas Expiry Date:	November 25, 2023									
Analyzer:												
ID# or Serial Number:	1236656188	Correction Factors:										
Measured Flow:	1.16 L/min	Previous C.F.:	As Found C.F.:	New C.F.:								
Last Calibration Date:	n/a	CH ₄ =	n/a	1.000								
Range ppm:	20 CH ₄ /20 NMHC/40 THC	NMHC =	n/a	0.999								
		THC =	n/a	0.999								
Calibration Standards:												
Low Flow Meter ID/Expiry Date:	Defender Low 153358 expires January 19, 2018	Standard Calibration Points for Analyzer Range of 20/20/40 ppm										
High Flow Meter ID/Expiry Date:	Defender High 152571 expires January 19, 2018	Point	CH4	NMHC	THC							
Calibrator ID/Expiry Date:	Sabio id# 17100415 expires May 16, 2018	High	13.00	13.00	26.00							
Cal Gas Cylinder I.D. #:	LL86139	Mid	7.00	7.00	14.00							
CH ₄ Cylinder Conc. =	599.0 211.0 =C ₂ H ₆ Cylinder Conc.	Low	3.00	3.00	6.00							
CH ₄ expressed as C ₂ H ₆ =	580.3 1179.3 =total CH ₄ equivalent											
ALL POINTS ARE 15 MINUTES OF STABILITY AS OF SEPTEMBER 23, 2015												
Calibrator Flow Rates (cc/min)				Correction Factors:								
Point	Diluent	Cal Gas	Total Flow			Calculated CH ₄ (ppm)	Calculated NMHC (ppm)	Calculated THC (ppm)	Indicated CH ₄ (ppm)	Indicated NMHC (ppm)	Indicated THC (ppm)	CH ₄
adjusted zero	2497	0.00	2497	0.00	0.00	0.00	0.00	0.00	0.00	n/a	n/a	n/a
adjusted high	2437	60.95	2498	14.62	14.16	28.77	14.61	14.17	28.80	1.000	0.999	0.999
mid	2470	30.65	2501	7.34	7.11	14.45	7.37	7.03	14.43	0.996	1.011	1.001
low	2489	15.91	2505	3.80	3.69	7.49	3.82	3.58	7.41	0.996	1.029	1.011
calibrator zero	2497	0.00	2497	0.00	0.00	0.00	0.00	0.00	0.00	n/a	n/a	n/a
									Average C.F.=	0.997	1.013	1.004
Linear Regression/Calibration Results:												
Correlation Coefficient =	1.000	NMHC	THC	LIMITS								
Slope =	0.999	1.003	1.002	> or = 0.995								
b (Intercept as % of full scale)=	0.07%	-0.30%	-0.10%	0.95-1.05								
% change in C.F. from last cal=	n/a	n/a	n/a	± 3% F.S.								
				n/a								
As Left Instrument Diagnostics:												
Interface Board Voltages:	Bias Supply:	-288.0	Calibration History cnt'd:	NM Peak Area:	88489							
Temperatures:	Detector Oven:	175.0	Crucial Settings:	Methane Start:	8.0							
	Filter:	175.1		Methane End:	16.0							
	Column Oven:	75.0		Backflush:	18.0							
	Internal:	31.5	Run History>1:	NMHV Start:	24.0							
Cylinder Pressures/reg.:	Carrier:	2500		NMHV End:	56.0							
	Fuel:	500		Date:	28Sep2017							
	Span Gas:	1800		Time:	09:09							
	Zero Air Generator:	45		CH ₄ PK HT:	0							
Internal Pressures:	Carrier:	30.5		CH ₄ RT:	12.0							
	Fuel:	40.0		CH ₄ Baseline:	1495							
	Air:	24.9		CH ₄ LOD:	45							
FID Status:	Status:	LIT		CH ₄ SD:	15							
	Counts:	~17500		CH ₄ CONC:	0.00							
	Flame:	346.4		NM PK HT:	0							
	Det Base:	175.0		NM Peak Area:	0							
Flame and Power Stats:	Last Power On:	27Sep2017@15:41		NM CONC:	0.00							
	Flameouts:	1		NM Base Start:	1408							
	Det Oven at Start:	170.0		NM Base End:	1359							
	Col Oven at Start:	74.5		NM LOD:	17							
Calibration History:	Time:	28Sep2017@07:37		NM Start IDX:	66							
	Type:	Span		NM End IDX:	37							
	Status:	Good		NM Max Slope:	2.0e-01							
	Check/Adjust:	Adjust		NM Min Slope:	-1.2e+00							
	CH ₄ Span Conc:	14.62	Expected Values:	NM PT Count:	0							
	CH ₄ SP Ratio:	0.000773		Previous CH4:	8.76							
	CH ₄ RT:	12.6		Previous NMHC:	10.08							
	CH ₄ PK IDX:	23		Previous THC:	18.85							
	CH ₄ PK HT:	18923		New CH4:	9.04							
	NM Span Conc:	14.16		New NMHC:	10.11							
	NM SP Ratio:	0.00016		New THC:	19.15							
Comments:												
The manifold blower was found to be working normally.												



Station: PRAMP_842 Daily: 17/09/28 Type: AVG 1 Min. [1 Min.]



WIND SYSTEM

AIR FCD-00055/1
Wind_0830_Install

Meteorological Sensor Audit/Calibration						
Location Information						
Company:	PRAMP	Performed By:	Chris Wesson			
Audit Location:	842b	Reviewed By:	Tom Bourque			
Audit Date:	August 30, 2017	Start /EndTime (mst):	10:50 / 11:20			
Calibration Purpose:	installation	Weather Conditions:	Cloudy/Overcast			
Wind Sensor Information						
Sensor ID Data:			Sensor Outputs:			
Sensor Make:	RM Young	Velocity Voltage Output Range:	0-1V			
Sensor Model:	05305VK	Velocity Unit Output Range:	0-200kmh			
Serial #:	124638	Direction Voltage Output Range:	0-1V			
Previous Cal/Audit Date:	n/a or unknown	Direction Unit Output Range:	0-360°			
Wind Calibrator Information						
Calibrator Make/ Model:	RM Young	Serial #:	CA 4039			
Maxxam Unit ID #:	n/a	Certification Date:	February 24, 2017			
Wind Speed Audit Data **+/- 2% of the average correction factor is the limit**						
RPM	Wind Speed Generated kph	Clockwise Wind Speed kph	Counter Clockwise Wind Speed kph	Correction Factor		
0	0	0.1	0.1	-		
1000	18.4	18.4	18.5	1.000		
2000	36.9	36.7	36.7	1.004		
3000	55.3	55.1	55.1	1.004		
4000	73.7	73.4	73.4	1.004		
5000	92.2	91.8	91.8	1.004		
6000	110.6	110.2	110.2	1.003		
7000	129.0	128.6	128.6	1.003		
8000	147.4	147.1	147.0	1.003		
9000	165.9	165.5	165.5	1.002		
10000	184.3	184.2	184.1	1.001		
			The audit meets AMD requirements.	Average Correction Factor= 1.003		
Wind Direction Audit Data **+/- 5° of the absolute average degrees difference for all points is the limit**						
Generated Wind Direction 0-360 (Up)	Generated Wind Direction 360-0 (Down)	Indicated Wind Direction 0-360 (Up)	Indicated Wind Direction 360-0 (Down)	Degrees Difference 0-360 (Up)	Degrees Difference 360-0 (Down)	Average Absolute Degrees Difference
0	355	1	353	0.5	2.0	1.3
30	330	28	329	2.0	1.0	1.5
60	300	59	299	1.0	1.0	1.0
90	270	89	269	1.0	1.0	1.0
120	240	118	239	2.0	1.0	1.5
150	210	149	209	1.0	1.0	1.0
180	180	178	179	2.0	1.0	1.5
210	150	208	150	2.0	0.0	1.0
240	120	238	119	2.0	1.0	1.5
270	90	268	89	2.0	1.0	1.5
300	60	298	59	2.0	1.0	1.5
330	30	328	29	2.0	1.5	1.8
355	0	353	0	2.0	0.4	1.2
			The audit meets AMD requirements.	Average Absolute Degrees Difference=		1.3
Comments:						

CALIBRATORS

Company <u>Maxxam</u>	Operator: <u>Mike</u>		
Calibrator:			
Make/Model	API 700		
Serial Number	690		
Last Verification Date	March 30, 2016\		
NO Cylinder S/N	EY0000597		
NO [PPM]	49.0	NOx [PPM]	49.0
Expiry Date	December 8, 2019		
Flow Measurement Device:			
Make/Model	Bios Defender 530		
Serial Number	Hi148944 Lo 152019		
Temperature (°C)	23.3		
Barometric Pressure	704.3mmHg		

Dilution Flow (sccm)								
Pt. #1	4898	Pt. #2	4942	Pt. #3	4953			
Gas Flow (sccm)								
Pt. #1	79.2	Pt. #2	38.6	Pt. #3	19.3			

Calibrator Flow (sccm)		Calculated Conc.(ppm)		Indicated Conc.(ppm)			% Difference vs Audit Gas	
Dilution	Gas	NO	NOx	NO	NO ₂	NOx	NO	NOx
4977	0.0	0.0000	0.0000	0.0000	0.0000	0.0000	Limit ± 10%	
4977	79.2	0.7792	0.7792	0.7841	0.0012	0.7854	1%	1%
4981	38.6	0.3797	0.3797	0.3813	0.0006	0.3819	0%	1%
492	19.3	0.1902	0.1902	0.1927	0.0002	0.1929	1%	1%
Absolute Average Percent Difference							1%	1%

LINEAR REGRESSION ANALYSIS								
y=mx+b (where x=calculated concentration, y=indicated concentration)								
NO			LIMITS			NOx		
Correlation= 1.0000			≥ 0.990			Correlation= 1.0000		
m (Slope)= 1.0056			0.90-1.10			m (Slope)= 1.0073		
b (Intercept % of FS)= 0.0357			± 3% F.S.			b (Intercept % of FS)= 0.0304		

Flow	O ₃ Conc	NO Decrease	NO	NO2	NOX	% Diff. Vs Audit gas	NO ₂	% Diff. Limit
4977	0.000	0.0000	0.7928	0.0014	0.7941	NO ₂	% Diff. Limit	
4977	0.500	0.5448	0.2480	0.5391	0.7871	-1%	± 10%	
4977	0.250	0.2862	0.5066	0.2861	0.7926	-1%	± 10%	
4977	0.100	0.1221	0.6707	0.1193	0.7914	-3%	± 10%	
Absolute Average Percent Difference							2%	± 10%

LINEAR REGRESSION ANALYSIS								
y=mx+b (where x=calculated concentration, y=indicated concentration)								
NO₂			LIMITS			NOx		
Correlation= 1.0000			≥ 0.995			Correlation= 1.0000		
m (Slope)= 0.9894			0.90-1.10			m (Slope)= 1.0073		
b (Intercept % of FS)= 0.0719			± 3% F.S.			b (Intercept % of FS)= 0.0304		

AENV Standards			NO_x Analyzer		
Audit Calibrator			Make/Model	Thermo 42i	
Make/Model	Thermo 146i		Serial/AMU Number	1868	
Serial/AMU Number	1809		Last Calibration Date	March 15, 2017	
SRM Gas Cylinder No.	CAL018140		Full Scale (ppm)	1.0	
Cylinder Conc. (ppm)	48.79		Cylinder Gas Expiry Date	March 28, 2019	

COMMENTS: Gas has ~50ppm SO₂

Auditor: Shea Beaton Date: March 17, 2017
 Operator Signature: SJB Location: McIntyre Center Edmonton

Company <u>Maxxam</u>	Operator: <u>Chris</u>	
Calibrator:		
Make/Model <u>API 700</u>	Flow Measurement Device:	
Serial Number <u>829</u>	Make/Model <u>Definer 530</u>	
Last Verification Date <u>February 3, 2016</u>	Serial Number <u>H-148944, L-152019</u>	
NO Cylinder S/N <u>EY0000597</u>	Temperature (°C) <u>23.5</u>	
NO [PPM] <u>49.0</u>	Barometric Pressure <u>707.1 mmHg</u>	
Expiry Date <u>December 8, 2019</u>		
Dilution Flow (sccm)		
Pt. #1 <u>4846</u>	Pt. #2 <u>4888</u>	Pt. #3 <u>4908</u>
Gas Flow (sccm)		
Pt. #1 <u>80.6</u>	Pt. #2 <u>38.9</u>	Pt. #3 <u>18.4</u>

Calibrator Flow (sccm)		Calculated Conc.(ppm)		Indicated Conc.(ppm)			% Difference vs Audit Gas	
Dilution	Gas	NO	NOx	NO	NO ₂	NOx	NO	NOx
	0.0	0.0000	0.0000	0.0000	0.0001	0.0001	Limit ± 10%	
4927	80.6	0.8016	0.8016	0.8021	0.0015	0.8036	0.1%	0.2%
4927	38.9	0.3867	0.3867	0.3862	0.0001	0.3863	-0.1%	-0.1%
4927	18.4	0.1830	0.1830	0.1820	0.0004	0.1824	-0.5%	-0.4%
Absolute Average Percent Difference							0.2%	0.1%

LINEAR REGRESSION ANALYSIS								
$y=mx+b$ (where x=calculated concentration, y=indicated concentration)								
NO			LIMITS			NOx		
Correlation=	1.0000		≥ 0.990			Correlation=	1.0000	
m (Slope)=	1.0010		0.90-1.10			m (Slope)=	1.0027	
b (Intercept % of FS)=	-0.0604		± 3% F.S.			b (Intercept % of FS)=	-0.0652	

Flow	O ₃ Conc	NO Decrease	NO	NO2	NOX	% Diff. Vs Audit gas
4927	0	0.0000	0.8016	0.0010	0.8020	NO ₂ % Diff. Limit
4927	500	0.5450	0.2566	0.5421	0.7987	-0.7% ± 10%
4927	250	0.2797	0.5219	0.2797	0.8016	-0.4% ± 10%
4927	80	0.0954	0.7062	0.0958	0.8026	-0.6% ± 10%
Absolute Average Percent Difference						1% ± 10%

LINEAR REGRESSION ANALYSIS								
$y=mx+b$ (where x=calculated concentration, y=indicated concentration)								
NO₂			LIMITS			NOx		
Correlation=	1.0000		≥ 0.995			Correlation=	1.0000	
m (Slope)=	0.9931		0.90-1.10			m (Slope)=	1.0027	
b (Intercept % of FS)=	0.1220		± 3% F.S.			b (Intercept % of FS)=	-0.0652	

AENV Standards			NO _x Analyzer		
Audit Calibrator			Make/Model		
Make/Model	Thermo 146i		Make/Model	Thermo 42i	
Serial/AMU Number	AMU1809		Serial/AMU Number	AMU 1868	
SRM Gas Cylinder No.	CAL018140		Last Calibration Date	January 25, 2017	
Cylinder Conc. (ppm)	48.79		Full Scale (ppm)	1.0	
			Cylinder Gas Expiry Date	March 25, 2019	

COMMENTS: Slight over reading at 0.0000 NO

Auditor: Sara Beaton Date: January 27, 2017

Operator Signature: [Signature] Location: McIntyre Center Edmonton

Calibrator Performance Audit
Oxides Of Nitrogen

File No. 2017-072A

Company	Maxxam	Operator:	Micheal Espiritu
Calibrator:		Flow Measurement Device:	
Make/Model	Sabio 2010	Make/Model	Mesa Defender 530
Serial Number	17100415	Serial Number	L-152019 H-148944
Last Verification Date	May 2016	Temperature (°C)	25.0 C
NO Cylinder S/N	EY0000597	Barometric Pressure	697 mmhg
NO [PPM]	49.0	NOx [PPM]	49.0
Expiry Date	December 2019		
Dilution Flow (sccm)			
Pt. #1	5000	Pt. #2	5000
Gas Flow (sccm)			
Pt. #1	80	Pt. #2	40
Pt. #3	20		

Calibrator Flow (sccm)		Calculated Conc.(ppm)		Indicated Conc.(ppm)			% Difference vs Audit Gas	
Dilution	Gas	NO	NOx	NO	NO ₂	NOx	NO	NOx
4996	0.0	0.000	0.000	0.000	0.000	0.000	Limit ± 10%	
5029	80.3	0.784	0.783	0.808	-0.013	0.794	3%	1%
5054	38.8	0.376	0.376	0.392	-0.006	0.386	4%	3%
5051	19.5	0.189	0.189	0.196	-0.003	0.193	4%	2%
Absolute Average Percent Difference							4%	2%

LINEAR REGRESSION ANALYSIS							
y=mx+b (where x=calculated concentration, y=indicated concentration)							
NO				LIMITS			
Correlation= 1.0000				≥ 0.990			
m (Slope)= 1.0311				0.90-1.10			
b (Intercept % of FS)= 0.1350				± 3% F.S.			
				NOx			
Correlation= 1.0000				Correlation= 1.0000			
m (Slope)= 1.0140				m (Slope)= 1.0140			
b (Intercept % of FS)= 0.1531				b (Intercept % of FS)= 0.1531			


Flow	O ₃ Conc (L.C)	NO Decrease	NO	NO2	NOX	% Diff. Vs Audit gas	% Diff. Limit
5029	0.000	0.000	0.803	-0.013	0.790	NO ₂	% Diff. Limit
5029	1.508	0.568	0.235	0.552	0.787	-1%	± 10%
5029	0.882	0.312	0.491	0.298	0.789	0%	± 10%
5029	0.390	0.108	0.695	0.095	0.789	0%	± 10%
Absolute Average Percent Difference							0% ± 10%

LINEAR REGRESSION ANALYSIS							
y=mx+b (where x=calculated concentration, y=indicated concentration)							
NO₂				LIMITS			
Correlation= 1.0000				≥ 0.995			
m (Slope)= 0.9945				0.90-1.10			
b (Intercept % of FS)= -1.2646				± 3% F.S.			

AENV Standards				NO _x Analyzer			
Audit Calibrator				Make/Model			
Make/Model				Teco 42i			
Serial/AMU Number				AMU 1868			
SRM Gas Cylinder No.				Last Calibration Date			
Cylinder Conc. (ppm)				May 16, 2017			
				Full Scale (ppm)			
				1.0			
				Cylinder Gas Expiry Date			
				March 2019			

COMMENTS: Contains 50.4 ppm SO₂.

Auditor: Al Clark Date: May 16, 2017

Operator Signature:  Location: McIntyre Center Edmonton

CALIBRATION GASES



Calibration Gas Audit

Single Component Cylinder Gas

File No. 2016-086CGA

Company: Maxxam

Operator's Name: Chris Wesson

Cylinder #: LL119513 Concentration PPM: 50.6 Tolerance(%) 1 Certified By: Praxair

Reference Calibrator and Gas:

Make/Model: Teco 146i

Serial Number: AMU 1809

Last Verification Date: June 17, 2016

Gas Type: SO₂ Conc. 98.07

Cylinder Number: CAL016625

Flow Measurement Device:

Make/Model: Bios DC2

Serial Number: AMU 1659

Temp.°C: 23.0 C

B.P. 700 mmhg

Reference Analyzer:

Make/Model: Teco 43C

Serial/AMU Number: 1623

Instrument Settings: Zero: 8.7 Span: 1.027 Range: 1.0

Last Calibration: Date: June 17/16 C.F. 1.000 Done By: Al Clark

Calibrator Flows (sccm)		Indicated Concentration (PPM)	Gas Flow/ Dilution Flow	Concentration Factor	Cylinder Concentration
Dilution	Gas				
5000	0.0	0.000			
4976	82.3	0.828	0.01654	60.462	50.1
4985	40.8	0.411	0.00818	122.181	50.2
4965	20.2	0.203	0.00407	245.792	49.9
Average Cylinder Concentration:					50.1

Previous Stated Concentration PPM: 50.6

Percent variance from Stated: 1.1

Meets Manufacturer Tolerance. Use manufacturers stated concentration COMMENTS: _____

<=5% Outside Manufacturer Tolerance. Use manufacturers concentration _____

> 5% Outside Manufacturer Tolerance. DO NOT USE this cylinder _____

Auditor: Al Clark

Date: June 17, 2016

Operator Signature: Al Clark

Location: McIntyre Center Edmonton



Calibration Gas Audit Single Component Cylinder Gas

File No. 2015-109CGA

Company: Maxxam

Operator's Name: Chris Wesson

Cylinder #: BLM001927 Concentration PPM: 10.3 Tolerance(%) 2 Certified By: Air Liquide

Reference Calibrator and Gas:

Make/Model: R&R MFC 201

Serial Number: AMU 1690

Last Verification Date: February 2, 2016

Gas Type: H2S Conc. 20.43

Cylinder Number: CAL015584

Flow Measurement Device:

Make/Model: Bios DC-2

Serial Number: Bios D

Temp. °C: 24.5

B.P. 702mmHg

Reference Analyzer:

Make/Model: Thermo 450i

Serial/AMU Number: 1980

Instrument Settings: Zero: 15.3 Span: 1.126 Range: 0.1

Last Calibration: Date: 1-Feb-16 C.F. 1.000 Done By: SB

Calibrator Flows (scm)		Indicated Concentration (PPM)	Gas Flow/ Dilution Flow	Concentration Factor	Cylinder Concentration
Dilution	Gas				
5025	0.0	0.000	X	X	X
5058	37.84	0.078	0.00748	133.668	10.4
5059	17.85	0.036	0.00353	283.417	10.3
5031	9.15	0.019	0.00182	549.836	10.2
Average Cylinder Concentration:					10.3

Previous Stated Concentration PPM: 10.3

Percent variance from Stated: 0.1

Meets Manufacturer Tolerance. Use manufacturers stated concentration COMMENTS: _____

<=5% Outside Manufacturer Tolerance. Use manufacturers concentration _____

> 5% Outside Manufacturer Tolerance. **DO NOT USE** this cylinder _____

Auditor: Shea Beaton

Date: February 2, 2016

Operator Signature:

Location: McIntyre Center Edmonton



Calibration Gas Audit

CH4 / C3H8 Cylinder Gas

File No. 2015-091CGA

Company:	Maxxam	Operators name:	Chris Wesson
Cylinder #:	LL86139	Conc CH4 (PPM)	599/211 Tolerance (%) 0.5 Certified By: Praxair

Reference Calibrator and Gas:		Flow Measurement Device:	
Make/Model	R&R MFC 201	Make/Model	Bios DC-2
Serial Number	AMU 1698	Serial Number	Bios D
Last Verification Date	January 18, 2016	Temp. °C	23
Gas Type	CH4	B.P.	599mmHg
Cylinder Number	D751932		
Gas Type	C3H8		
Cylinder Number	XF0037998		

Reference Analyzer:			
Make/Model	Thermo 55C	Serial/AMU Number:	1643
Instrument Settings	Zero: NA	Span:	NA Range: 20.0
Last Calibration:	Date: 18-Jan-16	C.F.	1.000 Done By: SB

Calibrator Flows (scfm)	Indicated Conc. (ppm)		Gas Flow/ Dilution Flow	Concentration Factor	Cylinder Concentration	
	Dilution	Gas	CH4	C3H8	CH4	C3H8
2583	0.00	0.00	0.00	X	X	X
2635	56.52	12.80	12.59	0.02145	46.621	597
2592	19.72	4.54	4.49	0.00761	131.440	597
2584	9.69	2.25	2.24	0.00375	266.667	600
Average Cylinder Concentration:					598	215

CH4

Previous Stated Concentration PPM: 599

C3H8

211

Percent variance from Stated: 0.2

1.9

Cylinder gas tolerances based on CH4 only

Meets Manufacturer Tolerance. Use manufacturers stated concentration

COMMENTS:


C3H8 manufacturers tolerance 1.1%

<=5% Outside Manufacturer Tolerance. Use manufacturers concentration

> 5% Outside Manufacturer Tolerance. DO NOT USE this cylinder

Auditor: 
Shea Beaton

Date: January 19, 2016

Operator Signature: 

Location: McIntyre Center Edmonton

APPENDIX III
REPORT CERTIFICATION FORM

Report Certification Form

Alberta Airshed (if applicable)	EPA Approval or Code of Practice Registration # (if applicable)
YES	NA
Company Name (if applicable)	Industrial Operation Name (if applicable)
Peace River Area Monitoring Program Committee	Three Creeks 842b Station
Name of the Representative of the Person Responsible (Last, First, Middle)	Position / Title of the Representative of the Person Responsible
Wunmi Adekanmbi	Project Manager, Customer Service, Air Services
Is an External Party Certifying the Report? (If 'Yes', fill in the fields below for the external person.)	
<input type="checkbox"/> Yes <input checked="" type="checkbox"/> No	
Name of External Person Certifying the Report (Last, First, Middle)	Position / Title of External Person Certifying the Report
NA	NA
Company Name for the External Person Certifying the Report	Identification of Qualifications / Professional Designations of the External Person Certifying the Report
NA	NA

I certify that I have reviewed and verified the submitted report. I also certify that the report presented with this certification form is complete, accurate and representative of the monitoring results and timeframe.



Signature of the Representative of the Person
Responsible / External Person Certifying the Report

October 27, 2017

Report Issued Date (dd-mm-yyyy)

APPENDIX IV
DATA VALIDATION CERTIFICATION FORM



Validation Certificate Form

Client: Peace River Area Monitoring Program Committee

Site: Three Creeks 842b Station

Project #: 8449-2017-09-80-C

Contact: Karla Reesor

Level 0 Preliminary Verification

Date October 13, 2017

Level 1 Primary Validation

Date October 13, 2017

Level 2 Final Validation

Date October 27, 2017

Level 3 Independent Data Review

Date October 27, 2017

Post-Final Validation

NA

Date NA

Notes

The Post-Final Validation step serves to re-evaluate the data that errors or omissions are discovered and/or suspected after the initial submittal of data. This validation is performed on an annual basis.