



MAXXAM ANALYTICS
#1 2080 39 Ave. NE, Calgary, AB
T2E 6P7

maxxam.ca
Toll Free 800-386-7247
Fax 403-219-3673

AMBIENT AIR MONITORING MONTHLY DATA REPORT
PEACE RIVER AREA MONITORING PROGRAM COMMITTEE
THREE CREEKS 986B STATION

JOB #: 8449-2018-04-67-C

April 2018

Prepared for:

PEACE RIVER AREA MONITORING PROGRAM COMMITTEE

Attention: LILY LIN

DATE: **May 24, 2018**

Prepared by:

Wunmi Adekanmbi, M.Sc., EPT, PMP
Project Team Lead, Customer Service, Air Services

Reviewed by:

Cheri Sinclair, B.Sc.
Supervisor, Customer Service, Air Services

SUMMARY

In April 2018, Maxxam Analytics was contracted to manage the ambient air quality monitoring and maintenance activities at the Three creeks 986b Station, near Peace River Oil Sands Area 2, Alberta. The monitoring station provides continuous meteorological measurements and air quality data for non-compliance parameters, as requested by the PRAMP Committee.

All data collected this month was compliant with the requirements outlined in the Air Monitoring Directive (Alberta Environment and Parks, 2016).

The operational time for all continuous ambient air analyzers, meteorological systems and data acquisition systems were above the 90% requirement.

TRS: An as-found response check was completed on April 30, to assess analyzer performance after a slight biased high drift in span response, incurring two hours of downtime.

THC/CH₄/NMHC: The NMHC span response exhibited an abrupt biased high drift on April 22. As corrective actions, a repeat zero-span check and a repeat calibration were performed on April 23 and April 24, respectively. Five hours of downtime were incurred due to the additional quality checks.

The summary of results is presented on the following pages.

Any deviations or modifications made to the sampling or analytical methods are outlined in Section 1.0, Discussion. On this basis, Maxxam Analytics is issuing this completed report to Peace River Area Monitoring Program Committee, Three Creeks 986b Station.

Should you have any questions concerning the results or if we can be of further assistance, please contact us at 403-219-3661 or toll-free at 1-800-386-7247.

Monthly Continuous Data Summary

Peace River Area Monitoring Program Committee						MAXIMUM VALUES							OPERATIONAL TIME (%)
Three Creeks 986b Station						1-HOUR					24-HOUR		
PARAMETER	OBJECTIVES		EXCEEDANCES		MONTHLY AVERAGE	READING	DAY	HOUR	WIND SPEED (kph)	WIND DIRECTION (sector)	READING	DAY	
	1-hr	24-hr	1-hr	24-hr									
SO ₂ (ppb)	172	48	0	0	0	2	1	8	0.3	NNW	1	1	100.0
TRS (ppb)	-	-	-	-	0.29	0.52	1	7	2.4	ESE	0.35	26	99.7
THC (ppm)	-	-	-	-	1.98	2.48	1	7	2.4	ESE	2.09	1	99.3
CH ₄ (ppm)	-	-	-	-	1.98	2.48	1	7	2.4	ESE	2.09	1	99.3
NMHC (ppm)	-	-	-	-	0.00	0.01	26	2	1	E	0.00	1	99.3
RELATIVE HUMIDITY (%)	-	-	-	-	59	99	16	22	6	N	93	17	100.0
BAROMETRIC PRESSURE (millibar)	-	-	-	-	942	960	5	9	8.6	NNW	959	5	100.0
AMBIENT TEMPERATURE (°C)	-	-	-	-	0.7	23.1	27	16	10.0	SSW	13.9	27	100.0
STATION TEMPERATURE (°C)	-	-	-	-	23.1	25.2	16	6	8.2	N	23.8	17	100.0
VECTOR WS (kph)	-	-	-	-	1.2	21.4	30	19	-	NW	10.6	19	100.0
VECTOR WD (sec)	-	-	-	-	219 (SW)	-	-	-	-	-	-	-	100.0

**SOUR GAS PROCESSING INDUSTRY
MONTHLY REPORT SUMMARY**

Three Creeks 986b Station

Peace River Area Monitoring Program Committee

Plant Name / Location

Company

Licence Number	Report Date	
	YEAR	MONTH
N/A	2018	April

CONTINUOUS AMBIENT MONITORING						
PARAMETER	% TIME OPERATIONAL	ONE - HOUR AVERAGE			24 - HOUR AVERAGE	
		MAXIMUM VALUES	NO. READINGS > REGULATION	MAXIMUM VALUES	NO. READINGS > REGULATION	
SO ₂	100.0	0.002 ppm	0	0.001 ppm	0	
TRS	99.7	0.001 ppm	-	0.000 ppm	-	
THC	99.3	2.48 ppm	-	2.09 ppm	-	
CH ₄	99.3	2.48 ppm	-	2.09 ppm	-	
NMHC	99.3	0.01 ppm	-	0.00 ppm	-	
RH	100.0	99 %	-	93 %	-	
BP	100.0	960 mb	-	959 mb	-	
Ambient TPX	100.0	23.1 °C	-	13.9 °C	-	
Station TPX	100.0	25.2 °C	-	23.8 °C	-	
Wind Speed	100.0	21.4 kph	-	10.6 kph	-	
Wind Direction	100.0	-	-	-	-	

SIGNATURE OF COMPANY REPRESENTATIVE

FOR ALBERTA ENVIRONMENT USE ONLY

Exceedance Summary Report

SO₂ 1-Hour Exceedances

Measured concentrations of sulphur dioxide were below the 1-hour AAAQO of 172 ppb.

SO₂ 24-Hour Exceedances

Measured concentrations of sulphur dioxide were below the 24-hour AAAQO of 48.0 ppb.

In accordance with EPEA and the Substance Release Regulation.

In accordance with A Guide to Release Reporting and the Alberta Ambient Air Quality Objectives and Guidelines Summary.

TABLE OF CONTENTS

<u>Title</u>	<u>Page</u>
SUMMARY	2
MONTHLY CONTINUOUS DATA SUMMARY REPORT	3
SOUR GAS SUMMARY REPORT	4
EXCEEDANCE SUMMARY REPORT	5
TABLE OF CONTENTS	6
<hr/>	
1.0 Discussion	8
<hr/>	
2.0 Project Personnel	10
<hr/>	
3.0 Plant Monthly Required AMD Summary	10
<hr/>	
4.0 Calculations and Results	10
<hr/>	
5.0 Methods and Procedures	11
<hr/>	
Appendix I	Continuous Monitoring Data Results 14
	Sulphur Dioxide 15
	Total Reduced Sulphur 21
	Total Hydrocarbon 27
	Methane 33
	Non-Methane Hydrocarbon 39
	Wind Speed 45
	Wind Direction 50
	Relative Humidity 53
	Barometric Pressure 56
	Ambient Temperature 59
	Station Temperature 62
Appendix II	Equipment Calibration Results 65
	Sulphur Dioxide 66
	Total Reduced Sulphur 69
	Total Hydrocarbon 74
	Wind System 81
	Meteorological System Check 84

	Calibrators	86
	Calibration Gases	90
Appendix III	Maximum Instantaneous Data	95
Appendix IV	Report Certification Form	108
Appendix V	Data Validation Certification Form	110

1.0 Discussion

This monthly report consists of continuous monitoring results for the following parameters: Sulphur Dioxide (SO₂), Total Reduced Sulphur (TRS), Total Hydrocarbon (THC), Methane (CH₄), Non-Methane Hydrocarbon (NMHC), Relative Humidity (RH), Barometric Pressure (BP), Ambient Temperature (AmbTPX), Station Temperature (StnTPX), Wind Speed (WS) and Wind Direction (WD).

The sample inlet filter for all continuous air analyzers are replaced before the calibration begins. The sample manifold is cleaned during the site visit each month.

Control checks, consisting of a zero and span, are conducted daily on all continuous air monitors. In place of the air sample, zero air (from scrubbed air or gas cylinders) is used for zero checks, and a known concentration of the pollutant being analyzed is used for span checks. These checks are controlled by automatic timers and valves. The total zero span cycle is completed within an hour, the commencement of the zero span cycle is at the beginning of the hour.

Multipoint calibrations are done a minimum of once a month for each continuous air monitor. An additional calibration is required under the following conditions: 1) within three days after the initial start-up and stabilization of a newly installed instrument, 2) prior to shut-down or moving of an instrument which has been working to specification, and 3) when major repair has been done on the instrument.

Time during the first multi-point calibration is not considered downtime (Data is flagged as C). If more than one calibration is performed during the month, the time during the additional calibration is considered as downtime (Data is flagged as C1).

Only one zero/span check is run per day. Time during the zero/span check is not considered as downtime (Data is flagged as S). If an extra zero/span check is performed, the time during the additional check is considered as downtime (Data is flagged as S1).

The AMD requires each instrument and accompanying data recording system to be operational 90% of the time, at a minimum, for each monthly monitoring period.

All sampling, analysis, and QA/QC for this project was performed by Maxxam Analytics and complies with the Alberta Air Monitoring Directive.

Data contained in this monthly report has undergone the verification and validation based on the requirements of the AMD Chapter 6: Ambient Data Quality (December, 2016). The descriptions of the data verification and validation process can be found in Section 5 of this report. Instantaneous data, where applicable, is provided for reference purposes and has not undergone zero correction. The minimum and maximum statistics are highlighted in the data table and are for reference only. The highlighted cells are based on the software's interpretation of the exact position of the minimum or maximum value. The visual presentation of these statistics may not be the obvious choice in a data range due to rounding, truncating or analyzer specifications.

Hourly/minute data have been reviewed based on daily zero/span results and multi-point calibration results. Data may be considered invalid if a zero-corrected span check in excess of +/- 10% of the span concentration (established by the previous multi-point calibration) is encountered and/or significant differences in the calibration factor occurs (greater than 10%).

SULPHUR DIOXIDE (SO₂)

- Operational time for the monitoring period was 100%.
- The routine monthly calibration was performed on April 10.

TOTAL REDUCED SULPHUR (TRS)

- Operational time for the monitoring period was 99.7%, equivalent to two hours of downtime.
- The routine monthly calibration was performed on April 10.
- An as-found response check was completed on April 30 to assess analyzer performance after a slight drift in span response towards the upper acceptance limit on April 28. The results met AMD requirements. Two hours of downtime were incurred due to the additional quality check.
- The case fan was replaced during the April 30 site visit.

TOTAL HYDROCARBONS (THC), METHANE (CH₄) and NON-METHANE HYDROCARBONS (NMHC)

- Operational time for the monitoring period was 99.3%, equivalent to five hours of downtime.
- The routine monthly calibration was performed on April 10. The fuel (H₂), carrier (N₂) and span gas cylinders were replaced during this site visit.
- The NMHC span response exhibited an abrupt biased high drift on April 22. A repeat zero-span check conducted on April 23 showed the same trend. This prompted a site visit on April 24 where a repeat calibration was successfully completed as a corrective action. Five hours of downtime were incurred due to this event.
- The canister sampler is programmed to draw in a whole air sample when the 5-minute average concentration of NMHC is above 0.30 ppm. A representative sample of ambient air is collected over a one-hour period when the canister event is triggered. No canister event was recorded this month. A trigger test was performed during the routine monthly calibration on April 10 to assess the effectiveness of the canister system.

WIND SPEED (WS) and WIND DIRECTION (WD)

- Operational time for the monitoring period was 100%.
- The routine annual wind system audit was performed on April 4.
- Wind data is reported as vector wind speed and vector wind direction. Wind direction is defined as the direction from which the wind is blowing from and is measured in degrees from true north.

RELATIVE HUMIDITY (RH)

- Operational time for the monitoring period was 100%.
- A humidity sensor audit was conducted on April 10. The result was satisfactory.

BAROMETRIC PRESSURE (BP)

- Operational time for the monitoring period was 100%.
- A pressure sensor audit was conducted on April 10. The result was satisfactory.

AMBIENT TEMPERATURE (AmbTPX)

- Operational time for the monitoring period was 100%.
- A temperature sensor audit was conducted on April 10. The result was satisfactory.

STATION TEMPERATURE (StnTPX)

- Operational time for the monitoring period was 100%.

2.0 Project Personnel

Karla Reesor was the contact for Peace River Area Monitoring Program Committee and the Maxxam field technicians were Christopher Wesson and Limin Li.

3.0 Plant Monthly Required AMD Summary

All data collected this month was compliant with the requirements outlined in the Air Monitoring Directive (Alberta Environment and Parks, 2016).

The operational time for all continuous ambient air analyzers, meteorological systems and data acquisition systems were above the 90% requirement.

4.0 Calculations and Results

All calculations and reporting of results follow the methods described in the AMD, 2016.

5.0 Methods and Procedures

The following methods and procedures were used to complete the monitoring program:

Maxxam AIR SOP-00001: Methane, Non-Methane Hydrocarbon Analyzer Monitoring

Maxxam AIR SOP-00013: RM Young Wind Monitor Calibration

Maxxam AIR SOP-00209: Ambient Sulphur Monitoring

There were no deviations from the prescribed methods.

The following instruments were used to perform the test program:

Sulphur Dioxide - Thermo 43C UV Fluorescent Analyzer

Total Reduced Sulphur - Thermo 43i - TLE UV Fluorescent Analyzer

Methane, Non-Methane Hydrocarbon - Thermo 55i FID Analyzer

Wind System - RM Young Unit

Relative Humidity - RM Young Unit

Barometric Pressure - Met One Unit

Ambient Temperature - RM Young Unit

Datalogger - ESC 8832

The following steps were used to complete the data verification and validation process:

Level 0 Preliminary Verification

Level 0 data are raw data obtained directly from the data acquisition system (DAS). Under the step of Level 0, these data undergo a certain amount of manual or automated screening and flagging. It included a) identification of periods of missing data; b) verification of time stamps against reference time; c) verification that instrument diagnostics/datalogger flags indicate normal operation; d) comparison of data to upper and lower limits; e) rate of change flagging indicating that data changed too rapidly or not at all; and f) verification that zero, span and multipoint performance checks are within specifications. This level of verification is performed on a daily basis.

Level 1 Primary Validation

Validation actions under the step of Level 1 include a) review of all screening flags assigned during preliminary verification; b) review of all supporting site information and documentation; c) review of operational acceptance limits for each parameter/analyzer; d) review of daily zero/span and monthly calibration results for all gaseous parameters; and e) application of any necessary adjustments to data (e.g. baseline adjustments, below zero adjustments). This level of validation is performed on a monthly basis.

Level 2 Final Validation


The purpose of Level 2 validation is to verify that there are no inconsistencies among related data, or among regional data measured at nearby sites.

Level 3 Independent Data Review

Level 3 validation is the last step of data review, and it is completed by an individual that is independent of both field operations and primary data validation. A final independent QA review and endorsement is performed during this step before data is submitted to Alberta Environment.

Post-Final Validation

The Post-Final Validation step serves to re-evaluate the data that errors or omissions are discovered and/or suspected after the initial submittal of data. Any data issues or patterns which were not clear on a monthly basis are highlighted during this step. This validation is performed on an annual basis.



Source: Air Monitoring Directive (December 2016), Chapter 6, Ambient Data Quality; Figure 1 Data Collection and Management Process Flow Chart

APPENDIX I
CONTINUOUS MONITORING DATA RESULTS

SULPHUR DIOXIDE



PEACE RIVER AREA MONITORING PROGRAM COMMITTEE
Three Creeks 986b Station - April 2018

SULPHUR DIOXIDE Hourly Averages (SO₂ ppb)

HR START (MST)	0:00	1:00	2:00	3:00	4:00	5:00	6:00	7:00	8:00	9:00	10:00	11:00	12:00	13:00	14:00	15:00	16:00	17:00	18:00	19:00	20:00	21:00	22:00	23:00	DAILY MIN.	DAILY MAX.	24-HR AVG.	RDGS.	
HR END (MST)	0:59	1:59	2:59	3:59	4:59	5:59	6:59	7:59	8:59	9:59	10:59	11:59	12:59	13:59	14:59	15:59	16:59	17:59	18:59	19:59	20:59	21:59	22:59	23:59					
DAY 1	1	1	1	1	1	1	1	1	2	2	1	1	1	1	1	1	1	1	0	S	0	1	1	0	2	1	24		
DAY 2	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	2	1	S	1	1	1	1	1	1	24		
DAY 3	1	1	1	1	1	1	0	0	0	0	0	0	0	0	0	0	0	0	S	0	0	0	0	0	0	0	24		
DAY 4	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	S	0	0	0	0	0	0	0	0	24		
DAY 5	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	S	0	1	1	1	0	1	0	0	1	24		
DAY 6	1	1	0	0	1	0	1	1	1	1	0	0	1	0	1	S	1	1	1	1	1	1	1	1	1	0	1	24	
DAY 7	1	1	1	1	1	1	1	1	1	1	1	1	2	1	S	1	1	1	1	1	1	1	1	0	1	0	2	1	24
DAY 8	1	1	0	1	1	1	1	1	1	1	1	1	1	S	1	1	1	1	1	1	1	1	1	1	1	0	1	1	24
DAY 9	1	1	1	1	1	1	1	1	2	1	1	1	S	1	1	1	1	1	1	1	0	1	1	1	0	0	2	1	24
DAY 10	0	0	0	0	0	0	0	0	0	0	0	S	0	C	C	C	C	0	0	0	0	0	0	0	0	0	0	0	24
DAY 11	0	0	0	0	0	0	0	0	0	0	S	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	24
DAY 12	0	0	0	0	0	0	1	1	0	S	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	1	0	24
DAY 13	0	0	0	0	0	0	0	0	S	0	0	0	1	0	0	0	0	0	0	0	0	0	0	0	1	0	1	0	24
DAY 14	1	0	0	0	0	0	0	S	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	1	0	24
DAY 15	0	0	0	0	0	0	S	0	0	0	0	0	0	0	0	1	0	0	0	0	0	1	1	0	0	0	1	0	24
DAY 16	0	0	0	0	0	S	0	0	0	0	0	0	0	1	0	0	0	0	0	1	0	0	0	0	0	0	1	0	24
DAY 17	0	0	0	0	S	0	0	1	1	0	0	1	0	0	0	0	0	1	0	1	1	1	1	0	0	0	1	0	24
DAY 18	1	1	0	S	0	0	0	0	0	1	1	1	1	0	0	0	0	0	0	0	0	0	0	0	0	0	1	0	24
DAY 19	0	0	S	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	24
DAY 20	0	S	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	24
DAY 21	S	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	S	0	0	24
DAY 22	0	0	0	0	0	0	0	0	0	0	0	1	0	1	1	0	1	1	1	1	1	1	1	S	1	0	1	0	24
DAY 23	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	S	1	1	1	1	1	24
DAY 24	1	1	1	1	1	1	1	1	1	1	1	1	0	0	0	0	0	0	0	0	S	0	0	0	0	0	1	1	24
DAY 25	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	S	0	0	0	0	0	0	0	0	24
DAY 26	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	S	0	0	0	0	0	0	0	0	0	0	24
DAY 27	0	0	0	0	0	0	0	0	0	0	1	0	0	0	0	0	0	S	0	0	0	0	0	0	1	0	1	0	24
DAY 28	1	1	1	1	0	0	0	1	1	1	1	1	1	0	0	1	S	0	0	0	0	0	0	0	0	0	1	0	24
DAY 29	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	S	0	0	0	0	0	0	0	0	0	0	0	0	24
DAY 30	0	0	0	0	0	0	0	0	0	0	1	1	1	S	1	1	1	1	1	1	1	1	1	1	1	0	1	1	24
HOURLY MAX	1	1	1	1	1	1	1	1	2	2	1	1	2	1	1	1	1	2	1	1	1	1	1	1	1	1	1	1	
HOURLY AVG	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	


STATUS FLAG CODES

C	- MONTHLY CALIBRATION	Q	- QUALITY ASSURANCE
C1	- REPEAT CALIBRATION	R	- RECOVERY
Y	- MAINTENANCE	X	- MACHINE MALFUNCTION
S	- DAILY ZERO/SPAN CHECK	G	- OUT FOR REPAIR
S1	- REPEAT ZERO/SPAN CHECK	P	- POWER FAILURE

OBJECTIVE LIMIT:

ALBERTA ENVIRONMENT: 1-HR 172 ppb 24-HR 48 ppb


24 HR AVERAGES April 2018



MONTHLY SUMMARY

NUMBER OF 1-HR EXCEEDANCES:	0
NUMBER OF 24-HR EXCEEDANCES:	0
NUMBER OF NON-ZERO READINGS:	232
MINIMUM 1-HR AVERAGE:	0 ppb @ HOUR 19 ON DAY 1
MAXIMUM 1-HR AVERAGE:	2 ppb @ HOUR 8 ON DAY 1
MAXIMUM 24-HR AVERAGE:	1 ppb ON DAY 1
IZS CALIBRATION TIME:	31 hrs
MONTHLY CALIBRATION TIME:	4 hrs
OPERATIONAL TIME:	720 hrs
AMD OPERATION UPTIME:	100.0 %
STANDARD DEVIATION:	0 ppb
MONTHLY AVERAGE:	0 ppb

SULPHUR DIOXIDE Hourly Averages (SO₂ ppb)



Wind: PRAMP_986
 Poll.: PRAMP_986-SO₂[ppb]
 Monthly: 18/04
 Type: PollutionRose
 Direction: Blowing From (Wind Frequency)
 Based On 1 Hr.

Calm: 6.73%

Calm Avg: 0.38 [ppb]

Direction	0-3	3-10	10-85	85-170	>170.0	Total
N	14.6	0.0	0.0	0.0	0.0	14.6
NE	5.9	0.0	0.0	0.0	0.0	5.9
E	11.8	0.0	0.0	0.0	0.0	11.8
SE	13.3	0.0	0.0	0.0	0.0	13.3
S	10.8	0.0	0.0	0.0	0.0	10.8
SW	20.6	0.0	0.0	0.0	0.0	20.6
W	7.2	0.0	0.0	0.0	0.0	7.2
NW	9.1	0.0	0.0	0.0	0.0	9.1
Summary	93.3	0.0	0.0	0.0	0.0	93.3

% Icon Classes (ppb)

93 0-3


0 3-10

0 10-85

0 85-170

0 >170.0

PRAMP_986 Poll.: PRAMP_986-SO2[ppb] 2018/04/01 00:00 - 2018/04/30 23:00 Calm: 6.73% Calm Poll Avg: 0.38[ppb]



SO2[ppb] Calibration: PRAMP_986 Monthly: 18/04 Type: Span



Span Meas Span Ref Span Low Span High

TOTAL REDUCED SULPHUR




TOTAL REDUCED SULPHUR Hourly Averages (TRS ppb)

HR START (MST)	0:00	1:00	2:00	3:00	4:00	5:00	6:00	7:00	8:00	9:00	10:00	11:00	12:00	13:00	14:00	15:00	16:00	17:00	18:00	19:00	20:00	21:00	22:00	23:00	DAILY	DAILY	24-HR	RDGS.
HR END (MST)	0:59	1:59	2:59	3:59	4:59	5:59	6:59	7:59	8:59	9:59	10:59	11:59	12:59	13:59	14:59	15:59	16:59	17:59	18:59	19:59	20:59	21:59	22:59	23:59	MIN.	MAX.	AVG.	
DAY																												
1	0.30	0.30	0.30	0.30	0.30	0.36	0.44	0.52	0.46	0.32	0.28	0.27	0.26	0.26	0.24	0.24	0.23	0.23	0.25	S	0.32	0.25	0.24	0.23	0.52	0.30	24	
2	0.22	0.24	0.24	0.23	0.24	0.23	0.26	0.26	0.27	0.26	0.28	0.27	0.26	0.28	0.29	0.28	0.26	0.28	0.28	S	0.35	0.30	0.27	0.30	0.22	0.35	0.27	24
3	0.28	0.31	0.34	0.31	0.31	0.31	0.29	0.30	0.32	0.31	0.30	0.30	0.29	0.29	0.30	0.29	0.28	0.30	S	0.37	0.29	0.29	0.30	0.29	0.28	0.37	0.30	24
4	0.29	0.29	0.29	0.29	0.30	0.29	0.29	0.29	0.30	0.30	0.32	0.30	0.30	0.30	0.31	0.30	0.30	S	0.40	0.32	0.30	0.28	0.29	0.28	0.28	0.40	0.30	24
5	0.27	0.27	0.27	0.26	0.26	0.27	0.27	0.27	0.28	0.26	0.29	0.27	0.27	0.27	0.26	0.25	S	0.33	0.27	0.26	0.26	0.27	0.26	0.25	0.25	0.33	0.27	24
6	0.28	0.30	0.27	0.27	0.28	0.31	0.41	0.37	0.31	0.27	0.26	0.27	0.26	0.26	0.26	S	0.35	0.28	0.28	0.27	0.27	0.26	0.27	0.26	0.26	0.41	0.29	24
7	0.26	0.27	0.28	0.27	0.28	0.28	0.29	0.31	0.28	0.30	0.28	0.27	0.29	0.28	S	0.35	0.27	0.27	0.26	0.27	0.27	0.27	0.27	0.25	0.25	0.35	0.28	24
8	0.26	0.26	0.25	0.26	0.26	0.25	0.25	0.28	0.27	0.27	0.27	0.26	S	0.34	0.28	0.27	0.27	0.25	0.26	0.26	0.25	0.27	0.27	0.25	0.34	0.27	24	
9	0.27	0.27	0.28	0.30	0.29	0.28	0.30	0.31	0.29	0.31	0.27	0.26	S	0.34	0.30	0.27	0.28	0.26	0.27	0.27	0.25	0.25	0.27	0.25	0.34	0.28	24	
10	0.31	0.28	0.26	0.27	0.26	0.26	0.26	0.25	0.26	C	C	C	C	C	C	0.42	0.34	0.18	0.19	0.20	0.20	0.21	0.21	0.21	0.18	0.42	0.25	24
11	0.21	0.26	0.26	0.26	0.25	0.25	0.29	0.28	0.29	0.31	S	0.35	0.29	0.28	0.29	0.28	0.29	0.29	0.30	0.30	0.28	0.28	0.29	0.27	0.21	0.35	0.28	24
12	0.29	0.29	0.29	0.33	0.29	0.29	0.28	0.27	0.29	S	0.33	0.27	0.26	0.25	0.25	0.26	0.26	0.26	0.27	0.27	0.27	0.26	0.27	0.27	0.25	0.33	0.28	24
13	0.27	0.29	0.30	0.28	0.28	0.29	0.29	0.28	S	0.34	0.31	0.29	0.28	0.28	0.27	0.28	0.27	0.28	0.26	0.25	0.27	0.27	0.28	0.27	0.25	0.34	0.28	24
14	0.27	0.31	0.32	0.31	0.29	0.29	0.31	S	0.36	0.31	0.29	0.30	0.29	0.28	0.28	0.27	0.27	0.26	0.27	0.27	0.28	0.31	0.30	0.28	0.26	0.36	0.29	24
15	0.29	0.30	0.29	0.29	0.29	0.28	S	0.37	0.31	0.29	0.30	0.30	0.31	0.30	0.32	0.35	0.34	0.34	0.32	0.28	0.29	0.30	0.30	0.32	0.28	0.37	0.31	24
16	0.31	0.29	0.30	0.29	0.30	S	0.38	0.32	0.32	0.33	0.33	0.32	0.30	0.29	0.30	0.29	0.29	0.28	0.29	0.29	0.30	0.30	0.30	0.29	0.28	0.38	0.30	24
17	0.28	0.28	0.28	0.28	S	0.35	0.30	0.31	0.32	0.32	0.32	0.35	0.32	0.31	0.31	0.32	0.31	0.33	0.33	0.30	0.29	0.30	0.27	0.27	0.35	0.31	24	
18	0.25	0.27	0.26	S	0.33	0.29	0.26	0.31	0.33	0.28	0.29	0.28	0.26	0.25	0.23	0.25	0.25	0.24	0.24	0.24	0.23	0.24	0.25	0.24	0.23	0.33	0.26	24
19	0.27	0.28	S	0.33	0.30	0.31	0.31	0.30	0.29	0.27	0.28	0.27	0.27	0.28	0.25	0.27	0.27	0.28	0.28	0.29	0.27	0.30	0.30	0.31	0.25	0.33	0.29	24
20	0.31	S	0.37	0.32	0.32	0.31	0.31	0.32	0.32	0.30	0.29	0.31	0.29	0.27	0.29	0.27	0.29	0.27	0.27	0.27	0.27	0.26	0.26	0.26	0.26	0.37	0.29	24
21	S	0.38	0.32	0.34	0.35	0.31	0.34	0.32	0.34	0.31	0.29	0.39	0.31	0.32	0.30	0.29	0.29	0.29	0.29	0.31	0.31	0.30	0.31	S	0.29	0.39	0.32	24
22	0.41	0.36	0.32	0.35	0.31	0.32	0.32	0.34	0.33	0.33	0.32	0.31	0.30	0.30	0.30	0.30	0.29	0.28	0.30	0.39	0.32	0.33	S	0.37	0.28	0.41	0.33	24
23	0.36	0.36	0.36	0.34	0.36	0.37	0.37	0.38	0.36	0.36	0.34	0.34	0.33	0.31	0.32	0.30	0.30	0.30	0.30	0.30	0.30	S	0.38	0.36	0.30	0.38	0.34	24
24	0.35	0.32	0.32	0.31	0.31	0.31	0.30	0.30	0.32	0.32	0.32	0.31	0.30	0.31	0.30	0.31	0.30	0.29	0.28	0.30	S	0.36	0.40	0.38	0.28	0.40	0.32	24
25	0.30	0.34	0.31	0.33	0.34	0.32	0.31	0.31	0.29	0.28	0.28	0.27	0.27	0.28	0.29	0.28	0.27	0.27	S	0.34	0.29	0.29	0.29	0.27	0.34	0.30	24	
26	0.30	0.29	0.34	0.37	0.39	0.40	0.37	0.39	0.34	0.32	0.33	0.35	0.33	0.33	0.33	0.32	0.31	0.30	S	0.37	0.37	0.45	0.35	0.33	0.29	0.45	0.35	24
27	0.35	0.36	0.37	0.36	0.33	0.33	0.30	0.33	0.32	0.32	0.33	0.32	0.33	0.33	0.35	0.33	0.33	S	0.36	0.33	0.36	0.36	0.34	0.32	0.30	0.37	0.34	24
28	0.33	0.31	0.30	0.29	0.28	0.29	0.31	0.30	0.32	0.32	0.33	0.34	0.31	0.33	0.36	0.31	S	0.32	0.29	0.30	0.30	0.30	0.34	0.33	0.28	0.36	0.31	24
29	0.34	0.33	0.34	0.32	0.32	0.33	0.30	0.30	0.29	0.31	0.28	0.28	0.26	0.27	S	0.31	0.27	0.26	0.26	0.25	0.25	0.26	0.25	0.25	0.34	0.29	24	
30	0.28	0.28	0.26	0.27	0.24	0.24	0.25	0.25	0.24	0.23	0.23	C1	C1	0.32	S	0.26	0.21	0.18	0.16	0.16	0.14	0.14	0.12	0.16	0.12	0.32	0.22	22
HOURLY MAX	0.41	0.38	0.37	0.37	0.39	0.40	0.44	0.52	0.46	0.36	0.34	0.39	0.34	0.34	0.36	0.42	0.35	0.34	0.40	0.39	0.37	0.45	0.40	0.38				
HOURLY AVG	0.29	0.30	0.30	0.30	0.30	0.31	0.32	0.31	0.30	0.30	0.30	0.30	0.29	0.29	0.30	0.29	0.28	0.28	0.29	0.28	0.29	0.29	0.29	0.28				

STATUS FLAG CODES

C	- MONTHLY CALIBRATION	Q	- QUALITY ASSURANCE
C1	- REPEAT CALIBRATION	R	- RECOVERY
Y	- MAINTENANCE	X	- MACHINE MALFUNCTION
S	- DAILY ZERO/SPAN CHECK	G	- OUT FOR REPAIR
S1	- REPEAT ZERO/SPAN CHECK	P	- POWER FAILURE


24 HR AVERAGES April 2018



MONTHLY SUMMARY

NUMBER OF NON-ZERO READINGS:	682			
MINIMUM 1-HR AVERAGE:	0.12 ppb	@ HOUR	22	ON DAY
MAXIMUM 1-HR AVERAGE:	0.52 ppb	@ HOUR	7	ON DAY
MAXIMUM 24-HR AVERAGE:	0.35 ppb			ON DAY
IZS CALIBRATION TIME:	30	hrs	OPERATIONAL TIME:	718 hrs
MONTHLY CALIBRATION TIME:	6	hrs	AMD OPERATION UPTIME:	99.7 %
STANDARD DEVIATION:	0.04		MONTHLY AVERAGE:	0.29 ppb

TOTAL REDUCED SULPHUR Hourly Averages (TRS ppb)



Wind: PRAMP_986
 Poll.: PRAMP_986-TRS[ppb]
 Monthly: 18/04
 Type: PollutionRose
 Direction: Blowing From (Wind Frequency)
 Based On 1 Hr.

Calm: 6.75% Calm Avg: 0.30 [ppb]

Direction	0-1	1-3	3-10	>10.0	Total
N	14.7	0.0	0.0	0.0	14.7
NE	5.9	0.0	0.0	0.0	5.9
E	11.8	0.0	0.0	0.0	11.8
SE	13.4	0.0	0.0	0.0	13.4
S	10.6	0.0	0.0	0.0	10.6
SW	20.7	0.0	0.0	0.0	20.7
W	7.2	0.0	0.0	0.0	7.2
NW	9.1	0.0	0.0	0.0	9.1
Summary	93.2	0.0	0.0	0.0	93.2

% Icon Classes (ppb)

93



0




0



0



PRAMP_986 Poll.: PRAMP_986-TRS[ppb] 2018/04/01 00:00 - 2018/04/30 23:00 Calm: 6.75% Calm Poll Avg: 0.30[ppb]



TRS[ppb] Calibration: PRAMP_986 Monthly: 18/04 Type: Span



Span Meas Span Ref Span Low Span High

TOTAL HYDROCARBON




PEACE RIVER AREA MONITORING PROGRAM COMMITTEE
Three Creeks 986b Station - April 2018

TOTAL HYDROCARBONS Hourly Averages (THC ppm)

HR START (MST)	0:00	1:00	2:00	3:00	4:00	5:00	6:00	7:00	8:00	9:00	10:00	11:00	12:00	13:00	14:00	15:00	16:00	17:00	18:00	19:00	20:00	21:00	22:00	23:00	DAILY MIN.	DAILY MAX.	24-HR AVG.	RDGS.	
HR END (MST)	0:59	1:59	2:59	3:59	4:59	5:59	6:59	7:59	8:59	9:59	10:59	11:59	12:59	13:59	14:59	15:59	16:59	17:59	18:59	19:59	20:59	21:59	22:59	23:59					
DAY																													
1	2.02	2.02	2.04	2.05	2.09	2.18	2.35	2.48	2.35	2.09	2.06	2.02	2.02	2.02	2.01	2.00	2.03	2.04	2.07	2.28	S	1.99	1.96	1.96	1.96	2.48	2.09	24	
2	1.97	2.01	1.98	1.99	1.98	2.03	1.98	1.96	1.97	1.97	1.98	1.95	2.00	1.98	1.98	1.98	1.96	1.96	S	1.95	1.96	2.05	1.97	1.95	2.05	1.98	24		
3	2.03	1.98	2.10	2.04	1.98	2.00	2.06	2.14	1.97	1.97	1.98	2.00	1.99	1.97	2.00	2.02	1.98	2.01	S	2.04	1.96	1.96	1.97	1.99	1.96	2.14	2.01	24	
4	1.97	1.97	1.98	1.98	1.98	1.97	1.97	1.98	1.97	2.00	2.00	2.04	2.03	2.02	2.02	2.01	2.02	S	1.94	2.08	2.00	1.98	2.20	1.97	1.94	2.20	2.00	24	
5	1.97	1.97	1.97	1.98	1.98	1.97	2.13	1.97	1.97	1.98	1.99	1.99	2.01	2.01	2.00	1.96	S	1.96	1.96	1.96	1.96	1.97	1.97	1.97	1.96	2.13	1.98	24	
6	1.98	1.98	1.97	1.97	1.98	1.99	2.00	1.99	1.98	1.99	1.97	1.97	1.98	1.95	S	1.95	1.93	1.96	1.95	1.95	1.96	1.98	1.97	1.93	2.00	1.97	24		
7	1.98	1.97	1.98	1.99	1.98	1.98	1.99	2.00	2.00	1.99	1.98	1.97	1.97	1.96	S	1.93	1.95	1.93	1.94	1.95	1.94	1.96	1.96	1.98	1.93	2.00	1.97	24	
8	1.99	1.98	1.98	1.98	1.99	1.99	2.01	2.03	2.03	2.02	2.02	2.02	2.01	S	1.99	2.01	1.99	1.98	1.98	1.98	2.01	2.02	2.02	2.03	1.98	2.03	2.00	24	
9	2.05	2.06	2.06	2.08	2.07	2.07	2.05	2.04	2.03	2.02	2.00	1.99	S	2.00	2.01	2.01	1.99	2.10	2.11	2.36	2.01	1.98	2.01	2.04	1.98	2.36	2.05	24	
10	2.01	1.99	1.96	1.96	1.96	1.96	1.95	1.94	1.94	C	C	C	C	C	C	1.97	1.96	1.96	1.95	1.94	1.94	1.96	1.97	1.96	1.94	2.01	1.96	24	
11	1.97	1.98	1.98	1.97	1.95	1.96	1.96	1.97	1.97	1.97	S	1.97	1.95	1.94	1.94	1.97	1.96	1.95	1.98	1.98	1.96	1.95	1.97	1.98	1.94	1.98	1.96	24	
12	1.99	1.98	1.96	1.96	1.97	1.98	1.98	1.98	1.98	S	1.97	1.94	1.95	1.97	1.96	1.97	1.98	2.00	2.00	1.99	1.98	2.00	2.03	2.04	1.94	2.04	1.98	24	
13	2.04	2.04	2.06	2.05	2.03	2.04	2.04	2.04	S	2.05	2.02	2.02	2.02	1.98	1.98	1.97	1.97	2.03	2.21	1.98	2.00	2.01	1.99	1.97	1.97	2.21	2.02	24	
14	1.96	2.02	2.05	2.00	1.94	1.94	1.95	S	1.93	1.94	1.95	1.95	1.95	1.95	1.95	1.95	1.95	1.95	1.95	1.95	1.96	1.97	1.97	1.96	1.94	1.93	2.05	1.96	24
15	1.95	1.97	1.98	1.96	1.97	1.98	S	1.98	2.01	1.98	2.01	2.00	2.01	2.01	2.03	2.00	2.00	1.99	1.98	1.98	1.96	1.95	1.97	1.98	1.95	2.03	1.98	24	
16	1.98	1.96	1.97	1.98	1.98	S	1.98	1.98	1.96	1.97	1.98	1.98	1.97	1.95	1.95	1.97	1.97	1.95	1.95	1.97	1.97	1.95	1.95	1.96	1.95	1.98	1.97	24	
17	1.96	1.96	1.94	1.96	S	2.04	2.04	2.04	2.08	1.99	1.98	1.97	1.95	1.97	1.98	1.97	1.96	1.98	1.97	1.97	1.97	1.97	1.98	1.98	1.94	2.08	1.98	24	
18	1.98	1.97	2.00	S	2.02	2.00	2.01	2.00	2.00	2.00	2.00	1.99	1.96	1.96	1.96	1.96	1.96	1.96	1.96	1.96	1.96	1.96	1.95	1.93	1.92	1.92	2.02	1.98	24
19	1.95	1.96	S	1.93	1.94	1.96	1.97	1.97	1.97	1.97	1.96	1.96	1.95	1.95	1.95	1.95	1.95	1.95	1.95	1.95	1.96	1.95	1.95	1.96	1.93	1.97	1.95	24	
20	1.95	S	1.92	1.93	1.96	1.96	1.94	1.93	1.96	1.96	1.97	1.97	1.96	1.96	1.96	1.96	1.96	1.96	1.96	1.96	1.96	1.96	1.96	1.98	2.01	1.92	2.01	1.96	24
21	S	2.02	2.00	2.14	2.22	2.12	2.06	2.00	2.00	1.97	1.96	1.96	1.95	1.96	1.95	1.96	1.95	1.96	1.95	2.03	1.95	1.95	1.95	S	1.95	2.22	2.00	24	
22	1.96	1.95	1.95	1.95	1.95	1.95	1.97	1.98	2.01	2.02	2.02	2.01	2.01	2.02	1.99	1.97	1.96	1.96	1.97	1.98	1.98	2.02	S	1.96	1.95	2.02	1.98	24	
23	1.98	2.01	2.02	2.00	2.00	2.02	2.02	2.03	S1	2.02	2.02	2.02	2.02	2.01	2.01	2.00	2.00	1.99	2.00	2.01	S	2.01	2.00	1.98	2.03	2.01	2.01	23	
24	1.98	1.98	1.96	1.96	1.99	1.99	1.99	1.98	1.98	1.99	1.98	1.97	1.96	C1	C1	C1	C1	1.94	1.94	1.94	S	1.95	2.00	2.00	1.94	2.00	1.97	20	
25	1.99	1.95	1.97	2.02	2.02	2.01	1.96	1.96	1.98	1.96	1.96	1.95	1.95	1.96	1.96	1.96	1.95	1.94	1.94	S	1.94	1.96	1.96	1.96	1.94	2.02	1.97	24	
26	1.95	1.95	1.97	1.95	1.97	1.98	2.08	2.00	1.95	1.97	1.98	1.97	1.96	1.95	1.96	1.97	1.95	1.94	S	1.95	1.95	1.96	1.99	2.03	1.94	2.08	1.97	24	
27	2.10	2.11	2.09	2.12	2.11	2.08	2.04	1.99	1.96	1.94	1.94	1.95	1.96	1.95	1.95	1.95	1.95	S	1.93	1.96	1.95	1.94	1.95	1.97	1.93	2.12	2.00	24	
28	1.99	1.99	2.04	2.00	1.93	1.91	1.92	1.93	1.95	1.95	1.96	1.95	1.94	1.94	1.96	1.94	S	1.94	1.95	1.94	1.94	1.95	1.95	1.95	1.91	2.04	1.95	24	
29	1.94	1.94	1.94	1.92	1.93	1.96	1.96	1.94	1.93	1.96	1.98	1.98	1.96	1.96	1.95	S	1.95	1.96	1.96	1.95	1.94	1.95	1.93	1.93	1.92	1.92	1.98	1.95	24
30	1.93	1.99	2.00	1.99	1.96	1.94	1.96	1.95	1.96	1.95	1.95	1.96	1.94	1.93	S	1.95	1.95	1.98	1.98	1.93	1.93	1.93	1.96	2.09	1.93	2.09	1.96	24	
HOURLY MAX	2.10	2.11	2.10	2.14	2.22	2.18	2.35	2.48	2.35	2.09	2.06	2.04	2.03	2.02	2.03	2.02	2.03	2.10	2.21	2.36	2.01	2.02	2.20	2.09					
HOURLY AVG	1.98	1.99	1.99	1.99	1.99	2.00	2.01	2.01	1.99	1.99	1.98	1.98	1.98	1.97	1.98	1.97	1.97	1.98	2.00	1.96	1.97	1.98	1.98						

STATUS FLAG CODES


C	- MONTHLY CALIBRATION	Q	- QUALITY ASSURANCE
C1	- REPEAT CALIBRATION	R	- RECOVERY
Y	- MAINTENANCE	X	- MACHINE MALFUNCTION
S	- DAILY ZERO/SPAN CHECK	G	- OUT FOR REPAIR
S1	- REPEAT ZERO/SPAN CHECK	P	- POWER FAILURE



MONTHLY SUMMARY

NUMBER OF NON-ZERO READINGS:	680					
MINIMUM 1-HR AVERAGE:	1.91	ppm	@ HOUR	5	ON DAY	28
MAXIMUM 1-HR AVERAGE:	2.48	ppm	@ HOUR	7	ON DAY	1
MAXIMUM 24-HR AVERAGE:	2.09	ppm			ON DAY	1
IZS CALIBRATION TIME:	30	hrs	OPERATIONAL TIME:	715	hrs	
MONTHLY CALIBRATION TIME:	5	hrs	AMD OPERATION UPTIME:	99.3	%	
STANDARD DEVIATION:	0.05		MONTHLY AVERAGE:	1.98	ppm	

TOTAL HYDROCARBONS Hourly Averages (THC ppm)



Wind: PRAMP_986
 Poll.: PRAMP_986-THC55[ppm]
 Monthly: 18/04
 Type: PollutionRose
 Direction: Blowing From (Wind Frequency)
 Based On 1 Hr.


Calm: 6.77%

Calm Avg: 2.01 [ppm]

Direction	0-2	2-3	3-5	5-10	>10.0	Total
N	12.4	2.4	0.0	0.0	0.0	14.7
NE	5.9	0.0	0.0	0.0	0.0	5.9
E	9.7	2.2	0.0	0.0	0.0	11.9
SE	8.0	5.5	0.0	0.0	0.0	13.4
S	9.0	1.9	0.0	0.0	0.0	10.9
SW	18.0	2.1	0.0	0.0	0.0	20.0
W	6.0	1.2	0.0	0.0	0.0	7.2
NW	3.4	5.7	0.0	0.0	0.0	9.1
Summary	72.3	20.9	0.0	0.0	0.0	93.2

% Icon Classes (ppm) 72 0-2 21 2-3 0 3-5 0 5-10 0 >10.0

PRAMP_986 Poll.: PRAMP_986-THC55[ppm] 2018/04/01 00:00 - 2018/04/30 23:00 Calm: 6.77% Calm Poll Avg: 2.01[ppm]



THC55[ppm] Calibration: PRAMP_986 Monthly: 18/04 Type: Span



■ Span Meas
 — Span Ref
 — Span Low
 — Span High

METHANE




METHANE Hourly Averages (CH₄ ppm)

HR START (MST)	0:00	1:00	2:00	3:00	4:00	5:00	6:00	7:00	8:00	9:00	10:00	11:00	12:00	13:00	14:00	15:00	16:00	17:00	18:00	19:00	20:00	21:00	22:00	23:00	DAILY	DAILY	24-HR	RDGS.	
HR END (MST)	0:59	1:59	2:59	3:59	4:59	5:59	6:59	7:59	8:59	9:59	10:59	11:59	12:59	13:59	14:59	15:59	16:59	17:59	18:59	19:59	20:59	21:59	22:59	23:59	MIN.	MAX.	AVG.		
DAY 1	2.02	2.02	2.04	2.05	2.09	2.18	2.35	2.48	2.35	2.09	2.06	2.02	2.02	2.02	2.01	2.00	2.03	2.04	2.07	2.28	S	1.99	1.96	1.96	1.96	2.48	2.09	2.4	
2	1.97	2.01	1.98	1.99	1.98	2.03	1.98	1.96	1.97	1.97	1.98	1.95	2.00	1.98	1.98	1.98	1.96	1.96	S	1.95	1.96	1.96	2.05	1.97	1.95	2.05	1.98	24	
3	2.03	1.98	2.10	2.04	1.98	2.00	2.06	2.14	1.97	1.97	1.98	2.00	1.99	1.97	2.00	2.02	1.98	2.01	S	2.04	1.96	1.96	1.97	1.99	1.96	2.14	2.01	24	
4	1.97	1.97	1.98	1.98	1.98	1.97	1.97	1.98	1.97	2.00	2.00	2.04	2.03	2.02	2.02	2.01	2.02	S	1.94	2.08	2.00	1.98	2.20	1.97	1.94	2.20	2.00	24	
5	1.97	1.97	1.97	1.98	1.98	1.97	2.13	1.97	1.97	1.98	1.99	1.99	2.01	2.01	2.00	1.96	S	1.96	1.96	1.96	1.96	1.97	1.97	1.97	1.96	2.13	1.98	24	
6	1.98	1.98	1.97	1.97	1.98	1.99	2.00	1.99	1.98	1.99	1.97	1.97	1.98	1.95	S	1.95	1.93	1.96	1.95	1.95	1.96	1.98	1.97	1.93	2.00	1.97	24		
7	1.98	1.97	1.98	1.99	1.98	1.98	1.99	2.00	2.00	1.99	1.98	1.97	1.97	1.96	S	1.93	1.95	1.93	1.94	1.95	1.94	1.96	1.96	1.98	1.93	2.00	1.97	24	
8	1.99	1.98	1.98	1.98	1.99	1.99	2.01	2.03	2.03	2.02	2.02	2.02	2.01	S	1.99	2.01	1.99	1.98	1.98	1.98	2.01	2.02	2.02	2.03	1.98	2.03	2.00	24	
9	2.05	2.06	2.06	2.08	2.07	2.07	2.05	2.04	2.03	2.02	2.00	1.99	S	2.00	2.01	2.01	1.99	2.10	2.11	2.36	2.01	1.98	2.01	2.04	1.98	2.36	2.05	24	
10	2.01	1.99	1.96	1.96	1.96	1.96	1.95	1.94	1.94	C	C	C	C	C	1.97	1.96	1.96	1.95	1.94	1.94	1.96	1.97	1.96	1.96	1.94	2.01	1.96	24	
11	1.97	1.98	1.98	1.97	1.95	1.96	1.96	1.97	1.97	1.97	S	1.97	1.95	1.94	1.94	1.97	1.96	1.95	1.98	1.98	1.96	1.95	1.97	1.98	1.94	1.98	1.96	24	
12	1.99	1.98	1.96	1.96	1.97	1.98	1.98	1.98	1.98	S	1.97	1.94	1.95	1.97	1.96	1.97	1.98	2.00	2.00	1.99	1.98	2.00	2.03	2.04	1.94	2.04	1.98	24	
13	2.04	2.04	2.06	2.05	2.03	2.04	2.04	2.04	S	2.05	2.02	2.02	2.02	1.98	1.98	1.97	1.97	2.03	2.21	1.98	2.00	2.01	1.99	1.97	1.97	2.21	2.02	24	
14	1.96	2.02	2.05	2.00	1.94	1.94	1.95	S	1.93	1.94	1.95	1.95	1.95	1.95	1.95	1.95	1.95	1.95	1.95	1.95	1.96	1.97	1.97	1.96	1.94	1.93	2.05	1.96	24
15	1.95	1.97	1.98	1.96	1.97	1.98	S	1.98	2.01	1.98	2.01	2.00	2.01	2.01	2.03	2.00	2.00	1.99	1.98	1.98	1.96	1.95	1.97	1.98	1.95	2.03	1.98	24	
16	1.98	1.96	1.97	1.98	1.98	S	1.98	1.98	1.96	1.97	1.98	1.98	1.97	1.95	1.95	1.97	1.97	1.95	1.95	1.97	1.97	1.95	1.95	1.96	1.95	1.98	1.97	24	
17	1.96	1.96	1.94	1.96	S	2.04	2.04	2.04	2.08	1.99	1.98	1.97	1.95	1.97	1.98	1.97	1.96	1.98	1.97	1.97	1.97	1.97	1.98	1.98	1.94	2.08	1.98	24	
18	1.98	1.97	2.00	S	2.02	2.00	2.01	2.00	2.00	2.00	2.00	1.99	1.96	1.96	1.96	1.96	1.96	1.96	1.96	1.96	1.96	1.96	1.95	1.93	1.92	1.92	2.02	1.98	24
19	1.95	1.96	S	1.93	1.94	1.96	1.97	1.97	1.97	1.97	1.96	1.96	1.95	1.95	1.95	1.95	1.95	1.95	1.95	1.95	1.96	1.95	1.95	1.96	1.93	1.97	1.95	24	
20	1.95	S	1.92	1.93	1.96	1.96	1.94	1.93	1.96	1.96	1.97	1.97	1.96	1.96	1.96	1.96	1.96	1.96	1.96	1.96	1.96	1.96	1.96	1.98	2.01	1.92	2.01	1.96	24
21	S	2.02	2.00	2.14	2.22	2.12	2.06	2.00	2.00	1.97	1.96	1.97	1.96	1.95	1.96	1.95	1.95	1.96	1.95	2.03	1.95	1.95	1.95	S	1.95	2.22	2.00	24	
22	1.96	1.95	1.95	1.95	1.95	1.95	1.97	1.98	2.01	2.02	2.02	2.01	2.01	2.02	1.99	1.97	1.96	1.97	1.98	1.98	2.02	S	1.96	1.95	2.02	1.98	2.02	24	
23	1.98	2.01	2.02	2.00	2.00	2.02	2.02	2.03	S1	2.02	2.02	2.02	2.02	2.01	2.01	2.00	2.00	1.99	2.00	2.01	S	2.01	2.00	1.98	2.03	2.01	2.01	23	
24	1.98	1.98	1.96	1.96	1.99	1.99	1.99	1.98	1.98	1.99	1.98	1.97	1.96	C1	C1	C1	C1	1.94	1.94	1.94	S	1.95	2.00	2.00	1.94	2.00	1.97	20	
25	1.99	1.95	1.97	2.02	2.02	2.01	1.96	1.96	1.98	1.96	1.96	1.95	1.95	1.96	1.96	1.96	1.95	1.94	1.94	S	1.94	1.96	1.96	1.96	1.94	2.02	1.97	24	
26	1.95	1.95	1.97	1.95	1.97	1.98	2.08	2.00	1.95	1.97	1.98	1.97	1.96	1.95	1.96	1.97	1.95	1.94	S	1.95	1.95	1.96	1.99	2.03	1.94	2.08	1.97	24	
27	2.10	2.11	2.09	2.12	2.11	2.08	2.04	1.99	1.96	1.94	1.94	1.95	1.96	1.95	1.95	1.95	1.95	S	1.93	1.96	1.95	1.94	1.95	1.97	1.93	2.12	2.00	24	
28	1.99	1.99	2.04	2.00	1.93	1.91	1.92	1.93	1.95	1.95	1.96	1.95	1.94	1.94	1.96	1.94	S	1.94	1.95	1.94	1.94	1.95	1.95	1.95	1.91	2.04	1.95	24	
29	1.94	1.94	1.94	1.92	1.93	1.96	1.96	1.94	1.93	1.96	1.98	1.98	1.96	1.96	1.95	S	1.95	1.96	1.96	1.95	1.94	1.95	1.93	1.92	1.92	1.98	1.95	24	
30	1.93	1.99	2.00	1.98	1.96	1.94	1.96	1.95	1.96	1.95	1.95	1.96	1.94	1.93	S	1.95	1.95	1.98	1.98	1.93	1.93	1.93	1.96	2.09	1.93	2.09	1.96	24	
HOURLY MAX	2.10	2.11	2.10	2.14	2.22	2.18	2.35	2.48	2.35	2.09	2.06	2.04	2.03	2.02	2.03	2.02	2.03	2.10	2.21	2.36	2.01	2.02	2.20	2.09					
HOURLY AVG	1.98	1.99	1.99	1.99	1.99	2.00	2.01	2.01	1.99	1.99	1.98	1.98	1.98	1.97	1.98	1.97	1.97	1.97	1.98	2.00	1.96	1.97	1.98	1.98					

STATUS FLAG CODES

C	- MONTHLY CALIBRATION	Q	- QUALITY ASSURANCE
C1	- REPEAT CALIBRATION	R	- RECOVERY
Y	- MAINTENANCE	X	- MACHINE MALFUNCTION
S	- DAILY ZERO/SPAN CHECK	G	- OUT FOR REPAIR
S1	- REPEAT ZERO/SPAN CHECK	P	- POWER FAILURE


24 HR AVERAGES April 2018



MONTHLY SUMMARY

NUMBER OF NON-ZERO READINGS:	680			
MINIMUM 1-HR AVERAGE:	1.91 ppm	@ HOUR	5	ON DAY 28
MAXIMUM 1-HR AVERAGE:	2.48 ppm	@ HOUR	7	ON DAY 1
MAXIMUM 24-HR AVERAGE:	2.09 ppm			ON DAY 1
IZS CALIBRATION TIME:	30 hrs	OPERATIONAL TIME:	715	hrs
MONTHLY CALIBRATION TIME:	5 hrs	AMD OPERATION UPTIME:	99.3	%
STANDARD DEVIATION:	0.05	MONTHLY AVERAGE:	1.98	ppm

METHANE Hourly Averages (CH₄ ppm)



Wind: PRAMP_986
 Poll.: PRAMP_986-CH₄[ppm]
 Monthly: 18/04
 Type: PollutionRose
 Direction: Blowing From (Wind Frequency)
 Based On 1 Hr.


Calm: 6.77%

Calm Avg: 2.01 [ppm]


Direction	0-2	2-3	3-5	5-10	>10.0	Total
N	12.4	2.4	0.0	0.0	0.0	14.7
NE	5.9	0.0	0.0	0.0	0.0	5.9
E	9.7	2.2	0.0	0.0	0.0	11.9
SE	8.0	5.5	0.0	0.0	0.0	13.4
S	9.0	1.9	0.0	0.0	0.0	10.9
SW	18.0	2.1	0.0	0.0	0.0	20.0
W	6.0	1.2	0.0	0.0	0.0	7.2
NW	3.4	5.7	0.0	0.0	0.0	9.1
Summary	72.3	20.9	0.0	0.0	0.0	93.2

% Icon Classes (ppm)	72	0-2	21	2-3	0	3-5	0	5-10	0	>10.0
										

PRAMP_986 Poll.: PRAMP_986-CH4[ppm] 2018/04/01 00:00 - 2018/04/30 23:00 Calm: 6.77% Calm Poll Avg: 2.01[ppm]



CH4[ppm] Calibration: PRAMP_986 Monthly: 18/04 Type: Span



Span Meas Span Ref Span Low Span High

NON-METHANE HYDROCARBON




NON-METHANE HYDROCARBONS Hourly Averages (NMHC ppm)

HR START (MST)	0:00	1:00	2:00	3:00	4:00	5:00	6:00	7:00	8:00	9:00	10:00	11:00	12:00	13:00	14:00	15:00	16:00	17:00	18:00	19:00	20:00	21:00	22:00	23:00	DAILY	DAILY	24-HR	RDGS.	
HR END (MST)	0:59	1:59	2:59	3:59	4:59	5:59	6:59	7:59	8:59	9:59	10:59	11:59	12:59	13:59	14:59	15:59	16:59	17:59	18:59	19:59	20:59	21:59	22:59	23:59	MIN.	MAX.	AVG.		
DAY 1	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	S	0.00	0.00	0.00	0.00	0.00	0.00	0.00	24
2	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	S	0.00	0.00	0.00	0.00	0.00	0.00	0.00	24
3	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	S	0.00	0.00	0.00	0.00	0.00	0.00	0.00	24
4	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	S	0.00	0.00	0.00	0.00	0.00	0.00	0.00	24
5	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	S	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	24
6	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	S	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	24
7	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	S	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	24
8	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	S	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	24
9	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	S	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	24
10	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	C	C	C	C	C	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	24
11	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	S	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	24
12	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	S	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	24
13	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	S	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	24
14	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	S	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	24
15	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	S	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	24
16	0.00	0.00	0.00	0.00	0.00	S	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	24
17	0.00	0.00	0.00	0.00	S	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	24
18	0.00	0.00	0.00	S	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	24
19	0.00	0.00	S	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	24
20	0.00	S	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	24
21	S	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	24
22	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	24
23	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	S1	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	S	0.00	0.00	0.00	0.00	0.00	23
24	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	C1	C1	C1	C1	0.00	0.00	0.00	0.00	S	0.00	0.00	0.00	0.00	0.00	0.00	20
25	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	S	0.00	0.00	0.00	0.00	0.00	0.00	24
26	0.00	0.00	0.01	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	S	0.00	0.00	0.00	0.00	0.00	0.00	0.01	0.00	24
27	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	S	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	24
28	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	S	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	24
29	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	S	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	24
30	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	S	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	24
HOURLY MAX	0.00	0.00	0.01	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	
HOURLY AVG	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	

STATUS FLAG CODES

C	- MONTHLY CALIBRATION	Q	- QUALITY ASSURANCE
C1	- REPEAT CALIBRATION	R	- RECOVERY
Y	- MAINTENANCE	X	- MACHINE MALFUNCTION
S	- DAILY ZERO/SPAN CHECK	G	- OUT FOR REPAIR
S1	- REPEAT ZERO/SPAN CHECK	P	- POWER FAILURE


24 HR AVERAGES April 2018



MONTHLY SUMMARY

NUMBER OF NON-ZERO READINGS:	1			
MINIMUM 1-HR AVERAGE:	0.00	ppm @ HOUR	0	ON DAY 1
MAXIMUM 1-HR AVERAGE:	0.01	ppm @ HOUR	2	ON DAY 26
MAXIMUM 24-HR AVERAGE:	0.00	ppm		ON DAY 1
IZS CALIBRATION TIME:	30	hrs	OPERATIONAL TIME:	715 hrs
MONTHLY CALIBRATION TIME:	5	hrs	AMD OPERATION UPTIME:	99.3 %
STANDARD DEVIATION:	0.00		MONTHLY AVERAGE:	0.00 ppm

NON-METHANE HYDROCARBONS Hourly Averages (NMHC ppm)



Wind: PRAMP_986
 Poll.: PRAMP_986-NMHC[ppm]
 Monthly: 18/04
 Type: PollutionRose
 Direction: Blowing From (Wind Frequency)
 Based On 1 Hr.


Calm: 6.77%

Calm Avg: 0.00 [ppm]

Direction	0-0.1	0.1-0.3	0.3-1	1-2	>2.0	Total
N	14.7	0.0	0.0	0.0	0.0	14.7
NE	5.9	0.0	0.0	0.0	0.0	5.9
E	11.9	0.0	0.0	0.0	0.0	11.9
SE	13.4	0.0	0.0	0.0	0.0	13.4
S	10.9	0.0	0.0	0.0	0.0	10.9
SW	20.0	0.0	0.0	0.0	0.0	20.0
W	7.2	0.0	0.0	0.0	0.0	7.2
NW	9.1	0.0	0.0	0.0	0.0	9.1
Summary	93.2	0.0	0.0	0.0	0.0	93.2

% Icon Classes (ppm) 93 0-0.1 0 0.1-0.3 0 0.3-1 0 1-2 0 >2.0

PRAMP_986 Poll.: PRAMP_986-NMHC[ppm] 2018/04/01 00:00 - 2018/04/30 23:00 Calm: 6.77% Calm Poll Avg: 0.00[ppm]



NMHC[ppm] Calibration: PRAMP_986 Monthly: 18/04 Type: Span



■ Span Meas
 — Span Ref
 — Span Low
 — Span High

WIND SPEED



WIND SPEED Hourly Averages (WS kph)

HR START (MST)	0:00	1:00	2:00	3:00	4:00	5:00	6:00	7:00	8:00	9:00	10:00	11:00	12:00	13:00	14:00	15:00	16:00	17:00	18:00	19:00	20:00	21:00	22:00	23:00	DAILY	DAILY	24-HR	RDGS.
HR END (MST)	0:59	1:59	2:59	3:59	4:59	5:59	6:59	7:59	8:59	9:59	10:59	11:59	12:59	13:59	14:59	15:59	16:59	17:59	18:59	19:59	20:59	21:59	22:59	23:59	MIN.	MAX.	AVG.	
DAY																												
1	3.8	2.5	2.5	2.6	2.6	2.3	2.8	2.4	0.3	2.2	6.8	7.2	11.0	10.1	7.9	7.9	8.5	7.0	4.9	3.9	2.0	2.5	2.3	1.7	0.3	11.0	2.6	24
2	1.8	0.9	2.0	0.9	1.2	4.9	4.2	4.9	7.2	5.4	6.4	5.3	7.3	11.5	7.8	8.7	8.1	7.0	6.2	3.7	1.6	1.6	0.6	1.6	0.6	11.5	3.7	24
3	1.1	1.8	0.0	2.9	3.7	2.7	3.2	4.4	5.8	6.5	5.8	6.6	6.9	7.7	6.8	7.7	7.9	8.8	6.5	5.5	3.9	4.9	3.3	4.5	0.0	8.8	4.6	24
4	5.2	2.1	3.1	4.0	1.8	2.7	4.2	6.2	8.5	8.6	11.6	15.3	14.2	12.7	12.4	19.0	C	C	5.1	2.1	0.9	1.2	2.5	1.2	0.9	19.0	5.9	24
5	0.3	0.5	1.1	0.3	3.4	5.5	5.0	6.8	7.8	8.6	7.8	7.4	7.4	9.4	7.2	7.0	4.9	6.8	4.4	4.3	6.4	3.7	3.5	4.3	0.3	9.4	3.9	24
6	5.2	4.5	4.0	4.3	5.3	6.3	5.6	5.3	5.7	7.9	10.6	9.8	9.2	9.1	11.9	10.2	9.1	10.4	9.8	6.1	6.0	6.4	7.1	6.6	4.0	11.9	7.3	24
7	5.5	5.4	5.2	5.9	6.9	7.6	5.9	6.2	4.9	14.3	16.0	15.5	14.7	12.5	11.7	13.5	12.0	14.5	12.7	9.7	8.5	6.5	5.6	4.7	4.7	16.0	9.2	24
8	4.7	4.7	4.8	4.6	4.1	5.4	5.0	4.7	2.6	1.9	8.5	5.6	5.0	6.4	9.1	5.8	2.9	5.4	4.9	4.3	2.2	1.8	3.3	2.7	1.8	9.1	1.1	24
9	3.5	3.4	3.1	5.5	4.9	4.0	2.4	4.4	2.4	5.1	6.6	5.7	5.7	4.7	7.2	6.9	3.5	2.6	4.6	3.4	0.9	2.6	1.5	0.4	0.4	7.2	1.4	24
10	3.6	6.6	7.4	6.0	9.1	9.0	9.4	12.0	12.3	9.5	10.6	11.7	10.5	9.2	10.0	11.3	11.3	9.8	10.9	9.2	8.2	8.9	7.7	8.4	3.6	12.3	9.0	24
11	6.4	6.7	5.9	8.7	8.5	7.6	5.5	5.9	6.2	4.8	6.5	6.3	5.8	7.5	8.5	8.0	6.9	6.3	5.8	5.4	5.7	3.3	2.7	3.7	2.7	8.7	5.8	24
12	3.6	6.4	4.4	5.5	6.3	5.1	6.6	5.8	7.8	6.4	7.9	7.0	6.8	5.0	2.4	8.7	9.4	8.8	14.2	11.6	6.5	8.4	6.2	5.4	2.4	14.2	6.2	24
13	10.4	7.8	5.6	5.4	4.9	5.2	7.8	11.4	12.1	12.4	11.2	9.7	9.6	10.3	9.3	9.4	7.4	2.5	2.7	3.0	9.5	15.2	21.2	15.5	2.5	21.2	7.4	24
14	9.8	7.1	9.4	11.2	9.4	9.5	9.8	10.2	11.8	10.8	11.7	13.2	14.1	14.1	13.7	10.9	11.8	10.7	7.6	5.8	7.2	5.6	5.3	4.9	4.9	14.1	8.4	24
15	4.0	3.1	3.1	2.2	0.8	1.0	0.3	3.8	7.4	6.6	9.2	9.8	10.5	11.7	12.0	12.2	11.1	11.6	11.2	10.2	11.1	11.1	9.0	9.1	0.3	12.2	6.5	24
16	9.5	9.8	8.6	8.3	8.5	7.2	8.2	8.5	6.8	9.9	11.1	10.3	10.9	9.9	11.6	9.9	9.8	7.7	8.4	8.2	7.2	5.6	6.0	5.9	5.6	11.6	8.3	24
17	5.5	6.0	5.0	5.6	4.8	5.5	5.7	5.8	5.6	11.6	10.7	7.0	6.3	7.5	8.0	8.9	8.5	9.8	7.4	10.1	5.9	5.2	4.6	6.2	4.6	11.6	5.2	24
18	4.8	2.7	1.5	1.6	1.0	1.4	3.8	4.3	5.9	8.2	10.4	10.8	11.8	11.6	12.1	10.4	9.7	11.2	9.5	5.8	7.2	8.2	8.6	7.4	1.0	12.1	6.6	24
19	5.5	5.5	6.1	6.7	10.5	10.5	11.1	15.2	14.9	14.8	15.3	15.1	12.8	16.0	14.0	13.6	11.0	10.5	9.3	9.0	9.9	8.8	7.1	5.3	5.3	16.0	10.6	24
20	7.4	10.0	9.9	7.5	9.1	10.8	9.9	10.0	11.1	13.9	13.3	10.5	10.8	12.1	12.2	11.4	11.2	6.8	4.9	1.5	3.5	4.7	4.2	3.5	1.5	13.9	8.2	24
21	3.9	2.7	3.0	2.5	1.6	1.8	0.3	1.1	2.2	3.0	5.2	6.5	6.2	6.7	8.9	9.7	9.2	7.9	7.7	4.6	5.8	6.3	7.5	7.7	0.3	9.7	3.7	24
22	7.6	14.5	12.0	13.6	16.7	14.5	16.9	14.7	13.8	19.5	19.2	16.4	16.4	12.3	13.2	10.4	12.2	9.5	6.9	4.5	2.7	2.6	2.7	2.7	2.6	19.5	10.4	24
23	4.4	4.3	4.5	3.9	4.2	4.9	4.2	8.2	7.5	8.4	9.3	10.0	12.0	12.3	13.6	14.7	14.4	9.3	8.8	7.1	5.7	8.1	5.7	5.1	3.9	14.7	7.5	24
24	6.7	6.4	7.2	8.9	9.9	7.6	8.5	12.0	10.5	8.4	9.7	10.4	12.7	12.3	11.9	12.4	11.8	9.2	8.5	7.0	4.4	5.2	7.8	3.7	3.7	12.7	8.4	24
25	0.8	2.0	0.7	0.2	0.5	1.7	1.6	3.3	4.5	4.7	7.1	11.5	14.3	15.5	15.2	15.8	12.6	13.0	9.6	6.2	5.5	5.6	6.9	6.3	0.2	15.8	6.4	24
26	3.9	1.4	1.0	1.0	1.6	2.4	0.5	1.6	6.2	6.2	7.7	6.8	4.8	5.3	6.5	5.8	6.2	6.7	5.8	3.9	2.8	1.9	2.9	3.9	0.5	7.7	2.1	24
27	4.6	3.7	5.1	5.1	4.1	5.6	7.1	7.3	6.8	8.9	8.9	6.9	8.7	11.4	11.9	10.0	10.0	8.8	8.4	7.4	3.1	3.4	4.5	6.0	3.1	11.9	4.9	24
28	7.1	6.8	6.1	5.8	5.9	5.2	5.5	6.3	6.3	6.0	6.6	7.0	7.1	12.8	9.3	10.4	11.3	11.0	10.1	10.3	10.7	10.2	6.7	10.0	5.2	12.8	2.1	24
29	12.6	13.2	11.2	8.7	8.3	5.8	5.3	6.8	6.1	4.0	7.7	8.5	3.3	3.5	6.3	4.6	6.0	8.2	7.3	5.6	5.3	4.8	4.2	2.9	2.9	13.2	4.2	24
30	3.9	3.1	5.6	6.0	7.3	7.3	8.8	12.0	10.5	11.9	12.5	15.2	15.3	11.4	9.1	7.5	13.8	11.7	13.0	21.4	12.5	8.3	6.1	5.1	3.1	21.4	4.0	24
HOURLY MAX	12.6	14.5	12.0	13.6	16.7	14.5	16.9	15.2	14.9	19.5	19.2	16.4	16.4	16.0	15.2	19.0	14.4	14.5	14.2	21.4	12.5	15.2	21.2	15.5				
HOURLY AVG	1.6	1.2	1.4	1.2	1.2	1.3	0.9	1.2	1.2	2.0	2.7	3.1	3.3	3.8	3.8	3.4	3.3	2.9	2.2	1.2	0.9	1.5	1.6	1.5				

STATUS FLAG CODES


C	- MONTHLY CALIBRATION	Q	- QUALITY ASSURANCE
C1	- REPEAT CALIBRATION	R	- RECOVERY
Y	- MAINTENANCE	X	- MACHINE MALFUNCTION
S	- DAILY ZERO/SPAN CHECK	G	- OUT FOR REPAIR
S1	- REPEAT ZERO/SPAN CHECK	P	- POWER FAILURE

LAST CALIBRATION:	April 4, 2018
DECLINATION:	MAGNETIC DECLINATION 15 DEGREE EAST


MONTHLY SUMMARY

NUMBER OF NON-ZERO READINGS:	717
MINIMUM 1-HR AVERAGE:	0.0 kph @ HOUR 2 ON DAY 3
MAXIMUM 1-HR AVERAGE:	21.4 kph @ HOUR 19 ON DAY 30
MAXIMUM 24-HR AVERAGE:	10.6 kph ON DAY 19
MONTHLY CALIBRATION TIME:	2 hrs
OPERATIONAL TIME:	720 hrs
AMD OPERATION UPTIME:	100.0 %
STANDARD DEVIATION:	3.8
MONTHLY AVERAGE:	1.2 kph

24 HR AVERAGES April 2018



WIND SPEED Hourly Averages (WS kph)




Wind: PRAMP_986
 Monitor: WSP [kph]
 Monthly: 18/04
 Type: WindRose
 Direction: Blowing From (Wind Frequency)
 Based On 1 Hr.

Calm: 6.69%

Direction	1.8-6.0	6.0-12.0	12.0-20.0	20.0-29.0	29.0-39.0	>39.0	Total
N	5.4	8.9	0.4	0.0	0.0	0.0	14.8
NE	1.3	4.5	0.3	0.0	0.0	0.0	6.0
E	7.9	3.8	0.3	0.0	0.0	0.0	12.0
SE	6.1	5.2	1.4	0.1	0.0	0.0	12.8
S	3.8	5.9	1.4	0.0	0.0	0.0	11.0
SW	4.5	12.5	3.8	0.0	0.0	0.0	20.8
W	2.7	3.5	0.8	0.0	0.0	0.0	7.0
NW	2.5	3.9	2.5	0.1	0.0	0.0	9.1
Summary	34.1	48.0	10.9	0.3	0.0	0.0	93.3

% Icon Classes (kph) 34 1.8-6.0 48 6.0-12.0 11 12.0-20.0 0 20.0-29.0 0 29.0-39.0 0 >39.0

PRAMP_986 2018/04/01 00:00 - 2018/04/30 23:00 Calm: 6.69% Calm Wind Avg Speed: 1.07(kph)



WIND DIRECTION



PEACE RIVER AREA MONITORING PROGRAM COMMITTEE
Three Creeks 986b Station - April 2018

WIND DIRECTION Hourly Averages (WD)

HR START (MST)	0:00	1:00	2:00	3:00	4:00	5:00	6:00	7:00	8:00	9:00	10:00	11:00	12:00	13:00	14:00	15:00	16:00	17:00	18:00	19:00	20:00	21:00	22:00	23:00	24-HOUR AVG	24-HR	
HR END (MST)	0:59	1:59	2:59	3:59	4:59	5:59	6:59	7:59	8:59	9:59	10:59	11:59	12:59	13:59	14:59	15:59	16:59	17:59	18:59	19:59	20:59	21:59	22:59	23:59	QUADRANT	RDGS.	
DAY																											
1	SE	ESE	ESE	ESE	ESE	ESE	ESE	ESE	NNW	NW	NW	NW	NW	NNW	NNW	NW	NW	NNW	NNW	NNW	N	NNE	SE	NNW	NNW	24	
2	ESE	NE	SE	ESE	SSE	NNW	NNW	N	NNW	N	N	NNW	N	NNW	NNW	NNW	N	N	N	NNE	E	SE	NW	SE	N	24	
3	E	SE	S	N	NNE	NNE	N	N	N	NNE	N	NNW	NNW	NNW	NNW	NNW	NNW	NNW	N	NNW	NNW	N	NNE	NNE	N	24	
4	NNE	ENE	NNE	NNE	NW	NW	N	NNW	N	NNW	NNW	NNW	NNW	NNW	NNW	NW	C	C	WSW	W	WNW	N	N	NNW	NNW	24	
5	WNW	S	ESE	N	N	N	NNW	N	N	NNW	N	NNW	NNW	NNW	NNW	N	N	NNW	N	E	E	ESE	E	E	N	24	
6	E	E	E	E	E	E	E	E	E	ESE	ESE	ESE	ESE	ESE	ESE	ESE	ESE	ESE	ESE	ESE	ESE	ESE	ESE	ESE	ESE	24	
7	ESE	ESE	ESE	ESE	ESE	ESE	E	E	E	SE	SE	SE	SE	SE	SE	ESE	ESE	ESE	ESE	SE	ESE	ESE	ESE	ESE	ESE	24	
8	ESE	ESE	ESE	ESE	ESE	ESE	ESE	ESE	ESE	SW	W	W	WNW	WNW	W	W	S	W	WSW	WSW	SW	SE	SE	ESE	SSW	24	
9	ESE	SE	ESE	ESE	ESE	SE	ESE	SE	SSE	SSW	SW	SW	WSW	W	W	WNW	WSW	WNW	WNW	WNW	ESE	ESE	ESE	SSE	SW	24	
10	ENE	E	E	E	ENE	NE	NE	ENE	NE	ENE	NE	NE	ENE	E	E	ENE	ENE	ENE	ENE	ENE	ENE	ENE	ENE	E	ENE	24	
11	E	E	ENE	NE	NE	NE	NE	NE	NE	ENE	NNE	NE	NE	ENE	NE	NE	ENE	ENE	ENE	ENE	E	E	ESE	E	E	ENE	24
12	ENE	ENE	ENE	E	E	E	E	E	ESE	ESE	ESE	ESE	ESE	S	SSE	SE	SE	SE	SE	SSE	SSE	SE	SE	ESE	ESE	24	
13	SE	SE	SE	ESE	ESE	ESE	SE	SSE	S	S	S	S	SW	SSW	SW	SSW	S	W	NW	E	ESE	SE	SE	SE	SSE	24	
14	SE	SE	SE	S	SSW	S	SSW	SSW	SSW	SW	SW	SW	SW	SW	WSW	SW	SW	WSW	SW	SSW	SSW	SSW	SSW	S	SSW	24	
15	SW	SSW	S	SSW	SE	E	NNE	NNW	NNW	NNW	NNW	NNW	NNW	NNW	NNW	N	N	N	N	N	N	N	N	N	NNW	24	
16	N	N	N	N	N	N	N	N	NNE	NE	NE	NE	NE	NNE	NE	NNE	NNE	NNE	NNE	NNE	N	N	N	N	N	NNE	24
17	N	N	N	NNW	NNW	NNW	NNW	NW	WNW	WNW	WNW	W	W	W	WSW	SW	SW	WSW	WSW	W	WSW	SW	SW	SW	W	24	
18	SW	SSW	SSW	SE	SW	S	SSE	SSE	S	S	S	SSW	SSW	SW	SW	SW	SW	SW	SW	SSW	S	S	SSW	S	SSW	24	
19	SSW	SSW	SSW	SSW	S	S	SSW	SSW	SSW	SSW	SSW	SSW	SSW	SSW	SSW	SSW	SSW	SSW	SSW	SSW	SSW	SSW	SSW	SSW	SSW	24	
20	SSW	SW	SW	SSW	SSW	SSW	SSW	SSW	SSW	SSW	SSW	SSW	SSW	SSW	SSW	SSW	SSW	SSW	SSW	SSW	S	SE	ESE	ESE	SSW	24	
21	ESE	E	E	NNW	NNE	NW	S	ENE	ESE	SW	SW	SSW	SW	WSW	SW	SW	WSW	WSW	W	WSW	SW	WSW	WSW	SW	24		
22	W	WNW	WNW	WNW	WNW	WNW	WNW	WNW	WNW	NW	NW	NW	NW	NW	WNW	WNW	WSW	WSW	W	WSW	SW	SW	SSE	SSE	WNW	24	
23	SSW	S	S	SSE	ESE	SE	SSE	S	SSW	SSW	SSW	SSW	SSW	SSW	SSW	SSW	SSW	SSW	SSW	SSW	SSW	S	S	SSW	SSW	24	
24	SSW	SSW	SW	SW	SSW	SSW	SSW	SSW	SW	SW	SW	SW	SW	SW	SW	WSW	WSW	SW	SW	SW	SW	WSW	W	NNW	SW	24	
25	WNW	SSW	WSW	S	SSW	SSE	SSE	W	W	SW	SSW	SW	SSW	SSW	SSW	SSW	SSW	SSW	SSW	SSW	SSW	SSW	S	SSW	SW	24	
26	S	SW	E	SE	ESE	SSE	ESE	W	W	NW	WNW	W	W	WSW	WSW	WNW	SSW	SSW	SW	SSW	SE	SE	ESE	ESE	WSW	24	
27	ESE	E	ESE	ESE	ESE	ESE	ESE	SE	S	SSW	S	SW	SW	SW	SSW	SSW	SSW	SW	SW	SSW	S	SE	ESE	SE	S	24	
28	SE	ESE	ESE	SE	SE	SSE	SSE	SSW	SW	SSW	SSW	SSW	SSW	WSW	WNW	NW	NW	WNW	NNW	N	N	N	N	NNW	NW	24	
29	NNW	NNW	NNW	NNW	NNW	NNW	NNW	NNW	NNW	NNW	WNW	WNW	NW	W	SSW	SW	WSW	W	W	WSW	SSW	SSW	S	SSE	WNW	24	
30	SE	ESE	SE	SE	SE	SE	SSE	S	SSW	SSW	SSW	SSW	SSW	SW	SSW	SSW	WNW	WNW	WNW	NW	NNW	NNW	NNW	NW	SW	24	


STATUS FLAG CODES

C	- MONTHLY CALIBRATION	Q	- QUALITY ASSURANCE
C1	- REPEAT CALIBRATION	R	- RECOVERY
Y	- MAINTENANCE	X	- MACHINE MALFUNCTION
S	- DAILY ZERO/SPAN CHECK	G	- OUT FOR REPAIR
S1	- REPEAT ZERO/SPAN CHECK	P	- POWER FAILURE

LAST CALIBRATION:	April 4, 2018
DECLINATION :	MAGNETIC DECLINATION 15 DEGREE EAST

MONTHLY CALIBRATION TIME:	2 hrs	OPERATIONAL TIME:	720 hrs
STANDARD DEVIATION:	97	AMD OPERATION UPTIME:	100.0 %
		MONTHLY AVERAGE:	219 (SW)

WIND DIRECTION Hourly Averages (WD)



RELATIVE HUMIDITY




RELATIVE HUMIDITY Hourly Averages (RH %)

HR START (MST)	0:00	1:00	2:00	3:00	4:00	5:00	6:00	7:00	8:00	9:00	10:00	11:00	12:00	13:00	14:00	15:00	16:00	17:00	18:00	19:00	20:00	21:00	22:00	23:00	DAILY	DAILY	24-HR	RDGS.
HR END (MST)	0:59	1:59	2:59	3:59	4:59	5:59	6:59	7:59	8:59	9:59	10:59	11:59	12:59	13:59	14:59	15:59	16:59	17:59	18:59	19:59	20:59	21:59	22:59	23:59	MIN.	MAX.	AVG.	
DAY 1	62	63	64	65	65	65	66	64	52	44	43	40	37	35	32	29	28	28	29	36	43	52	57	61	28	66	48	24
2	64	65	65	67	67	70	75	66	58	52	48	44	39	44	45	42	43	44	43	45	53	62	67	69	39	75	56	24
3	70	70	71	73	75	75	77	74	53	45	44	44	43	44	42	41	39	37	36	46	54	52	49	55	36	77	55	24
4	54	54	64	64	61	66	68	67	58	54	52	50	46	42	39	38	37	37	39	48	57	61	66	64	37	68	54	24
5	63	63	65	67	68	74	76	74	66	61	57	52	49	48	45	42	38	39	39	42	45	52	58	59	38	76	56	24
6	53	54	58	64	65	65	65	61	49	43	39	34	31	30	30	28	26	28	30	37	48	51	54	56	26	65	46	24
7	53	57	60	62	61	64	64	60	54	54	52	48	47	44	42	39	36	33	36	42	48	56	57	60	33	64	51	24
8	62	65	67	69	70	73	73	70	63	54	59	55	53	52	49	47	40	43	47	54	61	70	74	74	40	74	60	24
9	76	77	76	79	81	81	79	72	60	50	44	41	40	39	41	40	36	35	40	48	67	73	74	74	35	81	59	24
10	71	52	47	47	41	41	38	38	43	45	43	40	38	37	35	32	33	34	33	34	35	37	39	57	32	71	41	24
11	87	91	88	89	88	89	90	88	85	79	79	77	70	66	69	71	73	75	77	78	77	81	84	82	66	91	81	24
12	82	72	59	61	62	65	67	68	70	69	62	57	55	52	46	42	43	42	50	60	69	76	79	83	42	83	62	24
13	82	81	82	83	81	87	90	87	85	81	74	62	57	49	44	34	28	28	32	49	58	61	62	63	28	90	64	24
14	67	73	75	73	80	91	93	82	67	59	54	47	44	41	40	39	36	36	38	48	57	59	59	66	36	93	59	24
15	62	67	74	77	82	87	88	80	76	71	71	66	65	66	65	63	62	71	71	73	76	77	79	81	62	88	73	24
16	84	86	88	89	90	90	89	88	85	80	73	71	72	73	72	75	84	93	97	97	98	98	99	99	71	99	86	24
17	99	98	98	98	98	98	97	96	95	93	89	86	81	82	85	86	90	93	93	92	96	97	98	96	81	99	93	24
18	89	91	93	93	93	93	93	92	80	72	68	64	56	48	44	40	37	35	37	43	54	57	53	56	35	93	66	24
19	63	64	67	67	66	68	63	55	51	49	45	42	38	35	34	36	33	36	40	48	49	51	54	59	33	68	51	24
20	59	60	62	69	69	69	67	64	59	54	53	49	45	45	44	44	42	40	41	52	59	63	62	65	40	69	56	24
21	72	72	74	72	79	80	81	76	61	52	47	48	50	48	52	54	53	51	52	57	56	58	66	73	47	81	62	24
22	86	94	90	82	81	82	81	79	75	67	59	50	43	40	37	32	43	46	45	46	52	59	66	73	32	94	63	24
23	68	70	72	69	80	81	77	61	52	50	49	47	45	41	40	38	37	39	40	39	47	54	55	51	37	81	54	24
24	52	50	45	40	41	44	45	45	44	46	43	38	36	34	33	33	37	39	42	43	51	53	53	62	33	62	44	24
25	60	75	81	87	90	90	81	52	41	40	39	38	39	43	42	40	37	34	35	37	45	56	59	57	34	90	54	24
26	60	68	80	86	89	91	85	64	48	41	37	35	32	32	31	29	31	32	31	37	56	72	74	76	29	91	55	24
27	71	74	79	78	74	71	70	62	44	39	36	34	36	36	35	33	30	27	27	37	50	64	64	61	27	79	51	24
28	62	63	61	66	64	64	64	53	44	43	44	43	35	31	33	33	33	37	39	44	49	63	72	75	31	75	51	24
29	76	78	85	88	91	94	93	91	86	77	70	65	55	49	44	40	40	40	40	43	56	70	73	78	40	94	68	24
30	85	88	91	89	83	85	81	68	59	51	50	47	40	33	33	33	30	31	31	46	62	74	80	87	30	91	61	24
HOURLY MAX	99	98	98	98	98	98	97	96	95	93	89	86	81	82	85	86	90	93	97	97	98	98	99	99				
HOURLY AVG	70	71	73	74	75	76	76	70	62	57	54	50	47	45	44	42	42	43	44	50	58	64	66	69				

STATUS FLAG CODES

C	- MONTHLY CALIBRATION	Q	- QUALITY ASSURANCE
C1	- REPEAT CALIBRATION	R	- RECOVERY
Y	- MAINTENANCE	X	- MACHINE MALFUNCTION
S	- DAILY ZERO/SPAN CHECK	G	- OUT FOR REPAIR
S1	- REPEAT ZERO/SPAN CHECK	P	- POWER FAILURE


24 HR AVERAGES April 2018



MONTHLY SUMMARY

MINIMUM 1-HR AVERAGE:	26	%	@ HOUR	16	ON DAY	6
MAXIMUM 1-HR AVERAGE:	99	%	@ HOUR	22	ON DAY	16
MAXIMUM 24-HR AVERAGE:	93	%			ON DAY	17
OPERATIONAL TIME:					720	hrs
AMD OPERATION UPTIME:					100.0	%
STANDARD DEVIATION:	18				MONTHLY AVERAGE:	59 %

RELATIVE HUMIDITY Hourly Averages (RH %)



BAROMETRIC PRESSURE




BAROMETRIC PRESSURE Hourly Averages (BP mbar)

HR START (MST)	0:00	1:00	2:00	3:00	4:00	5:00	6:00	7:00	8:00	9:00	10:00	11:00	12:00	13:00	14:00	15:00	16:00	17:00	18:00	19:00	20:00	21:00	22:00	23:00	DAILY	DAILY	24-HR	RDGS.	
HR END (MST)	0:59	1:59	2:59	3:59	4:59	5:59	6:59	7:59	8:59	9:59	10:59	11:59	12:59	13:59	14:59	15:59	16:59	17:59	18:59	19:59	20:59	21:59	22:59	23:59	MIN.	MAX.	AVG.		
DAY 1	950	950	949	949	949	949	948	948	946	946	945	945	944	944	943	943	942	942	942	942	942	942	943	943	943	942	950	945	24
2	944	944	944	944	945	945	945	945	946	946	946	946	946	946	946	946	946	946	946	946	946	947	947	947	947	944	947	946	24
3	948	948	948	948	948	948	949	949	949	949	949	949	948	948	948	947	947	946	946	946	946	947	947	947	947	946	949	948	24
4	947	948	948	949	949	950	951	951	951	951	951	951	951	951	951	951	951	951	951	952	952	953	954	954	947	954	951	24	
5	955	956	957	957	957	958	958	959	959	960	960	960	960	960	960	960	959	958	958	958	959	959	959	959	955	960	959	24	
6	959	958	958	958	958	957	957	957	956	955	954	953	953	952	951	950	950	949	948	948	947	947	947	946	946	959	952	24	
7	946	945	945	945	944	943	943	942	941	940	939	939	939	938	938	937	936	936	936	936	937	937	937	937	936	946	946	940	24
8	937	937	937	938	938	938	938	939	939	939	940	940	940	941	941	941	941	942	942	942	942	942	943	943	937	943	940	24	
9	943	944	944	944	944	944	944	944	944	944	944	944	944	943	943	943	943	942	942	942	942	942	942	942	942	944	943	24	
10	942	943	943	943	943	943	944	944	944	944	944	944	943	942	941	941	939	939	938	937	937	938	938	937	937	944	941	24	
11	937	936	936	936	936	936	937	937	937	937	937	938	938	938	937	937	937	937	937	937	937	937	937	937	936	938	937	24	
12	937	936	936	936	936	936	936	936	936	936	936	937	937	937	937	937	937	937	937	937	938	938	938	938	936	938	937	24	
13	938	938	938	938	939	939	939	939	939	939	939	939	939	939	939	938	938	938	937	936	935	935	934	934	934	939	938	24	
14	934	934	935	935	936	936	936	936	937	937	937	937	937	937	938	938	938	938	939	939	939	939	939	939	934	939	937	24	
15	939	939	939	940	940	941	941	941	942	942	942	943	943	943	944	944	944	944	945	945	945	946	946	946	946	939	946	943	24
16	946	946	946	946	946	946	945	945	944	944	943	943	942	941	940	939	939	938	938	937	937	936	935	935	935	946	942	24	
17	934	933	932	932	932	932	932	932	932	932	933	933	933	934	934	935	935	935	936	936	937	937	937	938	932	938	934	24	
18	938	939	940	941	941	942	942	943	943	944	944	944	944	945	945	945	945	945	945	945	945	945	945	945	945	938	945	943	24
19	945	945	944	945	945	945	944	944	944	944	944	944	943	943	942	942	941	941	940	940	939	939	939	939	939	945	943	24	
20	939	939	939	940	940	940	941	941	941	941	941	941	941	941	941	941	941	940	940	940	940	940	939	939	939	941	940	24	
21	939	938	937	937	936	935	935	935	935	935	936	936	936	936	936	936	936	936	936	937	937	938	938	938	935	939	936	24	
22	939	940	940	941	942	943	944	945	945	946	947	947	947	948	948	948	949	949	950	950	950	950	950	950	939	950	946	24	
23	951	951	951	952	952	952	952	952	953	953	952	952	952	952	951	951	950	949	949	948	948	948	948	948	948	953	951	24	
24	948	947	947	947	947	946	946	946	945	945	945	944	944	944	944	944	943	943	942	942	942	943	943	944	942	948	945	24	
25	944	944	945	945	946	946	946	946	947	946	946	946	946	945	944	944	943	943	943	943	943	943	942	942	942	947	944	24	
26	942	943	944	944	945	945	946	947	947	948	948	948	948	947	947	946	946	946	945	945	945	945	944	944	942	948	946	24	
27	944	943	943	942	942	941	941	940	940	940	939	938	938	937	937	936	935	935	934	934	933	933	932	932	932	944	938	24	
28	931	930	930	929	929	929	928	928	928	928	928	928	928	928	928	928	928	928	928	929	930	932	934	935	936	928	936	930	24
29	937	938	939	940	941	941	942	943	943	944	944	944	943	943	943	942	942	942	942	942	942	942	942	941	941	937	944	942	24
30	940	940	939	939	939	938	938	937	937	936	936	935	934	933	933	932	931	931	932	933	934	935	936	936	931	940	936	24	
HOURLY MAX	959	958	958	958	958	958	958	959	959	960	960	960	960	960	960	960	959	958	958	958	959	959	959	959					
HOURLY AVG	942	942	942	943	943	943	943	943	943	943	943	943	943	942	942	942	942	942	942	942	942	942	942	942					

STATUS FLAG CODES

C	- MONTHLY CALIBRATION	Q	- QUALITY ASSURANCE
C1	- REPEAT CALIBRATION	R	- RECOVERY
Y	- MAINTENANCE	X	- MACHINE MALFUNCTION
S	- DAILY ZERO/SPAN CHECK	G	- OUT FOR REPAIR
S1	- REPEAT ZERO/SPAN CHECK	P	- POWER FAILURE


24 HR AVERAGES April 2018



MONTHLY SUMMARY

MINIMUM 1-HR AVERAGE:	928	mbar	@ HOUR	6	ON DAY	28
MAXIMUM 1-HR AVERAGE:	960	mbar	@ HOUR	9	ON DAY	5
MAXIMUM 24-HR AVERAGE:	959	mbar			ON DAY	5
OPERATIONAL TIME:					720	hrs
AMD OPERATION UPTIME:					100.0	%
STANDARD DEVIATION:	6				MONTHLY AVERAGE:	942 mbar

BAROMETRIC PRESSURE Hourly Averages (BP mbar)



AMBIENT TEMPERATURE




AMBIENT TEMPERATURE Hourly Averages (AmbTPX °C)

HR START (MST)	0:00	1:00	2:00	3:00	4:00	5:00	6:00	7:00	8:00	9:00	10:00	11:00	12:00	13:00	14:00	15:00	16:00	17:00	18:00	19:00	20:00	21:00	22:00	23:00	DAILY MIN.	DAILY MAX.	24-HR AVG.	RDGS.
DAY 1	-22.1	-22.8	-23.9	-24.9	-26.1	-26.7	-26.7	-24.3	-17.1	-10.2	-7.5	-4.9	-4.2	-3.3	-2.1	-1.7	-1.6	-1.8	-2.7	-6.0	-10.6	-15.4	-18.0	-19.8	-26.7	-1.6	-13.5	24
2	-20.8	-21.5	-23.4	-23.9	-24.8	-23.6	-16.7	-14.1	-11.8	-9.9	-7.4	-5.6	-4.3	-4.9	-4.8	-4.1	-3.8	-4.4	-5.2	-6.4	-10.7	-15.8	-15.5	-14.2	-24.8	-3.8	-12.4	24
3	-13.6	-13.2	-12.7	-11.8	-12.6	-15.7	-17.6	-14.7	-11.4	-9.4	-8.4	-7.2	-5.9	-5.5	-4.9	-4.8	-4.9	-5.1	-5.9	-8.8	-11.6	-12.3	-12.8	-14.8	-17.6	-4.8	-10.2	24
4	-14.4	-17.6	-22.2	-18.9	-22.8	-25.6	-24.0	-18.0	-15.2	-13.4	-12.3	-11.1	-10.0	-8.3	-6.9	-6.3	-5.3	-5.1	-5.8	-9.5	-14.9	-19.0	-20.6	-21.3	-25.6	-5.1	-14.5	24
5	-23.0	-23.8	-23.4	-21.5	-20.8	-15.8	-17.1	-15.8	-14.8	-14.1	-13.0	-11.3	-10.2	-9.6	-8.2	-7.1	-6.6	-6.8	-7.3	-9.2	-11.3	-15.6	-18.9	-18.6	-23.8	-6.6	-14.3	24
6	-15.1	-15.7	-18.9	-21.7	-21.7	-21.3	-21.4	-18.7	-13.5	-9.6	-7.3	-5.7	-4.2	-2.8	-1.8	-1.1	-0.3	-0.8	-2.7	-6.5	-11.4	-13.5	-15.4	-15.2	-21.7	-0.3	-11.1	24
7	-15.0	-17.3	-17.6	-17.0	-16.1	-16.8	-17.1	-14.6	-11.4	-7.6	-5.2	-3.3	-2.1	-0.8	0.1	0.3	0.7	0.5	-1.2	-3.5	-6.7	-9.1	-9.4	-10.6	-17.6	0.7	-8.4	24
8	-11.4	-13.6	-15.1	-15.6	-17.0	-16.7	-16.7	-12.3	-8.0	-4.1	-5.2	-3.8	-2.6	-1.8	-0.8	0.3	2.8	1.1	0.2	-2.2	-5.3	-9.7	-11.0	-11.9	-17.0	2.8	-7.5	24
9	-12.3	-11.7	-12.3	-13.4	-13.2	-12.8	-11.9	-7.9	-3.0	0.5	2.2	3.4	3.6	4.7	4.5	5.3	6.3	6.5	5.3	3.7	0.6	-2.1	-2.5	-1.3	-13.4	6.5	-2.4	24
10	-0.5	1.2	0.5	0.0	-0.1	-0.4	-0.7	-0.3	0.1	1.3	2.1	2.6	3.3	4.1	4.8	4.4	4.4	4.0	3.2	2.2	1.0	0.1	0.0	-0.8	-0.8	4.8	1.5	24
11	-1.2	-1.5	-2.6	-3.4	-4.2	-4.6	-4.8	-4.9	-4.5	-3.4	-2.9	-2.3	0.2	0.9	0.4	0.2	-0.1	-0.5	-0.7	-0.8	-0.5	-0.9	-2.0	-1.4	-4.9	0.9	-1.9	24
12	-1.6	-2.1	-2.1	-2.1	-2.2	-2.3	-2.2	-1.7	-0.8	0.2	1.6	2.9	3.6	5.1	6.3	7.2	7.2	7.3	5.2	2.7	0.4	-1.7	-2.6	-3.9	-3.9	7.3	1.0	24
13	-2.1	-1.6	-1.4	-1.8	-2.1	-1.8	-0.9	0.8	1.7	2.6	4.4	6.7	6.9	8.1	8.8	9.9	11.1	11.0	9.2	5.8	4.3	5.0	4.9	4.5	-2.1	11.1	3.9	24
14	3.0	1.7	1.8	2.8	3.1	2.4	2.6	3.8	5.2	6.2	6.9	7.5	8.1	8.7	8.9	9.6	10.0	9.6	9.1	6.4	4.0	3.3	2.6	1.4	1.4	10.0	5.4	24
15	2.4	1.1	-0.3	-0.7	-1.7	-3.2	-3.9	0.5	2.8	4.6	5.0	6.1	6.3	5.8	5.1	4.9	4.4	2.5	2.0	1.0	-0.3	-0.8	-1.0	-1.3	-3.9	6.3	1.7	24
16	-1.9	-2.4	-2.7	-2.8	-2.9	-3.0	-3.1	-2.6	-1.6	-0.8	-0.2	0.4	0.9	1.3	1.7	1.8	1.3	0.7	-0.1	-0.7	-1.0	-1.2	-1.4	-1.4	-3.1	1.8	-0.9	24
17	-1.4	-1.5	-1.5	-1.8	-1.9	-2.0	-2.1	-1.9	-1.8	-1.7	-1.1	-0.5	0.1	0.3	0.9	0.4	-0.3	-0.9	-1.0	-1.2	-1.4	-1.6	-1.4	-1.1	-2.1	0.9	-1.1	24
18	-1.3	-3.5	-5.8	-7.0	-6.2	-5.6	-4.9	-1.8	1.0	3.4	4.3	4.9	6.2	7.3	7.7	8.8	9.2	9.0	8.4	6.6	4.8	3.9	4.5	3.4	-7.0	9.2	2.4	24
19	2.4	2.8	2.4	2.5	2.5	1.7	3.0	4.9	6.7	7.4	8.3	8.9	10.2	10.8	11.0	10.4	11.5	10.0	9.0	7.3	7.1	6.8	5.8	4.8	1.7	11.5	6.6	24
20	5.3	5.8	5.5	3.8	3.7	3.9	4.3	5.0	6.3	7.7	8.4	9.5	10.4	10.2	10.2	10.4	10.9	11.2	10.9	8.1	6.3	5.6	5.6	4.9	3.7	11.2	7.2	24
21	4.1	3.8	3.6	3.4	2.6	1.7	1.9	3.5	6.7	8.0	9.1	9.2	10.0	10.7	10.5	10.5	10.2	10.0	9.0	7.7	7.5	7.0	6.3	5.3	1.7	10.7	6.8	24
22	4.7	4.7	4.7	4.7	4.1	3.8	3.6	3.4	3.9	5.2	6.4	7.5	8.4	9.2	9.4	9.7	8.8	8.9	8.1	7.7	5.8	4.3	2.8	1.3	1.3	9.7	5.9	24
23	1.6	1.0	0.3	0.5	-1.7	-1.7	-0.3	3.5	5.6	6.6	7.8	9.0	9.3	10.4	11.1	11.8	11.5	10.6	10.5	9.8	7.2	5.9	5.3	5.8	-1.7	11.8	5.9	24
24	5.7	5.8	6.7	7.5	7.2	6.5	7.0	7.5	8.3	8.2	9.8	12.1	12.9	13.4	13.8	13.7	13.3	12.5	11.6	11.1	8.7	8.3	8.0	5.8	5.7	13.8	9.4	24
25	5.6	2.4	1.3	0.0	-0.5	-0.4	1.6	6.7	8.1	10.0	11.5	11.7	12.3	12.7	13.4	14.1	14.2	14.2	14.5	13.7	10.9	8.3	7.5	8.2	-0.5	14.5	8.4	24
26	7.3	5.3	2.5	1.2	0.4	0.5	2.4	8.0	10.1	12.0	13.2	14.2	15.3	16.2	16.9	17.6	18.6	18.3	17.8	16.3	11.1	7.5	5.4	4.6	0.4	18.6	10.1	24
27	5.0	4.3	4.2	4.4	4.9	5.5	6.5	8.9	14.0	16.7	18.3	20.2	20.3	20.7	21.4	22.3	23.1	22.8	22.5	19.7	14.9	11.0	11.1	10.8	4.2	23.1	13.9	24
28	9.5	9.6	9.2	7.4	7.2	7.7	9.7	11.9	13.6	14.8	15.5	16.4	18.9	19.4	18.9	18.9	17.6	16.7	15.7	13.6	11.1	9.0	7.7	7.2	7.2	19.4	12.8	24
29	5.7	5.2	4.4	4.5	3.6	3.3	3.4	3.8	4.5	5.5	6.2	7.2	9.1	8.8	10.9	11.4	11.0	10.5	10.3	9.4	6.8	7.7	5.3	4.0	3.3	11.4	6.8	24
30	1.9	1.0	1.2	1.5	2.5	2.2	4.1	7.8	8.5	10.8	12.4	13.1	14.3	15.1	15.6	16.1	15.5	14.5	14.2	11.6	8.9	7.1	6.6	5.3	1.0	16.1	8.8	24
HOURLY MAX	9.5	9.6	9.2	7.5	7.2	7.7	9.7	11.9	14.0	16.7	18.3	20.2	20.3	20.7	21.4	22.3	23.1	22.8	22.5	19.7	14.9	11.0	11.1	10.8				
HOURLY AVG	-3.1	-3.8	-4.6	-4.8	-5.2	-5.4	-4.7	-2.5	-0.3	1.6	2.8	3.9	4.9	5.6	6.1	6.5	6.7	6.2	5.5	3.7	1.2	-0.6	-1.4	-2.0				

STATUS FLAG CODES

C	- MONTHLY CALIBRATION	Q	- QUALITY ASSURANCE
C1	- REPEAT CALIBRATION	R	- RECOVERY
Y	- MAINTENANCE	X	- MACHINE MALFUNCTION
S	- DAILY ZERO/SPAN CHECK	G	- OUT FOR REPAIR
S1	- REPEAT ZERO/SPAN CHECK	P	- POWER FAILURE


24 HR AVERAGES April 2018



MONTHLY SUMMARY

MINIMUM 1-HR AVERAGE:	-26.7 °C	@ HOUR	5	ON DAY	1
MAXIMUM 1-HR AVERAGE:	23.1 °C	@ HOUR	16	ON DAY	27
MAXIMUM 24-HR AVERAGE:	13.9 °C			ON DAY	27
OPERATIONAL TIME:					720 hrs
AMD OPERATION UPTIME:					100.0 %
STANDARD DEVIATION:	9.7	MONTHLY AVERAGE:			0.7 °C

AMBIENT TEMPERATURE Hourly Averages (AmbTPX °C)



STATION TEMPERATURE


STATION TEMPERATURE Hourly Averages (StnTPX °C)

HR START (MST)	0:00	1:00	2:00	3:00	4:00	5:00	6:00	7:00	8:00	9:00	10:00	11:00	12:00	13:00	14:00	15:00	16:00	17:00	18:00	19:00	20:00	21:00	22:00	23:00	DAILY MIN.	DAILY MAX.	24-HR AVG.	RDGS.		
DAY 1	23.7	23.7	23.6	23.6	23.6	23.6	23.7	23.7	23.8	24.2	24.3	24.7	23.4	22.8	22.4	22.2	22.0	21.5	22.9	24.2	24.5	24.2	23.6	23.7	21.5	24.7	23.5	24		
2	23.8	23.9	23.7	23.5	23.5	23.5	23.6	23.6	23.6	24.1	24.5	23.8	22.4	22.0	23.0	23.4	22.2	22.5	24.0	23.7	22.6	23.3	23.6	23.8	22.0	24.5	23.4	24		
3	23.8	23.7	23.7	23.7	23.6	23.7	23.7	23.7	23.7	23.9	24.3	24.1	22.6	22.5	23.5	22.4	21.7	23.2	23.3	22.2	23.2	23.6	23.7	23.7	21.7	24.3	23.4	24		
4	23.7	23.9	23.8	23.9	23.7	23.8	23.7	23.7	23.8	23.8	23.5	24.0	23.9	24.1	24.3	23.1	22.5	22.1	22.0	23.0	23.2	23.4	23.6	23.6	22.0	24.3	23.5	24		
5	23.7	23.6	23.6	23.6	23.8	23.6	23.6	23.6	23.6	23.6	23.6	23.7	24.0	24.2	23.2	22.1	23.6	22.0	23.0	23.2	23.2	23.3	23.4	23.5	22.0	24.2	23.4	24		
6	23.5	23.5	23.7	23.5	23.6	23.5	23.5	23.5	23.7	23.8	23.5	23.8	24.2	23.7	22.1	23.3	22.2	21.7	23.5	22.3	22.6	23.1	23.4	23.6	21.7	24.2	23.3	24		
7	23.6	23.5	23.6	23.6	23.6	23.5	23.5	23.6	23.7	23.5	23.3	23.7	24.2	23.1	21.9	22.6	23.0	21.8	23.3	22.3	22.6	23.1	23.2	23.5	21.8	24.2	23.2	24		
8	23.6	23.7	23.7	23.5	23.6	23.6	23.7	23.6	23.7	24.0	24.4	24.4	22.9	22.2	22.2	22.1	22.1	21.9	21.8	21.7	23.2	23.7	23.6	23.4	21.7	24.4	23.1	24		
9	23.4	23.4	23.3	23.5	23.5	23.6	23.6	23.3	23.3	24.2	24.5	24.2	23.6	22.6	22.5	22.4	23.0	23.4	23.6	23.4	23.5	22.0	22.5	22.3	22.9	23.9	22.0	24.2	23.2	24
10	24.4	22.8	22.9	23.9	24.3	24.3	22.6	22.8	23.0	22.5	22.6	22.8	22.8	22.6	22.9	22.7	22.1	21.1	22.0	22.4	24.1	24.4	22.4	23.3	21.1	24.4	23.0	24		
11	24.2	24.6	24.8	23.1	23.0	23.4	23.7	24.0	24.1	24.4	25.0	22.9	22.3	21.8	23.2	23.4	22.2	22.9	24.2	24.8	22.5	23.1	24.1	24.6	21.8	25.0	23.6	24		
12	24.9	24.9	22.6	23.4	23.9	24.2	24.4	24.5	24.7	24.8	22.7	22.1	23.0	23.1	22.0	22.3	22.8	22.7	22.4	22.1	22.3	22.6	23.9	24.4	22.0	24.9	23.4	24		
13	24.5	24.6	24.8	24.6	22.6	23.4	24.1	24.5	24.9	22.7	22.7	22.5	22.1	22.3	23.0	23.0	22.9	22.9	22.7	22.8	22.8	22.4	22.5	22.5	22.1	24.9	23.2	24		
14	22.7	24.3	23.8	22.7	23.1	24.5	23.6	23.0	22.7	22.3	22.2	23.2	23.2	23.0	23.0	22.9	22.9	22.9	22.9	22.9	22.9	23.1	22.8	22.6	22.5	22.2	24.5	23.0	24	
15	23.5	24.6	24.0	22.7	23.7	24.3	24.6	25.0	23.0	22.3	22.4	23.3	23.2	23.0	23.0	23.0	23.0	23.0	23.0	22.5	22.5	22.9	24.2	24.9	22.3	25.0	23.4	24		
16	23.7	22.7	23.8	24.3	24.7	25.0	25.2	23.8	22.7	23.7	24.2	24.8	22.5	22.2	23.5	24.6	22.5	22.4	23.6	24.7	24.2	22.5	23.4	24.4	22.2	25.2	23.7	24		
17	25.0	23.2	22.7	24.0	24.8	24.4	22.6	23.4	24.4	25.0	23.6	22.9	22.6	24.2	23.6	22.8	22.4	23.9	24.9	22.9	22.8	24.0	24.7	25.2	22.4	25.2	23.8	24		
18	22.9	23.0	23.9	24.2	24.3	24.5	24.5	24.7	25.1	22.5	22.4	22.2	22.2	22.1	22.0	22.9	23.0	22.9	22.9	22.8	23.1	22.3	22.8	22.6	22.0	25.1	23.2	24		
19	24.1	24.9	23.3	22.7	23.7	24.4	24.8	23.8	22.4	22.2	23.2	23.3	23.1	23.1	23.0	22.9	22.9	22.9	23.0	23.0	23.3	23.3	23.4	23.1	22.2	24.9	23.3	24		
20	22.5	22.9	22.5	23.5	24.7	23.1	22.7	22.9	24.5	23.1	23.4	23.1	23.1	23.0	23.0	23.0	23.0	22.9	22.9	22.9	23.0	23.2	23.3	23.3	22.5	24.7	23.1	24		
21	22.7	22.5	22.9	22.4	23.7	24.9	23.1	22.8	22.5	22.7	22.9	22.5	22.3	22.4	23.2	23.2	23.2	23.2	23.4	23.3	23.2	23.4	23.6	23.6	22.3	24.9	23.1	24		
22	23.1	22.7	23.1	23.1	23.0	22.9	24.0	24.8	24.5	22.9	23.3	22.7	22.5	22.4	22.4	22.3	22.3	23.1	23.4	23.4	23.5	23.0	22.8	22.8	22.3	24.8	23.1	24		
23	24.0	24.9	24.9	22.9	23.5	24.3	24.7	25.2	23.0	22.5	23.4	23.6	23.5	23.4	23.4	23.3	23.3	23.3	23.3	23.3	23.4	23.5	23.4	23.1	22.5	25.2	23.6	24		
24	22.8	23.2	22.8	23.0	24.7	23.2	23.2	23.1	23.0	23.0	22.7	22.4	22.2	22.1	22.3	22.1	21.9	21.9	21.9	22.0	22.0	22.1	22.2	22.2	21.9	24.7	22.6	24		
25	22.4	22.5	22.3	23.8	24.6	24.0	22.4	23.3	23.6	22.4	22.2	22.1	22.1	22.1	22.1	22.0	22.0	22.0	22.0	21.9	21.9	22.0	22.1	22.2	21.9	24.6	22.5	24		
26	22.3	22.3	22.6	22.2	23.8	24.1	22.6	22.1	22.3	22.3	22.3	22.2	22.2	22.2	22.2	22.2	22.2	22.1	22.1	21.9	21.8	22.0	22.1	22.3	21.8	24.1	22.4	24		
27	22.7	22.3	22.8	24.2	23.2	24.2	22.7	24.1	22.4	22.3	22.3	22.4	22.3	22.4	22.3	22.4	22.4	22.3	22.3	22.1	21.9	21.8	22.0	22.1	21.8	24.2	22.5	24		
28	22.1	22.2	22.2	22.3	22.5	22.6	22.4	22.2	22.1	22.1	22.0	22.0	22.2	22.2	22.3	22.2	22.2	22.1	22.0	21.9	21.9	22.0	22.1	22.2	21.9	22.6	22.2	24		
29	22.3	22.5	22.8	22.1	22.9	24.4	23.0	22.5	22.2	23.9	22.8	22.3	22.2	22.1	22.1	22.0	22.0	22.0	22.0	22.0	22.1	22.2	22.2	22.2	22.0	24.4	22.5	24		
30	23.3	24.2	24.6	24.9	22.6	22.9	23.7	24.5	23.0	22.2	22.1	22.0	21.9	22.0	21.8	22.0	21.9	22.0	21.9	22.0	22.1	22.0	22.1	22.2	21.8	24.9	22.7	24		
HOURLY MAX	25.0	24.9	24.9	24.9	24.8	25.0	25.2	25.2	25.1	25.0	25.0	24.8	24.2	24.2	24.3	24.6	23.6	23.9	24.9	24.8	24.5	24.4	24.7	25.2						
HOURLY AVG	23.4	23.5	23.4	23.4	23.7	23.8	23.6	23.6	23.4	23.2	23.2	23.0	22.8	22.7	22.7	22.7	22.5	22.9	22.7	22.8	22.9	23.1	23.3							

STATUS FLAG CODES

C	- MONTHLY CALIBRATION	Q	- QUALITY ASSURANCE
C1	- REPEAT CALIBRATION	R	- RECOVERY
Y	- MAINTENANCE	X	- MACHINE MALFUNCTION
S	- DAILY ZERO/SPAN CHECK	G	- OUT FOR REPAIR
S1	- REPEAT ZERO/SPAN CHECK	P	- POWER FAILURE


24 HR AVERAGES April 2018



MONTHLY SUMMARY

MINIMUM 1-HR AVERAGE:	21.1 °C	@ HOUR	17	ON DAY	10
MAXIMUM 1-HR AVERAGE:	25.2 °C	@ HOUR	6	ON DAY	16
MAXIMUM 24-HR AVERAGE:	23.8 °C			ON DAY	17
OPERATIONAL TIME:					720 hrs
AMTD OPERATION UPTIME:					100.0 %
STANDARD DEVIATION:	0.8	MONTHLY AVERAGE:			23.1 °C

STATION TEMPERATURE Hourly Averages (StnTPX °C)



APPENDIX II
EQUIPMENT CALIBRATION RESULTS

SULPHUR DIOXIDE



Thermo 43C Sulphur Dioxide Analyzer Calibration

Date:	April 10, 2018	Barometer/B.P./units:	Brunton 05490 expires December 11, 2018	944	millibars
Company/Airshed:	PRAMP	Thermometer/Station Temp:	F.S. 160459244 expires May 18, 2018	23	°C
Location/Station Name:	986b	Weather Conditions:	Sunny		
Parameter:	Sulphur Dioxide	Calibration Purpose:	routine monthly		
Start Time 24 hr. (mst):	13:19	Performed By/Reviewer:	Chris Wesson	Rob Fisher	
End Time 24 hr. (mst):	17:07	Cal Gas Expiry Date:	October 24, 2020		
Calibration Method:	Gas Dilution	Converter Model & s/n (if applicable):	n/a		
Analyzer:					
Serial Number/Owner:	43C-62339-335 Maxxam	Range ppb:	500		
Last Calibration Date:	March 6, 2018	As Found C.F.:	0.994		
Previous C.F.:	0.999	New C.F.:	1.000		

Calibration Standards:	Standard Calibration Points for Ranges
Low Flow Meter ID/Expiry Date: Defender Low 152020 expires November 22, 2018	
High Flow Meter ID/Expiry Date: Defender High 148943 expires November 21, 2018	
Calibrator ID/Expiry Date: Sabio id# 17100415 expires May 16, 2018	
Cal Gas Cylinder I.D. #: LL108015	
Cal Gas Conc. (ppm): 47.9	

Point	ppb
High	380
Mid	180
Low	90


ALL POINTS ARE 15 MINUTES OF STABILITY AS OF SEPTEMBER 23, 2015

Point	Calibrator Flow Rates (cc/min)			Calculated Concentration (ppb):	Indicated Concentration (ppb):	Correction Factors (C.F.):
	Diluent	Cal Gas	Total			
as found zero	6041	0.00	6041	0.0	0.0	n/a
as found high	5999	47.20	6046	373.9	376.0	0.994
adjusted zero	6041	0.00	6041	0.0	0.0	n/a
adjusted high	5999	47.20	6046	373.9	374.0	1.000
mid	6026	22.46	6048	177.8	177.0	1.005
low	6028	11.24	6039	89.1	89.0	1.001
calibrator zero	6041	0.00	6041	0.0	0.0	n/a
Average C.F. =						1.002

Linear Regression/Calibration Results:

Correlation Coefficient =	1.000	LIMITS	> or = 0.995
Slope =	1.000		0.95-1.05
b (Intercept as % of full scale) =	0.05%		± 3% F.S.
% change in C.F. from last cal =	0.45%		± 10%


Thermo 43C Sulphur Dioxide Analyzer Calibration



As found:		As left:	
Bkg:	78.2	Bkg:	77.6
Coef:	0.875	Coef:	0.867
Pmt:	-654	Pmt:	-654
0	Lamp=845	0	Lamp=845
Battery:	3.3	Battery:	3.3
Internal:	29.8	Internal:	29.6
Chamber:	45.4	Chamber:	45.3
Pressure:	423.3	Pressure:	422.5
Flow:	0.745	Flow:	0.745
Intensity:	~38000	Intensity:	~38000
Averaging Time:	120	Averaging Time:	120
Expected Value:	231.0	Expected Value:	231.0

Comments: The analyzer cooling fan filter(s) were cleaned.

SO2[ppb] Station: PRAMP_986 Daily: 18/04/10 Type: AVG 1 Min. [1 Min.]



— SO2[ppb]

TOTAL REDUCED SULPHUR



Thermo 431-TLE Total Reduced Sulphur Analyzer Calibration

Date:	April 10, 2018	Barometer/B.P./units:	Brunton 05490 expires December 11, 2018	946	millibars
Company/Airshed:	PRAMP	Thermometer/Station Temp:	F.S. 160459244 expires May 18, 2018		°C
Location/Station Name:	986b	Weather Conditions:	Mainly sunny		
Parameter:	Total Reduced Sulphur	Calibration Purpose:	routine monthly		
Start Time 24 hr. (mst):	9:04	Performed By/Reviewer:	Chris Wesson	Rob Fisher	
End Time 24 hr. (mst):	14:23	Cal Gas Expiry Date:	November 7, 2020		
Calibration Method:	Gas Dilution	Converter Model & s/n (if applicable):	CD-Nova CON101 #516		
Analyzer:					
Serial Number/Owner:	1152940011 Maxxam	Range ppb:	100		
Last Calibration Date:	March 6, 2018	As Found C.F.:	0.977		
Previous C.F.:	1.000	New C.F.:	1.000		

Calibration Standards: Low Flow Meter ID/Expiry Date: Defender Low 152020 expires November 22, 2018 High Flow Meter ID/Expiry Date: Defender High 148943 expires November 21, 2018 Calibrator ID/Expiry Date: API id# 829 expires January 31, 2019 Cal Gas Cylinder I.D. #: LL119432 Cal Gas Conc. (ppm): 10.3	Standard Calibration Points for Ranges <table border="1"> <tr><td>Point</td><td>ppb</td></tr> <tr><td>High</td><td>78</td></tr> <tr><td>Mid</td><td>38</td></tr> <tr><td>Low</td><td>19</td></tr> </table>	Point	ppb	High	78	Mid	38	Low	19	SO2 Scrubber Check (10 minutes): Start/End Time 24 hr.: 10:09 / 10:19 SO2 Analyzer Range: 500 Target Gas Response (ppb): 380 As Found Zero: 0.0 Analyzer Response (ppb): 0.0 Zero Corrected Result (ppb): 0.0
Point	ppb									
High	78									
Mid	38									
Low	19									


ALL POINTS ARE 15 MINUTES OF STABILITY AS OF SEPTEMBER 23, 2015

Calibrator Flow Rates (cc/min)				Calculated	Indicated Concentration (ppb):	Correction Factors (C.F.):
Point	Diluent	Cal Gas	Total	Concentration (ppb):		
as found zero	7481	0.00	7481	0.0	-0.2	n/a
as found high	7409	55.89	7465	77.0	78.6	0.977
adjusted zero	7481	0.00	7481	0.0	0.0	n/a
adjusted high	7409	55.89	7465	77.0	77.0	1.000
mid	7449	27.80	7477	38.2	38.2	1.000
low	7455	13.93	7469	19.2	18.9	1.014
calibrator zero	7481	0.00	7481	0.0	0.3	n/a
Average C.F.:						1.005

Linear Regression/Calibration Results:

Correlation Coefficient =	1.000	LIMITS	> or = 0.995
Slope =	0.998		0.95-1.05
b (Intercept as % of full scale) =	0.12%		± 3% F.S.
% change in C.F. from last cal =	2.35%		± 10%


Thermo 431-TLE Total Reduced Sulphur Analyzer Calibration



As found:	Bkg:	2.37	As left:	Bkg:	2.11
Coef:	0.985	Coef:	0.964		
Pmt:	-690.8	Pmt:	-690.8		
Flash:	971	Flash:	971		
Internal:	33.9	Internal:	33.7		
Chamber:	45.3	Chamber:	45.0		
Perm Oven Gas:	44.99	Perm Oven Gas:	45.00		
Perm Oven Heater:	44.24	Perm Oven Heater:	44.24		
Pressure:	656.8	Pressure:	655.3		
Sample Flow:	0.480	Sample Flow:	0.479		
Lamp Intensity:	92	Lamp Intensity:	91		
Converter:	820	Converter:	820		
Converter Set:	820	Converter Set:	820		
Averaging Time:	120	Averaging Time:	120		
Expected Value:	40.8	Expected Value:	41.7		

Comments:
The analyzer cooling fan filter(s) were cleaned.

The "As Found High" was interrupted by the daily zero/span. The point was restarted.



— TRS[ppb]



Thermo 431-TLE Total Reduced Sulphur Analyzer Calibration

Date: <u>April 30, 2018</u>	Barometer/B.P./units: <u>Brunton 05535 expires December 15, 2018</u>	<u>27.68</u>	<u>inHg</u>
Company/Airshed: <u>PRAMP</u>	Thermometer/Station Temp: <u>F.S. 160348895 expires March 29, 2019</u>	<u>22.5</u>	<u>°C</u>
Location/Station Name: <u>986B</u>	Weather Conditions: <u>Mainly sunny</u>		
Parameter: <u>Total Reduced Sulphur</u>	Calibration Purpose: <u>as found</u>		
Start Time 24 hr. (mst): <u>11:01</u>	Performed By/Reviewer: <u>Limin Li</u>	<u>Rob Fisher</u>	
End Time 24 hr. (mst): <u>12:52</u>	Cal Gas Expiry Date: <u>August 23, 2020</u>		
Calibration Method: <u>Gas Dilution</u>	Converter Model & s/n (if applicable): <u>CD-Nova CDN101#516</u>		
Analyzer: Serial Number/Owner: <u>1152940011</u> <u>Maxxam</u>	Range ppb: <u>100</u>		
Last Calibration Date: <u>April 10, 2018</u>	As Found C.F.: <u>1.002</u>		
Previous C.F.: <u>1.000</u>	New C.F.: <u>n/a</u>		

Calibration Standards: Low Flow Meter ID/Expiry Date: <u>Defender Low 156151 expires October 2, 2018</u> High Flow Meter ID/Expiry Date: <u>Defender High 156312 expires December 13, 2018</u> Calibrator ID/Expiry Date: <u>Envionics id# 1991 expires March 15, 2019</u> Cal Gas Cylinder I.D. #: <u>LL119500</u> Cal Gas Conc. (ppm): <u>9.8</u>	Standard Calibration Points for Ranges <table border="1" style="margin: auto;"> <tr><td>Point</td><td>ppb</td></tr> <tr><td>High</td><td>78</td></tr> <tr><td>Mid</td><td>38</td></tr> <tr><td>Low</td><td>19</td></tr> </table>	Point	ppb	High	78	Mid	38	Low	19	SO2 Scrubber Check (10 minutes): Start/End Time 24 hr.: <u>11:37/11:48</u> SO2 Analyzer Range: <u>500</u> Target Concentration (ppb): <u>380</u> As Found Zero: <u>-0.1</u> Analyzer Response: (ppb): <u>-0.1</u> Zero Corrected Result (ppb): <u>0.0</u>
Point	ppb									
High	78									
Mid	38									
Low	19									


ALL POINTS ARE 15 MINUTES OF STABILITY AS OF SEPTEMBER 23, 2015

Calibrator Flow Rates (cc/min)				Calculated	Indicated Concentration (ppb):	Correction Factors (C.F.):
Point	Diluent	Cal Gas	Total	Concentration (ppb):		
as found zero	7506	0.00	7506	0.0	-0.1	n/a
as found high	7416	60.28	7476	79.0	78.8	1.002
Average C.F. =						n/a

Linear Regression/Calibration Results:

Correlation Coefficient = <u>n/a</u>	LIMITS
Slope = <u>n/a</u>	<u>n/a</u>
b (Intercept as % of full scale) = <u>n/a</u>	<u>n/a</u>
% change in C.F. from last cal = <u>-0.15%</u>	<u>n/a</u>

Thermo 431-TLE Total Reduced Sulphur Analyzer Calibration




As found: Bkg: <u>2.12</u> Coef: <u>0.964</u> Pmt: <u>-690.4 V</u> Flash: <u>965 V</u> Internal: <u>34.1 °C</u> Chamber: <u>45.1 °C</u> Perm Oven Gas: <u>45.0 °C</u> Perm Oven Heater: <u>44.24 °C</u> Pressure: <u>648.9 mmHg</u> Sample Flow: <u>0.475 L/Min</u> Lamp Intensity: <u>91 %</u> Converter: <u>820 °C</u> Converter Set: <u>820 °C</u> Averaging Time: <u>120 SEC</u> Expected Value: <u>40.8</u>	As left: Bkg: <u>2.12</u> Coef: <u>0.964</u> Pmt: <u>-690.4 V</u> Flash: <u>965 V</u> Internal: <u>34.1 °C</u> Chamber: <u>45.1 °C</u> Perm Oven Gas: <u>45.0 °C</u> Perm Oven Heater: <u>44.24 °C</u> Pressure: <u>648.9 mmHg</u> Sample Flow: <u>0.475 L/Min</u> Lamp Intensity: <u>91 %</u> Converter: <u>820 °C</u> Converter Set: <u>820 °C</u> Averaging Time: <u>120 SEC</u> Expected Value: <u>40.8</u>
--	---

Comments:

The manifold blower was found to be working normally.

An As Found was performed due to a drift in the daily span.

TRS[ppb] Station: PRAMP_986 Daily: 18/04/30 Type: AVG 1 Min. [1 Min.]



— TRS[ppb]

TOTAL HYDROCARBON



Thermo 55i Methane/Non-Methane Analyzer Calibration

Date:	April 10, 2018	Barometer/B.P./units:	Brunton 05490 expires December 11, 2018	946	millibars
Company/Airshed:	PRAMP	Thermometer/Station Temp:	F.S. 160459244 expires May 18, 2018	23	°C
Location/Station Name:	986b	Weather Conditions:	Mainly sunny		
Parameter:	CH4 / NMHC / THC	Calibration Purpose:	routine monthly		
Start/End Time 24 hr. (mst):	09:04 / 14:00	Performed By/Reviewer:	Chris Wesson	Rob Fisher	
Calibration Method:	Gas Dilution	Cal Gas Expiry Date:	October 18, 2025		

Analyzer:		Correction Factors:			
Serial Number/Owner:	1022143392 Maxxam	Previous C.F.:	As Found C.F.:	New C.F.:	
Measured Flow:	0.97 L/min	CH ₄ =	0.999	1.014	1.000
Last Calibration Date:	March 6, 2018	NMHC =	1.006	1.003	0.995
Range ppm:	20 CH4/20 NMHC/40 THC	THC =	1.003	1.008	0.998

Calibration Standards:		Standard Calibration Points for Analyzer Range of 20/20/40 ppm			
Low Flow Meter ID/Expiry Date:	Defender Low 152020 expires November 22, 2018	Point	CH4	NMHC	THC
High Flow Meter ID/Expiry Date:	Defender High 148943 expires November 21, 2018	High	13.00	13.00	26.00
Calibrator ID/Expiry Date:	Sabio id# 17100415 expires May 16, 2018	Mid	7.00	7.00	14.00
Cal Gas Cylinder I.D. #:	LL107207	Low	3.00	3.00	6.00
CH4 Cylinder Conc.:	600.0				
CH ₄ expressed as C ₂ H ₆ :	207.0				
	569.3				
	1169.3				
	=total CH4 equivalent				

ALL POINTS ARE 15 MINUTES OF STABILITY AS OF SEPTEMBER 23, 2015										Correction Factors:		
Calibrator Flow Rates (cc/min)				Calculated CH ₄ (ppm)	Calculated NMHC (ppm)	Calculated THC (ppm)	Indicated CH ₄ (ppm)	Indicated NMHC (ppm)	Indicated THC (ppm)	CH ₄	NMHC	THC
Point	Diluent	Cal Gas	Total Flow									
as found zero	2506	0.00	2506	0.00	0.00	0.00	0.00	0.00	0.00	n/a	n/a	n/a
as found high	2442	60.68	2503	14.55	13.80	28.35	14.35	13.76	28.12	1.014	1.003	1.008
adjusted zero	2506	0.00	2506	0.00	0.00	0.00	0.00	0.00	0.00	n/a	n/a	n/a
adjusted high	2442	60.68	2503	14.55	13.80	28.35	14.54	13.87	28.41	1.000	0.995	0.998
mid	2481	30.48	2511	7.28	6.91	14.19	7.33	6.93	14.27	0.994	0.997	0.995
low	2500	15.78	2516	3.76	3.57	7.33	3.80	3.57	7.36	0.990	1.000	0.996
calibrator zero	2506	0.00	2506	0.00	0.00	0.00	0.00	0.00	0.00	n/a	n/a	n/a
										Average C.F.=		
										0.995		
										0.997		
										0.996		

Linear Regression/Calibration Results:				LIMITS	
CH ₄	NMHC	THC	Correlation Coefficient =	1.000	> or = 0.995
			Slope =	0.999	0.95-1.05
			b (Intercept as % of full scale)=	0.13%	± 3% F.S.
			% change in C.F. from last cal=	-1.46%	± 10%

As Left Instrument Diagnostics:						
Interface Board Voltages:	Bias Supply:	-312.1	Calibration History cnt'd:	NM Peak Area:	66976	
Temperatures:	Detector Oven:	175.1	Crucial Settings:	Methane Start:	n/a	
	Filter:	175.1		Methane End:	n/a	
	Column Oven:	75.5		Backflush:	n/a	
Cylinder Pressures/reg.:	Internal:	39.0	Run History>1:	NMHV Start:	n/a	
	Carrier:	150/2600		50	NMHC End:	n/a
	Fuel:	2000		50	Date:	10Apr2018
Internal Pressures:	Span Gas:	350/1900	28	Time:	12:47	
	Zero Air Generator:	45		CH ₄ PK HT:	0	
	Carrier:	31.4		CH ₄ RT:	8.0	
FID Status:	Fuel:	40.5		CH ₄ Baseline:	1728	
	Air:	32.4		CH ₄ LOD:	28	
	Status:	LIT		CH ₄ SD:	9	
Flame and Power Stats:	Counts:	21556		CH ₄ CONC:	0.00	
	Flame:	322.0		NM PK HT:	0	
	Det Base:	175.0		NM Peak Area:	0	
Calibration History:	Last Power On:	15Mar2018@09:19		NM CONC:	0.00	
	Flameouts:	1		NM Base Start:	1802	
	Det Oven at Start:	169.3		NM Base End:	1794	
Expected Values:	Col Oven at Start:	75.1		NM LOD:	10	
	Time:	10Apr2018@10:52		NM Start IDX:	22	
	Type:	Span		NM End IDX:	68	
	Status:	Good		NM Max Slope:	3.5e-01	
	Check/Adjust:	Adjust		NM Min Slope:	-3.7e-01	
	CH ₄ Span Conc:	14.55		NM PT Count:	0	
	CH ₄ SP Ratio:	0.000767		Previous CH ₄ :	9.06	
	CH ₄ RT:	12.4		Previous NMHC:	10.21	
	CH ₄ PK IDX:	22		Previous THC:	19.27	
	CH ₄ PK HT:	18965		New CH ₄ :	10.36	
NM Span Conc:	13.80		New NMHC:	11.18		
NM SP Ratio:	0.000206		New THC:	21.54		

Comments:

The analyzer sample inlet filter was changed. A new nitrogen cylinder was installed.

A new span gas cylinder was installed. No zero adjustment was required/made. As found zero values were copied to adjusted zero values for linearity calculation purposes.




The analyzer cooling fan filter(s) were cleaned.


The manifold blower was found to be working normally.

Calibration was assessed for stability. All points met Maxxam internal stability requirement.

Date: April 10, 2018
Company/Airshed: PRAMP
Location/Station Name: 986b

Start/End Time 24 hr. (mst): 09:04 / 14:00
Calibration Purpose: routine monthly
Calibration Method: Gas Dilution





— CH4[ppm] — NMHC[ppm]



Thermo 55i Methane/Non-Methane Analyzer Calibration

Date:	April 24, 2018	Barometer/B.P./units:	Defender High 148943 expires November 21, 2018	710	mmHg
Company/Airshed:	PRAMP	Thermometer/Station Temp:	Defender High 148943 expires November 21, 2018	23	°C
Location/Station Name:	986b	Weather Conditions:	Cloudy/Overcast		
Parameter:	CH4 / NMHC / THC	Calibration Purpose:	repeat		
Start/End Time 24 hr. (mst):	13:06 / 16:52	Performed By/Reviewer:	Chris Wesson	Rob Fisher	
Calibration Method:	Gas Dilution	Cal Gas Expiry Date:	October 18, 2025		

Analyzer:		Correction Factors:			
Serial Number/Owner:	1022143392 Maxxam	Previous C.F.:	As Found C.F.:	New C.F.:	
Measured Flow:	0.97 L/min	CH ₄ =	1.000	0.993	0.999
Last Calibration Date:	April 10, 2018	NMHC =	0.995	0.959	1.003
Range ppm:	20 CH4/20 NMHC/40 THC	THC =	0.998	0.978	1.001

Calibration Standards:

Low Flow Meter ID/Expiry Date: Defender Low 152020 expires November 22, 2018
 High Flow Meter ID/Expiry Date: Defender High 148943 expires November 21, 2018
 Calibrator ID/Expiry Date: Sabio id# 17100415 expires May 16, 2018
 Cal Gas Cylinder I.D. #: LL107207
 CH4 Cylinder Conc.= 600.0 | 207.0 =C₂H₆ Cylinder Conc.
 CH₄ expressed as C₂H₆= 569.3 | 1169.3 =total CH4 equivalent

Point	CH4	NMHC	THC
High	13.00	13.00	26.00
Mid	7.00	7.00	14.00
Low	3.00	3.00	6.00

ALL POINTS ARE 15 MINUTES OF STABILITY AS OF SEPTEMBER 23, 2015

Calibrator Flow Rates (cc/min)				Calculated CH ₄ (ppm)	Calculated NMHC (ppm)	Calculated THC (ppm)	Indicated CH ₄ (ppm)	Indicated NMHC (ppm)	Indicated THC (ppm)	Correction Factors:		
Point	Diluent	Cal Gas	Total Flow							CH ₄	NMHC	THC
as found zero	2503	0.00	2503	0.00	0.00	0.00	0.00	0.00	0.00	n/a	n/a	n/a
as found high	2442	60.77	2503	14.57	13.82	28.39	14.67	14.41	29.03	0.993	0.959	0.978
adjusted high	2442	60.77	2503	14.57	13.82	28.39	14.58	13.78	28.37	0.999	1.003	1.001
mid	2480	30.53	2511	7.30	6.92	14.22	7.35	6.90	14.26	0.993	1.003	0.997
low	2497	15.78	2513	3.77	3.57	7.34	3.81	3.56	7.37	0.989	1.004	0.996
calibrator zero	2503	0.00	2503	0.00	0.00	0.00	0.00	0.00	0.00	n/a	n/a	n/a
										Average C.F.=		
										0.993 1.003 0.998		

Linear Regression/Calibration Results:

	CH ₄	NMHC	THC	LIMITS
Correlation Coefficient =	1.000	1.000	1.000	> or = 0.995
Slope =	1.000	0.997	0.999	0.95-1.05
b (Intercept as % of full scale)=	0.13%	-0.01%	0.06%	± 3% F.S.
% change in C.F. from last cal=	0.70%	3.60%	2.01%	± 10%

As Left Instrument Diagnostics:

Interface Board Voltages:	Bias Supply:	-312.0	Calibration History cnt'd:	NM Peak Area:	70049
Temperatures:	Detector Oven:	175.0	Crucial Settings:	Methane Start:	n/a
	Filter:	175.0		Methane End:	n/a
	Column Oven:	75.2		Backflush:	n/a
	Internal:	38.3		NMHV Start:	n/a
Cylinder Pressures/reg.:	Carrier:	2100 50	Run History>1:	NMHC End:	n/a
	Fuel:	1100 50	Date:	24Apr2018	
	Span Gas:	1750 28	Time:	15:58	
	Zero Air Generator:	45	CH ₄ PK HT:	0	
Internal Pressures:	Carrier:	31.4	CH ₄ RT:	8.0	
	Fuel:	40.5	CH ₄ Baseline:	1723	
	Air:	32.4	CH ₄ LOD:	15	
	FID Status:	Status:	LIT	CH ₄ SD:	5
Flame and Power Stats:	Counts:	21204	CH ₄ CONC:	0.00	
	Flame:	322.2	NM PK HT:	0	
	Det Base:	175.0	NM Peak Area:	0	
	Last Power On:	15Mar2018@09:19	NM CONC:	0.00	
	Flameouts:	1	NM Base Start:	1755	
	Det Oven at Start:	169.3	NM Base End:	1752	
	Col Oven at Start:	75.1	NM LOD:	11	
	Calibration History:	Time:	24Apr2018@14:20	NM Start IDX:	42
Type:	Span		NM End IDX:	60	
Status:	Good		NM Max Slope:	3.3e01	
Check/Adjust:	Adjust		NM Min Slope:	-4.1e-01	
CH ₄ Span Conc:	14.57		NM PT Count:	0	
CH ₄ SP Ratio:	0.000763	Expected Values:	Previous CH ₄ :	10.36	
CH ₄ RT:	12.4		Previous NMHC:	11.18	
CH ₄ PK IDX:	22		Previous THC:	21.54	
CH ₄ PK HT:	19090		New CH ₄ :	10.26	
NM Span Conc:	13.82		New NMHC:	11.11	
NM SP Ratio:	0.000197		New THC:	21.37	

Comments:




No zero adjustment was required/made. As found zero values were copied to adjusted zero values for linearity calculation purposes.

The manifold blower was found to be working normally.

Repeat calibration due to NMHC span drift.

Date: April 24, 2018
Company/Airshed: PRAMP
Location/Station Name: 986b

Start/End Time 24 hr. (mst): 13:06 / 16:52
Calibration Purpose: repeat
Calibration Method: Gas Dilution





— CH4[ppm] — NMHC[ppm]

WIND SYSTEM



Meteorological Sensor Audit/Calibration

Location Information

Company: PRAMP
 Audit Location: 986b
 Audit Date: April 5, 2017

Performed By: Limin Li
 Reviewed By: Trina Whitsitt
 Start /EndTime (mst): 13:30/15:30

Wind Sensor Information

Sensor ID Data:		Sensor Outputs:	
Sensor Make:	RM Young	Velocity Voltage Output Range:	0-1 V
Sensor Model:	05305VK	Velocity Unit Output Range:	0-200 km/h
Serial #:	129612	Direction Voltage Output Range:	0-1 V
Previous Cal/Audit Date:	February 15, 2017	Direction Unit Output Range:	0-360 degrees

Wind Calibrator Information

Calibrator Make/ Model: RM Young 18802 Serial #: CA 0309
 Maxxam Unit ID #: 13-3357 Certification Date: October 6, 2016

Wind Speed Audit Data ****+/- 2% of the average correction factor is the limit****

RPM	Wind Speed Generated kph	Clockwise Wind Speed kph	Counter Clockwise Wind Speed kph	Correction Factor
0	0	0.1	0.1	-
1000	18.4	18.4	18.4	0.999
2000	36.9	36.8	36.8	1.003
3000	55.3	55.2	55.2	1.002
4000	73.7	73.6	73.6	1.002
5000	92.2	92.0	92.0	1.002
6000	110.6	110.5	110.5	1.001
7000	129.0	128.9	128.9	1.001
8000	147.4	147.4	147.4	1.000
9000	165.9	165.9	165.9	1.000
10000	184.3	184.5	184.5	0.999
The audit meets AMD requirements.			Average Correction Factor=	1.001

Wind Direction Audit Data ****+/- 5° of the absolute average degrees difference for all points is the limit****

Generated Wind Direction 0-360 (Up)	Generated Wind Direction 360-0 (Down)	Indicated Wind Direction 0-360 (Up)	Indicated Wind Direction 360-0 (Down)	Degrees Difference 0-360 (Up)	Degrees Difference 360-0 (Down)	Average Absolute Degrees Difference
0	355	0	353	0.4	2.5	1.5
30	330	29	329	1.4	1.3	1.3
60	300	60	300	0.4	0.3	0.3
90	270	89	271	0.8	-1.2	1.0
120	240	120	242	0.5	-1.6	1.1
150	210	150	212	0.1	-1.8	1.0
180	180	181	181	-0.9	-1.1	1.0
210	150	212	151	-1.5	-0.8	1.2
240	120	241	120	-1.3	-0.2	0.8
270	90	271	90	-0.5	-0.1	0.3
300	60	300	60	0.2	0.5	0.3
330	30	329	29	1.4	0.7	1.1
355	0	353	0	2.5	0.3	1.4
The audit meets AMD requirements.			Average Absolute Degrees Difference=		0.9	

Comments: Adjust wind speed gain before calibration.



Meteorological Sensor Audit/Calibration

Location Information

Company:	PRAMP	Performed By:	Limin Li
Audit Location:	986B	Reviewed By:	Tom Bourque
Audit Date:	April 4, 2018	Start/End Time (mst):	16:22/17:42
Calibration Purpose:	routine annual	Weather Conditions:	Mainly sunny

Wind Sensor Information

Sensor ID Data:		Sensor Outputs:	
Sensor Make:	RM Young	Velocity Voltage Output Range:	0-1V
Sensor Model:	5305VK	Velocity Unit Output Range:	0-200 KPH
Serial #:	129612	Direction Voltage Output Range:	0-1 V
Previous Cal/Audit Date:	April 5, 2017	Direction Unit Output Range:	0-360 DEG

Wind Calibrator Information

Calibrator I.D. and Expiry Date: RM Young 18802 id# CA03309 expires September 25, 2018

Wind Speed Audit Data ****+/- 2% of the average correction factor is the limit****

RPM	Wind Speed Generated kph	Clockwise Wind Speed kph	Counter Clockwise Wind Speed kph	Correction Factor
0	0	0.1	0.1	-
1000	18.4	18.5	18.5	0.996
2000	36.9	36.9	36.9	1.001
3000	55.3	55.3	55.3	1.000
4000	73.7	73.7	73.7	1.000
5000	92.2	92.2	92.2	1.000
6000	110.6	110.7	110.7	0.999
7000	129.0	129.2	129.2	0.998
8000	147.4	147.7	147.7	0.998
9000	165.9	166.2	166.2	0.998
10000	184.3	184.8	184.8	0.997
The audit meets AMD requirements.			Average Correction Factor=	0.999

Wind Direction Audit Data ****+/- 3° of the absolute average degrees difference for all points is the limit****

Generated Wind Direction 0-360 (Up)	Generated Wind Direction 360-0 (Down)	Indicated Wind Direction 0-360 (Up)	Indicated Wind Direction 360-0 (Down)	Degrees Difference 0-360 (Up)	Degrees Difference 360-0 (Down)	Average Absolute Degrees Difference
0	355	0	354	0.3	1.4	0.8
30	330	30	330	0.2	0.1	0.1
60	300	61	301	-0.7	-0.9	0.8
90	270	91	271	-1.1	-1.2	1.1
120	240	122	243	-1.7	-2.6	2.2
150	210	153	213	-2.5	-2.9	2.7
180	180	183	183	-2.6	-2.6	2.6
210	150	213	152	-2.6	-1.7	2.1
240	120	243	122	-2.8	-2.0	2.4
270	90	272	91	-2.3	-1.1	1.7
300	60	301	62	-0.8	-1.7	1.2
330	30	331	30	-0.5	-0.3	0.4
355	0	354	0	1.3	0.3	0.8
The audit meets AMD requirements.				Average Absolute Degrees Difference=		1.5

Comments:

METEOROLOGICAL SYSTEMS CHECK

Meteorological System Checklist

Date:	April 10, 2018		
Technician:	Chris Wesson		
Reviewer:	Rob Fisher		
Station:	PRAMP 986b		
Unit:	Make:	Model:	Serial #:
Temperature Sensor:	RM Young	43172VC	61012322
Barometric Pressure Sensor:	Met One	090D	3845
Relative Humidity Sensor:	RM Young	43172VC	61012322
AMBIENT TEMPERATURE SENSOR CHECK			
Previous check date:	March 6, 2018		
Parameter:	Temperature @ 2 metres (1 C tolerance)		
Reference Thermometer ID:	F.S. 160459244 expires May 18, 2018		
Reference Temperature (°C):	3.1		
Station - Ambient Temperature (°C):	3.9		
Temperature Difference (°C):	-0.8		
BAROMETRIC PRESSURE SENSOR CHECK			
Previous check date:	March 6, 2018		
Reference Barometer ID:	Brunton 05490 expires December 11, 2018		
Reference Pressure - Units/Reading:	mbar	942	
Station Pressure - Units/Reading:	mbar	940	
Pressure Tolerance +/- 15% of error:	801 - 1083	0.21%	
RELATIVE HUMIDITY (HYGROMETER) SENSOR CHECK			
Previous check date:	March 6, 2018		
Reference Hygrometer ID:	F.S. id# 160459244 expires May 18, 2018		
Reference Hygrometer % RH- Reading:	27.00		
Station Hygrometer % RH- Reading:	31.00		
RH Tolerance +/- 15% of error:	22.95 - 31.05	-4.0%	

CALIBRATORS

Company <u>Maxxam</u>		Operator: <u>Micheal Espiritu</u>	
Calibrator:		Flow Measurement Device:	
Make/Model	<u>Sabio 2010</u>	Make/Model	<u>Mesa Defender 530</u>
Serial Number	<u>17100415</u>	Serial Number	<u>L-152019 H-148944</u>
Last Verification Date	<u>May 2016</u>	Temperature (°C)	<u>25.0 C</u>
NO Cylinder S/N	<u>EY0000597</u>	Barometric Pressure	<u>697 mmhg</u>
NO [PPM]	<u>49.0</u>	NOx [PPM]	<u>49.0</u>
Expiry Date	<u>December 2019</u>		

Dilution Flow (sccm)		
Pt. #1 <u>5000</u>	Pt. #2 <u>5000</u>	Pt. #3 <u>5000</u>
Gas Flow (sccm)		
Pt. #1 <u>80</u>	Pt. #2 <u>40</u>	Pt. #3 <u>20</u>

Calibrator Flow (sccm)		Calculated Conc.(ppm)		Indicated Conc.(ppm)			% Difference vs Audit Gas	
Dilution	Gas	NO	NOx	NO	NO ₂	NOx	NO	NOx
4996	0.0	0.000	0.000	0.000	0.000	0.000	Limit ± 10%	
5029	80.3	0.784	0.783	0.808	-0.013	0.794	3%	1%
5054	38.8	0.376	0.376	0.392	-0.006	0.386	4%	3%
5051	19.5	0.189	0.189	0.196	-0.003	0.193	4%	2%
Absolute Average Percent Difference							4%	2%

LINEAR REGRESSION ANALYSIS			<i>y=mx+b (where x=calculated concentration, y=indicated concentration)</i>		
NO		LIMITS		NOx	
Correlation=	1.0000	≥	0.990	Correlation=	1.0000
m (Slope)=	1.0311		0.90-1.10	m (Slope)=	1.0140
b (Intercept % of FS)=	0.1350	±	3% F.S.	b (Intercept % of FS)=	0.1531

Flow	O ₂ Conc (LC)	NO Decrease	NO	NO ₂	NOX	% Diff. Vs Audit gas	
5029	0.000	0.000	0.803	-0.013	0.790	NO ₂	% Diff. Limit
5029	1.508	0.568	0.235	0.552	0.787	-1%	± 10%
5029	0.882	0.312	0.491	0.298	0.789	0%	± 10%
5029	0.390	0.108	0.695	0.095	0.789	0%	± 10%
Absolute Average Percent Difference						0%	± 10%

LINEAR REGRESSION ANALYSIS			<i>y=mx+b (where x=calculated concentration, y=indicated concentration)</i>		
NO₂		LIMITS			
Correlation=	1.0000	≥	0.995		
m (Slope)=	0.9945		0.90-1.10		
b (Intercept % of FS)=	-1.2646	±	3% F.S.		

AENV Standards		NO_x Analyzer	
Audit Calibrator		Make/Model	<u>Teco 42i</u>
Make/Model	<u>Teco 146i</u>	Serial/AMU Number	<u>AMU 1868</u>
Serial/AMU Number	<u>AMU 1809</u>	Last Calibration Date	<u>May 16, 2017</u>
SRM Gas Cylinder No.	<u>CAL018101</u>	Full Scale (ppm)	<u>1.0</u>
Cylinder Conc. (ppm)	<u>48.79</u>	Cylinder Gas Expiry Date	<u>March 2019</u>

COMMENTS: Contains 50.4 ppm SO₂.

Auditor: Al Clark
Operator Signature:

Date: May 16, 2017
Location: McIntyre Center Edmonton

Company <u>Maxxam</u>		Operator: <u>Christopher</u>	
Calibrator:		Flow Measurement Device:	
Make/Model	<u>API 700</u>	Make/Model	<u>Mesa 530+</u>
Serial Number	<u>829</u>	Serial Number	<u>H-156312 L-156151</u>
Last Verification Date	<u>January 2017</u>	Temperature (°C)	<u>N/A</u>
NO Cylinder S/N	<u>EY0000715</u>	Barometric Pressure	<u>N/A</u>
NO [PPM]	<u>50.7</u>	NOx [PPM]	<u>50.8</u>
Expiry Date	<u>May 2020</u>		

Dilution Flow (sccm)			
Pt. #1	<u>5000</u>	Pt. #2	<u>5000</u>
		Pt. #3	<u>5000</u>
Gas Flow (sccm)			
Pt. #1	<u>80</u>	Pt. #2	<u>40</u>
		Pt. #3	<u>20</u>

Calibrator Flow (sccm)		Calculated Conc.(ppm)		Indicated Conc.(ppm)			% Difference vs Audit Gas	
Dilution	Gas	NO	NOx	NO	NO ₂	NOx	NO	NOx
5000	0.0	0.000	0.000	0.000	0.000	0.000	Limit ± 10%	
4950	79.8	0.787	0.788	0.783	-0.001	0.781	-1%	-1%
4958	37.2	0.380	0.381	0.378	-0.001	0.377	-1%	-1%
4960	18.5	0.189	0.189	0.188	-0.001	0.187	-1%	-1%
Absolute Average Percent Difference							1%	1%

LINEAR REGRESSION ANALYSIS *y=mx+b (where x=calculated concentration, y=indicated concentration)*

NO		LIMITS		NOx
Correlation=	1.0000	≥ 0.990		Correlation= 1.0000
m (Slope)=	0.9949	0.90-1.10		m (Slope)= 0.9912
b (Intercept % of FS)=	-0.0030	± 3% F.S.		b (Intercept % of FS)= -0.0257


Flow	O ₃ Conc	NO Decrease	NO	NO ₂	NOX	% Diff. Vs Audit gas	
4950	0.000	0.000	0.776	0.000	0.776	NO ₂	% Diff. Limit
4950	0.500	0.476	0.300	0.474	0.775	0%	± 10%
4950	0.290	0.277	0.499	0.278	0.778	0%	± 10%
4950	0.095	0.090	0.686	0.093	0.779	3%	± 10%
Absolute Average Percent Difference						1%	± 10%

LINEAR REGRESSION ANALYSIS *y=mx+b (where x=calculated concentration, y=indicated concentration)*

NO₂		LIMITS
Correlation=	1.0000	≥ 0.995
m (Slope)=	0.9938	0.90-1.10
b (Intercept % of FS)=	0.1802	± 3% F.S.

AENV Standards	NO_x Analyzer
Audit Calibrator	
Make/Model	<u>Teco 146i</u>
Serial/AMU Number	<u>AMU 1809</u>
SRM Gas Cylinder No.	<u>APEX1170572</u>
Cylinder Conc. (ppm)	<u>49.99</u>
	Make/Model <u>Teco 42i</u>
	Serial/AMU Number <u>AMU 1868</u>
	Last Calibration Date <u>January 31, 2018</u>
	Full Scale (ppm) <u>1.0</u>
	Cylinder Gas Expiry Date <u>November 2020</u>

COMMENTS:

Auditor: Al Clark
Operator Signature: 

Date: January 31, 2018
Location: McIntyre Center Edmonton

Company: Maxxam		Operator: Chris W	
Calibrator:		Flow Measurement Device:	
Make/Model	Envionics 2000	Make/Model	Mesa Defender 530
Serial Number	1991	Serial Number	L-153351 H-152571
Last Verification Date	March 2017	Temperature (°C)	25.0 C
NO Cylinder S/N	LL108015	Barometric Pressure	695 mmHg
NO [PPM]	52.2	NOx [PPM]	52.3
Expiry Date	Oct 2020		

Dilution Flow (sccm)			
Pt. #1	5000	Pt. #2	5000
Pt. #3	5000		
Gas Flow (sccm)			
Pt. #1	80	Pt. #2	40
Pt. #3	20		

Calibrator Flow (sccm)		Calculated Conc.(ppm)		Indicated Conc.(ppm)			% Difference vs Audit Gas	
Dilution	Gas	NO	NOx	NO	NO ₂	NOx	NO	NOx
5000	0.0	0.000	0.000	0.000	0.000	0.000	Limit ± 10%	
4988	75.1	0.786	0.787	0.785	-0.002	0.783	0%	-1%
4988	36.5	0.382	0.383	0.382	0.001	0.383	0%	0%
4988	18.3	0.192	0.192	0.190	0.000	0.190	-1%	-1%
Absolute Average Percent Difference							0%	1%

LINEAR REGRESSION ANALYSIS *y=mx+b (where x=calculated concentration, y=indicated concentration)*

NO	LIMITS	NOx
Correlation= 1.0000	≥ 0.990	Correlation= 1.0000
m (Slope)= 0.9996	0.90-1.10	m (Slope)= 0.9956
b (Intercept % of FS)= -0.0599	± 3% F.S.	b (Intercept % of FS)= -0.0005

Flow	O ₃ Conc	NO Decrease	NO	NO ₂	NOX	% Diff. Vs Audit gas	
4988	0.000	0.000	0.788	-0.001	0.787	NO ₂	% Diff. Limit
4988	0.350	0.519	0.269	0.512	0.780	-1%	± 10%
4988	0.160	0.231	0.557	0.229	0.786	0%	± 10%
4988	0.070	0.099	0.689	0.097	0.787	-1%	± 10%
Absolute Average Percent Difference						1%	± 10%

LINEAR REGRESSION ANALYSIS *y=mx+b (where x=calculated concentration, y=indicated concentration)*

NO₂	LIMITS	
Correlation= 1.0000	≥ 0.995	
m (Slope)= 0.9885	0.90-1.10	
b (Intercept % of FS)= -0.0567	± 3% F.S.	

AENV Standards	NO_x Analyzer
Audit Calibrator	Make/Model <u>Teco 42i</u>
Make/Model <u>Teco 146i</u>	Serial/AMU Number <u>AMU 1868</u>
Serial/AMU Number <u>AMU 1809</u>	Last Calibration Date <u>March 14, 2018</u>
SRM Gas Cylinder No. <u>APEX1170572</u>	Full Scale (ppm) <u>1.0</u>
Cylinder Conc. (ppm) <u>49.99</u>	Cylinder Gas Expiry Date <u>November 2020</u>

COMMENTS: Cylinder contains 47.9 ppm SO₂.

Auditor: Al Clark
Operator Signature:

Date: March 15, 2018
Location: McIntyre Center Edmonton

CALIBRATION GASES



Calibration Gas Audit

Single Component Cylinder Gas

File No. 2017-486CGA

Company: Maxxam **Operator's Name:** Mike

Cylinder #: LL108015 Concentration PPM: 47.9 Tolerance(%) 2 Certified By: Praxair

Expiry Date: October 2020

Reference Calibrator and Gas:

Make/Model: R&R MFC 201

Serial Number: AMU 1690

Last Verification Date: December 13, 2017

Gas Type: SO2 Conc. 98.07

Cylinder Number: CAL016625

Expiry Date: January 2019

Flow Measurement Device:

Make/Model: Mesa Definer 220

Serial Number: H-133034 / L-132702

Temp. °C: 23.4 C

B.P. 707 mmHg

Reference Analyzer:

Make/Model: Teco 43C Serial/AMU Number: 1623

Instrument Settings: Zero: 10.0 Span: 1.006 Range: 1.0

Last Calibration: Date: Dec12/17 C.F. 1.000 Done By: Al Clark

Calibrator Flows (sccm)		Indicated Concentration (PPM)	Gas Flow/ Dilution Flow	Concentration Factor	Cylinder Concentration
Dilution	Gas				
5000	0.0	0.000	0.000	0.000	0.000
4989	79.5	0.760	0.01594	62.755	47.7
4995	39.6	0.374	0.00793	126.136	47.2
4992	19.6	0.183	0.00393	254.694	46.6
Average Cylinder Concentration:					47.2

Previous Stated Concentration PPM: 47.9

Percent variance from Stated: 2

Meets Manufacturer Tolerance. Use manufacturers stated concentration **COMMENTS:** _____

< =5% Outside Manufacturer Tolerance. Use manufacturers concentration _____

> 5% Outside Manufacturer Tolerance. **DO NOT USE** this cylinder _____

Auditor: Al Clark

Operator Signature:

Date: December 13, 2017

Location: McIntyre Center Edmonton



Calibration Gas Audit

Single Component Cylinder Gas

File No. 2017-137CGA

Company: Maxxam **Operator's Name:** Raja Abid Ashraf
Cylinder #: LL119432 **Concentration PPM:** 10.3 **Tolerance(%)** 2 **Certified By:** Praxair
Expiry Date: May 16, 2020

Reference Calibrator and Gas:
Make/Model: R&R MFC 201
Serial Number: AMU 1690
Last Verification Date: July 27, 2017
Gas Type: H2S **Conc.** 20.43
Cylinder Number: CAL015272
Expiry Date: Janaury 2019

Flow Measurement Device:
Make/Model: Mesa Definer 220
Serial Number: H-133034 L-132702
Temp. °C: 22.0 C
B.P. 700 mmhg

Reference Analyzer:
Make/Model: Teco 450i **Serial/AMU Number:** 1980
Instrument Settings: **Zero:** 21.9 **Span:** 1.069 **Range:** 0.1
Last Calibration: **Date:** July 27, 2017 **C.F.** 1.000 **Done By:** Al Clark

Calibrator Flows (scem)		Indicated Concentration (PPM)	Gas Flow/ Dilution Flow	Concentration Factor	Cylinder Concentration
Dilution	Gas				
5000	0.0	0.0000	0.0000	0.0000	0.0
5117	38.9	0.0595	0.00760	131.542	7.8
5103	18.4		0.00361	277.337	0.0
5097	9.4		0.00184	542.234	0.0
Average Cylinder Concentration:					2.6

Previous Stated Concentration PPM: 10.3
 Percent variance from Stated: 75

Meets Manufacturer Tolerance. Use manufacturers stated concentration **COMMENTS:** _____
 <=5% Outside Manufacturer Tolerance. Use manufacturers concentration Do not use. _____
 > 5% Outside Manufacturer Tolerance. **DO NOT USE** this cylinder _____

Auditor: Al Clark **Date:** July 27, 2017
Operator Signature: *Al Clark* **Location:** McIntyre Center Edmonton



Calibration Gas Audit

Single Component Cylinder Gas

File No. 2017-213CGA

Company: Maxxam **Operator's Name:** C. Wesson
 Cylinder #: LL119500 Concentration PPM: 9.8 Tolerance(%) 2 Certified By: Praxair
 Expiry Date: August 2020

Reference Calibrator and Gas:	Flow Measurement Device:
Make/Model: <u>R&R MFC 201</u>	Make/Model: <u>Mesa Definer 220</u>
Serial Number: <u>AMU 1690</u>	Serial Number: <u>H-133034 L-132702</u>
Last Verification Date: <u>September 22, 2017</u>	Temp. °C: <u>23.5 C</u>
Gas Type: <u>H2S</u> Conc. <u>20.43</u>	B.P. <u>705 mmhg</u>
Cylinder Number: <u>CAL015272</u>	
Expiry Date: <u>January 2019</u>	

Reference Analyzer:
 Make/Model: Teco 450i Serial/AMU Number: 1980
 Instrument Settings: Zero: 22.4 Span: 1.091 Range: 0.1
 Last Calibration: Date: Sep 22/17 C.F. 1.000 Done By: Al Clark

Calibrator Flows (sccm)		Indicated Concentration (PPM)	Gas Flow/ Dilution Flow	Concentration Factor	Cylinder Concentration
Dilution	Gas				
5000	0.0	0.0000	0.0000	0.0000	0.0000
5114	39.5	0.0734	0.00772	129.468	9.5
5096	18.5	0.0345	0.00363	275.459	9.5
5089	9.5	0.0178	0.00187	535.684	9.5
Average Cylinder Concentration:					9.5

Previous Stated Concentration PPM: 9.8
 Percent variance from Stated: 3

Meets Manufacturer Tolerance. Use manufacturers stated concentration **COMMENTS:** _____
 <=5% Outside Manufacturer Tolerance. Use manufacturers concentration
 > 5% Outside Manufacturer Tolerance. **DO NOT USE** this cylinder

Auditor: Al Clark Date: September 22, 2017
 Operator Signature: *Al Clark* Location: McIntyre Center Edmonton



Calibration Gas Audit

CH4 / C3H8 Cylinder Gas

File No. 2017-484CGA

Company: Maxxam **Operators name:** Mike
Cylinder #: LL107207 **Conc CH4 (PPM)** 600/207 **Tolerance (%)** 2 **Certified By:** Praxair
Expiry Date: October 2025

Reference Calibrator and Gas:				Flow Measurement Device:	
Make/Model	<u>R&R MFC 201</u>			Make/Model	<u>Mesa Definer 220</u>
Serial Number	<u>AMU 1690</u>			Serial Number	<u>H-133034 / L-132702</u>
Last Verification Date	<u>December 13, 2017</u>			Temp. °C	<u>23.1 C</u>
Gas Type	<u>CH4</u>	Conc.	<u>990.4</u>	B.P.	<u>707 mmHg</u>
Cylinder Number	<u>5604875</u>	Expiry Date	<u>July 2021</u>		
Gas Type	<u>C3H8</u>	Conc.	<u>246.5</u>		
Cylinder Number	<u>XF003845B</u>	Expiry Date	<u>July 2022</u>		

Reference Analyzer:
Make/Model Teco 55i **Serial/AMU Number:** 2108
Instrument Settings **Zero:** N/A **Span:** N/A **Range:** 20.0
Last Calibration: **Date:** Dec 12/17 **C.F.** 1.000 **Done By:** Al Clark

Calibrator Flows (sccm)		Indicated Conc. (ppm)		Gas Flow/ Dilution Flow	Concentration Factor	Cylinder Concentration	
Dilution	Gas	CH4	C3H8			CH4	C3H8
3500	0.0	0.00	0.00				
3618	80.4	13.28	12.77	0.02	45.00	598	209
3547	39.8	6.71	6.47	0.01	89.12	598	210
3560	19.8	3.35	3.26	0.01	179.80	602	213
Average Cylinder Concentration:						599	211

	CH4	C3H8
Previous Stated Concentration PPM:	<u>600</u>	<u>207</u>
Percent variance from Stated:	<u>0</u>	<u>2</u>

Cylinder gas tolerances based on CH4 only
 Meets Manufacturer Tolerance. Use manufacturers stated concentration **COMMENTS:**
 < =5% Outside Manufacturer Tolerance. Use manufacturers concentration
 > 5% Outside Manufacturer Tolerance. **DO NOT USE** this cylinder

Auditor: Al Clark **Date:** December 13, 2017
Operator Signature: **Location:** McIntyre Center Edmonton

***APPENDIX III
MAXIMUM INSTANTANEOUS DATA***



PEACE RIVER AREA MONITORING PROGRAM COMMITTEE
Three Creeks 986b Station - April 2018

SULPHUR DIOXIDE Instantaneous Maximum (SO₂ ppb)

HR START (MST)	0:00	1:00	2:00	3:00	4:00	5:00	6:00	7:00	8:00	9:00	10:00	11:00	12:00	13:00	14:00	15:00	16:00	17:00	18:00	19:00	20:00	21:00	22:00	23:00	DAILY MIN.	DAILY MAX.	24-HR AVG.	RDGS.	
HR END (MST)	0:59	1:59	2:59	3:59	4:59	5:59	6:59	7:59	8:59	9:59	10:59	11:59	12:59	13:59	14:59	15:59	16:59	17:59	18:59	19:59	20:59	21:59	22:59	23:59					
DAY 1	3	3	2	2	2	2	2	2	4	4	2	2	2	2	3	2	2	2	2	2	2	S	2	2	3	2	4	2	24
2	2	1	2	2	2	2	2	2	2	2	2	1	2	2	2	2	2	2	2	2	S	2	2	2	2	1	2	2	24
3	1	2	2	2	2	2	2	2	2	2	2	2	2	2	2	2	2	2	3	S	2	3	2	2	2	1	3	2	24
4	2	2	2	2	2	2	2	2	2	2	2	2	2	2	2	2	2	2	S	2	2	2	2	2	1	1	2	2	24
5	2	2	2	2	2	2	2	2	2	2	2	2	2	2	2	2	S	2	2	2	2	2	2	2	2	2	2	2	24
6	2	2	2	2	2	2	2	2	2	2	2	2	2	2	2	S	2	2	2	2	2	2	2	3	2	2	3	2	24
7	3	2	2	3	2	3	2	2	2	2	2	3	3	3	S	2	2	2	2	2	2	3	2	2	2	2	3	2	24
8	2	2	2	2	2	2	2	2	3	3	2	3	3	S	2	3	3	3	3	2	2	2	2	2	2	2	3	2	24
9	2	2	2	2	2	2	2	3	3	3	2	2	S	2	3	3	2	2	2	2	3	2	2	2	2	2	3	2	24
10	3	3	3	2	2	2	2	3	3	3	2	S	2	C	C	C	C	C	C	3	3	2	2	2	3	2	3	3	24
11	3	3	3	3	3	2	2	3	3	3	S	2	2	3	3	3	3	3	2	2	3	3	2	3	2	3	3	3	24
12	3	2	3	2	2	3	3	3	3	S	2	2	2	2	3	2	3	2	3	3	3	3	3	3	3	2	3	3	24
13	2	2	3	2	3	3	3	2	S	3	3	3	3	3	3	3	3	3	3	2	3	3	3	4	4	2	4	3	24
14	3	3	3	3	3	3	3	S	4	3	3	2	3	3	3	3	2	2	3	2	2	3	3	3	3	2	4	3	24
15	2	3	3	3	3	3	S	3	2	3	3	3	2	3	3	3	3	3	3	4	3	3	3	3	3	2	4	3	24
16	2	3	3	3	3	S	3	3	3	3	3	3	3	3	3	3	3	3	3	3	3	3	3	3	3	2	3	3	24
17	3	3	3	3	S	3	3	3	3	3	3	3	3	3	3	3	3	3	3	3	3	3	3	3	3	3	3	3	24
18	3	3	3	S	3	3	3	3	3	3	3	3	3	3	3	3	3	3	3	3	3	3	3	3	3	3	3	3	24
19	4	4	S	3	3	3	3	3	3	4	4	3	3	3	3	3	3	4	3	3	3	3	3	3	4	3	4	3	24
20	4	S	4	4	4	4	3	3	3	4	3	4	3	3	3	3	3	3	3	3	3	3	3	4	4	3	4	3	24
21	S	3	4	3	3	3	3	3	4	4	4	4	4	3	3	3	3	3	3	3	3	3	3	S	3	4	3	24	
22	3	3	4	4	4	3	3	4	3	3	3	3	3	3	3	3	4	3	3	3	4	3	S	4	3	4	3	24	
23	4	4	3	3	3	3	3	3	4	4	4	3	3	4	4	4	3	3	3	3	3	S	3	3	3	4	3	24	
24	3	3	4	4	4	3	4	4	4	4	4	4	4	3	4	3	3	4	3	3	S	3	4	3	3	4	4	24	
25	4	4	4	3	3	3	4	4	4	4	3	3	4	4	3	4	4	4	4	S	4	4	3	4	3	4	4	24	
26	4	4	3	4	4	4	4	3	4	3	4	4	3	3	4	3	4	3	S	4	4	3	4	3	3	4	4	24	
27	4	4	4	4	4	4	4	3	4	4	4	4	4	3	4	4	S	S	4	4	4	4	3	4	3	4	4	24	
28	4	4	4	4	4	4	4	4	4	4	4	4	4	4	4	4	S	4	4	3	4	4	4	4	3	4	4	24	
29	4	4	4	4	4	4	4	4	4	4	4	4	3	4	4	S	4	4	4	4	4	3	4	4	4	3	4	4	24
30	4	4	4	4	4	4	4	4	4	4	9	4	4	4	S	4	4	4	4	4	4	4	4	4	4	4	9	4	24
HOURLY MAX	4	4	4	4	4	4	4	4	4	4	9	4	4	4	4	4	4	4	4	4	4	4	4	4	4	4	4	4	24
HOURLY AVG	3	3	3	3	3	3	3	3	3	3	3	3	3	3	3	3	3	3	3	3	3	3	3	3	3	3	3	3	24


STATUS FLAG CODES

C	- MONTHLY CALIBRATION	Q	- QUALITY ASSURANCE
C1	- REPEAT CALIBRATION	R	- RECOVERY
Y	- MAINTENANCE	X	- MACHINE MALFUNCTION
S	- DAILY ZERO/SPAN CHECK	G	- OUT FOR REPAIR
S1	- REPEAT ZERO/SPAN CHECK	P	- POWER FAILURE

MONTHLY SUMMARY

NUMBER OF NON-ZERO READINGS:	684
MAXIMUM INSTANTANEOUS VALUE:	9 ppb @ HOUR 10 ON DAY 30
IZS CALIBRATION TIME:	31 hrs
MONTHLY CALIBRATION TIME:	5 hrs
STANDARD DEVIATION:	1
OPERATIONAL TIME:	720 hrs

SULPHUR DIOXIDE Instantaneous Maximum (SO₂ ppb)





PEACE RIVER AREA MONITORING PROGRAM COMMITTEE
Three Creeks 986b Station - April 2018

TOTAL REDUCED SULPHUR Instantaneous Maximum (TRS ppb)

HR START (MST)	0:00	1:00	2:00	3:00	4:00	5:00	6:00	7:00	8:00	9:00	10:00	11:00	12:00	13:00	14:00	15:00	16:00	17:00	18:00	19:00	20:00	21:00	22:00	23:00	DAILY	DAILY	24-HR	RDGS.	
HR END (MST)	0:59	1:59	2:59	3:59	4:59	5:59	6:59	7:59	8:59	9:59	10:59	11:59	12:59	13:59	14:59	15:59	16:59	17:59	18:59	19:59	20:59	21:59	22:59	23:59	MIN.	MAX.	AVG.		
DAY																													
1	0.22	0.27	0.22	0.24	0.22	0.29	0.50	0.45	0.43	0.27	0.22	0.19	0.19	0.16	0.19	0.19	0.16	0.21	0.19	0.19	S	0.32	0.19	0.19	0.16	0.50	0.25	24	
2	0.14	0.16	0.19	0.16	0.14	0.16	0.16	0.19	0.16	0.16	0.22	0.19	0.16	0.22	0.19	0.14	0.16	0.19	0.17	S	0.29	0.22	0.16	0.37	0.14	0.37	0.19	24	
3	0.19	0.25	0.24	0.19	0.19	0.19	0.19	0.19	0.22	0.22	0.17	0.19	0.16	0.19	0.17	0.16	0.22	0.19	S	0.37	0.22	0.16	0.19	0.19	0.16	0.37	0.20	24	
4	0.19	0.16	0.16	0.17	0.19	0.19	0.16	0.19	0.19	0.19	0.19	0.19	0.22	0.19	0.19	0.19	0.22	S	0.37	0.19	0.19	0.19	0.19	0.19	0.16	0.37	0.20	24	
5	0.19	0.19	0.22	0.19	0.19	0.22	0.22	0.22	0.24	0.22	0.22	0.24	0.22	0.22	0.19	0.22	S	0.35	0.19	0.24	0.22	0.22	0.22	0.22	0.19	0.35	0.22	24	
6	0.22	0.22	0.21	0.27	0.24	0.27	0.37	0.32	0.29	0.22	0.22	0.19	0.22	0.19	0.16	S	0.42	0.19	0.19	0.19	0.22	0.19	0.22	0.16	0.16	0.42	0.23	24	
7	0.16	0.22	0.19	0.22	0.19	0.19	0.19	0.24	0.19	0.19	0.19	0.16	0.19	0.22	S	0.30	0.17	0.14	0.14	0.16	0.19	0.19	0.19	0.16	0.14	0.30	0.19	24	
8	0.19	0.19	0.16	0.16	0.16	0.14	0.16	0.22	0.16	0.17	0.19	0.19	0.19	S	0.32	0.19	0.22	0.22	0.16	0.18	0.16	0.17	0.22	0.19	0.14	0.32	0.19	24	
9	0.22	0.16	0.19	0.19	0.19	0.19	0.19	0.24	0.19	0.22	0.16	0.14	S	0.27	0.22	0.19	0.19	0.17	0.19	0.22	0.19	0.16	0.19	0.19	0.14	0.27	0.19	24	
10	0.22	0.19	0.19	0.19	0.16	0.19	0.16	0.16	0.16	C	C	C	C	C	C	0.82	0.95	0.42	0.40	0.43	0.40	0.40	0.40	0.40	0.16	0.95	0.35	24	
11	0.37	0.48	0.45	0.46	0.37	0.37	0.40	0.37	0.43	0.37	S	0.53	0.37	0.35	0.37	0.38	0.40	0.37	0.42	0.42	0.40	0.35	0.37	0.35	0.35	0.53	0.40	24	
12	0.40	0.37	0.38	0.40	0.40	0.40	0.35	0.35	0.35	S	0.45	0.35	0.37	0.32	0.32	0.34	0.37	0.35	0.35	0.35	0.32	0.35	0.35	0.37	0.32	0.45	0.36	24	
13	0.37	0.37	0.38	0.37	0.34	0.40	0.37	0.38	S	0.45	0.40	0.37	0.35	0.35	0.35	0.37	0.34	0.34	0.35	0.32	0.32	0.33	0.32	0.37	0.32	0.45	0.36	24	
14	0.33	0.37	0.35	0.35	0.32	0.34	0.35	S	0.40	0.37	0.35	0.35	0.32	0.35	0.35	0.30	0.32	0.29	0.35	0.34	0.32	0.35	0.35	0.32	0.29	0.40	0.34	24	
15	0.32	0.34	0.32	0.35	0.35	0.35	S	0.43	0.40	0.32	0.35	0.37	0.37	0.37	0.37	0.40	0.37	0.40	0.40	0.35	0.35	0.34	0.35	0.37	0.32	0.43	0.36	24	
16	0.34	0.38	0.35	0.32	0.37	S	0.43	0.37	0.35	0.37	0.37	0.38	0.32	0.34	0.34	0.32	0.32	0.32	0.29	0.34	0.34	0.32	0.32	0.32	0.29	0.43	0.34	24	
17	0.29	0.29	0.27	0.30	S	0.37	0.32	0.35	0.33	0.37	0.35	0.43	0.35	0.37	0.35	0.35	0.40	0.35	0.35	0.35	0.35	0.35	0.37	0.32	0.27	0.43	0.34	24	
18	0.32	0.32	0.32	S	0.40	0.35	0.32	0.38	0.40	0.34	0.33	0.32	0.32	0.27	0.29	0.29	0.30	0.32	0.29	0.29	0.32	0.29	0.32	0.29	0.27	0.40	0.32	24	
19	0.35	0.32	S	0.43	0.37	0.35	0.37	0.38	0.35	0.32	0.33	0.35	0.30	0.30	0.29	0.30	0.29	0.30	0.30	0.27	0.32	0.29	0.30	0.27	0.43	0.32	24		
20	0.29	S	0.38	0.30	0.35	0.33	0.30	0.30	0.32	0.27	0.29	0.32	0.32	0.29	0.27	0.27	0.27	0.29	0.27	0.29	0.27	0.29	0.27	0.27	0.27	0.38	0.30	24	
21	S	0.45	0.32	0.33	0.40	0.30	0.37	0.32	0.38	0.29	0.27	0.40	0.35	0.33	0.30	0.24	0.27	0.27	0.29	0.32	0.29	0.27	0.27	S	0.24	0.45	0.32	24	
22	0.42	0.37	0.30	0.35	0.29	0.27	0.32	0.34	0.32	0.32	0.27	0.30	0.27	0.30	0.27	0.27	0.25	0.32	0.43	0.30	0.37	S	0.37	0.25	0.43	0.32	24		
23	0.35	0.35	0.35	0.32	0.34	0.37	0.40	0.37	0.35	0.35	0.34	0.38	0.35	0.30	0.30	0.30	0.27	0.30	0.30	0.29	0.32	S	0.38	0.35	0.27	0.40	0.34	24	
24	0.33	0.30	0.30	0.29	0.30	0.32	0.30	0.29	0.32	0.32	0.32	0.32	0.29	0.32	0.32	0.27	0.30	0.27	0.29	0.32	S	0.40	0.40	0.40	0.27	0.40	0.32	24	
25	0.33	0.35	0.29	0.35	0.35	0.30	0.32	0.35	0.30	0.29	0.29	0.27	0.27	0.35	0.30	0.29	0.29	0.27	S	0.38	0.30	0.29	0.29	0.27	0.38	0.31	24		
26	0.32	0.30	0.38	0.42	0.40	0.48	0.40	0.43	0.35	0.32	0.32	0.33	0.32	0.30	0.30	0.30	0.29	0.27	S	0.40	0.33	0.45	0.35	0.29	0.27	0.48	0.35	24	
27	0.30	0.34	0.34	0.35	0.30	0.30	0.27	0.30	0.25	0.27	0.29	0.27	0.29	0.26	0.29	0.29	0.27	S	0.32	0.29	0.32	0.32	0.32	0.24	0.24	0.35	0.30	24	
28	0.29	0.24	0.24	0.27	0.22	0.22	0.27	0.24	0.27	0.27	0.27	0.29	0.27	0.33	0.29	0.27	S	0.30	0.24	0.27	0.24	0.24	0.32	0.29	0.22	0.33	0.27	24	
29	0.27	0.29	0.32	0.37	0.29	0.29	0.30	0.27	0.26	0.29	0.29	0.29	0.27	0.22	0.24	S	0.34	0.27	0.24	0.22	0.24	0.24	0.24	0.25	0.24	0.22	0.37	0.27	24
30	0.32	0.27	0.24	0.29	0.24	0.24	0.27	0.27	0.27	0.24	0.24	C1	C1	0.42	S	0.40	0.27	0.27	0.24	0.24	0.24	0.27	0.27	0.32	0.24	0.42	0.28	22	
HOURLY MAX	0.42	0.48	0.45	0.46	0.40	0.48	0.50	0.45	0.43	0.45	0.45	0.53	0.37	0.42	0.37	0.82	0.95	0.42	0.42	0.43	0.40	0.45	0.40	0.40					
HOURLY AVG	0.28	0.29	0.28	0.30	0.29	0.29	0.30	0.31	0.30	0.29	0.28	0.30	0.28	0.29	0.28	0.30	0.31	0.28	0.28	0.30	0.29	0.29	0.29	0.29					


STATUS FLAG CODES

C	- MONTHLY CALIBRATION	Q	- QUALITY ASSURANCE
C1	- REPEAT CALIBRATION	R	- RECOVERY
Y	- MAINTENANCE	X	- MACHINE MALFUNCTION
S	- DAILY ZERO/SPAN CHECK	G	- OUT FOR REPAIR
S1	- REPEAT ZERO/SPAN CHECK	P	- POWER FAILURE

MONTHLY SUMMARY

NUMBER OF NON-ZERO READINGS:	682
MAXIMUM INSTANTANEOUS VALUE:	0.95 ppb @ HOUR 16 ON DAY 10
IZS CALIBRATION TIME:	30 hrs
MONTHLY CALIBRATION TIME:	6 hrs
STANDARD DEVIATION:	0.08
OPERATIONAL TIME:	718 hrs

TOTAL REDUCED SULPHUR Instantaneous Maximum (TRS ppb)





PEACE RIVER AREA MONITORING PROGRAM COMMITTEE
Three Creeks 986b Station - April 2018

TOTAL HYDROCARBONS Instantaneous Maximum (THC ppm)

HR START (MST)	0:00	1:00	2:00	3:00	4:00	5:00	6:00	7:00	8:00	9:00	10:00	11:00	12:00	13:00	14:00	15:00	16:00	17:00	18:00	19:00	20:00	21:00	22:00	23:00	DAILY	DAILY	24-HR	RDGS.	
HR END (MST)	0:59	1:59	2:59	3:59	4:59	5:59	6:59	7:59	8:59	9:59	10:59	11:59	12:59	13:59	14:59	15:59	16:59	17:59	18:59	19:59	20:59	21:59	22:59	23:59	MIN.	MAX.	AVG.		
DAY																													
1	2.03	2.04	2.04	2.06	2.11	2.22	2.45	2.56	2.53	2.51	2.34	2.25	2.13	2.20	2.26	2.18	2.37	2.56	2.57	3.71	S	2.07	1.97	1.97	1.97	3.71	2.31	2.4	
2	1.98	2.50	1.99	2.01	2.01	2.73	2.72	2.04	2.04	1.99	2.02	2.20	2.00	2.21	2.25	2.12	2.12	2.22	2.00	S	1.96	1.96	2.37	1.99	1.96	2.73	2.15	24	
3	2.86	1.99	4.35	2.92	2.00	2.15	2.71	4.09	1.98	1.98	2.06	2.26	2.19	2.13	2.20	2.22	2.20	2.22	S	2.61	1.97	1.99	1.98	2.77	1.97	4.35	2.43	24	
4	1.97	1.98	2.23	2.11	2.04	2.07	1.97	2.08	2.01	2.08	2.13	2.16	2.28	2.25	2.23	2.34	2.21	S	2.09	5.10	2.09	1.98	4.18	1.98	1.97	5.10	2.33	24	
5	1.98	1.98	1.98	1.99	1.99	1.98	2.54	2.00	1.98	2.19	2.12	2.15	2.27	2.43	2.27	2.08	S	2.09	2.04	1.97	1.98	1.98	2.00	1.98	1.97	2.54	2.09	24	
6	1.99	1.99	1.98	1.99	1.99	2.00	2.01	1.99	1.99	1.99	1.98	1.99	1.98	1.98	1.97	S	1.97	1.94	1.97	1.98	1.96	1.98	2.01	2.00	1.94	2.01	1.98	24	
7	2.00	1.97	2.00	2.00	1.99	1.98	2.00	2.00	2.00	2.00	1.98	1.98	1.98	1.97	S	1.95	1.97	1.95	1.96	1.97	1.96	1.98	1.97	1.99	1.95	2.00	1.98	24	
8	2.00	1.99	1.98	1.99	2.00	1.99	2.03	2.04	2.04	2.07	2.03	2.29	2.15	S	2.01	2.22	2.03	1.99	1.99	2.00	2.05	2.02	2.02	2.04	1.98	2.29	2.04	24	
9	2.06	2.07	2.08	2.08	2.07	2.08	2.06	2.06	2.03	2.04	2.02	2.01	S	2.04	2.15	2.12	2.00	2.79	2.56	3.10	2.05	1.99	2.50	2.41	1.99	3.10	2.19	24	
10	2.02	2.02	1.97	1.97	1.97	1.97	1.97	1.95	1.95	C	C	C	C	C	C	1.97	1.97	2.21	1.95	1.95	1.98	1.97	1.97	1.98	1.95	2.21	1.99	24	
11	1.98	1.98	1.99	1.98	1.96	1.97	1.97	1.98	1.97	S	1.99	1.97	1.95	1.97	1.95	1.97	1.98	1.97	1.97	1.99	1.99	1.98	1.97	2.00	1.99	1.95	2.00	1.98	24
12	2.01	1.98	1.98	1.97	1.98	1.99	1.99	1.98	1.99	S	1.99	1.96	1.97	1.98	1.97	2.08	1.99	2.00	2.00	2.00	1.99	2.04	2.04	2.05	1.96	2.08	2.00	24	
13	2.04	2.05	2.16	2.06	2.05	2.05	2.05	S	2.05	2.03	2.04	2.03	2.00	1.99	1.98	1.99	2.50	3.65	2.05	2.03	2.02	2.00	1.98	1.98	3.65	2.12	24		
14	2.01	2.02	2.07	2.04	1.95	1.95	1.96	S	2.01	1.95	1.97	1.96	1.96	1.97	1.97	1.96	1.96	1.96	1.96	1.96	1.96	1.97	1.97	1.97	1.95	1.95	2.07	1.98	24
15	1.97	1.99	1.99	1.99	1.98	2.00	S	2.00	2.15	2.14	2.20	2.23	2.19	2.17	2.20	2.18	2.07	2.00	1.99	1.99	1.97	1.97	1.98	1.99	1.97	2.23	2.06	24	
16	1.99	1.97	1.98	1.99	1.99	S	1.99	1.99	1.97	1.98	1.99	1.98	1.99	1.97	1.97	1.98	1.98	1.96	1.97	1.98	1.98	1.98	1.97	1.97	1.96	1.99	1.98	24	
17	1.97	1.97	1.96	2.09	S	2.41	2.30	2.20	2.50	2.12	2.06	2.04	1.96	1.99	1.99	1.99	1.99	1.98	1.98	1.98	1.99	1.99	1.99	1.99	1.96	2.50	2.06	24	
18	1.99	1.99	2.01	S	2.07	2.01	2.01	2.02	2.01	2.01	2.01	2.00	2.00	1.97	1.97	1.97	1.97	1.97	1.97	1.97	1.97	1.96	1.94	1.94	1.94	2.07	1.99	24	
19	1.97	1.97	S	1.95	1.96	1.97	1.97	1.98	1.98	1.97	1.97	1.97	1.96	1.96	1.96	1.96	1.96	1.96	1.97	1.98	1.99	1.97	1.98	1.97	1.95	1.99	1.97	24	
20	1.96	S	1.93	1.96	1.98	1.97	1.95	1.95	1.97	1.98	2.02	1.98	1.97	1.97	1.98	1.98	1.99	1.97	1.97	1.97	2.00	1.98	1.99	2.03	1.93	2.03	1.98	24	
21	S	2.03	2.02	2.68	2.96	2.55	2.14	2.03	2.13	2.00	1.97	2.04	2.04	1.96	1.96	1.97	1.97	1.97	1.96	2.67	1.95	1.95	1.96	S	1.95	2.96	2.13	24	
22	1.96	1.96	1.97	1.96	1.97	2.03	2.03	2.12	2.17	2.24	2.12	2.14	2.13	2.16	2.14	2.08	1.97	1.97	1.99	2.00	1.99	2.14	S	1.99	1.96	2.24	2.05	24	
23	2.00	2.04	2.05	2.03	2.02	2.03	2.03	2.03	S1	S1	2.02	2.02	2.02	2.01	2.01	2.03	2.01	2.11	2.02	2.01	2.06	S	2.06	2.01	2.00	2.11	2.03	22	
24	2.00	2.02	1.98	1.97	1.99	2.00	2.00	2.00	1.98	2.03	1.99	1.98	1.97	C1	C1	C1	C1	1.94	1.95	1.97	S	2.01	2.13	2.27	1.94	2.27	2.01	20	
25	2.13	2.02	2.29	2.13	2.11	2.11	2.02	2.10	2.10	1.97	1.97	1.96	1.96	2.06	1.99	1.97	2.01	1.95	1.95	S	1.96	2.14	1.98	1.97	1.95	2.29	2.04	24	
26	1.97	1.98	2.64	2.04	2.04	2.01	2.17	2.10	2.03	2.04	2.11	2.02	2.01	1.99	2.00	2.07	1.99	1.96	S	2.02	1.98	2.15	2.07	2.08	1.96	2.64	2.06	24	
27	2.14	2.13	2.14	2.14	2.12	2.09	2.07	2.01	1.98	1.95	1.98	1.96	1.97	1.96	1.96	1.97	1.97	S	1.94	2.04	2.02	1.97	1.97	2.01	1.94	2.14	2.02	24	
28	2.01	2.03	2.05	2.05	1.97	1.93	1.94	1.95	2.04	1.96	1.96	1.96	1.95	1.99	2.02	2.00	S	1.95	1.95	1.95	1.95	1.96	1.96	1.96	1.96	1.93	2.05	1.98	24
29	1.95	1.97	1.96	1.93	1.95	1.97	1.98	1.95	1.93	2.02	2.03	2.09	2.00	2.02	1.98	S	2.08	2.04	2.06	1.98	1.95	2.01	1.96	1.93	1.93	2.09	1.99	24	
30	1.97	2.01	2.04	2.00	1.98	1.97	1.97	1.97	1.97	1.97	1.99	1.97	1.95	S	1.96	1.96	2.01	2.08	2.08	1.96	1.94	1.94	2.26	2.80	1.94	2.80	2.03	24	
HOURLY MAX	2.86	2.50	4.35	2.92	2.96	2.73	2.72	4.09	2.53	2.51	2.34	2.29	2.28	2.43	2.27	2.34	2.37	2.79	3.65	5.10	2.09	2.15	4.18	2.80					
HOURLY AVG	2.03	2.02	2.13	2.07	2.04	2.07	2.10	2.11	2.05	2.04	2.04	2.05	2.04	2.05	2.05	2.05	2.03	2.08	2.09	2.25	1.99	2.00	2.11	2.07					


STATUS FLAG CODES

C	- MONTHLY CALIBRATION	Q	- QUALITY ASSURANCE
C1	- REPEAT CALIBRATION	R	- RECOVERY
Y	- MAINTENANCE	X	- MACHINE MALFUNCTION
S	- DAILY ZERO/SPAN CHECK	G	- OUT FOR REPAIR
S1	- REPEAT ZERO/SPAN CHECK	P	- POWER FAILURE

MONTHLY SUMMARY

NUMBER OF NON-ZERO READINGS:	678
MAXIMUM INSTANTANEOUS VALUE:	5.10 ppm @ HOUR 19 ON DAY 4
IZS CALIBRATION TIME:	30 hrs
MONTHLY CALIBRATION TIME:	6 hrs
OPERATIONAL TIME:	714 hrs
STANDARD DEVIATION:	0.25

TOTAL HYDROCARBONS Instantaneous Maximum (THC ppm)





PEACE RIVER AREA MONITORING PROGRAM COMMITTEE
Three Creeks 986b Station - April 2018

METHANE MAX Instantaneous Maximum (CH₄ ppm)

HR START (MST)	0:00	1:00	2:00	3:00	4:00	5:00	6:00	7:00	8:00	9:00	10:00	11:00	12:00	13:00	14:00	15:00	16:00	17:00	18:00	19:00	20:00	21:00	22:00	23:00	DAILY	DAILY	24-HR	RDGS.	
HR END (MST)	0:59	1:59	2:59	3:59	4:59	5:59	6:59	7:59	8:59	9:59	10:59	11:59	12:59	13:59	14:59	15:59	16:59	17:59	18:59	19:59	20:59	21:59	22:59	23:59	MIN.	MAX.	AVG.		
DAY																													
1	2.04	2.05	2.05	2.07	2.12	2.24	2.46	2.53	2.48	2.52	2.35	2.26	2.15	2.21	2.27	2.19	2.38	2.57	2.58	3.73	S	2.08	1.97	1.98	1.97	3.73	2.32	24	
2	1.98	2.51	2.00	2.02	2.02	2.74	2.74	2.05	2.05	2.00	2.02	2.21	2.00	2.22	2.26	2.13	2.13	2.24	2.01	S	1.97	1.96	2.38	2.00	1.96	2.74	2.16	24	
3	2.87	2.00	4.37	2.93	2.01	2.16	2.73	4.11	1.99	1.99	2.07	2.27	2.20	2.14	2.21	2.24	2.22	2.24	S	2.62	1.97	1.99	1.99	2.79	1.97	4.37	2.44	24	
4	1.98	1.99	2.25	2.12	2.04	2.08	1.98	2.09	2.02	2.08	2.14	2.17	2.29	2.25	2.24	2.35	2.22	S	2.10	S	5.10	2.10	1.99	4.19	1.99	1.98	5.10	2.34	24
5	1.99	1.99	1.99	2.00	2.00	1.99	2.55	2.01	1.99	2.20	2.13	2.16	2.28	2.44	2.28	2.09	S	2.10	2.05	1.98	1.98	1.98	2.00	1.98	1.98	2.55	2.09	24	
6	1.99	2.00	1.99	1.99	1.99	2.01	2.01	2.00	2.00	2.00	1.99	1.99	1.99	1.99	1.97	S	1.98	1.94	1.97	1.98	1.96	1.98	2.01	2.00	1.94	2.01	1.99	24	
7	2.01	1.98	2.00	2.00	1.99	1.99	2.01	2.01	2.01	2.01	1.99	1.99	1.99	1.98	S	1.95	1.97	1.95	1.96	1.97	1.96	1.98	1.98	1.99	1.95	2.01	1.99	24	
8	2.00	2.00	1.99	1.99	2.01	2.00	2.03	2.04	2.04	2.08	2.03	2.30	2.16	S	2.01	2.24	2.04	1.99	1.99	2.01	2.07	2.03	2.03	2.05	1.99	2.30	2.05	24	
9	2.07	2.08	2.09	2.09	2.08	2.08	2.06	2.06	2.04	2.04	2.02	2.02	S	2.04	2.16	2.13	2.00	2.80	2.57	3.12	2.06	2.00	2.50	2.43	2.00	3.12	2.20	24	
10	2.03	2.03	1.98	1.98	1.98	1.97	1.97	1.96	1.96	C	C	C	C	C	C	1.98	1.98	1.99	1.96	1.96	1.98	1.98	1.98	1.98	1.96	2.03	1.98	24	
11	1.98	1.99	2.00	1.99	1.96	1.97	1.97	1.98	1.98	1.98	S	1.99	1.97	1.95	1.97	1.99	1.98	1.98	2.00	1.99	1.98	1.97	2.00	2.00	1.95	2.00	1.98	24	
12	2.01	1.99	1.99	1.98	1.99	1.99	1.99	1.99	2.00	S	1.99	1.96	1.98	1.98	1.98	2.09	2.00	2.01	2.01	2.01	2.00	2.04	2.05	2.06	1.96	2.09	2.00	24	
13	2.05	2.06	2.17	2.06	2.06	2.06	2.06	S	2.06	2.04	2.05	2.04	2.01	2.00	1.99	2.00	2.51	3.67	2.06	2.04	2.02	2.01	1.99	1.99	3.67	2.13	24		
14	2.01	2.03	2.07	2.04	1.96	1.96	1.96	S	2.02	1.95	1.97	1.96	1.96	1.98	1.98	1.96	1.96	1.97	1.96	1.97	1.98	1.98	1.98	1.96	1.95	2.07	1.98	24	
15	1.97	1.99	1.99	2.00	1.99	2.01	S	2.01	2.16	2.15	2.21	2.24	2.21	2.18	2.21	2.19	2.08	2.01	2.00	1.99	1.98	1.98	1.98	2.00	1.97	2.24	2.07	24	
16	2.00	1.97	1.99	2.00	2.00	S	2.00	2.00	1.98	1.99	2.00	1.99	1.99	1.98	1.98	1.99	1.99	1.97	1.98	1.98	1.98	1.98	1.98	1.97	1.98	2.00	1.99	24	
17	1.98	1.98	1.96	2.10	S	2.43	2.32	2.21	2.50	2.14	2.07	2.05	1.96	2.00	2.00	2.00	1.98	1.99	1.99	1.98	1.98	1.99	1.99	2.00	1.96	2.50	2.07	24	
18	1.99	1.99	2.01	S	2.07	2.02	2.02	2.03	2.02	2.02	2.01	2.01	2.01	1.98	1.98	1.98	1.97	1.98	1.98	1.98	1.97	1.96	1.95	1.94	1.94	2.07	1.99	24	
19	1.97	1.98	S	1.95	1.96	1.98	1.98	1.99	1.98	1.98	1.98	1.98	1.97	1.96	1.96	1.96	1.96	1.97	1.97	1.99	2.00	1.97	1.99	1.98	1.95	2.00	1.97	24	
20	1.96	S	1.93	1.96	1.99	1.97	1.96	1.95	1.98	1.99	2.03	1.98	1.98	1.98	1.98	1.98	1.99	1.97	1.98	1.97	2.01	1.99	2.00	2.04	1.93	2.04	1.98	24	
21	S	2.03	2.02	2.69	2.98	2.57	2.16	2.04	2.14	2.00	1.98	2.05	2.05	1.96	1.96	1.97	1.96	1.97	1.97	2.69	1.96	1.96	1.96	S	1.96	2.98	2.14	24	
22	1.96	1.96	1.98	1.96	1.98	2.03	2.03	2.13	2.18	2.25	2.13	2.15	2.14	2.17	2.16	2.09	1.97	1.98	1.99	2.01	2.00	2.15	S	1.99	1.96	2.25	2.06	24	
23	2.00	2.05	2.05	2.03	2.03	2.03	2.04	2.04	S1	S1	2.03	2.03	2.03	2.02	2.02	2.04	2.02	2.12	2.03	2.02	2.07	S	2.06	2.02	2.00	2.12	2.04	22	
24	2.01	2.02	1.98	1.98	2.00	2.00	2.01	2.00	1.99	2.04	2.00	1.99	1.97	C1	C1	C1	C1	1.94	1.95	1.97	S	2.02	2.14	2.28	1.94	2.28	2.02	20	
25	2.14	2.03	2.30	2.14	2.12	2.12	2.02	2.11	2.11	1.98	1.97	1.96	1.96	2.07	2.00	1.98	2.01	1.95	1.96	S	1.96	2.15	1.99	1.98	1.95	2.30	2.04	24	
26	1.98	1.98	2.65	2.04	2.05	2.02	2.18	2.11	2.04	2.05	2.13	2.03	2.02	1.99	2.01	2.08	2.00	1.97	S	2.03	1.99	2.16	2.08	2.09	1.97	2.65	2.07	24	
27	2.15	2.14	2.15	2.15	2.14	2.10	2.07	1.99	1.95	1.99	1.96	1.97	1.96	1.97	1.96	1.96	1.97	1.98	S	1.94	2.05	2.03	1.98	1.97	2.02	1.94	2.15	2.03	24
28	2.02	2.03	2.06	2.06	1.98	1.94	1.94	1.95	2.04	1.96	1.97	1.96	1.96	2.00	2.03	2.00	S	1.95	1.96	1.95	1.95	1.96	1.96	1.96	1.94	2.06	1.98	24	
29	1.95	1.97	1.96	1.94	1.96	1.98	1.98	1.96	1.94	2.02	2.03	2.10	2.01	2.02	1.99	S	2.09	2.04	2.07	1.99	1.96	2.02	1.96	1.93	1.93	2.10	1.99	24	
30	1.98	2.02	2.04	2.01	1.99	1.98	1.98	1.97	1.98	1.97	2.00	1.98	1.96	S	1.96	2.02	2.02	2.09	1.96	1.94	1.94	2.27	2.81	1.94	2.81	2.04	24		
HOURLY MAX	2.87	2.51	4.37	2.93	2.98	2.74	2.74	4.11	2.50	2.52	2.35	2.30	2.29	2.44	2.28	2.35	2.38	2.80	3.67	5.10	2.10	2.16	4.19	2.81					
HOURLY AVG	2.04	2.03	2.14	2.08	2.05	2.08	2.11	2.12	2.06	2.05	2.05	2.06	2.04	2.05	2.06	2.06	2.03	2.08	2.10	2.25	1.99	2.01	2.12	2.08					


STATUS FLAG CODES

C	- MONTHLY CALIBRATION	Q	- QUALITY ASSURANCE
C1	- REPEAT CALIBRATION	R	- RECOVERY
Y	- MAINTENANCE	X	- MACHINE MALFUNCTION
S	- DAILY ZERO/SPAN CHECK	G	- OUT FOR REPAIR
S1	- REPEAT ZERO/SPAN CHECK	P	- POWER FAILURE

MONTHLY SUMMARY

NUMBER OF NON-ZERO READINGS:	678
MAXIMUM INSTANTANEOUS VALUE:	5.10 ppm @ HOUR 19 ON DAY 4
IZS CALIBRATION TIME:	30 hrs
MONTHLY CALIBRATION TIME:	6 hrs
STANDARD DEVIATION:	0.26
OPERATIONAL TIME:	714 hrs

METHANE MAX Instantaneous Maximum (CH₄ ppm)





PEACE RIVER AREA MONITORING PROGRAM COMMITTEE
Three Creeks 986b Station - April 2018

NON-METHANE HYDROCARBONS Instantaneous Maximum (NMHC ppm)

HR START (MST)	0:00	1:00	2:00	3:00	4:00	5:00	6:00	7:00	8:00	9:00	10:00	11:00	12:00	13:00	14:00	15:00	16:00	17:00	18:00	19:00	20:00	21:00	22:00	23:00	DAILY MIN.	DAILY MAX.	24-HR AVG.	RDGS.	
HR END (MST)	0:59	1:59	2:59	3:59	4:59	5:59	6:59	7:59	8:59	9:59	10:59	11:59	12:59	13:59	14:59	15:59	16:59	17:59	18:59	19:59	20:59	21:59	22:59	23:59					
DAY 1	0.00	0.00	0.00	0.00	0.00	0.00	0.08	0.08	0.07	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	S	0.00	0.00	0.00	0.00	0.08	0.01	24	
2	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	S	0.00	0.00	0.00	0.00	0.00	0.00	24	
3	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	S	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	24	
4	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	S	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	24	
5	0.00	0.00	0.01	0.01	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	S	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.01	0.00	24	
6	0.00	0.01	0.00	0.00	0.00	0.00	0.01	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	S	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.01	0.00	24	
7	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.01	0.00	0.00	0.00	0.00	0.00	S	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.01	0.00	24	
8	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	S	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	24	
9	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	S	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	24	
10	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	C	C	C	C	C	C	0.00	0.00	0.23	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.23	0.01	24
11	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	S	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	24	
12	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	S	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	24	
13	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	S	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	24	
14	0.00	0.00	0.00	0.00	0.00	0.00	S	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	24	
15	0.00	0.00	0.00	0.00	0.00	0.00	S	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	24	
16	0.00	0.00	0.00	0.00	0.00	S	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	24	
17	0.00	0.00	0.00	0.00	S	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	24	
18	0.00	0.00	0.00	S	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	24	
19	0.00	0.00	S	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	24	
20	0.00	S	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	24	
21	S	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.02	0.00	0.00	0.00	0.00	0.00	0.00	0.00	S	0.00	0.02	0.00	24	
22	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	S	0.02	0.00	0.02	0.00	24	
23	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	S1	S1	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	S	0.00	0.00	0.00	0.00	0.00	22	
24	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	C1	C1	C1	C1	0.00	0.00	0.00	S	0.00	0.00	0.00	0.00	0.00	0.00	20	
25	0.10	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	S	0.00	0.00	0.00	0.00	0.00	0.10	0.00	24	
26	0.00	0.00	0.13	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	S	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.13	0.01	24	
27	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	S	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	24	
28	0.00	0.00	0.00	0.00	0.00	0.01	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	S	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.01	0.00	24	
29	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	S	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	24	
30	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	S	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	24	
HOURLY MAX	0.10	0.01	0.13	0.01	0.00	0.01	0.08	0.08	0.07	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.02	0.23	0.00	0.00	0.00	0.00	0.00	0.02	0.00	0.00	0.00	24	
HOURLY AVG	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.01	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	24	


STATUS FLAG CODES

C	- MONTHLY CALIBRATION	Q	- QUALITY ASSURANCE
C1	- REPEAT CALIBRATION	R	- RECOVERY
Y	- MAINTENANCE	X	- MACHINE MALFUNCTION
S	- DAILY ZERO/SPAN CHECK	G	- OUT FOR REPAIR
S1	- REPEAT ZERO/SPAN CHECK	P	- POWER FAILURE

MONTHLY SUMMARY

NUMBER OF NON-ZERO READINGS:	14
MAXIMUM INSTANTANEOUS VALUE:	0.23 ppm @ HOUR 17 ON DAY 10
IZS CALIBRATION TIME:	30 hrs
MONTHLY CALIBRATION TIME:	6 hrs
STANDARD DEVIATION:	0.01
OPERATIONAL TIME:	714 hrs

NON-METHANE HYDROCARBONS Instantaneous Maximum (NMHC ppm)





PEACE RIVER AREA MONITORING PROGRAM COMMITTEE
Three Creeks 986b Station - April 2018

WIND SPEED Instantaneous Maximum (WS kph)

HR START (MST)	0:00	1:00	2:00	3:00	4:00	5:00	6:00	7:00	8:00	9:00	10:00	11:00	12:00	13:00	14:00	15:00	16:00	17:00	18:00	19:00	20:00	21:00	22:00	23:00	DAILY	DAILY	24-HR	RDGS.
HR END (MST)	0:59	1:59	2:59	3:59	4:59	5:59	6:59	7:59	8:59	9:59	10:59	11:59	12:59	13:59	14:59	15:59	16:59	17:59	18:59	19:59	20:59	21:59	22:59	23:59	MIN.	MAX.	AVG.	
DAY 1	5.9	4.6	4.4	4.5	4.5	4.6	4.6	4.9	4.3	12.1	15.3	17.8	23.3	20.7	24.8	21.0	19.8	17.1	16.8	9.0	4.3	6.8	7.2	3.3	3.3	24.8	10.9	24
2	4.2	11.0	4.0	3.4	5.0	9.3	10.9	14.0	19.0	15.2	16.1	20.4	26.2	31.4	27.7	25.5	21.5	22.0	19.2	10.6	4.8	5.6	5.6	3.8	3.4	31.4	14.0	24
3	3.8	3.9	5.3	6.3	8.0	8.5	8.4	14.8	17.6	20.7	19.2	19.7	23.2	25.9	20.8	18.6	23.4	23.0	16.7	12.2	8.6	14.3	10.2	14.5	3.8	25.9	14.5	24
4	14.3	12.1	9.8	12.5	5.7	6.7	8.3	23.5	22.9	21.8	29.7	36.8	38.8	37.8	32.7	36.5	C	C	18.6	7.3	5.0	3.8	10.7	7.8	3.8	38.8	18.3	24
5	3.8	2.3	3.2	2.7	11.2	22.2	13.2	23.7	25.3	20.1	24.1	20.7	21.9	24.7	24.4	21.6	17.7	19.2	13.5	19.0	20.2	11.4	11.7	14.7	2.3	25.3	16.4	24
6	17.1	15.4	13.0	10.9	12.4	17.4	16.1	12.2	14.1	24.8	27.6	26.9	26.1	24.7	26.9	25.4	23.6	26.9	23.3	18.1	18.2	14.7	14.8	15.1	10.9	27.6	19.4	24
7	13.2	11.4	12.4	12.2	15.9	17.8	13.8	15.7	14.5	27.9	28.9	32.6	30.4	30.0	32.3	32.2	32.3	35.3	30.0	21.1	17.2	18.2	16.2	14.1	11.4	35.3	21.9	24
8	11.9	11.3	11.3	8.6	11.6	9.9	9.6	8.8	6.2	13.7	19.0	12.9	12.8	17.2	20.1	15.8	9.9	13.1	12.6	14.7	6.9	5.0	5.5	5.4	5.0	20.1	11.4	24
9	5.4	5.8	6.7	9.4	7.8	7.4	5.9	8.3	7.8	14.2	17.6	16.5	14.8	13.6	17.0	15.2	13.9	9.1	8.5	6.8	3.5	6.7	9.3	9.7	3.5	17.6	10.0	24
10	12.9	21.4	20.1	18.5	23.3	21.7	19.7	32.7	34.4	26.5	30.5	33.2	33.6	26.7	28.4	31.6	30.7	24.7	27.4	29.6	23.7	23.4	23.8	30.9	12.9	34.4	26.2	24
11	22.5	20.3	16.0	20.7	19.6	18.9	14.3	16.1	17.1	16.2	17.6	16.1	17.4	21.9	23.7	21.0	18.1	17.0	16.7	18.7	18.3	8.4	9.7	9.3	8.4	23.7	17.3	24
12	9.9	16.2	15.3	18.4	19.1	18.9	16.9	15.7	20.3	16.3	19.8	15.0	16.5	13.0	8.3	23.2	18.6	17.6	21.7	19.9	13.1	13.3	11.5	11.9	8.3	23.2	16.3	24
13	20.3	14.1	11.5	9.7	8.5	10.7	13.5	17.7	22.2	20.8	23.5	26.8	25.8	26.0	24.0	21.1	18.4	12.0	7.2	14.4	21.3	34.7	33.0	26.5	7.2	34.7	19.3	24
14	19.7	17.3	16.8	21.9	20.0	20.4	26.6	21.7	27.9	33.6	32.6	41.2	36.7	36.4	35.7	34.3	29.8	28.6	21.3	18.2	12.9	13.4	12.2	13.9	12.2	41.2	24.7	24
15	15.6	10.0	7.5	9.6	3.3	2.9	3.0	18.3	17.7	19.2	23.4	26.3	28.1	28.7	32.1	32.9	31.3	30.9	31.7	27.1	33.1	31.3	25.2	21.7	2.9	33.1	21.3	24
16	26.6	27.8	23.9	23.0	23.8	18.8	23.5	22.5	19.5	27.0	26.6	30.2	31.2	29.0	29.8	32.4	26.2	23.7	24.2	28.9	27.1	15.7	16.9	16.9	15.7	32.4	24.8	24
17	14.2	15.4	13.3	14.6	13.5	14.7	14.0	15.4	13.2	20.0	18.8	17.7	16.1	18.0	25.4	24.3	24.1	25.5	22.5	23.2	16.4	13.3	12.6	21.6	12.6	25.5	17.8	24
18	16.9	10.6	5.8	4.6	6.5	6.0	7.2	9.3	14.0	15.1	20.9	27.4	28.3	31.6	33.6	26.9	26.4	31.3	27.8	20.4	13.2	14.0	17.2	15.8	4.6	33.6	18.0	24
19	15.6	19.7	14.2	18.4	22.2	19.6	25.3	34.3	32.0	33.6	34.5	35.0	32.8	42.5	37.2	36.1	28.0	28.6	24.8	23.1	24.7	24.2	20.8	14.7	14.2	42.5	26.7	24
20	22.7	28.4	27.7	25.0	21.0	25.3	23.8	22.4	27.1	35.4	32.2	28.4	33.6	37.4	31.9	28.0	26.5	19.7	14.2	9.2	8.6	9.4	8.9	7.7	7.7	37.4	23.1	24
21	8.4	8.8	9.1	8.8	6.9	5.9	4.7	3.8	6.2	14.1	14.5	18.7	16.9	18.7	20.8	30.2	27.9	22.5	24.5	13.5	17.5	19.7	21.4	22.5	3.8	30.2	15.3	24
22	18.1	25.2	24.6	28.3	36.6	27.6	40.7	32.7	35.1	40.8	44.6	42.6	38.5	31.1	37.1	24.8	34.3	24.0	18.4	13.1	8.9	9.0	6.1	7.7	6.1	44.6	27.1	24
23	10.2	7.9	7.9	8.4	7.1	8.1	11.1	16.0	18.8	18.6	22.1	23.9	28.5	27.6	32.8	36.9	33.4	22.1	25.7	21.8	12.9	11.9	11.6	12.8	7.1	36.9	18.3	24
24	15.5	17.3	22.0	25.2	30.5	19.3	23.8	28.1	28.1	23.9	26.0	28.7	31.9	33.2	29.8	33.7	29.9	30.0	26.0	22.7	13.1	18.8	16.5	18.3	13.1	33.7	24.7	24
25	12.0	6.2	5.7	3.5	7.1	4.0	4.1	10.1	12.2	12.8	20.6	30.8	34.9	37.0	38.0	33.8	33.1	38.6	30.4	20.3	13.2	13.4	13.3	16.7	3.5	38.6	18.8	24
26	12.3	7.6	4.6	3.5	5.2	7.5	3.5	10.3	13.0	16.7	18.8	17.6	14.0	17.9	20.9	19.4	21.6	18.7	13.5	10.1	6.3	5.6	6.3	7.8	3.5	21.6	11.8	24
27	8.5	7.9	9.6	9.6	8.1	10.8	11.2	12.3	16.1	19.6	19.5	19.2	23.9	28.4	30.1	24.2	27.2	26.6	24.2	21.6	8.1	6.3	8.1	8.9	6.3	30.1	16.3	24
28	10.7	9.7	10.9	10.1	9.9	9.3	11.9	14.7	20.7	15.1	16.3	19.1	20.2	28.9	25.7	29.7	26.4	28.5	27.2	32.9	35.9	32.1	27.2	32.1	9.3	35.9	21.1	24
29	33.0	37.0	31.6	23.6	21.5	15.5	14.0	17.6	18.2	13.4	20.6	21.1	16.4	15.6	18.1	17.3	18.7	22.0	18.3	18.4	13.6	9.8	9.3	6.2	6.2	37.0	18.8	24
30	6.9	6.5	9.9	10.2	12.4	11.7	14.6	21.2	24.6	28.6	29.0	32.0	37.4	29.4	27.0	22.5	29.9	24.3	46.4	52.0	42.5	26.5	19.0	10.3	6.5	52.0	24.0	24
HOURLY MAX	33.0	37.0	31.6	28.3	36.6	27.6	40.7	34.3	35.1	40.8	44.6	42.6	38.8	42.5	38.0	36.9	34.3	38.6	46.4	52.0	42.5	34.7	33.0	32.1				
HOURLY AVG	13.7	13.8	12.6	12.9	13.6	13.4	13.9	17.3	19.0	21.3	23.7	25.2	26.0	26.8	27.2	26.5	24.9	23.5	21.8	19.3	15.8	14.7	14.1	14.1				


STATUS FLAG CODES

C	- MONTHLY CALIBRATION	Q	- QUALITY ASSURANCE
C1	- REPEAT CALIBRATION	R	- RECOVERY
Y	- MAINTENANCE	X	- MACHINE MALFUNCTION
S	- DAILY ZERO/SPAN CHECK	G	- OUT FOR REPAIR
S1	- REPEAT ZERO/SPAN CHECK	P	- POWER FAILURE

MONTHLY SUMMARY

MAXIMUM INSTANTANEOUS VALUE:	52.0	kph	@ HOUR	19	ON DAY	30	
OPERATIONAL TIME:						720	hrs

WIND SPEED Instantaneous Maximum (WS kph)




APPENDIX IV
REPORT CERTIFICATION FORM

Report Certification Form

Alberta Airshed (if applicable)	EPA Approval or Code of Practice Registration # (if applicable)
YES	NA
Company Name (if applicable)	Industrial Operation Name (if applicable)
Peace River Area Monitoring Program Committee	Three Creeks 986b Station
Name of the Representative of the Person Responsible	Position / Title of the Representative of the Person Responsible
Mike Bisaga / Lily Lin	Technical Program Managers
Is an External Party Certifying the Report? (If 'Yes', fill in the fields below for the external person.)	
<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No	
Name of External Person Certifying the Report	Position / Title of External Person Certifying the Report
Cheri Sinclair	Supervisor, Customer Service, Air Services
Company Name for the External Person Certifying the Report	Identification of Qualifications / Professional Designations of the External Person Certifying the Report
Maxxam Analytics, A Bureau Veritas Group Company	B.Sc.

Maxxam Analytics is the designated contractor conducting monitoring and reporting activities. I certify that the submitted data has been (a) reviewed and validated as per the AMD Chapter 6: Ambient Data Quality. I certify that the submitted report (b) accurately reflects the monitoring results and reporting timeframe and (c) meets the specified analysis, summarization and reporting requirements as per the AMD Chapter 9: Reporting.



Signature of the External Person Certifying the Report

24-May-2018





Report Issued Date (dd-mon-yyyy)

APPENDIX V
DATA VALIDATION CERTIFICATION FORM



Validation Certificate Form

Client: <u>Peace River Area Monitoring Program Committee</u>	Project #: <u>8449-2018-04-67-C</u>
Site: <u>Three Creeks 986b Station</u>	Contact: <u>Karla Reesor</u>

Level 0 Preliminary Verification	<u></u>	Date <u>08-May-2018</u>
Level 1 Primary Validation	<u></u>	Date <u>22-May-2018</u>
Level 2 Final Validation	<u></u>	Date <u>23-May-2018</u>
Level 3 Independent Data Review	<u></u>	Date <u>24-May-2018</u>
Post-Final Validation	<u>NA</u>	Date <u>NA</u>

Notes
The Post-Final Validation step serves to re-evaluate the data that errors or omissions are discovered and/or suspected after the initial submittal of data. This validation is performed on an annual basis.