



MAXXAM ANALYTICS
#1 2080 39 Ave. NE, Calgary, AB
T2E 6P7

maxxam.ca
Toll Free 800-386-7247
Fax 403-219-3673

AMBIENT AIR MONITORING MONTHLY DATA REPORT
PEACE RIVER AREA MONITORING PROGRAM COMMITTEE
THREE CREEKS 986B STATION

JOB #: 8449-2018-12-67-C

December 2018

Prepared for:

PEACE RIVER AREA MONITORING PROGRAM COMMITTEE

Attention: LILY LIN

DATE: **January 23, 2019**

Prepared by:

Wunmi Adekanmbi, M.Sc., EPt, PMP
Project Team Lead, Customer Service, Air Services

Reviewed by:

Cheri Sinclair, B.Sc.
Supervisor, Customer Service, Air Services

SUMMARY

In December 2018, Maxxam Analytics was contracted to manage the ambient air quality monitoring and maintenance activities at the Three Creeks 986b Station, near Peace River Oil Sands Area 2, Alberta. The monitoring station provides continuous meteorological measurements and air quality data for non-compliance parameters, as requested by the PRAMP Committee.

All data collected this month was compliant with the requirements outlined in the Air Monitoring Directive (Alberta Environment and Parks, 2016).

The operational time for all continuous ambient air analyzers, meteorological systems and data acquisition systems were above the 90% requirement.

All data collected this month were within the Alberta Ambient Air Quality Objectives and Guidelines (November, 2018).

Canister System: A canister event was recorded on December 9 at 20:25, at an initial concentration of 0.63 ppm. The sample was processed for analysis as outlined in the PRAMP Chain of Custody.

THC/CH₄/NMHC:

- An additional zero-span verification was performed as a quality check following a span gas replacement on December 4, incurring one hour of downtime.
- The sample pump was rebuilt on December 12, in response to a "low flow" alarm. Four hours of downtime were incurred as a result.

Station Temperature: The station temperature exhibited instability beginning in the November monitoring period. An HVAC company was engaged and the faulty components were ordered. On December 11, the HVAC system was repaired and the temporary portable heaters were removed from site.

The summary of results is presented on the following pages.

Any deviations or modifications made to the sampling or analytical methods are outlined in Section 1.0, Discussion. On this basis, Maxxam Analytics is issuing this completed report to Peace River Area Monitoring Program Committee.

Should you have any questions concerning the results or if we can be of further assistance, please contact us at 403-219-3661 or toll-free at 1-800-386-7247.

Monthly Continuous Data Summary

Peace River Area Monitoring Program Committee Three Creeks 986b Station						MAXIMUM VALUES						OPERATIONAL TIME (%)	
						1-HOUR			24-HOUR				
PARAMETER	OBJECTIVES		EXCEEDANCES		MONTHLY AVERAGE	READING	DAY	HOUR	WIND SPEED (kph)	WIND DIRECTION (sector)	READING	DAY	
	1-hr	24-hr	1-hr	24-hr									
SO ₂ (ppb)	172	48	0	0	0	2	17	17	6.1	SSW	1	8	100.0
TRS (ppb)	-	-	-	-	0.31	0.63	24	21	2.8	E	0.40	24	100.0
THC (ppm)	-	-	-	-	2.04	2.40	9	20	1.7	W	2.17	28	99.3
CH ₄ (ppm)	-	-	-	-	2.04	2.40	28	12	0.3	NNE	2.17	28	99.3
NMHC (ppm)	-	-	-	-	0.00	0.10	9	20	1.7	W	0.00	1	99.3
RELATIVE HUMIDITY (%)	-	-	-	-	79	100	21	3	11.8	WNW	95	1	100.0
BAROMETRIC PRESSURE (millibar)	-	-	-	-	938	955	30	18	6.7	S	951	5	100.0
AMBIENT TEMPERATURE (°C)	-	-	-	-	-9.3	3.6	18	13	5.6	SW	0.8	13	100.0
STATION TEMPERATURE (°C)	-	-	-	-	21.9	26.7	5	11	3.8	WSW	25.6	9	100.0
VECTOR WS (kph)	-	-	-	-	2.4	18.2	11	14	-	SSE	11.7	11	100.0
VECTOR WD (sec)	-	-	-	-	192 (S)	-	-	-	-	-	-	-	100.0

SOUR GAS PROCESSING INDUSTRY
MONTHLY REPORT SUMMARY

Three Creeks 986b Station

Peace River Area Monitoring Program Committee

Plant Name / Location

Company

Licence Number	Report Date	
	YEAR	MONTH
N/A	2018	December

CONTINUOUS AMBIENT MONITORING						
PARAMETER	% TIME OPERATIONAL	ONE - HOUR AVERAGE		24 - HOUR AVERAGE		
		MAXIMUM VALUES	NO. READINGS > REGULATION	MAXIMUM VALUES	NO. READINGS > REGULATION	
SO ₂	100.0	0.002 ppm	0	0.001 ppm	0	
TRS	100.0	0.001 ppm	-	0.000 ppm	-	
THC	99.3	2.40 ppm	-	2.17 ppm	-	
CH ₄	99.3	2.40 ppm	-	2.17 ppm	-	
NMHC	99.3	0.10 ppm	-	0.00 ppm	-	
RH	100.0	100 %	-	95 %	-	
BP	100.0	955 mb	-	951 mb	-	
Ambient TPX	100.0	3.6 °C	-	0.8 °C	-	
Station TPX	100.0	26.7 °C	-	25.6 °C	-	
Wind Speed	100.0	18.2 kph	-	11.7 kph	-	
Wind Direction	100.0	-	-	-	-	

	FOR ALBERTA ENVIRONMENT USE ONLY

SIGNATURE OF COMPANY REPRESENTATIVE

Exceedance Summary Report

SO₂ 1-Hour Exceedances

Measured concentrations of sulphur dioxide were below the 1-hour AAAQO of 172 ppb.

SO₂ 24-Hour Exceedances

Measured concentrations of sulphur dioxide were below the 24-hour AAAQO of 48.0 ppb.

In accordance with EPEA and the Substance Release Regulation.

In accordance with A Guide to Release Reporting and the Alberta Ambient Air Quality Objectives and Guidelines Summary.

TABLE OF CONTENTS

<u>Title</u>	<u>Page</u>
SUMMARY	2
MONTHLY CONTINUOUS DATA SUMMARY REPORT	3
SOUR GAS SUMMARY REPORT	4
EXCEEDANCE SUMMARY REPORT	5
TABLE OF CONTENTS	6
1.0 Discussion	8
2.0 Project Personnel	11
3.0 Plant Monthly Required AMD Summary	11
4.0 Calculations and Results	11
5.0 Methods and Procedures	12
Appendix I	Continuous Monitoring Data Results
	15
	Sulphur Dioxide
	16
	Total Reduced Sulphur
	21
	Total Hydrocarbon
	26
	Methane
	31
	Non-Methane Hydrocarbon
	36
	Wind Speed
	41
	Wind Direction
	45
	Relative Humidity
	49
	Barometric Pressure
	52
	Ambient Temperature
	55
	Station Temperature
	58
Appendix II	Equipment Calibration Results
	61
	Sulphur Dioxide
	62
	Total Reduced Sulphur
	65
	Total Hydrocarbon
	68
	Wind System
	74
	Meteorological System Check
	76
	Calibrators
	78
	Calibration Gases
	81

Appendix III	Maximum Instantaneous Data	85
Appendix IV	Report Certification Form	98
Appendix V	Data Validation Certification Form	100

1.0 Discussion

This monthly report consists of continuous monitoring results for the following parameters: Sulphur Dioxide (SO₂), Total Reduced Sulphur (TRS), Total Hydrocarbon (THC), Methane (CH₄), Non-Methane Hydrocarbon (NMHC), Relative Humidity (RH), Barometric Pressure (BP), Ambient Temperature (AmbTPX), Station Temperature (StnTPX), Wind Speed (WS) and Wind Direction (WD).

The sample inlet filter for all continuous air analyzers are replaced before the calibration begins. The sample manifold is cleaned during the site visit each month.

Control checks, consisting of a zero and span, are conducted daily on all continuous air monitors. In place of the air sample, zero air (from scrubbed air or gas cylinders) is used for zero checks, and a known concentration of the pollutant being analyzed is used for span checks. These checks are controlled by automatic timers and valves. The total zero span cycle is completed within an hour, the commencement of the zero span cycle is at the beginning of the hour.

Multipoint calibrations are done a minimum of once a month for each continuous air monitor. An additional calibration is required under the following conditions: 1) within three days after the initial start-up and stabilization of a newly installed instrument, 2) prior to shut-down or moving of an instrument which has been working to specification, and 3) when major repair has been done on the instrument.

Time during the first multi-point calibration is not considered downtime (Data is flagged as C). If more than one calibration is performed during the month, the time during the additional calibration is considered as downtime (Data is flagged as C1).

Only one zero/span check is run per day. Time during the zero/span check is not considered as downtime (Data is flagged as S). If an extra zero/span check is performed, the time during the additional check is considered as downtime (Data is flagged as S1).

The AMD requires each instrument and accompanying data recording system to be operational 90% of the time, at a minimum, for each monthly monitoring period.

All sampling, analysis, and QA/QC for this project was performed by Maxxam Analytics and complies with the Alberta Air Monitoring Directive.

Data contained in this monthly report has undergone verification and validation based on the requirements of the AMD (December, 2016) Chapter 6: Ambient Data Quality and Chapter 9: Reporting. The descriptions of the data verification and validation process can be found in Section 5 of this report. Instantaneous data, where applicable, is provided for reference purposes and has not undergone zero correction. The minimum and maximum statistics are highlighted in the data table and are for reference only. The highlighted cells are based on the software's interpretation of the exact position of the minimum or maximum value. The visual presentation of these statistics may not be the obvious choice in a data range due to rounding, truncating or analyzer specifications.

Hourly/minute data have been reviewed based on daily zero/span results and multi-point calibration results. Data may be considered invalid if a zero-corrected span check in excess of +/- 10% of the span concentration (established by the previous multi-point calibration) is encountered and/or significant differences in the calibration factor occurs (greater than 10%).

SULPHUR DIOXIDE (SO₂)

- Operational time for the monitoring period was 100%.
- The routine monthly calibration was performed on December 12.
- Unstable station temperature was exhibited early in the month. Based on instrument diagnostic information, no impact on data quality was demonstrated. The analyzer manual states the operating temperature thus: *20-30 °C (may be safely operated over the range of 0-45 °C)*. Data collected at a station temperature range of 0-45°C is therefore considered valid. However, data collected outside the optimum range of 20-30°C should be used with caution.

TOTAL REDUCED SULPHUR (TRS)

- Operational time for the monitoring period was 100%.
- The routine monthly calibration was performed on December 12.
- Unstable station temperature was exhibited early in the month. Based on instrument diagnostic information, no impact on data quality was demonstrated. The analyzer manual states the operating temperature thus: *15-35 °C (may be safely operated over the range of 0-45 °C)*. Data collected at a station temperature range of 0-45°C is therefore considered valid. All data recorded in the December monitoring period was collected within the optimum temperature range.

TOTAL HYDROCARBONS (THC), METHANE (CH₄) and NON-METHANE HYDROCARBONS (NMHC)

- Operational time for the monitoring period was 99.3%, equivalent to five hours of downtime.
- The span gas cylinder was replaced on December 4. A repeat zero-span check was subsequently completed as a quality check. The expected span value was not updated as the span gas concentration did not change significantly. One hour of downtime was recorded due to the additional quality check.
- An immediate site visit was scheduled on December 12 in response to a "low flow" alarm. Following a successful shut-down calibration, the sample pump was rebuilt. A successful post-repair calibration was subsequently completed. As the shut-down calibration met AMD requirements, no data was discarded due to this event. However four hours of downtime were incurred due to the maintenance activity.
- The carrier gas (N₂) cylinder was replaced during the December 12 site visit.
- Unstable station temperature was exhibited early in the month. Based on instrument diagnostic information, no impact on data quality was demonstrated. The analyzer manual states the operating temperature thus: *15-35 °C*. All monthly data was collected within the operating temperature range and is therefore considered valid.
- The canister sampler is programmed to draw in a whole air sample when the 5-minute average concentration of NMHC is above 0.30 ppm. A representative sample of ambient air is collected over a one-hour period when the canister event is triggered. One canister event was recorded on December 9 at 20:25, at an initial concentration of 0.63 ppm. The sample was processed for analysis as outlined in the PRAMP Chain of Custody.

WIND SPEED (WS) and WIND DIRECTION (WD)

- Operational time for the monitoring period was 100%.
- An anemometer sensor check was conducted on December 12. The result was satisfactory.
- Wind data is reported as vector wind speed and vector wind direction. Wind direction is defined as the direction from which the wind is blowing from and is measured in degrees from true north.

RELATIVE HUMIDITY (RH)

- Operational time for the monitoring period was 100%.
- A humidity sensor check was conducted on December 12. The result was satisfactory.

BAROMETRIC PRESSURE (BP)

- Operational time for the monitoring period was 100%.
- A pressure sensor check was conducted on December 12. The result was satisfactory.

AMBIENT TEMPERATURE (AmbTPX)

- Operational time for the monitoring period was 100%.
- A temperature sensor check was conducted on December 12. The result was satisfactory.

STATION TEMPERATURE (StnTPX)

- Operational time for the monitoring period was 100%.
- The station temperature exhibited instability beginning in the November monitoring period. An HVAC company was engaged and the faulty components were ordered. On December 11, the HVAC system was repaired and the temporary portable heaters were removed from site.

2.0 Project Personnel

Karla Reesor was the contact for Peace River Area Monitoring Program Committee and the Maxxam field technician was Christopher Wesson.

3.0 Plant Monthly Required AMD Summary

All data collected this month was compliant with the requirements outlined in the Air Monitoring Directive (Alberta Environment and Parks, 2016).

The operational time for all continuous ambient air analyzers, meteorological systems and data acquisition systems were above the 90% requirement.

All data collected this month were within the Alberta Ambient Air Quality Objectives and Guidelines (November, 2018).

4.0 Calculations and Results

All calculations and reporting of results follow the methods described in the AMD, 2016.

5.0 Methods and Procedures

The following methods and procedures were used to complete the monitoring program:

- Maxxam AIR SOP-00001: Methane, Non-Methane Hydrocarbon Analyzer Monitoring
- Maxxam AIR SOP-00013: RM Young Wind Monitor Calibration
- Maxxam AIR SOP-00209: Ambient Sulphur Monitoring

There were no deviations from the prescribed methods.

The following instruments were used to perform the test program:

- Sulphur Dioxide - Thermo 43C UV Fluorescent Analyzer
- Total Reduced Sulphur - Thermo 43i - TLE UV Fluorescent Analyzer
- Methane, Non-Methane Hydrocarbon - Thermo 55i FID Analyzer
- Wind System - RM Young Unit
- Relative Humidity - RM Young Unit
- Barometric Pressure - Met One Unit
- Ambient Temperature - RM Young Unit
- Datalogger - Envidas Ultimate

The following steps were used to complete the data verification and validation process:

Level 0 Preliminary Verification

Level 0 data are raw data obtained directly from the data acquisition system (DAS). Under the step of Level 0, these data undergo a certain amount of manual or automated screening and flagging. It included a) identification of periods of missing data; b) verification of time stamps against reference time; c) verification that instrument diagnostics/datalogger flags indicate normal operation; d) comparison of data to upper and lower limits; e) rate of change flagging indicating that data changed too rapidly or not at all; and f) verification that zero, span and multipoint performance checks are within specifications. This level of verification is performed on a daily basis.

Level 1 Primary Validation

Validation actions under the step of Level 1 include a) review of all screening flags assigned during preliminary verification; b) review of all supporting site information and documentation; c) review of operational acceptance limits for each parameter/analyser; d) review of daily zero/span and monthly calibration results for all gaseous parameters; and e) application of any necessary adjustments to data (e.g. baseline adjustments, below zero adjustments). This level of validation is performed on a monthly basis.

Level 2 Final Validation


The purpose of Level 2 validation is to verify that there are no inconsistencies among related data, or among regional data measured at nearby sites.

Level 3 Independent Data Review

Level 3 validation is the last step of data review, and it is completed by an individual that is independent of both field operations and primary data validation. A final independent QA review and endorsement is performed during this step before data is submitted to Alberta Environment.

Post-Final Validation

The Post-Final Validation step serves to re-evaluate the data that errors or omissions are discovered and/or suspected after the initial submittal of data. Any data issues or patterns which were not clear on a monthly basis are highlighted during this step. This validation is performed on an annual basis.



Source: Air Monitoring Directive (December 2016), Chapter 6, Ambient Data Quality; Figure 1 Data Collection and Management Process Flow Chart

APPENDIX I
CONTINUOUS MONITORING DATA RESULTS

SULPHUR DIOXIDE

SULPHUR DIOXIDE Hourly Averages (SO₂ ppb)

	HR START (MST)	0:00	1:00	2:00	3:00	4:00	5:00	6:00	7:00	8:00	9:00	10:00	11:00	12:00	13:00	14:00	15:00	16:00	17:00	18:00	19:00	20:00	21:00	22:00	23:00	DAILY MIN.	DAILY MAX.	24-HR AVG.	RDGS.	
	HR END (MST)	0:59	1:59	2:59	3:59	4:59	5:59	6:59	7:59	8:59	9:59	10:59	11:59	12:59	13:59	14:59	15:59	16:59	17:59	18:59	19:59	20:59	21:59	22:59	23:59					
DAY																														
1		0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	24	
2		0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	1	1	0	1	1	0	1	0	24	
3		1	1	1	1	1	S	1	1	1	0	1	1	0	1	1	0	0	0	0	0	0	0	0	0	0	0	1	0	24
4		0	0	0	0	S	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	24	
5		0	0	0	S	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	24	
6		0	S	0	0	0	0	1	0	0	0	0	0	0	0	0	0	0	0	0	1	1	0	0	0	0	0	1	0	24
7		S	0	1	1	1	1	0	1	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	1	0	24
8		1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	0	0	0	0	0	0	0	1	1	24
9		0	1	1	1	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	S	0	0	0	0	1	0	24
10		0	0	0	1	1	0	0	0	0	0	0	0	0	0	0	0	0	0	0	1	S	0	0	0	0	0	1	0	24
11		0	0	0	1	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	S	0	0	0	0	0	1	0	24
12		0	0	0	0	0	0	0	0	0	0	0	0	0	0	C	C	C	C	0	0	0	0	0	0	0	0	0	24	
13		0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	S	0	0	0	0	0	0	0	0	24	
14		0	0	0	0	0	0	0	0	0	0	0	0	0	0	1	0	1	S	0	1	0	1	0	0	0	1	0	24	
15		0	1	1	0	0	0	0	1	0	1	0	0	0	0	0	0	S	0	0	0	0	0	0	0	0	1	0	24	
16		0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	S	0	0	0	0	0	0	0	0	0	0	24		
17		0	0	0	0	0	0	0	0	1	0	0	0	0	0	S	1	1	1	2	1	1	1	1	1	0	2	1	24	
18		1	1	1	1	1	1	0	0	0	0	1	0	0	0	S	1	0	1	1	1	1	2	2	2	0	2	1	24	
19		2	1	1	1	1	1	1	1	1	1	1	1	1	1	S	1	0	0	0	0	0	0	0	0	0	2	1	24	
20		0	0	0	0	0	0	0	0	0	0	0	0	0	0	S	0	0	0	0	0	0	0	0	0	0	0	0	24	
21		0	0	0	0	0	0	0	0	0	0	0	0	0	0	S	0	0	0	0	0	0	0	0	0	0	0	0	24	
22		0	0	0	0	0	0	0	0	0	0	0	0	0	0	S	0	0	0	0	0	0	0	0	0	0	0	0	24	
23		0	0	0	0	0	0	0	0	0	0	0	0	0	0	S	0	0	0	0	0	0	0	0	1	1	0	1	24	
24		0	0	0	0	0	0	0	0	0	0	0	0	0	0	S	0	1	1	1	1	1	1	1	1	0	1	1	24	
25		1	1	1	1	1	S	1	0	1	0	0	1	0	0	0	0	0	0	0	0	0	0	0	0	0	1	0	24	
26		0	0	0	0	S	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	24	
27		0	0	0	S	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	24	
28		0	0	S	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	24	
29		0	S	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	24	
30		S	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	S	0	0	0	0	24		
31		0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	S	0	0	0	0	24		
HOURLY MAX		2	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	2	1	1	2	2	2	2	2				
HOURLY AVG		0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0			

STATUS FLAG CODES


C	- MONTHLY CALIBRATION	Q	- QUALITY ASSURANCE
C1	- REPEAT CALIBRATION	R	- RECOVERY
Y	- MAINTENANCE	X	- MACHINE MALFUNCTION
S	- DAILY ZERO/SPAN CHECK	G	- OUT FOR REPAIR
S1	- REPEAT ZERO/SPAN CHECK	P	- POWER FAILURE

OBJECTIVE LIMIT:


ALBERTA ENVIRONMENT: 1-HR 172 ppb 24-HR 48 ppb

MONTHLY SUMMARY

NUMBER OF 1-HR EXCEEDANCES:	0
NUMBER OF 24-HR EXCEEDANCES:	0
NUMBER OF NON-ZERO READINGS:	113
MINIMUM 1-HR AVERAGE	0 ppb @ HOUR
MAXIMUM 1-HR AVERAGE:	2 ppb @ HOUR
MAXIMUM 24-HR AVERAGE:	1 ppb
I2S CALIBRATION TIME:	32 hrs
MONTHLY CALIBRATION TIME:	4 hrs
OPERATIONAL TIME:	AMD OPERATION UPTIME:
STANDARD DEVIATION:	0
MONTHLY AVERAGE:	0 ppb




SULPHUR DIOXIDE Hourly Averages (SO₂ ppb)




Peace River Area Monitoring Program Committee_Three Creeks 986b Station_SO₂ (ppb)_18/12

Pollutant Rose_Wind Frequency (Blowing From)_CALM Avg = 0.1_CALM % = 11.9%



Direction	0-3	3-10	10-85	85-170	>170	TOTAL
N	5.9	0.0	0.0	0.0	0.0	5.9
NE	5.2	0.0	0.0	0.0	0.0	5.2
E	13.0	0.0	0.0	0.0	0.0	13.0
SE	10.3	0.0	0.0	0.0	0.0	10.3
S	23.9	0.0	0.0	0.0	0.0	23.9
SW	17.5	0.0	0.0	0.0	0.0	17.5
W	3.4	0.0	0.0	0.0	0.0	3.4
NW	8.9	0.0	0.0	0.0	0.0	8.9
Summary	88.1	0.0	0.0	0.0	0.0	88.1
CALM	11.9	0.0	0.0	0.0	0.0	11.9

SO₂[ppb] Calibration: PRAMP_986 Monthly: 18/12 Type: Span



TOTAL REDUCED SULPHUR




PEACE RIVER AREA MONITORING PROGRAM COMMITTEE

Three Creeks 986b Station - December 2018

TOTAL REDUCED SULPHUR Hourly Averages (TRS ppb)

	HR START (MST)	0:00	1:00	2:00	3:00	4:00	5:00	6:00	7:00	8:00	9:00	10:00	11:00	12:00	13:00	14:00	15:00	16:00	17:00	18:00	19:00	20:00	21:00	22:00	23:00	DAILY MIN.	DAILY MAX.	24-HR AVG.	RDGS.	
	HR END (MST)	0:59	1:59	2:59	3:59	4:59	5:59	6:59	7:59	8:59	9:59	10:59	11:59	12:59	13:59	14:59	15:59	16:59	17:59	18:59	19:59	20:59	21:59	22:59	23:59					
DAY																														
1		0.30	0.33	0.36	0.34	0.32	0.32	S	0.36	0.36	0.31	0.27	0.26	0.26	0.26	0.26	0.26	0.26	0.27	0.30	0.30	0.30	0.31	0.30	0.28	0.26	0.36	0.30	24	
2		0.27	0.28	0.28	0.26	0.25	S	0.32	0.27	0.27	0.26	0.26	0.25	0.26	0.25	0.25	0.27	0.28	0.28	0.30	0.30	0.29	0.30	0.29	0.29	0.25	0.32	0.27	24	
3		0.29	0.29	0.28	0.28	S	0.35	0.30	0.30	0.31	0.30	0.30	0.32	0.37	0.37	0.49	0.39	0.30	0.29	0.29	0.29	0.29	0.27	0.27	0.49	0.31	24			
4		0.28	0.29	0.27	S	0.33	0.29	0.28	0.27	0.27	0.26	0.24	0.25	0.32	0.25	0.26	0.25	0.27	0.28	0.27	0.29	0.27	0.28	0.24	0.33	0.27	24			
5		0.28	0.28	S	0.35	0.30	0.28	0.30	0.30	0.32	0.32	0.28	0.28	0.31	0.29	0.30	0.30	0.30	0.29	0.29	0.29	0.31	0.31	0.28	0.35	0.30	24			
6		0.33	S	0.41	0.36	0.35	0.34	0.34	0.33	0.33	0.34	0.34	0.32	0.34	0.32	0.30	0.31	0.33	0.30	0.31	0.31	0.30	0.29	0.29	0.29	0.41	0.32	24		
7		S	0.40	0.34	0.34	0.33	0.33	0.33	0.31	0.31	0.31	0.32	0.31	0.32	0.30	0.28	0.29	0.29	0.30	0.30	0.29	0.28	0.28	S	0.28	0.40	0.31	24		
8		0.40	0.35	0.31	0.33	0.32	0.31	0.31	0.29	0.28	0.29	0.29	0.27	0.29	0.29	0.27	0.28	0.28	0.25	0.26	0.26	0.26	S	0.33	0.25	0.40	0.30	24		
9		0.29	0.26	0.28	0.29	0.28	0.29	0.25	0.26	0.27	0.28	0.26	0.25	0.24	0.25	0.24	0.25	0.24	0.23	0.24	0.23	0.59	S	0.33	0.28	0.59	0.28	24		
10		0.27	0.27	0.27	0.31	0.31	0.30	0.29	0.29	0.29	0.29	0.29	0.28	0.27	0.29	0.29	0.29	0.31	0.33	0.31	0.31	0.30	0.27	0.37	0.30	0.37	0.30	24		
11		0.31	0.32	0.31	0.32	0.32	0.30	0.31	0.28	0.29	0.28	0.28	0.26	0.28	0.28	0.28	0.28	0.28	0.26	0.26	0.26	S	0.37	0.31	0.30	0.27	0.34	0.29	24	
12		0.27	0.26	0.28	0.27	0.27	0.27	0.26	0.29	0.27	C	C	C	C	C	C	C	C	C	C	C	C	C	C	C	C	0.34	0.27	24	
13		0.26	0.25	0.25	0.25	0.26	0.26	0.26	0.26	0.26	0.26	0.26	0.26	0.26	0.26	0.26	0.26	0.26	0.25	0.25	0.25	0.27	0.26	0.26	0.27	0.25	0.32	0.26	24	
14		0.26	0.26	0.27	0.27	0.26	0.26	0.26	0.26	0.28	0.33	0.29	0.25	0.26	0.27	0.25	0.27	S	0.30	0.27	0.30	0.32	0.25	0.26	0.26	0.32	0.25	0.33	0.27	24
15		0.32	0.31	0.32	0.28	0.23	0.26	0.27	0.23	0.23	0.26	0.20	0.20	0.22	0.20	0.20	S	0.31	0.25	0.24	0.24	0.43	0.42	0.38	0.34	0.20	0.43	0.27	24	
16		0.37	0.30	0.31	0.29	0.28	0.28	0.27	0.28	0.27	0.31	0.29	0.27	0.28	0.31	S	0.41	0.32	0.28	0.28	0.28	0.31	0.32	0.29	0.27	0.41	0.30	24		
17		0.29	0.30	0.31	0.30	0.30	0.31	0.31	0.32	0.31	0.32	0.32	0.33	S	0.38	0.36	0.34	0.35	0.34	0.32	0.35	0.38	0.36	0.36	0.29	0.38	0.33	24		
18		0.36	0.34	0.35	0.35	0.36	0.34	0.34	0.35	0.35	0.35	0.36	0.33	S	0.38	0.32	0.32	0.31	0.32	0.30	0.30	0.31	0.30	0.30	0.30	0.38	0.33	24		
19		0.31	0.30	0.30	0.32	0.32	0.31	0.31	0.26	0.25	0.25	0.24	S	0.30	0.27	0.26	0.25	0.25	0.25	0.27	0.24	0.26	0.27	0.27	0.24	0.27	0.24	24		
20		0.26	0.27	0.26	0.27	0.27	0.31	0.29	0.31	0.33	0.29	S	0.36	0.29	0.31	0.28	0.26	0.27	0.27	0.30	0.28	0.27	0.27	0.26	0.24	0.36	0.28	24		
21		0.25	0.26	0.27	0.29	0.28	0.27	0.29	0.28	0.29	S	0.35	0.30	0.28	0.29	0.27	0.28	0.27	0.27	0.26	0.28	0.26	0.26	0.26	0.25	0.35	0.28	24		
22		0.30	0.30	0.31	0.27	0.27	0.32	0.32	0.33	S	0.42	0.34	0.34	0.32	0.33	0.32	0.33	0.32	0.32	0.32	0.32	0.32	0.32	0.35	0.32	0.40	0.33	24		
23		0.42	0.56	0.53	0.44	0.44	0.39	0.33	S	0.50	0.38	0.36	0.36	0.35	0.33	0.33	0.32	0.32	0.31	0.31	0.33	0.34	0.33	0.37	0.37	0.31	0.56	0.38	24	
24		0.41	0.33	0.31	0.29	0.27	0.35	S	0.40	0.40	0.31	0.31	0.30	0.31	0.31	0.32	0.39	0.45	0.54	0.53	0.51	0.63	0.60	0.47	0.63	0.40	24			
25		0.41	0.38	0.30	0.32	0.33	S	0.41	0.34	0.35	0.35	0.36	0.32	0.33	0.34	0.31	0.29	0.30	0.28	0.27	0.33	0.36	0.53	0.27	0.53	0.34	24			
26		0.54	0.44	0.41	0.41	S	0.46	0.34	0.31	0.33	0.36	0.49	0.29	0.32	0.33	0.30	0.32	0.36	0.29	0.29	0.32	0.34	0.33	0.31	0.29	0.54	0.36	24		
27		0.30	0.33	0.35	S	0.62	0.50	0.45	0.36	0.33	0.33	0.33	0.32	0.33	0.31	0.33	0.30	0.31	0.30	0.30	0.29	0.30	0.31	0.32	0.29	0.62	0.34	24		
28		0.34	0.33	S	0.38	0.31	0.34	0.32	0.31	0.31	0.31	0.33	0.33	0.33	0.33	0.30	0.31	0.31	0.31	0.31	0.31	0.31	0.29	0.30	0.38	0.32	24			
29		0.30	S	0.42	0.35	0.43	0.63	0.38	0.41	0.35	0.37	0.40	0.38	0.39	0.37	0.35	0.35	0.35	0.33	0.32	0.31	0.39	0.40	0.36	0.30	0.63	0.38	24		
30		S	0.40	0.35	0.35	0.35	0.34	0.33	0.33	0.34	0.34	0.34	0.34	0.34	0.34	0.33	0.34	0.31	0.33	0.34	0.34	0.34	S	0.31	0.40	0.34	24			
31		0.43	0.35	0.35	0.37	0.37	0.36	0.34	0.34	0.31	0.32	0.32	0.30	0.30	0.29	0.30	0.30	0.29	0.27	0.29	0.26	S	0.34	0.26	0.43	0.32	24			
HOURLY MAX		0.54	0.56	0.53	0.44	0.62	0.63	0.45	0.41	0.50	0.42	0.49	0.38	0.39	0.38	0.49	0.41	0.39	0.45	0.54	0.53	0.59	0.63	0.60	0.53					
HOURLY AVG		0.32	0.32	0.32	0.32	0.32	0.33	0.32	0.31	0.31	0.31	0.31	0.30	0.30	0.30	0.30	0.30	0.30	0.30	0.30	0.32	0.32	0.32	0.32	0.32	0.32	0.32			




NUMBER OF NON-ZERO READINGS:	705	MINIMUM 1-HR AVERAGE:	0.20	ppb @ HOUR	10	ON DAY	15
MAXIMUM 1-HR AVERAGE:	0.63	ppb @ HOUR	21	ON DAY	24		
MAXIMUM 24-HR AVERAGE:	0.40	ppb		ON DAY	24		
I2S CALIBRATION TIME:	33	hrs		OPERATIONAL TIME:		744	hrs
MONTHLY CALIBRATION TIME:	6	hrs		AMD OPERATION UPTIME:		100.0	%
STANDARD DEVIATION:	0.06			MONTHLY AVERAGE:		0.31	ppb



PEACE RIVER AREA MONITORING PROGRAM COMMITTEE


Three Creeks 986b Station - December 2018

TOTAL REDUCED SULPHUR Hourly Averages (TRS ppb)




Peace River Area Monitoring Program Committee_Three Creeks 986b StationTRS (ppb)_18/12

Pollutant Rose_Wind Frequency (Blowing From)_CALM Avg = 0.4_CALM % = 11.9%




TRS[ppb] Calibration: PRAMP_986 Monthly: 18/12 Type: Span



TOTAL HYDROCARBON

TOTAL HYDROCARBONS Hourly Averages (THC ppm)

	HR START (MST)	0:00	1:00	2:00	3:00	4:00	5:00	6:00	7:00	8:00	9:00	10:00	11:00	12:00	13:00	14:00	15:00	16:00	17:00	18:00	19:00	20:00	21:00	22:00	23:00	DAILY MIN.	DAILY MAX.	24-HR AVG.	RDGS.
	HR END (MST)	0:59	1:59	2:59	3:59	4:59	5:59	6:59	7:59	8:59	9:59	10:59	11:59	12:59	13:59	14:59	15:59	16:59	17:59	18:59	19:59	20:59	21:59	22:59	23:59				
DAY																													
1		2.07	1.97	1.98	2.01	2.04	2.05	S	2.03	2.05	2.05	2.04	2.03	2.01	2.01	2.02	2.02	2.09	2.05	2.02	2.02	2.02	2.04	2.03	1.97	2.09	2.03	24	
2		2.03	2.01	2.02	2.02	2.07	S	2.01	2.01	2.00	1.98	1.99	1.96	1.96	1.97	1.97	1.97	1.97	1.98	1.99	2.00	2.01	2.02	2.02	1.96	2.07	2.00	24	
3		2.03	2.03	2.02	2.02	S	2.01	2.01	2.01	2.01	2.01	2.10	2.17	2.09	2.04	1.98	1.99	1.98	1.97	1.97	1.97	1.97	1.96	1.96	1.96	2.17	2.01	24	
4		1.95	1.95	1.96	S	1.98	1.98	1.98	1.96	S1	1.97	1.97	1.99	2.00	1.99	1.98	1.98	1.98	1.99	1.99	2.00	2.01	2.02	2.04	2.06	1.95	2.06	1.99	23
5		2.05	2.13	S	1.99	1.99	1.99	2.03	2.00	2.06	2.01	1.99	1.98	1.99	1.98	1.99	1.98	1.97	1.97	1.97	1.97	1.99	2.04	1.97	1.98	1.97	2.13	2.00	24
6		1.98	S	2.00	2.00	1.99	2.00	1.99	2.00	2.00	2.01	2.02	2.01	2.01	2.00	2.00	2.01	2.00	1.99	1.99	2.01	2.02	2.02	2.02	1.98	2.02	2.00	24	
7		S	2.04	2.04	2.05	2.07	2.07	2.06	2.04	2.03	2.02	2.02	2.03	2.01	2.01	2.02	2.03	2.02	2.02	2.02	2.01	2.01	1.99	2.00	2.07	2.03	24		
8		2.02	2.05	2.03	2.05	2.03	2.06	2.05	2.04	2.04	2.02	2.04	2.03	2.04	2.06	2.05	2.05	2.05	2.04	2.01	2.01	2.02	2.00	2.06	2.03	24			
9		2.01	1.99	2.01	2.04	2.04	2.03	2.04	2.04	2.03	2.03	2.01	2.01	2.01	2.00	2.01	2.02	2.05	2.00	2.00	2.40	S	2.01	2.03	1.99	2.40	2.04	24	
10		2.04	2.05	2.03	2.04	2.05	2.04	2.01	2.01	2.00	1.98	1.98	1.97	1.96	1.96	1.96	1.96	1.97	1.97	1.95	1.94	S	1.95	1.95	1.96	1.94	2.05	1.99	24
11		1.96	1.96	1.95	1.95	1.95	1.95	1.94	1.96	1.95	1.94	1.94	1.93	1.90	1.90	1.90	1.90	1.90	1.89	S	1.89	1.88	1.88	1.88	1.88	1.96	1.92	24	
12		1.88	1.89	1.89	1.89	1.89	1.90	1.89	1.89	C	C	Y	Y	Y	C	C	C	C	1.98	1.99	1.99	1.98	1.98	1.98	1.98	1.98	1.92	20	
13		1.98	1.98	1.98	1.97	1.98	1.97	1.97	1.98	1.97	1.97	1.98	1.98	2.00	2.01	1.99	1.99	S	1.99	1.99	1.98	1.98	1.98	1.97	2.01	1.98	24		
14		1.98	1.98	1.98	1.98	1.98	1.98	1.97	1.99	1.99	1.98	1.97	1.97	1.97	1.96	1.97	S	1.97	1.98	1.99	2.18	2.00	2.01	2.05	1.96	2.18	1.99	24	
15		2.05	2.04	2.06	2.05	2.04	2.04	2.06	2.03	2.03	2.02	2.03	2.04	2.14	2.10	2.13	S	2.02	2.01	2.01	2.03	2.07	2.11	2.15	2.19	2.01	2.19	2.06	24
16		2.15	2.10	2.07	2.21	2.07	2.02	2.02	2.00	2.00	2.01	2.01	2.01	2.00	2.00	S	2.00	2.00	2.00	2.00	2.01	2.01	2.01	2.02	2.00	2.21	2.03	24	
17		2.02	2.03	2.04	2.04	2.05	2.09	2.08	2.04	2.04	2.04	2.05	2.06	S	2.08	2.08	2.09	2.09	2.08	2.06	2.09	2.10	2.08	2.07	2.02	2.10	2.06	24	
18		2.08	2.10	2.09	2.07	2.13	2.04	2.09	2.05	2.03	2.03	2.02	S	2.00	2.01	2.02	2.02	2.03	2.05	2.06	2.07	2.06	2.06	2.00	2.13	2.06	24		
19		2.06	2.07	2.06	2.06	2.06	2.08	2.08	2.05	2.03	2.03	S	2.01	2.00	2.00	2.00	1.99	1.99	2.00	1.99	1.99	2.00	2.00	2.00	1.99	2.08	2.03	24	
20		2.00	2.00	2.01	2.00	2.01	2.02	2.02	2.06	2.04	2.03	S	2.02	1.99	1.98	1.98	1.99	1.99	1.98	1.98	1.98	1.98	1.98	1.98	1.97	2.06	2.00	24	
21		2.03	2.05	2.07	2.04	2.03	2.02	2.07	2.02	2.02	S	2.05	2.20	2.06	2.08	2.09	2.09	2.06	2.33	2.30	2.21	2.02	2.30	2.02	2.33	2.10	24		
22		2.05	2.02	2.01	2.01	2.06	2.09	2.12	S	2.09	2.09	2.10	2.05	2.04	2.04	2.04	2.04	2.05	2.06	2.06	2.07	2.12	2.11	2.01	2.12	2.06	24		
23		2.14	2.14	2.14	2.19	2.30	2.23	2.05	S	2.07	2.06	2.07	2.06	2.07	2.05	2.05	2.04	2.04	2.03	2.03	2.03	2.02	2.02	2.02	2.30	2.08	24		
24		2.02	2.03	2.03	2.04	2.04	2.04	S	2.05	2.05	2.06	2.07	2.08	2.09	2.11	2.13	2.14	2.14	2.16	2.18	2.19	2.19	2.17	2.14	2.02	2.19	2.10	24	
25		2.13	2.11	2.07	2.08	2.04	S	2.05	2.06	2.10	2.12	2.21	2.11	2.08	2.12	2.18	2.15	2.15	2.06	2.05	2.05	2.06	2.08	2.12	2.17	2.04	2.21	2.10	24
26		2.19	2.16	2.13	2.12	S	2.13	2.10	2.14	2.19	2.10	2.07	2.06	2.05	2.05	2.05	2.06	2.07	2.17	2.08	2.10	2.09	2.09	2.10	2.05	2.19	2.10	24	
27		2.14	2.25	2.09	S	2.14	2.12	2.15	2.13	2.09	2.13	2.09	2.10	2.09	2.08	2.08	2.08	2.08	2.08	2.07	2.07	2.07	2.07	2.07	2.25	2.10	24		
28		2.07	2.07	S	2.09	2.09	2.10	2.11	2.12	2.11	2.11	2.11	2.11	2.40	2.32	2.22	2.29	2.21	2.36	2.20	2.15	2.15	2.35	2.10	2.07	2.40	2.17	24	
29		2.08	S	2.13	2.11	2.12	2.12	2.12	2.10	2.09	2.09	2.10	2.10	2.12	2.12	2.11	2.11	2.11	2.11	2.07	2.07	2.06	2.05	2.05	2.13	2.10	24		
30		S	2.07	2.09	2.10	2.09	2.10	2.09	2.10	2.14	2.19	2.19	2.22	2.14	2.14	2.11	2.12	2.12	2.12	2.14	2.13	2.13	S	2.07	2.22	2.13	24		
31		2.13	2.14	2.13	2.14	2.14	2.13	2.13	2.14	2.13	2.12	2.12	2.11	2.10	2.09	2.08	2.06	2.05	2.05	2.04	2.04	S	2.03	2.03	2.14	2.10	24		
HOURLY MAX		2.19	2.25	2.14	2.21	2.30	2.23	2.15	2.13	2.14	2.19	2.21	2.22	2.24	2.32	2.22	2.29	2.21	2.36	2.30	2.21	2.40	2.18	2.35	2.30				
HOURLY AVG		2.05	2.05	2.04	2.05	2.05	2.04	2.04	2.05	2.05	2.05	2.05	2.05	2.05	2.05	2.04	2.04	2.04	2.05	2.04	2.05	2.03	2.04	2.05	2.05	2.05			




NUMBER OF NON-ZERO READINGS:	701
MINIMUM 1-HR AVERAGE	1.88 ppm @ HOUR
MAXIMUM 1-HR AVERAGE:	2.40 ppm @ HOUR
MAXIMUM 24-HR AVERAGE:	2.17 ppm
IZS CALIBRATION TIME:	32 hrs
MONTHLY CALIBRATION TIME:	6 hrs
OPERATIONAL TIME:	AMD OPERATION UPTIME:
STANDARD DEVIATION:	0.07
MONTHLY AVERAGE:	2.04 ppm



PEACE RIVER AREA MONITORING PROGRAM COMMITTEE


Three Creeks 986b Station - December 2018

TOTAL HYDROCARBONS Hourly Averages (THC ppm)




Peace River Area Monitoring Program Committee_Three Creeks 986b Station_THC (ppm)_18/12

Pollutant Rose_Wind Frequency (Blowing From)_CALM Avg = 2.1_CALM % = 12.0%



Direction	0-2	2-3	3-5	5-10	>10	TOTAL
N	0.6	5.4	0.0	0.0	0.0	6.0
NE	1.1	4.1	0.0	0.0	0.0	5.3
E	2.1	11.0	0.0	0.0	0.0	13.1
SE	1.7	8.7	0.0	0.0	0.0	10.4
S	8.7	15.4	0.0	0.0	0.0	24.1
SW	7.4	9.7	0.0	0.0	0.0	17.1
W	1.0	2.0	0.0	0.0	0.0	3.0
NW	0.4	8.6	0.0	0.0	0.0	9.0
Summary	23.1	64.9	0.0	0.0	0.0	88.0
CALM	1.0	11.0	0.0	0.0	0.0	12.0

THC55[ppm] Calibration: PRAMP_986 Monthly: 18/12 Type: Span



METHANE


METHANE Hourly Averages (CH₄ ppm)

	HR START (MST)	0:00	1:00	2:00	3:00	4:00	5:00	6:00	7:00	8:00	9:00	10:00	11:00	12:00	13:00	14:00	15:00	16:00	17:00	18:00	19:00	20:00	21:00	22:00	23:00	DAILY MIN.	DAILY MAX.	24-HR AVG.	RDGS.		
	HR END (MST)	0:59	1:59	2:59	3:59	4:59	5:59	6:59	7:59	8:59	9:59	10:59	11:59	12:59	13:59	14:59	15:59	16:59	17:59	18:59	19:59	20:59	21:59	22:59	23:59						
DAY																															
1		2.07	1.97	1.98	2.01	2.04	2.05	S	2.03	2.05	2.04	2.03	2.02	2.01	2.01	2.02	2.02	2.09	2.05	2.02	2.02	2.02	2.04	2.03	1.97	2.09	2.03	24			
2		2.03	2.01	2.02	2.02	2.07	S	2.01	2.01	2.01	2.01	2.01	2.10	2.17	2.09	2.04	1.98	1.99	1.97	1.97	1.97	1.98	1.99	2.00	2.01	2.02	2.02	1.96	2.07	2.00	24
3		2.03	2.03	2.02	2.02	S	2.01	2.01	2.01	2.01	2.01	2.10	2.17	2.09	2.04	1.98	1.99	1.98	1.97	1.97	1.97	1.97	1.96	1.96	1.96	1.96	2.17	2.01	24		
4		1.95	1.95	1.96	S	1.98	1.98	1.98	1.96	S1	1.97	1.97	1.99	2.00	1.99	1.98	1.98	1.98	1.98	1.99	1.99	2.00	2.01	2.02	2.04	2.06	1.95	2.06	1.99	23	
5		2.05	2.13	S	1.99	1.99	1.99	2.03	2.00	2.06	2.01	1.99	1.98	1.99	1.99	1.98	1.98	1.98	1.97	1.97	1.97	1.99	2.04	1.97	1.98	1.97	2.13	2.00	24		
6		1.98	S	2.00	2.00	1.99	2.00	1.99	2.00	2.00	2.01	2.02	2.01	2.01	2.00	2.00	2.00	2.01	1.99	1.99	2.01	2.02	2.02	2.02	1.98	2.02	2.00	24			
7		S	2.04	2.04	2.05	2.07	2.07	2.06	2.04	2.03	2.02	2.02	2.03	2.01	2.01	2.02	2.03	2.02	2.02	2.02	2.01	2.01	1.99	2.00	S	1.99	2.07	2.03	24		
8		2.02	2.05	2.03	2.05	2.03	2.06	2.05	2.04	2.04	2.02	2.03	2.04	2.06	2.05	2.05	2.05	2.05	2.05	2.04	2.01	2.00	2.00	S	2.02	2.06	2.03	24			
9		2.01	1.99	2.01	2.04	2.04	2.03	2.04	2.04	2.03	2.03	2.01	2.01	2.01	2.00	2.01	2.02	2.05	2.00	2.00	2.30	S	2.01	2.03	1.99	2.30	2.03	24			
10		2.04	2.05	2.03	2.04	2.05	2.04	2.01	2.01	2.00	1.98	1.98	1.97	1.96	1.96	1.96	1.97	1.97	1.95	1.94	S	1.95	1.95	1.96	1.94	2.05	1.99	24			
11		1.96	1.96	1.95	1.95	1.95	1.94	1.96	1.95	1.94	1.94	1.93	1.90	1.90	1.90	1.90	1.90	1.90	1.89	S	1.88	1.88	1.88	1.88	1.88	1.96	1.92	24			
12		1.88	1.89	1.89	1.89	1.90	1.89	1.89	1.89	C	C	Y	Y	Y	C	C	C	C	1.98	1.99	1.99	1.98	1.98	1.98	1.98	1.99	1.92	20			
13		1.98	1.98	1.98	1.97	1.97	1.98	1.97	1.97	1.97	1.98	1.98	2.00	2.01	1.99	1.99	S	1.99	1.99	1.98	1.98	1.98	1.98	1.98	1.97	2.01	1.98	24			
14		1.98	1.98	1.98	1.98	1.98	1.98	1.97	1.97	1.99	1.99	1.98	1.97	1.97	1.96	1.97	S	1.97	1.98	1.99	2.18	2.00	2.01	2.05	1.96	2.18	1.99	24			
15		2.05	2.04	2.06	2.05	2.04	2.04	2.06	2.03	2.03	2.02	2.03	2.04	2.14	2.10	2.13	S	2.02	2.01	2.03	2.07	2.11	2.15	2.19	2.01	2.19	2.06	24			
16		2.15	2.10	2.07	2.21	2.07	2.02	2.02	2.00	2.01	2.01	2.01	2.00	S	2.00	2.00	2.00	2.00	2.00	2.01	2.01	2.01	2.01	2.02	2.00	2.21	2.03	24			
17		2.02	2.03	2.04	2.04	2.05	2.09	2.08	2.04	2.04	2.05	2.06	S	2.08	2.08	2.09	2.09	2.08	2.08	2.06	2.09	2.10	2.08	2.07	2.02	2.10	2.06	24			
18		2.08	2.10	2.09	2.07	2.13	2.04	2.09	2.05	2.03	2.03	2.02	S	2.00	2.01	2.02	2.02	2.03	2.05	2.05	2.06	2.07	2.06	2.06	2.06	2.00	2.13	2.06	24		
19		2.06	2.07	2.06	2.06	2.06	2.08	2.08	2.05	2.03	2.03	2.03	S	2.01	2.00	2.00	2.00	1.99	1.99	1.99	1.99	1.99	1.99	2.00	2.00	1.99	2.08	2.03	24		
20		2.00	2.00	2.01	2.00	2.01	2.02	2.02	2.06	2.04	2.03	S	2.02	1.99	1.98	1.99	1.99	1.98	1.97	1.98	1.98	1.98	1.99	1.97	2.06	2.00	24				
21		2.03	2.05	2.07	2.04	2.03	2.02	2.07	2.02	2.02	S	2.05	2.20	2.06	2.08	2.09	2.09	2.06	2.33	2.30	2.21	2.08	2.02	2.30	2.02	2.33	2.10	24			
22		2.05	2.02	2.01	2.01	2.06	2.09	2.12	S	2.09	2.09	2.10	2.05	2.04	2.04	2.04	2.04	2.05	2.06	2.06	2.07	2.12	2.11	2.01	2.12	2.06	24				
23		2.14	2.14	2.14	2.19	2.30	2.23	2.05	S	2.07	2.06	2.07	2.06	2.07	2.05	2.05	2.04	2.04	2.03	2.03	2.03	2.03	2.02	2.02	2.30	2.08	24				
24		2.02	2.03	2.03	2.04	2.04	2.04	S	2.05	2.05	2.06	2.07	2.08	2.09	2.11	2.13	2.14	2.14	2.16	2.18	2.19	2.19	2.17	2.14	2.02	2.19	2.10	24			
25		2.13	2.11	2.07	2.08	2.04	S	2.05	2.06	2.10	2.12	2.21	2.11	2.08	2.12	2.18	2.15	2.15	2.06	2.05	2.05	2.06	2.17	2.12	2.17	2.04	2.21	2.10	24		
26		2.19	2.16	2.13	2.12	S	2.13	2.10	2.14	2.19	2.10	2.07	2.06	2.05	2.05	2.05	2.06	2.07	2.17	2.08	2.10	2.09	2.10	2.05	2.19	2.10	24				
27		2.14	2.25	2.09	S	2.14	2.12	2.15	2.13	2.09	2.10	2.09	2.08	2.08	2.08	2.08	2.08	2.08	2.08	2.07	2.07	2.07	2.07	2.07	2.25	2.10	24				
28		2.07	2.07	S	2.09	2.09	2.10	2.11	2.12	2.11	2.11	2.11	S	2.40	2.32	2.22	2.29	2.21	2.36	2.20	2.15	2.15	2.11	2.35	2.10	2.07	S	2.40	2.17	24	
29		2.08	S	2.13	2.11	2.12	2.12	2.12	2.10	2.09	2.09	2.10	2.10	2.10	2.12	2.12	2.11	2.11	2.11	2.11	2.07	2.07	2.06	2.05	2.05	2.13	2.10	24			
30		S	2.07	2.09	2.10	2.09	2.10	2.09	2.10	2.14	2.19	2.19	2.22	2.14	2.14	2.11	2.12	2.12	2.12	2.12	2.14	2.13	2.13	S	2.07	2.22	2.13	24			
31		2.13	2.14	2.13	2.14	2.14	2.13	2.13	2.14	2.13	2.12	2.12	2.12	2.12	2.11	2.10	2.09	2.08	2.06	2.05	2.05	2.04	S	2.03	2.14	2.10	24				
HOURLY MAX		2.19	2.25	2.14	2.21	2.30	2.23	2.15	2.14	2.19	2.21	2.22	2.40	2.32	2.22	2.29	2.21	2.36	2.30	2.21	2.30	2.18	2.35	2.30							
HOURLY AVG		2.05	2.05	2.04	2.05	2.05	2.04	2.04	2.05	2.05	2.05	2.05	2.05	2.05	2.04	2.04	2.04	2.05	2.04	2.04	2.05	2.03	2.04	2.05							

STATUS FLAG CODES

C	- MONTHLY CALIBRATION	Q	- QUALITY ASSURANCE
C1	- REPEAT CALIBRATION	R	- RECOVERY
Y	- MAINTENANCE	X	- MACHINE MALFUNCTION
S	- DAILY ZERO/SPAN CHECK	G	- OUT FOR REPAIR
S1	- REPEAT ZERO/SPAN CHECK	P	- POWER FAILURE


24 HR AVERAGES December 2018



MONTHLY SUMMARY


NUMBER OF NON-ZERO READINGS:	701		
</tr

METHANE Hourly Averages (CH₄ ppm)




Peace River Area Monitoring Program Committee_Three Creeks 986b Station_CH4 (ppm)_18/12

Pollutant Rose_Wind Frequency (Blowing From)_CALM Avg = 2.1_CALM % = 12.0%



Direction	0-2	2-3	3-5	5-10	>10	TOTAL
N	0.6	5.4	0.0	0.0	0.0	6.0
NE	1.1	4.1	0.0	0.0	0.0	5.3
E	2.1	11.0	0.0	0.0	0.0	13.1
SE	1.7	8.7	0.0	0.0	0.0	10.4
S	8.7	15.4	0.0	0.0	0.0	24.1
SW	7.4	9.7	0.0	0.0	0.0	17.1
W	1.0	2.0	0.0	0.0	0.0	3.0
NW	0.4	8.6	0.0	0.0	0.0	9.0
Summary	23.1	64.9	0.0	0.0	0.0	88.0
CALM	1.0	11.0	0.0	0.0	0.0	12.0

CH4[ppm] Calibration: PRAMP_986 Monthly: 18/12 Type: Span



NON-METHANE HYDROCARBON




PEACE RIVER AREA MONITORING PROGRAM COMMITTEE

Three Creeks 986b Station - December 2018

NON-METHANE HYDROCARBONS Hourly Averages (NMHC ppm)

	HR START (MST)	0:00	1:00	2:00	3:00	4:00	5:00	6:00	7:00	8:00	9:00	10:00	11:00	12:00	13:00	14:00	15:00	16:00	17:00	18:00	19:00	20:00	21:00	22:00	23:00	DAILY MIN.	DAILY MAX.	24-HR AVG.	RDGS.
	HR END (MST)	0:59	1:59	2:59	3:59	4:59	5:59	6:59	7:59	8:59	9:59	10:59	11:59	12:59	13:59	14:59	15:59	16:59	17:59	18:59	19:59	20:59	21:59	22:59	23:59				
DAY																													
1		0.00	0.00	0.00	0.00	0.00	0.00	S	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	24
2		0.00	0.00	0.00	0.00	0.00	S	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	24
3		0.00	0.00	0.00	0.00	S	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	24
4		0.00	0.00	0.00	S	0.00	0.00	0.00	0.00	0.00	S1	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	23	
5		0.00	0.00	S	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	24
6		S	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	24	
7		S	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	24	
8		0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	24	
9		0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.10	S	0.00	0.00	0.00	0.10	0.00	0.00	24	
10		0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	S	0.00	0.00	0.00	0.00	0.00	0.00	0.00	24
11		0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	S	0.00	0.00	0.00	0.00	0.00	0.00	0.00	24
12		0.00	0.00	0.00	0.00	0.00	0.00	0.00	C	C	C	Y	Y	Y	C	C	C	C	C	C	C	0.00	0.00	0.00	0.00	0.00	0.00	0.00	20
13		0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	S	0.00	0.00	0.00	0.00	0.00	0.00	0.00	24		
14		0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	S	0.00	0.00	0.00	0.00	0.00	0.00	0.00	24		
15		0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	S	0.00	0.00	0.00	0.00	0.00	0.00	0.00	24		
16		0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	S	0.00	0.00	0.00	0.00	0.00	0.00	0.00	24		
17		0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	S	0.00	0.00	0.00	0.00	0.00	0.00	0.00	24		
18		0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	S	0.00	0.00	0.00	0.00	0.00	0.00	0.00	24		
19		0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	S	0.00	0.00	0.00	0.00	0.00	0.00	0.00	24		
20		0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	S	0.00	0.00	0.00	0.00	0.00	0.00	0.00	24		
21		0.00	0.00	0.00	0.00	0.00	0.00	0.00	S	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	24		
22		0.00	0.00	0.00	0.00	0.00	0.00	0.00	S	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	24		
23		0.00	0.00	0.00	0.00	0.00	0.00	0.00	S	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	24		
24		0.00	0.00	0.00	0.00	0.00	0.00	0.00	S	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	24		
25		0.00	0.00	0.00	0.00	0.00	0.00	S	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	24		
26		0.00	0.00	0.00	0.00	S	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	24		
27		0.00	0.00	S	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	24		
28		0.00	S	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	24		
29		S	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	24	
30		S	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	S	0.00	0.00	0.00	0.00	0.00	0.00	0.00	24
31		0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	S	0.00	0.00	0.00	0.00	0.00	0.00	0.00	24
HOURLY MAX		0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.10	0.00	0.00	0.00	0.00	0.00	0.00	0.00		
HOURLY AVG		0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00		




NUMBER OF NON-ZERO READINGS:	1
MINIMUM 1-HR AVERAGE	0.00 ppm @ HOUR
MAXIMUM 1-HR AVERAGE:	0.10 ppm @ HOUR
MAXIMUM 24-HR AVERAGE:	0.00 ppm
I2S CALIBRATION TIME:	32 hrs
MONTHLY CALIBRATION TIME:	6 hrs
AMD OPERATION UPTIME:	99.3 %
STANDARD DEVIATION:	0.00 ppm
MONTHLY AVERAGE:	0.00 ppm



PEACE RIVER AREA MONITORING PROGRAM COMMITTEE


Three Creeks 986b Station - December 2018

NON-METHANE HYDROCARBONS Hourly Averages (NMHC ppm)




Peace River Area Monitoring Program Committee_Three Creeks 986b Station_NMHC (ppm)_18/12

Pollutant Rose_Wind Frequency (Blowing From)_CALM Avg = 0.0_CALM % = 12.0%



Direction	0-0.1	0.1-0.3	0.3-1	1-2	2-4	>4	TOTAL
N	6.0	0.0	0.0	0.0	0.0	0.0	6.0
NE	5.3	0.0	0.0	0.0	0.0	0.0	5.3
E	13.1	0.0	0.0	0.0	0.0	0.0	13.1
SE	10.4	0.0	0.0	0.0	0.0	0.0	10.4
S	24.1	0.0	0.0	0.0	0.0	0.0	24.1
SW	17.1	0.0	0.0	0.0	0.0	0.0	17.1
W	3.0	0.0	0.0	0.0	0.0	0.0	3.0
NW	9.0	0.0	0.0	0.0	0.0	0.0	9.0
Summary	88.0	0.0	0.0	0.0	0.0	0.0	88.0
CALM	11.8	0.1	0.0	0.0	0.0	0.0	12.0


NMHC[ppm] Calibration: PRAMP_986 Monthly: 18/12 Type: Span



WIND SPEED

WIND SPEED Hourly Averages (WS kph)

	HR START (MST) HR END (MST)	0:00 0:59	1:00 1:59	2:00 2:59	3:00 3:59	4:00 4:59	5:00 5:59	6:00 6:59	7:00 7:59	8:00 8:59	9:00 9:59	10:00 10:59	11:00 11:59	12:00 12:59	13:00 13:59	14:00 14:59	15:00 15:59	16:00 16:59	17:00 17:59	18:00 18:59	19:00 19:59	20:00 20:59	21:00 21:59	22:00 22:59	23:00 23:59	DAILY MIN.	DAILY MAX.	24-HR AVG.	RDGS.	
DAY																														
1		0.7	1.8	1.8	1.7	1.6	2.0	2.7	2.0	3.3	3.1	3.0	2.9	3.0	4.2	4.4	3.2	2.5	1.8	2.4	2.9	4.3	0.8	0.7	0.6	0.6	4.4	1.5	24	
2		1.9	1.7	1.9	2.1	1.0	3.1	3.6	3.4	2.1	1.7	2.9	4.6	5.5	4.5	5.4	4.4	4.7	4.7	4.4	5.7	6.3	5.7	5.5	4.8	1.0	6.3	3.5	24	
3		4.1	4.4	3.6	3.5	3.1	3.6	1.9	1.1	1.3	0.8	0.7	0.4	3.9	0.6	0.1	0.1	0.1	0.1	0.1	4.2	5.9	5.4	6.1	6.8	0.1	6.8	2.0	24	
4		3.5	3.9	5.6	4.2	3.5	1.8	4.8	2.5	4.5	5.5	6.4	7.3	5.1	4.9	12.6	15.0	10.0	11.5	11.0	10.0	7.5	10.4	10.2	8.5	1.8	15.0	4.8	24	
5		6.3	1.9	1.8	1.7	1.1	2.0	1.2	1.9	2.3	1.9	2.6	3.8	4.3	5.5	8.4	5.6	6.5	8.0	7.4	6.8	5.0	2.8	4.8	5.6	1.1	8.4	3.0	24	
6		4.6	5.0	5.0	3.6	4.3	4.0	3.6	3.9	3.6	3.9	4.2	6.0	7.5	7.5	8.0	10.1	8.5	8.5	6.7	7.4	6.9	5.3	6.6	4.8	3.6	10.1	5.7	24	
7		8.6	10.7	9.9	8.1	7.2	7.1	6.1	4.7	4.5	5.2	5.7	6.3	6.9	6.5	6.9	6.6	6.5	5.5	5.2	4.7	5.2	5.9	5.8	7.3	4.5	10.7	5.9	24	
8		5.8	4.7	6.2	6.6	7.1	7.0	9.3	8.8	9.0	11.2	10.8	11.6	12.1	11.8	11.3	9.0	8.3	5.9	7.4	6.1	7.6	6.6	8.2	10.6	4.7	12.1	8.0	24	
9		6.9	6.7	6.9	7.4	7.2	8.6	10.0	10.3	9.8	7.0	5.9	4.3	7.6	5.9	5.8	4.7	4.3	1.3	2.3	2.0	1.7	3.9	5.2	6.0	1.3	10.3	4.9	24	
10		6.0	5.7	7.5	8.6	7.0	8.2	9.5	12.1	8.6	6.0	8.6	10.2	10.7	10.4	10.8	7.4	6.9	6.1	7.5	9.2	10.3	7.6	8.6	8.0	5.7	12.1	8.1	24	
11		6.6	6.9	5.9	10.7	14.9	12.0	12.5	13.3	11.1	13.1	13.6	15.9	10.7	15.8	18.2	15.9	16.4	12.7	12.9	11.7	12.3	15.4	13.1	11.6	5.9	18.2	11.7	24	
12		10.2	10.4	12.0	12.4	13.5	11.1	8.4	9.4	9.4	9.2	10.5	11.8	14.2	13.3	14.7	12.6	12.2	10.3	8.1	7.4	9.6	7.2	8.5	10.1	7.2	14.7	10.2	24	
13		12.2	14.8	14.3	16.8	15.5	12.2	8.8	7.3	10.0	10.7	11.3	11.7	11.7	10.6	13.4	12.6	12.3	12.3	9.7	6.9	5.3	7.1	9.3	9.6	5.3	16.8	9.7	24	
14		5.3	6.8	10.5	6.8	6.5	6.5	7.7	4.0	3.2	3.3	4.9	6.4	6.0	6.3	4.5	3.2	3.9	4.4	5.1	4.8	2.4	4.2	4.4	4.7	2.4	10.5	1.9	24	
15		5.4	5.6	7.7	14.2	5.1	10.0	12.6	9.2	10.5	15.4	16.6	14.3	16.6	13.4	6.6	3.1	2.0	2.3	3.5	3.9	2.5	3.0	2.6	5.3	2.0	16.6	5.7	24	
16		4.7	4.5	2.8	1.4	1.4	5.1	9.4	13.7	13.0	10.0	9.6	9.4	7.6	7.4	6.1	5.8	7.2	6.6	6.2	6.9	4.0	4.0	5.6	4.9	1.4	13.7	6.4	24	
17		6.7	4.0	5.3	4.8	7.5	6.6	2.9	5.1	6.2	9.6	11.1	11.7	10.5	10.1	9.0	8.2	7.7	6.1	4.7	6.0	6.5	7.1	2.0	5.2	2.0	11.7	5.8	24	
18		7.7	4.7	6.6	5.9	5.7	3.8	6.0	1.2	1.7	4.0	5.3	6.8	5.8	5.6	8.2	6.9	6.1	4.2	6.2	5.1	5.7	6.5	6.6	7.7	1.2	8.2	4.4	24	
19		7.5	5.0	7.1	8.5	8.2	7.2	5.1	6.0	9.0	10.0	9.4	8.7	10.7	10.0	10.3	8.4	9.3	9.4	8.0	7.6	6.4	7.1	8.1	5.8	5.0	10.7	8.0	24	
20		5.1	6.1	4.5	5.1	3.5	3.6	6.6	4.9	5.8	5.9	4.5	5.3	4.9	7.8	7.3	8.3	10.9	9.6	8.8	7.9	5.2	5.4	3.2	3.1	3.1	10.9	4.2	24	
21		4.9	6.3	8.0	11.8	13.4	13.6	14.3	15.8	12.7	10.1	10.4	10.8	15.9	12.5	13.2	7.1	4.5	4.7	4.1	4.7	2.7	1.8	0.6	1.1	0.6	15.9	8.3	24	
22		1.5	2.1	1.2	1.9	1.9	2.2	1.6	1.5	1.6	2.4	1.3	1.0	1.7	2.3	2.2	3.0	2.5	1.9	0.9	1.4	1.0	1.3	0.7	0.6	0.6	3.0	1.3	24	
23		0.9	1.0	0.7	0.1	0.5	1.5	2.2	2.2	2.1	2.2	1.7	2.0	2.4	2.5	1.9	2.0	1.8	2.0	1.8	1.4	2.8	1.8	1.3	1.5	0.1	2.8	1.4	24	
24		1.3	1.6	2.4	3.2	2.3	2.6	1.8	2.1	2.5	2.9	4.1	3.3	3.1	3.3	3.4	2.2	3.4	3.9	2.3	1.4	2.8	1.9	2.8	1.3	4.1	2.4	24		
25		4.0	4.4	4.5	6.1	5.5	5.9	4.1	3.3	2.4	3.3	4.3	1.8	3.1	6.2	6.7	3.1	1.7	1.8	1.9	1.5	2.3	2.0	2.1	1.5	6.7	2.4	24		
26		1.6	1.5	2.2	3.5	3.9	3.5	3.3	2.7	1.5	1.8	2.5	3.8	2.3	3.6	4.9	3.9	4.1	2.9	2.4	1.9	2.1	5.7	5.7	2.6	1.5	5.7	2.7	24	
27		1.2	2.1	1.7	2.1	1.0	0.7	1.3	3.1	0.5	2.4	0.9	1.0	2.4	2.5	2.7	2.0	1.9	1.6	1.9	1.9	2.0	1.8	1.9	2.6	0.5	3.1	1.5	24	
28		2.0	2.3	2.0	1.1	1.1	3.1	5.6	6.1	5.2	4.7	4.3	3.5	0.3	1.8	3.1	2.0	2.1	1.6	2.4	1.3	0.7	0.5	0.3	1.0	0.3	6.1	1.2	24	
29		2.9	3.2	1.8	1.1	1.4	0.6	3.0	3.5	4.6	4.8	7.9	7.0	6.5	6.8	8.0	9.1	9.2	7.9	7.3	6.3	6.9	9.0	9.4	8.2	0.6	9.4	4.3	24	
30		8.1	7.5	8.1	7.6	7.2	7.6	6.1	4.7	5.5	4.5	4.1	3.9	4.9	4.7	1.3	2.8	5.8	6.1	6.7	7.3	6.5	8.3	5.9	1.3	8.3	1.9	24		
31		4.3	4.7	6.2	5.5	6.9	7.6	8.2	7.3	10.2	6.4	8.4	10.0	9.4	10.5	10.3	10.0	9.4	9.0	9.6	12.5	12.8	14.9	16.3	11.8	4.3	16.3	8.7	24	
	HOURLY MAX	12.2	14.8	14.3	16.8	15.5	13.6	14.3	15.8	13.0	15.4	16.6	15.9	16.6	15.8	18.2	18.2	15.9	16.4	12.7	12.9	12.5	12.8	15.4	16.3	11.8				


24 HR AVERAGES December 2018

MONTHLY SUMMARY

NUMBER OF NON-ZERO READINGS:	744
MINIMUM 1-HR AVERAGE	0.1 kph @ HOUR
MAXIMUM 1-HR AVERAGE:	18.2 kph @ HOUR
MAXIMUM 24-HR AVERAGE:	11.7 kph
OPERATIONAL TIME:	744 hrs
MONTHLY CALIBRATION TIME:	0 hrs
AMD OPERATION UPTIME:	100.0 %
STANDARD DEVIATION:	3.7
MONTHLY AVERAGE:	2.4 kph




PEACE RIVER AREA MONITORING PROGRAM COMMITTEE
Three Creeks 986b Station - December 2018

WIND SPEED Hourly Averages (WS kph)



Peace River Area Monitoring Program Committee_Three Creeks 986b Station_18/12

Wind Rose_Wind Frequency (Blowing From)_CALM Avg = 1.1_CALM % = 11.4%



Direction	1.8-6	6-12	12-20	20-29	29-39	>39	TOTAL
N	4.7	1.3	0.0	0.0	0.0	0.0	6.0
NE	3.6	1.3	0.0	0.0	0.0	0.0	5.0
E	10.9	2.7	0.0	0.0	0.0	0.0	13.6
SE	6.2	3.0	0.7	0.0	0.0	0.0	9.8
S	7.3	13.6	3.0	0.0	0.0	0.0	23.8
SW	6.3	10.4	1.1	0.0	0.0	0.0	17.8
W	1.8	1.1	1.2	0.0	0.0	0.0	4.0
NW	3.5	3.2	1.9	0.0	0.0	0.0	8.6
Summary	44.2	36.6	7.8	0.0	0.0	0.0	88.6
CALM	11.4	0.0	0.0	0.0	0.0	0.0	11.4

WIND DIRECTION



PEACE RIVER AREA MONITORING PROGRAM COMMITTEE

Three Creeks 986b Station - December 2018

WIND DIRECTION Hourly Averages (WD)

HR START (MST)	0:00	1:00	2:00	3:00	4:00	5:00	6:00	7:00	8:00	9:00	10:00	11:00	12:00	13:00	14:00	15:00	16:00	17:00	18:00	19:00	20:00	21:00	22:00	23:00	24-HOUR AVG QUADRANT	24-HR RDGS.	
HR END (MST)	0:59	1:59	2:59	3:59	4:59	5:59	6:59	7:59	8:59	9:59	10:59	11:59	12:59	13:59	14:59	15:59	16:59	17:59	18:59	19:59	20:59	21:59	22:59	23:59			
DAY																											
1	E	ESE	E	E	ESE	ENE	E	ESE	ESE	ESE	ESE	S	S	SSW	SSW	SSW	WNW	WSW	SSW	S	S	SW	SW	SSE	24		
2	W	SW	SW	WSW	W	WNW	WSW	SW	WSW	SW	WSW	S	S	SSW	24												
3	SSW	SSW	SSW	SW	SW	SSW	SSW	S	SE	SE	WNW	WNW	W	WSW	S	SE	ESE	SE	SSE	SSE	S	SSE	S	S	24		
4	S	SSW	SSW	SW	SW	SSW	S	SSW	SW	WSW	W	W	WSW	WSW	NWW	NNW	NNW	NNW	NNW	NW	NW	NW	NW	WNW	24		
5	WNW	WSW	SSW	S	SSW	SW	WSW	SW	W	SSW	SSW	WSW	SSW	S	S	SSE	SSE	SE	SE	SSE	S	S	S	S	24		
6	SSW	S	SSE	S	S	S	S	SSE	SE	S	S	S	S	S	SSE	SSE	SSE	SSE	SSE	S	S	S	S	S	24		
7	SSW	SSW	SSW	SW	SSW	SSW	SSW	S	S	SE	S	SSW	SSW	SSW	S	SSE	SSE	SSE	SSE	SSE	S	S	S	S	24		
8	S	SE	SE	S	S	SSW	SSW	S	S	SSW	SSW	SSW	SSW	S	S	SSW	24										
9	SW	SW	SSW	SE	SSE	ESE	SSE	SE	SE	ESE	ESE	ESE	SE	W	SSW	SSW	SSW	SSE	24								
10	SSW	SSW	S	SSE	SSW	SSW	S	S	SSE	SSW	SSW	SSW	SSW	S	S	SSW	24										
11	S	SSE	SE	SSE	SSE	SSE	SE	SE	SSE	SSE	SSE	S	S	SSE	SSE	S	SSE	SSE	SSE	S	SSW	SSW	SSW	S	24		
12	SW	SSW	SW	SW	SW	WSW	SW	24																			
13	S	SSE	SSE	S	SSW	SSW	SSW	SSW	SSW	SSW	SSW	SSW	SSW	SSW	SSW	SSW	SSW	SSW	24								
14	SSW	SSW	SSW	SSW	S	SSW	S	SSE	E	ENE	E	E	E	E	E	E	E	E	E	ESE	N	NNW	NNW	SE	24		
15	NNW	NNW	NW	NW	WNW	NW	NW	WNW	WNW	WNW	WNW	NNW	WNW	24													
16	E	ESE	ESE	NE	ESE	ESE	ESE	ESE	ESE	ESE	E	E	E	E	E	E	E	E	E	E	E	E	E	E	24		
17	ESE	E	SSE	ESE	ESE	SSE	SSW	SW	SSW	S	SSW	SSW	SSW	SSW	S	SSW	SSW	SSW	SSW	SSW	SE	SE	ESE	SE	S	24	
18	SSE	ESE	SE	ESE	ESE	ESE	S	SW	SE	ESE	SE	SSE	SSW	SW	SSW	S	S	SSW	SSE	S	SSW	SSW	SW	S	24		
19	SW	SW	SW	SW	SSW	SSW	SSW	SW	SW	SW	SW	SW	SW	SW	SW	SW	SW	SW	SW	SW	SW	SW	SW	SW	24		
20	SW	SW	S	SSE	E	E	ESE	ESE	ESE	ESE	E	E	E	E	E	E	E	E	E	E	E	E	NE	N	24		
21	NNW	NW	NW	WNW	NW	NW	NW	NW	NNW	NNW	NNW	NNW	NNW	NNW	NNW	NNW	NNW	NNW	NNW	NNW	NNW	NNW	NNW	NNW	24		
22	WSW	S	SSE	ESE	SE	ESE	ESE	E	SE	ESE	ESE	SSE	SSE	SSE	SSE	SSE	SSE	ESE	ESE	ESE	ESE	ESE	ESE	SE	24		
23	E	ESE	SE	N	NNW	NNW	N	N	N	NNW	N	N	N	N	N	N	N	N	N	N	N	N	NNE	NNE	24		
24	E	ENE	E	ESE	E	E	E	E	ENE	E	E	E	E	E	E	E	E	E	E	E	E	E	E	ENE	24		
25	N	N	NNW	NNW	N	N	N	NNW	NNW	NNW	NNW	NNW	NNW	NNW	NNW	NNW	NNW	NNW	NNW	NNW	NNW	NNW	NNW	NNW	24		
26	E	E	E	ESE	E	E	E	ESE	E	E	E	E	E	E	E	E	E	E	E	E	E	E	E	ENE	24		
27	ENE	NNE	E	ESE	NNW	NE	NW	NNW	NE	NW	E	NNE	24														
28	NE	E	ESE	ENE	E	ESE	ESE	ESE	ESE	ESE	ESE	NNE	WNW	NNE	NW	S	E	E	24								
29	ESE	ESE	ENE	ENE	ESE	E	E	ENE	NE	ESE	E	E	ENE	ENE	ENE	NE	NE	NE	NNE	NNE	NNW	NNW	NNW	NE	24		
30	NNW	NNW	NNW	NNW	NNW	NNW	NNW	NNW	NNW	NNW	NNW	NNW	NNW	NNW	WNW	24											
31	SE	SSE	SE	SSE	SSE	SSE	SE	SSE	SSE	SSE	SSE	SSE	SSE	SSE	SSE	SSE	SSE	SSE	SSE	SSE	SSE	SSE	SSE	S	24		


STATUS FLAG CODES

C	- MONTHLY CALIBRATION	Q	- QUALITY ASSURANCE
C1	- REPEAT CALIBRATION	R	- RECOVERY
Y	- MAINTENANCE	X	- MACHINE MALFUNCTION
S	- DAILY ZERO/SPAN CHECK	G	- OUT FOR REPAIR
S1	- REPEAT ZERO/SPAN CHECK	P	- POWER FAILURE

LAST CALIBRATION: April 4, 2018
 DECLINATION : MAGNETIC DECLINATION 15 DEGREE EAST

MONTHLY CALIBRATION TIME:	0 hrs	OPERATIONAL TIME:	744 hrs
STANDARD DEVIATION:	85	AMD OPERATION UPTIME:	100.0 %
		MONTHLY AVERAGE:	192 (\$)

WIND DIRECTION Hourly Averages (WD)



WDR[degwdr] Station: PRAMP_986 Monthly: 18/12 Type: AVG 1 Hr. [1 Hr.]



RELATIVE HUMIDITY


RELATIVE HUMIDITY Hourly Averages (RH %)

	HR START (MST)	0:00	1:00	2:00	3:00	4:00	5:00	6:00	7:00	8:00	9:00	10:00	11:00	12:00	13:00	14:00	15:00	16:00	17:00	18:00	19:00	20:00	21:00	22:00	23:00	DAILY MIN.	DAILY MAX.	24-HR AVG.	RDGS.
	HR END (MST)	0:59	1:59	2:59	3:59	4:59	5:59	6:59	7:59	8:59	9:59	10:59	11:59	12:59	13:59	14:59	15:59	16:59	17:59	18:59	19:59	20:59	21:59	22:59	23:59				
DAY																													
1		94	94	94	93	93	93	93	92	93	93	92	91	92	95	96	98	99	99	99	99	98	98	98	91	99	95	24	
2		97	97	96	92	90	93	96	96	95	95	94	95	96	96	96	96	95	94	93	93	93	92	90	97	95	24		
3		92	91	90	91	92	90	88	87	86	84	85	88	89	91	91	89	87	84	83	85	86	88	90	90	83	92	88	24
4		92	93	93	93	93	93	94	95	96	96	94	92	91	90	88	88	92	91	88	88	89	86	87	88	86	96	91	24
5		88	88	89	89	90	90	88	88	89	89	88	85	82	82	84	90	91	91	89	89	89	91	90	82	91	88	24	
6		90	89	87	86	86	84	84	83	82	81	84	87	86	83	80	79	81	76	76	79	80	77	75	75	90	82	24	
7		74	72	69	66	72	74	73	74	74	76	76	71	63	61	63	65	71	77	78	77	79	78	79	69	61	72	24	
8		67	74	71	68	66	69	62	58	61	57	54	50	52	56	55	57	63	68	68	67	69	70	70	50	74	63	24	
9		70	69	71	75	77	78	80	82	81	80	74	68	66	63	65	70	76	80	84	84	85	78	76	75	63	85	75	24
10		73	72	70	73	70	71	69	73	72	71	69	64	60	59	60	62	67	69	66	66	68	69	68	59	73	68	24	
11		69	77	77	69	68	68	63	61	62	55	54	51	54	60	61	63	64	65	62	65	66	63	64	64	51	77	64	24
12		62	63	63	64	62	61	64	64	63	60	57	52	50	50	49	48	48	52	57	61	63	63	61	48	64	58	24	
13		60	86	90	68	59	68	69	74	75	75	72	68	64	59	55	53	53	52	53	55	56	56	55	56	52	90	64	24
14		56	58	57	57	60	59	67	84	96	98	95	92	89	82	79	81	85	84	83	84	87	91	93	93	56	98	80	24
15		92	91	90	88	87	86	83	82	75	67	67	66	65	63	63	64	72	78	79	77	76	76	78	63	92	77	24	
16		80	79	80	81	83	83	76	64	58	58	63	64	60	62	65	66	69	69	70	68	68	72	72	58	83	70	24	
17		72	71	73	77	79	79	82	86	89	91	91	95	98	89	82	83	85	87	91	89	86	86	90	92	71	98	85	24
18		84	86	86	86	84	80	73	78	81	86	86	83	72	65	67	68	72	77	76	80	83	79	78	77	65	86	79	24
19		77	79	82	82	83	84	86	84	80	79	75	71	69	68	67	69	69	68	67	66	67	70	71	66	86	74	24	
20		73	73	78	81	85	89	90	90	89	86	82	75	72	70	74	78	85	96	97	95	95	96	97	70	97	85	24	
21		97	98	99	100	100	99	97	96	94	94	93	93	92	91	89	90	91	91	92	92	91	90	90	89	100	94	24	
22		90	86	83	82	80	79	79	77	78	82	83	85	86	87	87	86	86	84	81	78	76	75	74	72	72	90	81	24
23		72	73	75	76	75	75	76	76	76	76	75	75	76	77	78	79	79	79	80	81	80	80	80	72	81	77	24	
24		80	80	80	80	80	81	82	82	82	82	81	81	83	83	82	83	83	81	80	79	81	81	82	79	83	81	24	
25		83	84	86	84	82	81	80	79	79	78	78	77	77	77	77	76	73	71	70	68	68	69	68	86	77	24		
26		69	67	67	67	69	70	71	71	68	68	71	74	75	74	75	75	74	74	74	74	74	74	74	67	75	72	24	
27		74	74	74	74	76	79	79	78	78	78	79	79	79	79	79	79	80	81	81	80	80	81	82	74	82	78	24	
28		82	82	81	79	79	80	82	82	82	82	82	80	81	82	83	84	84	84	84	84	84	82	79	79	84	82	24	
29		77	78	77	77	78	80	82	81	81	82	82	82	83	82	81	81	82	82	83	84	82	80	78	77	84	81	24	
30		77	77	77	77	77	78	80	80	81	82	81	81	80	80	80	80	80	79	78	78	78	79	77	82	79	24		
31		80	80	79	80	80	82	83	83	84	83	81	76	73	75	77	80	80	79	79	92	95	96	93	73	96	82	24	
HOURLY MAX		97	98	99	100	100	99	97	96	96	98	95	98	96	96	98	99	99	99	99	99	98	98	98	98				
HOURLY AVG		79	80	80	79	79	80	80	80	80	80	79	78	76	75	75	76	78	78	79	80	79	80	79					

STATUS FLAG CODES

C	- MONTHLY CALIBRATION	Q	- QUALITY ASSURANCE
C1	- REPEAT CALIBRATION	R	- RECOVERY
Y	- MAINTENANCE	X	- MACHINE MALFUNCTION
S	- DAILY ZERO/SPAN CHECK	G	- OUT FOR REPAIR
S1	- REPEAT ZERO/SPAN CHECK	P	- POWER FAILURE


24 HR AVERAGES December 2018



MONTHLY SUMMARY

MINIMUM 1-HR AVERAGE:	48	% @ HOUR	16	ON DAY	12
MAXIMUM 1-HR AVERAGE:	100	% @ HOUR	3	ON DAY	21
MAXIMUM 24-HR AVERAGE:	95	%		ON DAY	1
OPERATIONAL TIME:					744 hrs
AMD OPERATION UPTIME:					100.0 %
STANDARD DEVIATION:	11			MONTHLY AVERAGE:	79 %

RELATIVE HUMIDITY Hourly Averages (RH %)



BAROMETRIC PRESSURE

BAROMETRIC PRESSURE Hourly Averages (BP mbar)

	HR START (MST)	0:00	1:00	2:00	3:00	4:00	5:00	6:00	7:00	8:00	9:00	10:00	11:00	12:00	13:00	14:00	15:00	16:00	17:00	18:00	19:00	20:00	21:00	22:00	23:00	DAILY MIN.	DAILY MAX.	24-HR AVG.	RDGS.
	HR END (MST)	0:59	1:59	2:59	3:59	4:59	5:59	6:59	7:59	8:59	9:59	10:59	11:59	12:59	13:59	14:59	15:59	16:59	17:59	18:59	19:59	20:59	21:59	22:59	23:59				
	DAY																												
1	939	939	940	940	940	941	941	941	941	942	942	942	943	943	943	944	944	945	945	945	945	946	946	946	939	946	943	24	
2	947	947	947	947	947	947	947	947	947	948	948	948	948	948	948	948	948	948	948	948	948	948	948	948	947	948	948	24	
3	949	949	949	949	949	949	949	950	950	950	950	950	950	950	951	951	950	950	949	949	948	947	947	948	948	951	949	24	
4	944	943	942	942	941	941	941	941	941	941	941	941	941	941	942	943	945	947	948	949	950	951	952	941	952	944	24		
5	953	953	953	953	953	953	953	953	952	952	952	951	950	950	949	948	948	947	947	946	946	946	953	951	951	951	951	24	
6	946	946	946	946	946	946	946	946	946	945	945	944	944	944	944	944	944	944	943	943	942	941	941	940	946	946	944	24	
7	940	939	939	939	939	939	939	939	939	939	939	939	939	939	939	940	940	940	940	941	941	941	941	941	939	941	940	24	
8	942	942	942	942	942	942	942	942	942	942	942	942	942	943	943	943	944	944	944	944	944	945	945	943	945	943	943	24	
9	944	944	944	944	944	943	943	942	942	940	940	938	937	936	935	935	936	935	935	935	935	934	934	933	936	935	939	24	
10	936	935	935	935	935	935	935	935	935	935	935	935	935	935	935	935	935	935	935	935	935	934	934	933	936	935	935	24	
11	933	932	931	930	929	927	926	924	923	921	920	919	918	917	916	915	915	914	914	914	914	914	914	914	933	921	24		
12	914	914	915	916	916	917	918	918	919	920	921	921	922	923	924	925	926	927	927	927	926	926	925	914	927	921	24		
13	924	922	921	921	920	921	921	921	922	923	924	925	926	926	927	928	928	928	928	929	929	929	920	929	925	24			
14	929	929	929	929	928	928	927	927	926	925	924	923	921	920	919	918	917	917	916	916	916	916	916	916	929	922	24		
15	916	917	918	921	922	924	926	928	930	932	934	935	937	938	940	941	943	944	944	945	945	946	946	946	946	946	934	24	
16	946	945	945	945	944	943	942	941	940	939	938	937	935	935	934	934	934	933	932	932	931	931	930	930	946	938	24		
17	930	929	930	929	929	928	929	929	928	928	928	928	928	928	928	928	928	927	927	926	926	925	925	925	930	928	24		
18	924	923	923	923	922	922	921	921	921	921	921	921	921	921	921	921	922	922	923	923	923	924	924	925	921	925	922	24	
19	926	927	927	928	928	929	929	930	930	931	931	932	932	933	934	934	935	935	936	936	937	937	937	938	926	938	932	24	
20	938	938	938	938	937	937	936	935	934	933	932	931	929	927	925	923	921	920	919	918	918	918	918	918	938	928	24		
21	919	920	921	922	923	924	926	927	928	930	932	933	934	936	937	938	939	940	940	941	941	942	943	919	943	932	24		
22	943	943	944	944	945	945	945	945	945	945	945	945	944	944	944	944	944	944	945	944	944	944	944	944	945	944	944	24	
23	944	943	943	943	943	943	943	942	942	943	943	942	942	942	942	943	943	943	943	943	943	943	942	944	943	943	24		
24	943	944	944	944	944	943	943	943	943	943	943	943	943	943	942	942	943	943	944	944	945	945	942	945	943	943	24		
25	946	946	947	947	948	948	949	950	950	951	952	951	951	952	952	952	953	952	952	953	953	953	952	946	953	950	24		
26	952	952	952	952	951	951	950	950	950	950	950	948	947	946	946	946	947	948	948	948	948	949	949	946	952	949	24		
27	949	950	950	950	950	949	950	950	950	951	951	950	950	950	950	951	950	950	950	949	949	948	948	951	950	950	24		
28	948	947	946	945	944	943	942	941	940	939	939	937	936	936	936	936	936	936	936	936	936	936	936	948	939	24			
29	936	936	935	935	933	932	931	930	929	928	928	926	924	924	925	926	927	928	929	931	932	934	924	936	929	24			
30	936	938	940	941	942	944	945	946	947	949	950	951	952	953	954	954	955	955	955	955	955	955	936	955	949	24			
31	955	955	954	954	953	952	951	950	950	949	948	947	946	945	944	943	942	940	939	937	935	932	932	955	946	24			
	HOURLY MAX	955	955	954	954	953	953	953	952	952	952	952	951	952	953	953	954	954	955	955	955	955	955	955					
	HOURLY AVG	938	938	938	939	938	938	938	938	938	938	938	938	938	938	938	938	938	938	938	938	938	938	938					

STATUS FLAG CODES

C - MONTHLY CALIBRATION

C1 - REPEAT CALIBRATION

Y - MAINTENANCE

S - DAILY ZERO/SPAN CHECK

S1 - REPEAT ZERO/SPAN CHECK


Q - QUALITY ASSURANCE

R - RECOVERY

X - MACHINE MALFUNCTION

G - OUT FOR REPAIR

P - POWER FAILURE



MINIMUM 1-HR AVERAGE:

914 mbar @ HOUR

17 ON DAY

11

MAXIMUM 1-HR AVERAGE:

955 mbar @ HOUR

18 ON DAY

30

MAXIMUM 24-HR AVERAGE:

951 mbar

ON DAY

5

OPERATIONAL TIME:

744 hrs

AMD OPERATION UPTIME:

100.0 %


STANDARD DEVIATION:

10

MONTHLY AVERAGE:

938 mbar

BAROMETRIC PRESSURE Hourly Averages (BP mbar)



AMBIENT TEMPERATURE


AMBIENT TEMPERATURE Hourly Averages (AmbTPX °C)

HR START (MST) HR END (MST)	0:00 0:59	1:00 1:59	2:00 2:59	3:00 3:59	4:00 4:59	5:00 5:59	6:00 6:59	7:00 7:59	8:00 8:59	9:00 9:59	10:00 10:59	11:00 11:59	12:00 12:59	13:00 13:59	14:00 14:59	15:00 15:59	16:00 16:59	17:00 17:59	18:00 18:59	19:00 19:59	20:00 20:59	21:00 21:59	22:00 22:59	23:00 23:59	DAILY MIN.	DAILY MAX.	24-HR AVG.	RDGS.
DAY																												
1	-7.8	-8.5	-8.0	-7.5	-7.6	-7.3	-7.0	-6.6	-6.5	-6.2	-5.6	-4.7	-3.9	-3.6	-3.9	-4.5	-4.9	-4.9	-5.0	-5.2	-5.5	-5.5	-5.4	-5.4	-8.5	-3.6	-5.9	24
2	-5.5	-5.5	-5.5	-5.4	-5.6	-6.1	-6.7	-6.9	-7.1	-7.0	-6.8	-6.7	-7.1	-7.1	-7.1	-7.3	-7.7	-8.3	-8.8	-9.1	-9.4	-9.6	-9.6	-9.6	-5.4	-7.2	24	
3	-9.8	-10.5	-10.9	-10.3	-10.0	-11.2	-12.2	-13.3	-14.3	-15.1	-13.7	-11.5	-10.8	-9.3	-9.6	-11.3	-13.2	-15.8	-16.2	-14.8	-13.6	-12.0	-10.8	-10.5	-16.2	-9.3	-12.1	24
4	-9.5	-8.5	-8.7	-8.5	-8.4	-8.2	-7.7	-6.8	-6.1	-5.6	-4.6	-4.3	-5.3	-4.7	-4.1	-5.1	-6.6	-7.2	-7.7	-8.0	-8.1	-8.2	-8.3	-9.5	-4.1	-7.0	24	
5	-8.5	-8.9	-10.2	-12.2	-10.7	-9.3	-9.1	-9.0	-8.8	-8.6	-8.2	-7.4	-7.0	-7.3	-7.3	-8.4	-10.7	-10.5	-11.7	-11.9	-11.9	-10.5	-11.4	-12.2	-7.0	-9.6	24	
6	-10.6	-11.8	-12.9	-14.0	-14.1	-15.3	-15.3	-16.2	-16.7	-16.7	-13.8	-11.5	-10.4	-9.6	-9.2	-9.9	-11.5	-10.0	-11.1	-11.5	-11.0	-10.7	-10.4	-16.7	-9.2	-12.3	24	
7	-9.3	-8.3	-7.5	-7.2	-8.9	-9.9	-10.0	-10.5	-10.8	-11.6	-11.2	-9.2	-5.9	-5.2	-5.4	-6.4	-8.3	-9.7	-10.6	-10.6	-11.9	-11.3	-11.5	-7.8	-11.9	-5.2	-9.1	24
8	-8.4	-11.9	-10.7	-9.6	-9.4	-9.0	-7.0	-6.3	-7.0	-5.1	-4.5	-2.8	-3.1	-4.2	-3.7	-3.4	-3.4	-4.8	-5.0	-2.8	-2.1	-2.1	-2.5	-11.9	-2.1	-5.4	24	
9	-1.8	-1.0	-2.2	-4.2	-4.8	-4.9	-5.6	-5.6	-5.6	-5.9	-4.3	-3.0	-1.7	-1.2	-1.1	-2.7	-3.9	-6.5	-7.3	-8.1	-6.9	-4.1	-3.6	-2.9	-8.1	-1.0	-4.1	24
10	-2.5	-2.2	-1.7	-1.9	-0.3	0.5	1.3	0.1	0.3	0.2	0.5	1.7	2.6	2.5	2.1	1.3	-0.2	-1.0	-1.3	-1.7	-1.6	-2.3	-2.6	-2.5	-2.6	2.6	-0.4	24
11	-3.3	-6.1	-6.4	-3.5	-3.0	-3.5	-2.7	-2.2	-2.0	-0.7	-0.5	0.6	0.8	1.2	2.3	2.6	2.6	1.7	2.6	2.4	2.5	3.1	2.3	1.4	-6.4	3.1	-0.3	24
12	0.9	0.4	0.1	0.0	0.2	0.1	-0.8	-0.7	-0.6	0.4	0.8	1.9	2.2	2.0	1.4	0.7	0.3	-0.5	-1.7	-2.6	-3.1	-2.6	-1.6	-3.1	2.2	-0.2	24	
13	-1.1	-2.0	-1.7	0.2	1.0	1.2	0.2	0.2	1.2	1.3	1.9	2.0	1.9	2.2	2.6	2.5	2.0	1.8	1.2	0.5	0.1	0.2	0.4	0.2	-2.0	2.6	0.8	24
14	0.1	-0.2	0.1	0.3	-0.2	0.2	0.0	-1.1	-1.8	-1.6	-1.3	-1.0	-0.6	-0.3	-0.4	-0.9	-1.7	-2.2	-2.3	-2.7	-3.9	-4.6	-5.8	-6.8	-6.8	0.3	-1.6	24
15	-7.3	-7.9	-8.4	-9.5	-9.7	-9.7	-10.0	-9.9	-9.8	-10.0	-10.7	-10.5	-10.3	-9.9	-9.8	-10.2	-14.2	-17.4	-19.8	-19.6	-20.3	-20.8	-21.8	-19.9	-21.8	-7.3	-12.8	24
16	-17.7	-18.1	-17.2	-16.2	-14.3	-12.4	-10.5	-8.7	-8.5	-8.7	-9.9	-10.1	-8.9	-9.3	-9.4	-9.3	-9.8	-9.9	-10.1	-8.9	-9.2	-10.4	-10.0	-9.6	-18.1	-8.5	-11.1	24
17	-8.6	-8.0	-7.8	-8.1	-7.8	-7.6	-7.6	-7.2	-6.8	-6.4	-5.7	-5.3	-5.1	-3.9	-3.6	-3.9	-4.4	-4.8	-4.7	-3.1	-3.2	-3.6	-6.4	-5.9	-8.6	-3.1	-5.8	24
18	-3.9	-4.6	-4.7	-3.9	-3.2	-1.3	-2.8	-3.1	-3.3	-1.8	0.1	2.7	3.6	2.3	0.9	-1.0	-2.4	-1.7	-4.0	-4.7	-3.3	-2.5	-2.1	-4.7	3.6	-2.1	24	
19	-1.8	-2.3	-2.4	-2.4	-2.6	-3.3	-3.7	-2.6	-1.5	-1.2	-0.3	0.8	0.9	1.1	1.7	1.1	0.7	0.7	0.2	0.0	-0.5	-1.6	-1.8	-2.5	-3.7	1.7	-1.0	24
20	-2.8	-2.9	-4.7	-6.0	-7.8	-9.6	-8.6	-7.6	-7.3	-6.2	-5.1	-3.7	-1.9	-2.0	-2.1	-2.8	-2.7	-2.6	-2.7	-2.8	-2.8	-2.9	-2.9	-9.6	-1.9	-4.3	24	
21	-3.1	-3.4	-3.4	-3.3	-4.2	-5.1	-5.4	-5.8	-6.4	-6.9	-7.0	-6.9	-7.1	-7.3	-7.8	-8.0	-8.1	-8.7	-8.9	-8.6	-8.9	-9.2	-9.5	-9.6	-9.6	-3.1	-6.8	24
22	-11.2	-14.1	-16.2	-17.2	-18.0	-19.3	-19.2	-20.2	-19.9	-16.9	-15.4	-13.0	-11.2	-10.1	-9.9	-10.1	-12.4	-15.4	-18.2	-20.1	-21.4	-22.2	-22.9	-23.8	-23.8	-9.9	-16.6	24
23	-23.8	-22.7	-22.0	-21.4	-21.7	-21.4	-20.9	-20.7	-20.8	-20.9	-21.2	-19.7	-19.7	-18.7	-18.1	-17.8	-18.0	-17.6	-16.9	-16.6	-16.5	-16.6	-16.9	-23.8	-16.5	-19.5	24	
24	-16.7	-16.3	-16.1	-15.9	-15.5	-15.2	-14.6	-14.0	-13.7	-13.2	-13.0	-12.4	-11.4	-10.6	-10.2	-10.8	-13.1	-14.8	-15.8	-16.7	-18.7	-16.8	-15.3	-14.4	-18.7	-10.2	-14.4	24
25	-14.0	-13.9	-14.0	-14.6	-15.6	-16.5	-16.9	-17.7	-17.5	-17.6	-17.9	-17.3	-16.6	-16.8	-17.0	-17.1	-18.9	-21.7	-23.7	-25.4	-26.2	-27.0	-27.3	-25.8	-27.3	-13.9	-19.0	24
26	-26.5	-28.1	-28.1	-27.9	-26.1	-26.1	-25.0	-25.0	-26.9	-27.1	-23.2	-19.7	-18.2	-17.6	-18.0	-18.6	-19.3	-20.2	-21.3	-20.6	-20.0	-19.5	-19.3	-19.5	-28.1	-17.6	-22.6	24
27	-20.4	-21.8	-20.3	-19.7	-19.3	-19.0	-18.7	-18.8	-18.9	-18.9	-18.6	-17.9	-17.5	-17.1	-16.8	-16.8	-16.9	-16.9	-17.0	-17.1	-17.1	-17.0	-21.8	-16.8	-18.2	24		
28	-16.9	-16.7	-16.7	-18.3	-18.1	-17.0	-15.5	-14.7	-13.8	-13.1	-12.5	-12.7	-11.8	-12.7	-12.7	-13.4	-13.9	-13.8	-13.7	-14.6	-16.4	-18.9	-18.9	-11.8	-14.8	24		
29	-20.5	-20.1	-19.8	-19.7	-18.9	-17.2	-15.5	-15.3	-15.3	-14.8	-14.6	-14.3	-13.7	-14.0	-14.4	-14.4	-14.1	-13.8	-13.4	-13.0	-15.1	-17.3	-18.8	-19.4	-20.5	-13.0	-16.1	24
30	-19.5	-19.1	-18.8	-18.6	-18.2	-18.0	-17.7	-17.4	-17.2	-17.0	-16.7	-16.3	-16.0	-15.9	-15.5	-15.3	-15.3	-15.4	-15.5	-15.4	-15.3	-15.3	-16.5	-19.5	-15.3	-16.7	24	
31	-18.0	-19.1	-19.3	-18.6	-18.0	-17.6	-15.7	-14.6	-13.9	-13.9	-12.5	-11.3	-9.5	-9.2	-9.3	-10.0	-10.8	-8.4	-6.3	-5.9	-4.5	-3.3	-3.0	-19.3	-3.0	-11.8	24	
HOURLY MAX	0.9	0.4	0.1	0.3	1.0	1.2	1.3	0.2	1.2	1.3	1.9	2.0	2.7	3.6	2.6	2.6	2.6	1.8	2.6	2.4	2.5	3.1	2.3	1.4				
HOURLY AVG	-10.0	-10.5	-10.5	-10.5	-10.4	-10.4	-10.0	-9.9	-9.9	-9.7	-9.0	-8.0	-7.2	-6.9	-6.9	-7.3	-8.3	-9.1	-9.5	-9.6	-9.9	-9.8	-9.9	-9.9				

STATUS FLAG CODES

C	- MONTHLY CALIBRATION	Q	- QUALITY ASSURANCE
C1	- REPEAT CALIBRATION	R	- RECOVERY
Y	- MAINTENANCE	X	- MACHINE MALFUNCTION
S	- DAILY ZERO/SPAN CHECK	G	- OUT FOR REPAIR
S1	- REPEAT ZERO/SPAN CHECK	P	- POWER FAILURE


24 HR AVERAGES December 2018



MONTHLY SUMMARY

MINIMUM 1-HR AVERAGE:	-28.1	°C	@ HOUR	1	ON DAY	26
MAXIMUM 1-HR AVERAGE:	3.6	°C	@ HOUR	13	ON DAY	18
MAXIMUM 24-HR AVERAGE:	0.8	°C			ON DAY	13
OPERATIONAL TIME:	744	hrs				
AMD OPERATION UPTIME:	100.0	%				
STANDARD DEVIATION:	7.1					
MONTHLY AVERAGE:	-9.3	°C				

AMBIENT TEMPERATURE Hourly Averages (AmbTPX °C)




STATION TEMPERATURE

STATION TEMPERATURE Hourly Averages (StnTPX °C)

	HR START (MST) HR END (MST)	0:00 0:59	1:00 1:59	2:00 2:59	3:00 3:59	4:00 4:59	5:00 5:59	6:00 6:59	7:00 7:59	8:00 8:59	9:00 9:59	10:00 10:59	11:00 11:59	12:00 12:59	13:00 13:59	14:00 14:59	15:00 15:59	16:00 16:59	17:00 17:59	18:00 18:59	19:00 19:59	20:00 20:59	21:00 21:59	22:00 22:59	23:00 23:59	DAILY MIN.	DAILY MAX.	24-HR AVG.	RDGS.
DAY																													
1		21.5	21.1	20.9	21.0	21.0	21.1	21.2	21.3	21.4	21.6	22.0	22.4	22.8	23.0	22.8	22.6	22.4	22.3	22.1	22.0	21.9	22.0	22.2	20.9	23.0	21.8	24	
2		22.2	22.1	22.1	22.2	22.1	22.0	21.8	21.7	21.4	21.5	21.6	21.7	21.8	21.8	21.6	21.3	20.9	20.6	20.2	19.8	19.4	19.2	19.1	19.1	22.2	21.2	24	
3		19.0	18.9	18.6	18.5	19.0	18.8	17.9	17.0	17.0	16.3	16.7	17.7	18.9	20.3	20.9	20.2	18.8	17.4	17.2	19.0	20.5	19.2	19.2	19.9	16.3	20.9	18.6	24
4		19.9	20.6	21.3	22.8	22.6	22.3	22.4	22.4	21.9	22.9	23.5	26.2	26.1	25.7	25.4	25.3	25.1	25.4	24.7	24.3	23.9	23.9	23.8	19.9	26.2	23.6	24	
5		23.8	24.3	24.1	23.6	23.6	23.9	24.1	24.3	24.4	24.4	26.7	26.0	25.6	25.0	25.0	23.8	22.8	22.3	24.1	23.0	22.2	22.4	22.9	22.2	26.7	24.0	24	
6		21.9	21.5	22.5	21.5	20.6	20.2	20.2	21.2	21.1	23.1	25.2	25.3	24.4	24.5	24.6	23.8	22.7	21.8	21.9	21.9	23.4	24.1	24.3	24.2	20.2	25.3	22.7	24
7		22.8	23.0	23.1	23.7	23.4	23.2	22.7	22.6	22.7	22.3	23.0	23.1	23.9	25.3	25.1	24.8	24.1	23.3	22.9	22.8	21.8	21.5	21.1	21.0	25.3	23.0	24	
8		21.5	21.8	21.8	22.1	21.3	22.8	22.5	22.9	22.6	23.0	23.6	24.3	25.5	25.5	25.8	25.5	25.3	25.0	24.6	24.4	24.7	25.0	25.0	24.9	21.3	25.8	23.8	24
9		25.0	25.5	25.7	25.4	25.0	24.5	25.1	24.7	23.9	25.9	26.1	26.1	25.6	25.7	25.5	25.6	25.9	26.2	26.2	26.3	26.0	25.8	25.9	23.9	26.3	25.6	24	
10		25.8	25.7	25.7	25.7	25.7	25.7	25.4	25.6	25.8	25.7	25.3	25.4	25.2	25.1	25.3	25.1	24.7	22.0	22.6	23.1	23.3	23.1	23.2	22.0	25.8	24.8	24	
11		23.0	22.7	22.2	22.3	22.2	22.3	22.7	23.1	23.2	24.1	21.9	20.8	21.0	20.8	20.9	21.0	20.9	21.0	21.3	21.2	21.2	20.8	24.1	21.8	24			
12		21.3	21.4	21.5	21.5	21.4	21.5	21.4	21.3	21.3	21.8	23.0	23.9	23.7	23.6	24.0	23.9	23.8	23.7	22.4	21.3	21.2	21.2	21.0	24.0	22.2	24		
13		21.0	20.9	20.8	20.9	20.9	21.0	21.1	21.2	21.1	21.4	21.6	21.8	21.9	21.8	22.1	21.8	21.9	21.6	21.5	21.4	21.3	21.3	20.8	22.1	21.4	24		
14		21.3	21.5	21.2	21.3	21.3	21.4	21.3	21.3	21.5	21.4	21.5	21.6	21.4	21.3	21.5	21.5	21.4	21.3	21.2	21.6	21.6	21.5	21.6	21.2	21.6	21.4	24	
15		21.5	21.5	21.5	21.4	21.5	21.5	21.3	21.6	21.6	21.8	22.1	22.0	21.4	21.2	21.2	21.2	21.2	21.2	21.4	21.3	21.3	21.3	21.2	22.1	21.4	24		
16		21.3	21.2	21.4	21.5	21.4	21.3	20.9	20.9	20.8	20.8	21.1	20.9	20.9	21.1	21.2	21.1	21.1	21.2	21.0	21.1	21.3	21.4	21.2	21.3	20.8	21.5	21.1	24
17		21.1	21.2	21.2	21.0	21.1	21.3	21.4	21.3	21.1	20.1	21.0	21.0	21.0	21.0	21.1	21.2	21.3	21.3	21.2	21.2	21.0	21.0	21.0	21.6	21.2	24		
18		21.3	21.5	21.4	21.4	21.5	21.3	21.5	21.6	21.5	21.4	21.3	21.4	21.9	21.9	21.2	21.4	21.3	21.5	21.2	21.3	21.3	21.4	21.2	21.2	21.9	21.4	24	
19		21.4	21.4	21.5	21.3	21.4	21.4	21.3	21.4	21.3	21.4	21.4	21.7	21.6	21.4	21.2	21.3	21.4	21.5	21.5	21.3	21.4	21.3	21.2	21.7	21.4	24		
20		21.5	21.4	21.4	21.3	21.4	21.2	21.3	21.3	21.4	21.3	21.3	21.4	21.4	21.4	21.3	21.4	21.4	21.2	21.2	21.3	21.3	21.2	21.1	21.6	21.4	24		
21		21.6	21.6	21.5	21.6	21.4	21.3	21.3	21.3	21.4	21.4	21.5	21.6	21.5	21.4	21.5	21.6	21.5	21.6	21.6	21.5	21.6	21.6	21.3	21.6	21.5	24		
22		21.5	21.4	21.3	21.4	21.3	21.4	21.4	21.4	21.4	21.6	21.3	20.9	21.0	21.2	21.3	21.4	21.3	21.2	21.4	21.3	21.3	21.4	20.9	21.6	21.3	24		
23		21.4	21.4	21.4	21.5	21.5	21.4	21.6	21.4	21.5	21.4	21.4	21.7	21.5	21.5	21.5	21.5	21.5	21.5	21.5	21.3	21.3	21.3	21.7	21.5	21.5	24		
24		21.4	21.4	21.3	21.2	21.3	21.3	21.4	21.3	21.3	21.2	21.1	21.4	21.4	21.1	21.3	21.4	21.4	21.4	21.2	21.2	21.3	21.1	21.5	21.3	24			
25		21.5	21.5	21.5	21.4	21.5	21.5	21.4	21.5	21.5	21.6	21.5	21.5	21.3	21.3	21.3	21.3	21.3	21.3	21.4	21.5	21.6	21.5	21.6	21.4	24			
26		21.3	21.3	21.4	21.4	21.4	21.3	21.3	21.5	21.4	21.4	21.5	21.5	21.5	21.2	21.2	21.2	21.3	21.4	21.4	21.3	21.5	21.1	21.5	21.3	24			
27		21.5	21.4	21.5	21.4	21.4	21.5	21.5	21.5	21.4	21.4	21.5	21.6	21.6	21.5	21.5	21.5	21.5	21.4	21.4	21.5	21.5	21.4	21.4	21.6	21.5	24		
28		21.4	21.3	21.3	21.4	21.5	21.4	21.1	21.0	21.2	21.2	21.3	21.5	21.4	21.5	21.5	21.5	21.5	21.5	21.4	21.4	21.4	21.4	21.0	21.5	21.4	24		
29		21.3	21.4	21.4	21.4	21.5	21.5	21.3	21.3	21.3	21.1	21.2	21.2	21.3	21.3	21.3	21.4	21.4	21.4	21.3	21.4	21.4	21.4	21.1	21.5	21.3	24		
30		21.4	21.3	21.5	21.4	21.4	21.5	21.4	21.4	21.4	21.5	21.5	21.5	21.5	21.5	21.5	21.5	21.5	21.4	21.2	21.1	21.1	21.2	21.1	21.5	21.3	24		
31		21.3	21.3	21.3	21.3	21.3	21.3	21.1	21.1	21.0	21.2	21.1	21.0	21.1	21.0	21.0	21.1	21.0	21.0	21.1	21.1	20.9	21.2	20.9	21.3	21.1	24		
HOURLY MAX		25.8	25.7	25.7	25.7	25.7	25.7	25.4	25.6	25.9	26.1	26.7	26.1	25.7	25.8	25.6	25.9	25.9	26.2	26.2	26.2	26.3	26.0	25.8	25.9				
HOURLY AVG		21.7	21.8	21.8	21.8	21.7	21.7	21.7	21.7	21.9	22.1	22.3	22.3	22.4	22.3	22.3	22.1	21.9	21.7	21.8	21.8	21.8	21.7	21.8					


24 HR AVERAGES December 2018



MONTHLY SUMMARY

MINIMUM 1-HR AVERAGE:	16.3	°C	@ HOUR	9	ON DAY	3
MAXIMUM 1-HR AVERAGE:	26.7	°C	@ HOUR	11	ON DAY	5
MAXIMUM 24-HR AVERAGE:	25.6	°C			ON DAY	9
OPERATIONAL TIME:					744	hrs
AMD OPERATION UPTIME:					100.0	%
STANDARD DEVIATION:	1.5				MONTHLY AVERAGE:	21.9 °C

STATION TEMPERATURE Hourly Averages (StnTPX °C)



APPENDIX II
EQUIPMENT CALIBRATION RESULTS

SULPHUR DIOXIDE



Thermo 43C Sulphur Dioxide Analyzer Calibration

Date:	December 12, 2018	Barometer/B.P./units:	F.S. 10528 expires January 15, 2019	925	millibars
Company/Airshed:	PRAMP	Thermometer/Station Temp:	F.S. 10528 expires January 15, 2019	23	°C
Location/Station Name:	986b	Weather Conditions:	Mainly sunny		
Parameter:	Sulphur Dioxide	Calibration Purpose:	routine monthly		
Start Time 24 hr. (mst):	14:01	Performed By/Reviewer:	Chris Wesson	Rob Fisher	
End Time 24 hr. (mst):	18:05	Cal Gas Expiry Date:	October 24, 2020		
Calibration Method:	Gas Dilution	Converter Model & s/n (if applicable):	n/a		

Analyzer:

Serial Number/Owner:	43C-62339-335	Maxxam	Range ppb:	500
Last Calibration Date:	November 6, 2018		As Found C.F.:	0.995
Previous C.F.:	0.999		New C.F.:	1.000

Calibration Standards:

Low Flow Meter ID/Expiry Date: N/A
High Flow Meter ID/Expiry Date: N/A
Calibrator ID/Expiry Date: Sabio id# 3860808 expires August 21, 2019
Cal Gas Cylinder I.D. #: LL108015
Cal Gas Conc. (ppm): 47.9

Standard Calibration Points for Ranges


Point	ppb
High	380
Mid	180
Low	90

ALL POINTS ARE 15 MINUTES OF STABILITY AS OF SEPTEMBER 23, 2015

Calibrator Flow Rates (cc/min)				Calculated Concentration (ppb):	Indicated Concentration (ppb):	Correction Factors (C.F.):
Point	Diluent	Cal Gas	Total			
as found zero	6000	0.00	6000	0.0	-0.1	n/a
as found high	5951	47.60	5999	380.1	382	0.995
adjusted zero	5999	0.00	5999	0.0	0	n/a
adjusted high	5951	47.60	5999	380.1	380	1.000
mid	5974	22.50	5996	179.7	181	0.993
low	5987	11.30	5998	90.2	90	1.003
calibrator zero	6000	0.00	6000	0.0	0.2	n/a
				Average C.F. =	0.999	

Linear Regression/Calibration Results:

Correlation Coefficient = 1.000
Slope = 1.000
b (Intercept as % of full scale) = -0.04%
% change in C.F. from last cal = 0.43%
LIMITS
> or = 0.995
0.95-1.05
± 3% F.S.
± 10%

Thermo 43C Sulphur Dioxide Analyzer Calibration

As found:		As left:	
Bkg:	84.7	Bkg:	85.3
Coef:	0.921	Coef:	0.921
Pmt:	-654	Pmt:	-654
0	Lamp=846	0	Lamp=846
Battery:	3.3	Battery:	3.3
Internal:	29.5	Internal:	29.9
Chamber:	45.3	Chamber:	45.3
Pressure:	402.0	Pressure:	404.2
Flow:	0.708	Flow:	0.712
Intensity:	~38000	Intensity:	~38000
Averaging Time:	120	Averaging Time:	120
Expected Value:	246.8	Expected Value:	256.0


Comments:

The analyzer sample inlet filter was changed.

The analyzer cooling fan filter(s) were cleaned.

The manifold blower was found to be working normally.

SO2[ppb] Station: PRAMP_986 Daily: 18/12/12 Type: AVG 1 Min. [1 Min.]




TOTAL REDUCED SULPHUR



Thermo 43I-TLE Total Reduced Sulphur Analyzer Calibration

Date: December 12, 2018	Barometer/B.P./units: F.S. 10528 expires January 15, 2019	921	millibars
Company/Airshed: PRAMP	Thermometer/Station Temp: F.S. 10528 expires January 15, 2019	21	°C
Location/Station Name: 986b	Weather Conditions: Mix of sun and clouds		
Parameter: Total Reduced Sulphur	Calibration Purpose: routine monthly		
Start Time 24 hr. (mst): 9:21	Performed By/Reviewer: Chris Wesson	Rob Fisher	
End Time 24 hr. (mst): 14:21	Cal Gas Expiry Date: November 7, 2020		
Calibration Method: Gas Dilution	Converter Model & s/n (if applicable): CD Nova CDN-101 #516		
Analyzer:			
Serial Number/Owner: 1152940011 Maxxam	Range ppb: 100		
Last Calibration Date: November 6, 2018	As Found C.F.: 0.999		
Previous C.F.: 1.001	New C.F.: 1.000		
Calibration Standards:	Standard Calibration Points for Ranges		
Low Flow Meter ID/Expiry Date: N/A	Point: High	ppb: 78	SO2 Scrubber Check (10 minutes):
High Flow Meter ID/Expiry Date: N/A	Point: Mid	ppb: 38	Start/End Time 24 hr.: 09:37 / 09:52
Calibrator ID/Expiry Date: Sabio id# 3860808 expires August 21, 2019	Point: Low	ppb: 19	SO2 Analyzer Range: 500
Cal Gas Cylinder I.D. #: LL119432			Target Concentration (ppb): 380
Cal Gas Conc. (ppm): 10.3			As Found Zero: -0.2
			Analyzer Response (ppb): -0.2
			Zero Corrected Result (ppb): 0.0
ALL POINTS ARE 15 MINUTES OF STABILITY AS OF SEPTEMBER 23, 2015			
Calibrator Flow Rates (cc/min)	Calculated Concentration (ppb):	Indicated Concentration (ppb):	Correction Factors (C.F.):
Point Diluent Cal Gas Total			
as found zero 7499 0.00 7499	0.0	-0.2	n/a
as found high 7441 56.90 7498	78.0	77.92	0.999
adjusted zero 7499 0.00 7499	0.0	0	n/a
adjusted high 7442 56.90 7499	78.0	78	1.000
mid 7473 27.70 7501	38.0	37.8	1.004
low 7487 13.80 7501	18.9	19.01	0.995
calibrator zero 7504 0.00 7504	0.0	0.1	n/a
	Average C.F.=	1.000	
Linear Regression/Calibration Results:			
Correlation Coefficient = 1.000	LIMITS		
Slope = 1.001	> or = 0.995		
b (Intercept as % of full scale) = -0.01%	0.95-1.05		
% change in C.F. from last cal= 0.24%	± 3% F.S.		
± 10%			
Thermo 43I-TLE Total Reduced Sulphur Analyzer Calibration			
As found:		As left:	
Bkg: 2.23	Bkg: 2.21		
Coef: 0.979	Coef: 0.984		
Pmt: -690.8	Pmt: -690.4		
Flash: 964	Flash: 965		
Internal: 30.7	Internal: 32.2		
Chamber: 45.1	Chamber: 45.1		
Perm Oven Gas: 45.00	Perm Oven Gas: 45.00		
Perm Oven Heater: 44.25	Perm Oven Heater: 44.25		
Pressure: 639.9	Pressure: 641.7		
Sample Flow: 0.468	Sample Flow: 0.470		
Lamp Intensity: 91	Lamp Intensity: 92		
Converter: 820	Converter: 820		
Converter Set: 820	Converter Set: 820		
Averaging Time: 120	Averaging Time: 120		
Expected Value: 41.2	Expected Value: 43.1		
Comments:			
The analyzer sample inlet filter was changed.			
The analyzer cooling fan filter(s) were cleaned.			
The manifold blower was found to be working normally.			

TRS[ppb] Station: PRAMP_986 Daily: 18/12/12 Type: AVG 1 Min. [1 Min.]




TOTAL HYDROCARBON



Thermo 55i Methane/Non-Methane Analyzer Calibration


Date: December 12, 2018		Barometer/B.P./units: F.S. 10528 expires January 15, 2019		921	millibars
Company/Airshed: PRAMP		Thermometer/Station Temp: F.S. 10528 expires January 15, 2019		21	°C
Location/Station Name: 986b		Weather Conditions: Mix of sun and clouds			
Parameter: CH ₄ / NMHC / THC		Calibration Purpose: shut down			
Start/End Time 24 hr. (mst): 09:21 / 11:19		Performed By/Reviewer: Chris Wesson		Rob Fisher	
Calibration Method: Gas Dilution		Cal Gas Expiry Date: October 18, 2025			
Analyzer:					
Correction Factors:					
Serial Number/Owner: 1022143392 Maxxam		Previous C.F.:		As Found C.F.:	
Measured Flow: 823 mL/min		CH ₄ = 1.001	1.050	n/a	
Last Calibration Date: November 6, 2018		NMHC = 0.998	1.071	n/a	
Range ppm: 20 CH ₄ /20 NMHC/40 THC		THC = 1.000	1.059	n/a	
Calibration Standards:					
Low Flow Meter ID/Expiry Date: N/A		Standard Calibration Points for Analyzer Range of 20/20/40 ppm			
High Flow Meter ID/Expiry Date: N/A		Point	CH ₄	NMHC	THC
Calibrator ID/Expiry Date: Sabio id# 17100415 expires August 21, 2019		High	13.00	13.00	26.00
Cal Gas Cylinder I.D. #: LL107207		Mid	7.00	7.00	14.00
CH ₄ Cylinder Conc.= 600.0 207.0 =C ₂ H ₆ Cylinder Conc.		Low	3.00	3.00	6.00
CH ₄ expressed as C ₂ H ₆ = 569.3 1169.3 =total CH ₄ equivalent					
<i>ALL POINTS ARE 15 MINUTES OF STABILITY AS OF SEPTEMBER 23, 2015</i>					
Calibrator Flow Rates (cc/min)					
Point	Diluent	Cal Gas	Total Flow	Calculated CH ₄ (ppm)	Calculated NMHC (ppm)
as found zero	2998	0.00	2998	0.00	0.00
as found high	2924	73.40	2997	14.69	13.94
mid	2963	36.80	3000	7.36	6.98
low	2981	18.50	2999	3.70	3.51
Calculated THC (ppm)	Indicated CH ₄ (ppm)	Indicated NMHC (ppm)	Indicated THC (ppm)	Correction Factors:	
				CH ₄	NMHC
				THC	
				n/a	n/a
				1.071	1.059
				1.050	1.040
				1.040	1.068
				1.040	1.052
				1.040	1.074
				1.040	1.054
				Average C.F.= 1.043	1.071
					1.055
Linear Regression/Calibration Results:					
Correlation Coefficient = 1.000 CH ₄ NMHC THC			LIMITS		
Slope = 0.952 0.934 0.944			> or = 0.995		
b (Intercept as % of full scale)= 0.14% 0.00% 0.08%			0.90-1.10		
% change in C.F. from last cal= -4.86% -7.29% -5.94%			± 3% F.S.		
			± 10%		
As Left Instrument Diagnostics:					
Interface Board Voltages:			Calibration History cnt'd:		
Temperatures:			Crucial Settings:		
Detector Oven: 175.0			NM Peak Area: n/a		
Filter: 175.0			Methane Start: n/a		
Column Oven: 74.9			Methane End: n/a		
Internal: 36.6			Backflush: n/a		
Carrier: 600 50			NMHV Start: n/a		
Fuel: 1200 50			NMHC End: n/a		
Span Gas: 1850 18			Run History>1:		
Zero Air Generator: 50			Date: n/a		
Carrier: 31.3			Time: n/a		
Fuel: 40.5			CH ₄ PK HT: n/a		
Air: 30.9			CH ₄ RT: n/a		
FID Status:			CH ₄ Baseline: n/a		
Status: LIT			CH ₄ LOD: n/a		
Counts: 19787			CH ₄ SD: n/a		
Flame: 321.0			CH ₄ CONC: n/a		
Det Base: 175.0			NM PK HT: n/a		
Flameouts: 1			NM Peak Area: n/a		
Det Oven at Start: 169.5			NM CONC: n/a		
Col Oven at Start: 74.8			NM Base Start: n/a		
Calibration History:			NM Base End: n/a		
Time: n/a			NM LOD: n/a		
Type: n/a			NM Start IDX: n/a		
Status: n/a			NM End IDX: n/a		
Check/Adjust: n/a			NM Max Slope: n/a		
CH ₄ Span Conc: n/a			NM Min Slope: n/a		
CH ₄ SP Ratio: n/a			NM PT Count: n/a		
CH ₄ RT: n/a			Previous CH4: n/a		
CH ₄ PK IDX: n/a			Previous NMHC: n/a		
CH ₄ PK HT: n/a			Previous THC: n/a		
NM Span Conc: n/a			New CH4: n/a		
NM SP Ratio: n/a			New NMHC: n/a		
			New THC: n/a		
Comments:					
The manifold blower was found to be working normally.					
A Shut-down calibration was performed to rebuild the pump .					






Thermo 55i Methane/Non-Methane Analyzer Calibration

Thermo 55i Methane/Non-Methane Analyzer Calibration																																																																																									
Date: December 12, 2018 Company/Airshed: PRAMP Location/Station Name: 986b Parameter: CH ₄ / NMHC / THC Start/End Time 24 hr. (mst): 16:00 / 18:45 Calibration Method: Gas Dilution				Barometer/B.P./units: F.S. 10528 expires January 15, 2019 Thermometer/Station Temp: F.S. 10528 expires January 15, 2019 Weather Conditions: Mainly clear Calibration Purpose: post repair Performed By/Reviewer: Chris Wesson / Rob Fisher Cal Gas Expiry Date: October 18, 2025																																																																																					
Analyzer: Serial Number/Owner: 1022143392 Maxxam Measured Flow: 903 mL/min Last Calibration Date: n/a Range ppm: 20 CH ₄ /20 NMHC/40 THC				Correction Factors: <table border="1" style="margin-left: auto; margin-right: auto;"> <tr> <th></th> <th>Previous C.F.:</th> <th>As Found C.F.:</th> <th>New C.F.:</th> </tr> <tr> <td>CH₄ =</td> <td>n/a</td> <td>n/a</td> <td>0.999</td> </tr> <tr> <td>NMHC =</td> <td>n/a</td> <td>n/a</td> <td>1.000</td> </tr> <tr> <td>THC =</td> <td>n/a</td> <td>n/a</td> <td>0.999</td> </tr> </table>									Previous C.F.:	As Found C.F.:	New C.F.:	CH ₄ =	n/a	n/a	0.999	NMHC =	n/a	n/a	1.000	THC =	n/a	n/a	0.999																																																														
	Previous C.F.:	As Found C.F.:	New C.F.:																																																																																						
CH ₄ =	n/a	n/a	0.999																																																																																						
NMHC =	n/a	n/a	1.000																																																																																						
THC =	n/a	n/a	0.999																																																																																						
Calibration Standards:																																																																																									
Low Flow Meter ID/Expiry Date: N/A High Flow Meter ID/Expiry Date: N/A Calibrator ID/Expiry Date: Sabio id# 17100415 expires August 21, 2019 Cal Gas Cylinder I.D. #: LL107207 CH ₄ Cylinder Conc.= 600.0 207.0 =C ₃ H ₈ Cylinder Conc. CH ₄ expressed as C ₃ H ₈ = 569.3 1169.3 =total CH ₄ equivalent				Standard Calibration Points for Analyzer Range of 20/20/40 ppm <table border="1" style="margin-left: auto; margin-right: auto;"> <tr> <th>Point</th> <th>CH₄</th> <th>NMHC</th> <th>THC</th> </tr> <tr> <td>High</td> <td>13.00</td> <td>13.00</td> <td>26.00</td> </tr> <tr> <td>Mid</td> <td>7.00</td> <td>7.00</td> <td>14.00</td> </tr> <tr> <td>Low</td> <td>3.00</td> <td>3.00</td> <td>6.00</td> </tr> </table>								Point	CH ₄	NMHC	THC	High	13.00	13.00	26.00	Mid	7.00	7.00	14.00	Low	3.00	3.00	6.00																																																														
Point	CH ₄	NMHC	THC																																																																																						
High	13.00	13.00	26.00																																																																																						
Mid	7.00	7.00	14.00																																																																																						
Low	3.00	3.00	6.00																																																																																						
ALL POINTS ARE 15 MINUTES OF STABILITY AS OF SEPTEMBER 23, 2015																																																																																									
Calibrator Flow Rates (cc/min)				Correction Factors: <table border="1" style="margin-left: auto; margin-right: auto;"> <tr> <th>Point</th> <th>Diluent</th> <th>Cal Gas</th> <th>Total Flow</th> <th>Calculated CH₄ (ppm)</th> <th>Calculated NMHC (ppm)</th> <th>Calculated THC (ppm)</th> <th>Indicated CH₄ (ppm)</th> <th>Indicated NMHC (ppm)</th> <th>Indicated THC (ppm)</th> <th>CH₄</th> <th>NMHC</th> <th>THC</th> </tr> <tr> <td>adjusted zero</td> <td>2998</td> <td>0.00</td> <td>2998</td> <td>0.00</td> <td>0.00</td> <td>0.00</td> <td>0.00</td> <td>0.00</td> <td>0.00</td> <td>n/a</td> <td>n/a</td> <td>n/a</td> </tr> <tr> <td>adjusted high</td> <td>2924</td> <td>73.40</td> <td>2997</td> <td>14.69</td> <td>13.94</td> <td>28.64</td> <td>14.71</td> <td>13.94</td> <td>28.66</td> <td>0.999</td> <td>1.000</td> <td>0.999</td> </tr> <tr> <td>mid</td> <td>2963</td> <td>36.80</td> <td>3000</td> <td>7.36</td> <td>6.98</td> <td>14.34</td> <td>7.43</td> <td>6.98</td> <td>14.42</td> <td>0.991</td> <td>1.000</td> <td>0.995</td> </tr> <tr> <td>low</td> <td>2981</td> <td>18.50</td> <td>2999</td> <td>3.70</td> <td>3.51</td> <td>7.21</td> <td>3.73</td> <td>3.49</td> <td>7.23</td> <td>0.992</td> <td>1.006</td> <td>0.998</td> </tr> <tr> <td>calibrator zero</td> <td>2998</td> <td>0.00</td> <td>2998</td> <td>0.00</td> <td>0.00</td> <td>0.00</td> <td>0.00</td> <td>0.00</td> <td>0.00</td> <td>n/a</td> <td>n/a</td> <td>n/a</td> </tr> </table>								Point	Diluent	Cal Gas	Total Flow	Calculated CH ₄ (ppm)	Calculated NMHC (ppm)	Calculated THC (ppm)	Indicated CH ₄ (ppm)	Indicated NMHC (ppm)	Indicated THC (ppm)	CH ₄	NMHC	THC	adjusted zero	2998	0.00	2998	0.00	0.00	0.00	0.00	0.00	0.00	n/a	n/a	n/a	adjusted high	2924	73.40	2997	14.69	13.94	28.64	14.71	13.94	28.66	0.999	1.000	0.999	mid	2963	36.80	3000	7.36	6.98	14.34	7.43	6.98	14.42	0.991	1.000	0.995	low	2981	18.50	2999	3.70	3.51	7.21	3.73	3.49	7.23	0.992	1.006	0.998	calibrator zero	2998	0.00	2998	0.00	0.00	0.00	0.00	0.00	0.00	n/a	n/a	n/a
Point	Diluent	Cal Gas	Total Flow	Calculated CH ₄ (ppm)	Calculated NMHC (ppm)	Calculated THC (ppm)	Indicated CH ₄ (ppm)	Indicated NMHC (ppm)	Indicated THC (ppm)	CH ₄	NMHC	THC																																																																													
adjusted zero	2998	0.00	2998	0.00	0.00	0.00	0.00	0.00	0.00	n/a	n/a	n/a																																																																													
adjusted high	2924	73.40	2997	14.69	13.94	28.64	14.71	13.94	28.66	0.999	1.000	0.999																																																																													
mid	2963	36.80	3000	7.36	6.98	14.34	7.43	6.98	14.42	0.991	1.000	0.995																																																																													
low	2981	18.50	2999	3.70	3.51	7.21	3.73	3.49	7.23	0.992	1.006	0.998																																																																													
calibrator zero	2998	0.00	2998	0.00	0.00	0.00	0.00	0.00	0.00	n/a	n/a	n/a																																																																													
Average C.F.= 0.994 1.002 0.997																																																																																									
Linear Regression/Calibration Results:																																																																																									
Correlation Coefficient = 1.000 Slope = 1.001 b (Intercept as % of full scale)= 0.11% % change in C.F. from last cal= n/a				LIMITS > or = 0.995 0.95-1.05 ± 3% F.S. n/a																																																																																					
As Left Instrument Diagnostics:																																																																																									
Interface Board Voltages: Temperatures: Column Oven: Cylinder Pressures/reg.: Internal Pressures: FID Status: Flame and Power Stats:				Calibration History crnt'd: Crucial Settings: Run History>1: Expected Values:																																																																																					
Bias Supply: -311.9 Detector Oven: 175.0 Filter: 175.1 Carrier: 2626 50 Fuel: 1200 50 Span Gas: 1850 18 Zero Air Generator: 50 Carrier: 31.3 Fuel: 40.5 Air: 30.7 Status: LIT Counts: 20400 Flame: 321.9 Det Base: 175.0 Last Power On: 01Oct2018 Flameouts: 1 Det Oven at Start: 169.5 Col Oven at Start: 74.8 Time: 12Dec2018@16:52 Type: Span Status: Good Check/Adjust: Adjust CH ₄ Span Conc: 14.69 CH ₄ SP Ratio: 0.00077 CH ₄ RT: 12.2 CH ₄ PK IDX: 21 CH ₄ PK HT: 19088 NM Span Conc: 13.94 NM SP Ratio: 0.000185				NM Peak Area: 75168 Methane Start: n/a Methane End: n/a Backflush: n/a NMHV Start: n/a NMHC End: n/a Date: 12Dec2018 Time: 17:43 CH ₄ PK HT: 0 CH ₄ RT: 12.2 CH ₄ Baseline: 1636 CH ₄ LOD: 19 CH ₄ SD: 6 CH ₄ CONC: 0.00 NM PK HT: 0 NM Peak Area: 0 NM CONC: 0.00 NM Base Start: 1642 NM Base End: 1631 NM LOD: 9 NM Start IDX: 40 NM End IDX: 42 NM Max Slope: 2.1e-01 NM Min Slope: -2.9e01 NM PT Count: 0 Previous CH4: 10.21 Previous NMHC: 11.08 Previous THC: 21.3 New CH4: 10.21 New NMHC: 11.08 New THC: 21.30																																																																																					
Comments: The analyzer sample inlet filter was changed. A new nitrogen cylinder was installed. No zero adjustment was required/made. As found zero values were copied to adjusted zero values for linearity calculation purposes. The analyzer cooling fan filter(s) were cleaned. The manifold blower was found to be working normally.																																																																																									
A Post-repair calibration was performed following the pump rebuild.																																																																																									



Station: PRAMP_986 Daily: 18/12/12 Type: AVG 1 Min. [1 Min.]



WIND SYSTEM



Meteorological Sensor Audit/Calibration

Location Information

Company:	PRAMP	Performed By:	Limin Li
Audit Location:	986B	Reviewed By:	Tom Bourque
Audit Date:	April 4, 2018	Start/End Time (mst):	16:22/17:42
Calibration Purpose:	routine annual	Weather Conditions:	Mainly sunny

Wind Sensor Information

Sensor ID Data:		Sensor Outputs:	
Sensor Make:	RM Young	Velocity Voltage Output Range:	0-1V
Sensor Model:	5305VK	Velocity Unit Output Range:	0-200 KPH
Serial #:	129612	Direction Voltage Output Range:	0-1 V
Previous Cal/Audit Date:	April 5, 2017	Direction Unit Output Range:	0-360 DEG

Wind Calibrator Information

Calibrator I.D. and Expiry Date: RM Young 18802 id# CA03309 expires September 25, 2018

Wind Speed Audit Data **+/- 2% of the average correction factor is the limit**

RPM	Wind Speed Generated kph	Clockwise Wind Speed kph	Counter Clockwise Wind Speed kph	Correction Factor
0	0	0.1	0.1	-
1000	18.4	18.5	18.5	0.996
2000	36.9	36.9	36.9	1.001
3000	55.3	55.3	55.3	1.000
4000	73.7	73.7	73.7	1.000
5000	92.2	92.2	92.2	1.000
6000	110.6	110.7	110.7	0.999
7000	129.0	129.2	129.2	0.998
8000	147.4	147.7	147.7	0.998
9000	165.9	166.2	166.2	0.998
10000	184.3	184.8	184.8	0.997

The audit meets AMD requirements. Average Correction Factor= 0.999

Wind Direction Audit Data **+/- 3° of the absolute average degrees difference for all points is the limit**

Generated Wind Direction 0-360 (Up)	Generated Wind Direction 360-0 (Down)	Indicated Wind Direction 0-360 (Up)	Indicated Wind Direction 360-0 (Down)	Degrees Difference 0-360 (Up)	Degrees Difference 360-0 (Down)	Average Absolute Degrees Difference
0	355	0	354	0.3	1.4	0.8
30	330	30	330	0.2	0.1	0.1
60	300	61	301	-0.7	-0.9	0.8
90	270	91	271	-1.1	-1.2	1.1
120	240	122	243	-1.7	-2.6	2.2
150	210	153	213	-2.5	-2.9	2.7
180	180	183	183	-2.6	-2.6	2.6
210	150	213	152	-2.6	-1.7	2.1
240	120	243	122	-2.8	-2.0	2.4
270	90	272	91	-2.3	-1.1	1.7
300	60	301	62	-0.8	-1.7	1.2
330	30	331	30	-0.5	-0.3	0.4
355	0	354	0	1.3	0.3	0.8

The audit meets AMD requirements. Average Absolute Degrees Difference= 1.5

Comments:

METEOROLOGICAL SYSTEMS CHECK



Meteorological System Checklist

Date:	December 12, 2018				
Technician:	Chris Wesson				
Reviewer:	Rob Fisher				
Station:	PRAMP 986b				
Unit:	Make:	Model:	Serial #:		
Temperature Sensor:	RM Young	43172VC	61012322		
Barometric Pressure Sensor:	MetOne	090D	F3845		
Relative Humidity Sensor:	RM Young	43172VC	61012322		
Anemometer:	RM Young	05305VK	129612		
AMBIENT TEMPERATURE SENSOR CHECK					
Previous check date:	November 1, 2018				
Parameter:	Temperature @ 2 metres (1 C tolerance)				
Reference Thermometer ID:	F.S. 160348895 expires June 19, 2020				
Reference Temperature (°C):	0.0				
Station - Ambient Temperature (°C):	1.0				
Temperature Difference (°C):	-1.0				
BAROMETRIC PRESSURE SENSOR CHECK					
Previous check date:	November 1, 2018				
Reference Barometer ID:	F.S. 10528 expires January 15, 2019				
Reference Pressure - Units/Reading:	mmHg	927			
Station Pressure - Units/Reading:	mmHg	926			
Pressure Tolerance +/- 15% of error:	788 - 1066	0.11%			
RELATIVE HUMIDITY (HYGROMETER) SENSOR CHECK					
Previous check date:	November 1, 2018				
Reference Hygrometer ID:	F.S. id# 160459244 expires June 19, 2020				
Reference Hygrometer % RH- Reading:	51.90				
Station Hygrometer % RH- Reading:	47.10				
RH Tolerance +/- 15% of difference:	44.12 - 59.69	9.2%			
ANEMOMETER - WIND SPEED & WIND DIRECTION SENSOR CHECK					
WIND SPEED		WIND DIRECTION			
Previous check date:	November 1, 2018	Previous check date:	November 1, 2018		
Wind Speed Observed (kph):	20-Oct	Wind Direction Observed:	W		
Wind speed on Data Logger (kph):	12	Wind Direction on Data Logger:	W		
		Wind Direction Pass/Fail?:	Pass		

CALIBRATORS

Calibrator Performance Audit

Oxides Of Nitrogen

File No. 2018-132A

Company <u>Maxxam</u>	Operator: <u>Mike</u>																																																																
Calibrator: Make/Model <u>Sabio</u> Serial Number <u>3860808</u> Last Verification Date <u>26-Jan-17</u> NO Cylinder S/N <u>LL104183</u> NO [PPM] <u>50.8</u> NOx [PPM] <u>50.9</u> Expiry Date <u>October 24, 2020</u>																																																																	
Flow Measurement Device: Make/Model <u>Bios Definer 220</u> Serial Number <u>H=128686; L=129069</u> Temperature (°C) <u>22.2 C</u> Barometric Pressure <u>706.1mmHg</u>																																																																	
Dilution Flow (sccm) Pt. #1 <u>5065</u> Pt. #2 <u>5085</u> Pt. #3 <u>5079</u> Gas Flow (sccm) Pt. #1 <u>78.3</u> Pt. #2 <u>38.1</u> Pt. #3 <u>19.2</u>																																																																	
<table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th colspan="2">Calibrator Flow (sccm)</th> <th colspan="2">Calculated Conc.(ppm)</th> <th colspan="3">Indicated Conc.(ppm)</th> <th colspan="2">% Difference vs Audit Gas</th> </tr> <tr> <th>Dilution</th> <th>Gas</th> <th>NO</th> <th>NOx</th> <th>NO</th> <th>NO₂</th> <th>NOx</th> <th>NO</th> <th>NOx</th> </tr> </thead> <tbody> <tr> <td>5095</td> <td>0.0</td> <td>0.0000</td> <td>0.0000</td> <td>0.0000</td> <td>0.0001</td> <td>0.0001</td> <td colspan="2">Limit ± 10%</td> </tr> <tr> <td>5065</td> <td>78.3</td> <td>0.7853</td> <td>0.7869</td> <td>0.7957</td> <td>0.0005</td> <td>0.7962</td> <td>1%</td> <td>1%</td> </tr> <tr> <td>5085</td> <td>38.1</td> <td>0.3806</td> <td>0.3814</td> <td>0.3850</td> <td>0.0001</td> <td>0.3851</td> <td>1%</td> <td>1%</td> </tr> <tr> <td>5079</td> <td>19.2</td> <td>0.1920</td> <td>0.1924</td> <td>0.1936</td> <td>-0.0001</td> <td>0.1935</td> <td>1%</td> <td>1%</td> </tr> <tr> <td align="center" colspan="8">Absolute Average Percent Difference</td> <td>1%</td> <td>1%</td> </tr> </tbody> </table>		Calibrator Flow (sccm)		Calculated Conc.(ppm)		Indicated Conc.(ppm)			% Difference vs Audit Gas		Dilution	Gas	NO	NOx	NO	NO ₂	NOx	NO	NOx	5095	0.0	0.0000	0.0000	0.0000	0.0001	0.0001	Limit ± 10%		5065	78.3	0.7853	0.7869	0.7957	0.0005	0.7962	1%	1%	5085	38.1	0.3806	0.3814	0.3850	0.0001	0.3851	1%	1%	5079	19.2	0.1920	0.1924	0.1936	-0.0001	0.1935	1%	1%	Absolute Average Percent Difference								1%	1%
Calibrator Flow (sccm)		Calculated Conc.(ppm)		Indicated Conc.(ppm)			% Difference vs Audit Gas																																																										
Dilution	Gas	NO	NOx	NO	NO ₂	NOx	NO	NOx																																																									
5095	0.0	0.0000	0.0000	0.0000	0.0001	0.0001	Limit ± 10%																																																										
5065	78.3	0.7853	0.7869	0.7957	0.0005	0.7962	1%	1%																																																									
5085	38.1	0.3806	0.3814	0.3850	0.0001	0.3851	1%	1%																																																									
5079	19.2	0.1920	0.1924	0.1936	-0.0001	0.1935	1%	1%																																																									
Absolute Average Percent Difference								1%	1%																																																								

LINEAR REGRESSION ANALYSIS							
<i>y=mx+b</i> (where x=calculated concentration, y=indicated concentration)							
NO		LIMITS		NOx			
Correlation=	1.0000	≥ 0.990		Correlation=	1.0000		
m (Slope)=	1.0136	0.90-1.10		m (Slope)=	1.0121		
b (Intercept % of FS)=	-0.0509	± 3% F.S.		b (Intercept % of FS)=	-0.0577		

LINEAR REGRESSION ANALYSIS							
<i>y=mx+b</i> (where x=calculated concentration, y=indicated concentration)							
NO₂		LIMITS					
Correlation=	1.0000	≥ 0.995					
m (Slope)=	1.0014	0.90-1.10					
b (Intercept % of FS)=	-0.1619	± 3% F.S.					
Absolute Average Percent Difference							
1% ± 10%							

AENV Standards				NO_x Analyzer			
Audit Calibrator							
Make/Model	Thermo 146i			Make/Model	Thermo 42i		
Serial/AMU Number	1809			Serial/AMU Number	1868		
SRM Gas Cylinder No.	APEX1170572			Last Calibration Date	August 16, 2018		
Cylinder Conc. (ppm)	49.99			Full Scale (ppm)	1.0		
				Cylinder Gas Expiry Date	November 15, 2020		

COMMENTS:

 Auditor: Shea Beaton

 Date: August 21, 2018

 Operator Signature: SB

 Location: McIntyre Center Edmonton

Calibrator Performance Audit

Oxides Of Nitrogen

File No. 2018-131A

Company Maxxam	Operator: Mike
Calibrator: Make/Model Sabio Serial Number 17100415 Last Verification Date May 16, 2017 NO Cylinder S/N LL104183 NO [PPM] 50.8 NOx [PPM] 50.9 Expiry Date October 24, 2020	
Flow Measurement Device: Make/Model Bios Definer 220 Serial Number H=128686; L=129069 Temperature (°C) 22.2 C Barometric Pressure 706.1mmHg	
Dilution Flow (sccm) Pt. #1 5120 Pt. #2 5121 Pt. #3 5128 Gas Flow (sccm) Pt. #1 77.4 Pt. #2 37.8 Pt. #3 19	

Calibrator Flow (sccm)		Calculated Conc.(ppm)		Indicated Conc.(ppm)			% Difference vs Audit Gas	
Dilution	Gas	NO	NOx	NO	NO ₂	NOx	NO	NOx
5136	0.0	0.0000	0.0000	0.0001	-0.0002	0.0001	Limit ± 10%	
5120	77.4	0.7680	0.7695	0.7793	0.0003	0.7796	1%	1%
5121	37.8	0.3750	0.3757	0.3802	0.0000	0.3802	1%	1%
5128	19.0	0.1882	0.1885	0.1908	0.0005	0.1909	1%	1%
Absolute Average Percent Difference							1%	1%

LINEAR REGRESSION ANALYSIS								
$y=mx+b$ (where x=calculated concentration, y=indicated concentration)								
NO			LIMITS			NOx		
Correlation= 1.0000			≥ 0.990			Correlation= 1.0000		
m (Slope)= 1.0146			0.90-1.10			m (Slope)= 1.0130		
b (Intercept % of FS)= -0.0074			± 3% F.S.			b (Intercept % of FS)= -0.0059		


Flow	O ₃ Conc	NO Decrease	NO	NO ₂	NOX	% Diff. Vs Audit gas
5120	0.0	0.0000	0.7794	0.0005	0.7799	NO ₂ % Diff. Limit
5120	500.0	0.4827	0.2967	0.4854	0.7806	0% ± 10%
5120	275.0	0.2672	0.5122	0.2676	0.7798	0% ± 10%
5120	90.0	0.0896	0.6898	0.0890	0.7787	-1% ± 10%
Absolute Average Percent Difference						0% ± 10%

LINEAR REGRESSION ANALYSIS								
$y=mx+b$ (where x=calculated concentration, y=indicated concentration)								
NO₂			LIMITS					
Correlation= 1.0000			≥ 0.995					
m (Slope)= 1.0053			0.90-1.10					
b (Intercept % of FS)= -0.0370			± 3% F.S.					

AENV Standards			NO _x Analyzer		
Audit Calibrator			Make/Model Thermo 42i		
Make/Model Thermo 146i			Serial/AMU Number 1868		
Serial/AMU Number 1809			Last Calibration Date August 16, 2018		
SRM Gas Cylinder No. APEX1170572			Full Scale (ppm) 1.0		
Cylinder Conc. (ppm) 49.99			Cylinder Gas Expiry Date November 15, 2020		

COMMENTS: _____

Auditor: Shea Beaton _____ Date: August 21, 2018

Operator Signature:  Location: McIntyre Center Edmonton

CALIBRATION GASES



Calibration Gas Audit

Single Component Cylinder Gas

File No. 2017-486CGA

Company: Maxxam

Operator's Name: Mike

Cylinder #: LL108015 Concentration PPM: 47.9 Tolerance(%) 2 Certified By: Praxair

Expiry Date: October 2020

Reference Calibrator and Gas:

Make/Model: R&R MFC 201

Serial Number: AMU 1690

Last Verification Date: December 13, 2017

Gas Type: SO₂ Conc. 98.07

Cylinder Number: CAL016625

Expiry Date: January 2019

Flow Measurement Device:

Make/Model: Mesa Definer 220

Serial Number: H-133034 / L-132702

Temp. °C: 23.4 C

B.P. 707 mmHg

Reference Analyzer:

Make/Model: Teco 43C Serial/AMU Number: 1623

Instrument Settings: Zero: 10.0 Span: 1.006 Range: 1.0

Last Calibration: Date: Dec12/17 C.F. 1.000 Done By: Al Clark

Calibrator Flows (sccm)		Indicated Concentration (PPM)	Gas Flow/ Dilution Flow	Concentration Factor	Cylinder Concentration
Dilution	Gas				
5000	0.0	0.000	X	X	X
4989	79.5	0.760	0.01594	62.755	47.7
4995	39.6	0.374	0.00793	126.136	47.2
4992	19.6	0.183	0.00393	254.694	46.6
Average Cylinder Concentration:					47.2

Previous Stated Concentration PPM: 47.9

Percent variance from Stated: 2

Meets Manufacturer Tolerance. Use manufacturers stated concentration COMMENTS: _____

<=5% Outside Manufacturer Tolerance. Use manufacturers concentration _____

>5% Outside Manufacturer Tolerance. **DO NOT USE** this cylinder _____

Auditor: Al Clark

Date: December 13, 2017

Operator Signature:

Location: McIntyre Center Edmonton



Calibration Gas Audit Single Component Cylinder Gas

File No. 2017-137CGA

Company: Maxxam

Operator's Name: Raja Abid Ashraf

Cylinder #: LL119432 Concentration PPM: 10.3 Tolerance(%) 2 Certified By: Praxair

Expiry Date: May 16, 2020

Reference Calibrator and Gas:

Make/Model: R&R MFC 201

Serial Number: AMU 1690

Last Verification Date: July 27, 2017

Gas Type: H2S Conc. 20.43

Cylinder Number: CAL015272

Expiry Date: Janauary 2019

Flow Measurement Device:

Make/Model: Mesa Definer 220

Serial Number: H-133034 L-132702

Temp. °C: 22.0 C

B.P. 700 mmhg

Reference Analyzer:

Make/Model: Teco 450i

Serial/AMU Number: 1980

Instrument Settings: Zero: 21.9

Span: 1.069 Range: 0.1

Last Calibration: Date: July 27, 2017

C.F. 1.000 Done By: Al Clark

Calibrator Flows (sccm)		Indicated Concentration (PPM)	Gas Flow/ Dilution Flow	Concentration Factor	Cylinder Concentration
Dilution	Gas				
5000	0.0	0.0000	X	X	X
5117	38.9	0.0595	0.00760	131.542	7.8
5103	18.4		0.00361	277.337	0.0
5097	9.4		0.00184	542.234	0.0
Average Cylinder Concentration:					2.6

Previous Stated Concentration PPM: 10.3

Percent variance from Stated: 75

Meets Manufacturer Tolerance. Use manufacturers stated concentration COMMENTS: _____

<=5% Outside Manufacturer Tolerance. Use manufacturers concentration Do not use.

> 5% Outside Manufacturer Tolerance. DO NOT USE this cylinder X

Auditor: Al Clark

Date: July 27, 2017

Operator Signature:

Location: McIntyre Center Edmonton



Calibration Gas Audit

CH4 / C3H8 Cylinder Gas

Form No. F-GAS-004
Version No. 1.1

File No. 2017-484CGA

Company: Maxxam Operators name: Mike
Cylinder #: LL107207 Conc CH4 (PPM) 600/207 Tolerance (%) 2 Certified By: Praxair

Expiry Date: October 2025

Reference Calibrator and Gas:		Flow Measurement Device:	
Make/Model	R&R MFC 201	Make/Model	Mesa Definer 220
Serial Number	AMU 1690	Serial Number	H-133034 / L-132702
Last Verification Date	December 13, 2017	Temp. °C	23.1 C
Gas Type	CH4	B.P.	707 mmHg
Cylinder Number	5604875	Expiry Date	July 2021
Gas Type	C3H8	Conc.	246.5
Cylinder Number	XF003845B	Expiry Date	July 2022

Reference Analyzer:			
Make/Model	Teco 55i	Serial/AMU Number:	2108
Instrument Settings	Zero: N/A	Span: N/A	Range: 20.0
Last Calibration:	Date: Dec 12/17	C.F.	1.000 Done By: Al Clark

Calibrator Flows (sccm)	Indicated Conc. (ppm)		Gas Flow/ Dilution Flow	Concentration Factor	Cylinder Concentration	
	Dilution	Gas	CH4	C3H8	CH4	C3H8
3500	0.0	0.00	0.00	X	X	X
3618	80.4	13.28	12.77	0.02	45.00	598
3547	39.8	6.71	6.47	0.01	89.12	598
3560	19.8	3.35	3.26	0.01	179.80	602
Average Cylinder Concentration:					599	211

CH4

Previous Stated Concentration PPM: 600

C3H8

207

Percent variance from Stated: 0

2

Cylinder gas tolerances based on CH4 only

Meets Manufacturer Tolerance. Use manufacturers stated concentration COMMENTS:

<=5% Outside Manufacturer Tolerance. Use manufacturers concentration

>5% Outside Manufacturer Tolerance. DO NOT USE this cylinder

Auditor: Al Clark

Date: December 13, 2017

Operator Signature:

Location: McIntyre Center Edmonton

APPENDIX III
MAXIMUM INSTANTANEOUS DATA



PEACE RIVER AREA MONITORING PROGRAM COMMITTEE

Three Creeks 986b Station - December 2018

SULPHUR DIOXIDE Instantaneous Maximum (SO₂ ppb)

HR START (MST) HR END (MST)	0:00 0:59	1:00 1:59	2:00 2:59	3:00 3:59	4:00 4:59	5:00 5:59	6:00 6:59	7:00 7:59	8:00 8:59	9:00 9:59	10:00 10:59	11:00 11:59	12:00 12:59	13:00 13:59	14:00 14:59	15:00 15:59	16:00 16:59	17:00 17:59	18:00 18:59	19:00 19:59	20:00 20:59	21:00 21:59	22:00 22:59	23:00 23:59	DAILY MIN.	DAILY MAX.	24-HR AVG.	RDGS.
DAY																												
1	2	2	2	1	2	2	S	2	2	2	2	2	2	2	1	1	2	2	2	2	2	2	2	2	1	2	2	24
2	2	2	2	2	2	S	1	2	2	2	2	2	2	2	2	2	2	2	2	2	3	2	2	2	1	3	2	24
3	2	2	2	2	S	2	2	2	2	2	2	2	2	2	2	2	2	2	2	2	2	1	2	2	2	1	2	24
4	2	1	2	S	2	2	2	2	1	2	2	2	1	1	2	2	1	2	2	2	2	2	2	2	1	2	2	24
5	2	2	S	2	1	2	2	1	1	2	2	2	2	2	2	2	2	2	2	2	2	2	1	1	1	2	2	24
6	2	S	2	1	2	2	2	1	1	1	2	1	1	2	1	2	2	2	2	2	2	1	2	1	1	2	2	24
7	S	2	2	2	2	2	2	2	1	1	2	1	1	2	1	1	2	1	1	2	1	S	1	1	1	2	2	24
8	2	2	2	2	2	1	2	2	3	2	2	2	2	2	2	2	2	2	2	2	1	S	1	1	1	3	2	24
9	2	2	2	2	2	2	2	1	1	2	2	2	2	2	2	2	1	2	1	1	2	S	2	2	1	2	2	24
10	2	2	2	3	2	2	2	2	2	1	2	2	1	2	2	2	2	2	2	2	S	2	2	2	1	3	2	24
11	2	2	2	2	2	2	2	2	1	2	2	2	2	2	2	2	2	3	2	2	2	2	2	1	1	3	2	24
12	1	2	1	2	2	2	2	2	2	2	2	2	2	2	2	C	C	C	C	C	2	1	1	1	1	1	2	24
13	1	1	1	1	2	1	1	2	1	1	1	2	2	1	1	2	1	S	1	1	1	1	2	1	1	2	1	24
14	1	1	1	1	1	1	1	1	1	2	1	1	2	1	2	1	S	1	3	1	2	1	1	1	1	3	1	24
15	1	1	1	1	1	1	1	2	1	1	1	1	1	1	1	1	1	0	1	1	1	1	0	2	1	24		
16	1	1	1	0	1	1	0	1	1	0	1	1	1	1	S	1	1	1	1	1	1	1	0	0	1	1	24	
17	1	1	1	0	1	1	1	1	1	1	1	1	1	1	S	2	2	1	2	2	1	1	1	1	0	2	1	24
18	1	1	1	1	1	1	1	1	1	1	1	1	1	S	1	1	1	1	1	2	2	3	3	2	1	3	1	24
19	2	2	2	1	2	1	1	2	1	1	S	2	1	1	1	1	1	1	1	1	1	1	1	1	2	1	24	
20	1	1	1	1	1	1	1	1	S	2	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	2	1	24
21	1	1	1	1	1	1	1	1	2	S	1	2	1	1	2	1	1	1	1	1	1	1	2	1	1	2	1	24
22	1	1	1	1	1	1	1	1	S	1	1	1	1	1	2	2	1	1	1	1	1	1	1	0	0	2	1	24
23	1	0	1	0	0	1	0	S	1	1	1	1	1	1	1	0	0	1	1	1	0	1	1	0	1	1	24	
24	1	0	1	0	0	0	1	S	1	1	0	1	1	1	1	1	1	0	0	1	1	1	1	0	1	1	24	
25	1	1	0	0	0	0	S	1	0	0	0	1	0	0	0	0	0	0	1	0	0	1	1	0	1	0	24	
26	0	0	0	1	S	0	1	0	0	0	0	1	0	0	0	0	0	0	0	0	0	0	0	0	1	0	24	
27	1	0	1	S	1	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	1	0	24	
28	0	0	S	0	0	0	0	1	0	1	1	0	0	0	0	0	0	0	0	0	0	0	0	0	1	0	24	
29	0	S	0	0	0	0	0	0	0	0	1	0	0	0	1	0	0	0	0	0	0	0	0	0	1	0	24	
30	S	0	1	1	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	1	0	0	S	0	1	0	24	
31	0	0	0	0	0	1	1	1	0	0	0	0	0	0	1	0	0	0	0	0	S	0	0	1	0	24		
HOURLY MAX	2	2	2	3	2	2	2	2	3	2	2	2	2	2	2	2	2	2	3	3	2	3	3	3	2			
HOURLY AVG	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1		

STATUS FLAG CODES

C	- MONTHLY CALIBRATION	Q	- QUALITY ASSURANCE
C1	- REPEAT CALIBRATION	R	- RECOVERY
Y	- MAINTENANCE	X	- MACHINE MALFUNCTION
S	- DAILY ZERO/SPAN CHECK	G	- OUT FOR REPAIR
S1	- REPEAT ZERO/SPAN CHECK	P	- POWER FAILURE

MONTHLY SUMMARY


NUMBER OF NON-ZERO READINGS:	554
MAXIMUM INSTANTANEOUS VALUE:	3 ppb
@ HOUR	20
ON DAY	2
IIZS CALIBRATION TIME:	32 hrs
MONTHLY CALIBRATION TIME:	5 hrs
STANDARD DEVIATION:	1
OPERATIONAL TIME:	744 hrs



PEACE RIVER AREA MONITORING PROGRAM COMMITTEE

Three Creeks 986b Station - December 2018

SULPHUR DIOXIDE Instantaneous Maximum (SO₂ ppb)





PEACE RIVER AREA MONITORING PROGRAM COMMITTEE

Three Creeks 986b Station - December 2018

TOTAL REDUCED SULPHUR Instantaneous Maximum (TRS ppb)

HR START (MST) HR END (MST)	0:00 0:59	1:00 1:59	2:00 2:59	3:00 3:59	4:00 4:59	5:00 5:59	6:00 6:59	7:00 7:59	8:00 8:59	9:00 9:59	10:00 10:59	11:00 11:59	12:00 12:59	13:00 13:59	14:00 14:59	15:00 15:59	16:00 16:59	17:00 17:59	18:00 18:59	19:00 19:59	20:00 20:59	21:00 21:59	22:00 22:59	23:00 23:59	DAILY MIN.	DAILY MAX.	24-HR AVG.	RDGS.	
DAY																													
1	0.22	0.26	0.26	0.28	0.23	0.23	S	0.32	0.28	0.28	0.19	0.17	0.18	0.17	0.17	0.19	0.21	0.22	0.28	0.20	0.28	0.23	0.24	0.21	0.21	0.17	0.32	0.23	24
2	0.22	0.24	0.23	0.20	0.21	S	0.33	0.23	0.21	0.20	0.21	0.21	0.21	0.21	0.20	0.21	0.24	0.23	0.23	0.24	0.25	0.27	0.22	0.21	0.20	0.33	0.23	24	
3	0.23	0.24	0.24	0.21	S	0.34	0.25	0.23	0.27	0.24	0.26	0.27	0.33	0.34	0.48	0.50	0.24	0.26	0.21	0.23	0.22	0.25	0.23	0.22	0.21	0.50	0.27	24	
4	0.21	0.24	0.23	S	0.30	0.26	0.21	0.21	0.22	0.26	0.18	0.21	0.36	0.20	0.22	0.21	0.27	0.23	0.23	0.25	0.26	0.25	0.18	0.36	0.24	24			
5	0.26	0.25	S	0.37	0.30	0.27	0.25	0.29	0.26	0.27	0.26	0.25	0.26	0.29	0.23	0.25	0.24	0.24	0.22	0.23	0.24	0.25	0.30	0.25	0.22	0.37	0.26	24	
6	0.26	S	0.38	0.30	0.30	0.25	0.28	0.30	0.24	0.31	0.27	0.25	0.28	0.27	0.28	0.24	0.26	0.25	0.24	0.26	0.25	0.24	0.23	0.23	0.38	0.27	24		
7	S	0.45	0.29	0.29	0.30	0.27	0.26	0.28	0.28	0.26	0.29	0.30	0.24	0.24	0.24	0.27	0.26	0.29	0.26	0.27	0.24	S	0.24	0.45	0.28	24			
8	0.40	0.35	0.32	0.31	0.30	0.29	0.30	0.32	0.28	0.26	0.30	0.28	0.31	0.28	0.29	0.31	0.25	0.25	0.24	0.25	S	0.36	0.24	0.40	0.30	24			
9	0.28	0.29	0.28	0.29	0.27	0.29	0.28	0.23	0.25	0.26	0.28	0.27	0.22	0.25	0.25	0.25	0.22	0.23	0.31	0.26	2.20	S	0.41	0.26	0.22	2.20	0.35	24	
10	0.26	0.23	0.21	0.28	0.27	0.27	0.25	0.25	0.25	0.22	0.22	0.23	0.19	0.21	0.23	0.22	0.20	0.21	0.24	0.22	S	0.35	0.23	0.25	0.19	0.35	0.24	24	
11	0.22	0.25	0.24	0.22	0.26	0.26	0.20	0.21	0.19	0.17	0.21	0.15	0.19	0.18	0.17	0.18	0.19	0.15	S	0.26	0.19	0.16	0.15	0.26	0.20	24			
12	0.20	0.16	0.17	0.19	0.18	0.16	0.17	0.21	0.16	C	C	C	C	C	C	C	0.35	0.32	0.25	S	0.43	0.25	0.22	0.21	0.23	0.16	0.43	0.23	24
13	0.22	0.18	0.23	0.18	0.23	0.20	0.25	0.19	0.24	0.23	0.20	0.19	0.22	0.23	0.24	0.19	0.17	S	0.31	0.21	0.19	0.21	0.20	0.21	0.17	0.31	0.21	24	
14	0.20	0.18	0.19	0.20	0.23	0.20	0.22	0.23	0.23	0.29	0.26	0.22	0.22	0.22	0.20	0.23	S	0.32	0.24	0.36	0.29	0.22	0.26	0.34	0.18	0.36	0.24	24	
15	0.29	0.28	0.30	0.39	0.23	0.36	0.35	0.28	0.28	0.38	0.19	0.21	0.26	0.23	0.24	S	0.39	0.34	0.30	0.25	0.54	0.57	0.46	0.39	0.19	0.57	0.33	24	
16	0.43	0.30	0.31	0.32	0.32	0.27	0.28	0.27	0.25	0.28	0.30	0.26	0.32	0.32	S	0.42	0.34	0.24	0.26	0.25	0.34	0.43	0.30	0.31	0.43	0.31	24		
17	0.28	0.29	0.26	0.28	0.27	0.33	0.29	0.26	0.27	0.27	0.30	0.28	0.31	S	0.36	0.34	0.31	0.39	0.29	0.24	0.30	0.34	0.32	0.33	0.24	0.39	0.30	24	
18	0.30	0.28	0.30	0.29	0.30	0.27	0.25	0.27	0.31	0.29	0.29	0.25	S	0.32	0.26	0.22	0.24	0.22	0.36	0.22	0.23	0.27	0.28	0.22	0.36	0.27	24		
19	0.26	0.26	0.27	0.27	0.28	0.30	0.30	0.27	0.23	0.27	0.21	S	0.31	0.26	0.23	0.26	0.20	0.22	0.23	0.23	0.25	0.26	0.20	0.31	0.25	24			
20	0.22	0.23	0.23	0.22	0.22	0.26	0.24	0.28	0.28	0.24	S	0.40	0.28	0.26	0.21	0.22	0.23	0.23	0.25	0.22	0.22	0.23	0.22	0.21	0.21	0.40	0.25	24	
21	0.21	0.22	0.22	0.25	0.29	0.23	0.27	0.27	0.22	S	0.33	0.28	0.23	0.23	0.25	0.23	0.24	0.20	0.21	0.26	0.21	0.20	0.23	0.21	0.20	0.33	0.24	24	
22	0.28	0.39	0.38	0.24	0.21	0.29	0.47	0.30	S	0.51	0.30	0.34	0.30	0.29	0.29	0.28	0.29	0.33	0.28	0.33	0.35	0.30	0.39	0.21	0.51	0.32	24		
23	0.45	0.55	0.63	0.44	0.45	0.41	S	0.66	0.37	0.33	0.37	0.35	0.33	0.32	0.35	0.31	0.32	0.30	0.36	0.35	0.42	0.42	0.43	0.30	0.66	0.40	24		
24	0.46	0.35	0.35	0.31	0.31	0.49	S	0.51	0.42	0.35	0.36	0.34	0.35	0.36	0.37	0.44	0.43	0.53	0.58	0.58	0.58	0.70	0.71	0.59	0.31	0.71	0.46	24	
25	0.44	0.43	0.30	0.39	0.37	S	0.52	0.39	0.39	0.41	0.39	0.42	0.38	0.36	0.38	0.37	0.35	0.37	0.35	0.33	0.36	0.43	0.43	0.72	0.30	0.72	0.40	24	
26	0.68	0.53	0.50	0.55	S	0.78	0.44	0.41	0.41	0.52	0.61	0.39	0.43	0.42	0.35	0.37	0.40	0.47	0.37	0.42	0.42	0.44	0.38	0.38	0.78	0.46	24		
27	0.38	0.42	0.49	S	0.73	0.60	0.56	0.43	0.41	0.39	0.39	0.39	0.40	0.37	0.37	0.40	0.38	0.39	0.36	0.38	0.39	0.39	0.38	0.36	0.73	0.43	24		
28	0.40	0.41	S	0.55	0.38	0.44	0.41	0.39	0.39	0.37	0.42	0.39	0.41	0.42	0.43	0.39	0.36	0.41	0.38	0.36	0.36	0.38	0.34	0.34	0.55	0.40	24		
29	0.38	S	0.59	0.40	0.63	0.74	0.54	0.55	0.41	0.45	0.45	0.47	0.45	0.47	0.43	0.44	0.42	0.41	0.39	0.39	0.45	0.47	0.45	0.38	0.74	0.47	24		
30	S	0.55	0.46	0.44	0.44	0.47	0.44	0.50	0.44	0.44	0.43	0.44	0.44	0.45	0.43	0.42	0.43	0.43	0.44	0.45	0.47	0.45	S	0.42	0.55	0.45	24		
31	0.58	0.47	0.46	0.47	0.48	0.47	0.45	0.44	0.43	0.40	0.42	0.39	0.37	0.35	0.40	0.39	0.39	0.35	0.38	0.36	S	0.45	0.35	0.58	0.42	24			
HOURLY MAX	0.68	0.55	0.63	0.55	0.73	0.78	0.56	0.55	0.66	0.52	0.61	0.47	0.45	0.47	0.48	0.50	0.43	0.53	0.58	0.58	2.20	0.70	0.71	0.72					
HOURLY AVG	0.32	0.32	0.32	0.31	0.32	0.35	0.33	0.31	0.31	0.32	0.30	0.30	0.30	0.30	0.30	0.29	0.30	0.30	0.30	0.30	0.38	0.33	0.31	0.32					


STATUS FLAG CODES

C	- MONTHLY CALIBRATION	Q	- QUALITY ASSURANCE
C1	- REPEAT CALIBRATION	R	- RECOVERY
Y	- MAINTENANCE	X	- MACHINE MALFUNCTION
S	- DAILY ZERO/SPAN CHECK	G	- OUT FOR REPAIR
S1	- REPEAT ZERO/SPAN CHECK	P	- POWER FAILURE

MONTHLY SUMMARY

NUMBER OF NON-ZERO READINGS:	705
MAXIMUM INSTANTANEOUS VALUE:	2.20 ppb
@ HOUR	20
ON DAY	9
Izs Calibration Time:	33 hrs
Monthly Calibration Time:	6 hrs
Standard Deviation:	0.13
Operational Time:	744 hrs

TOTAL REDUCED SULPHUR Instantaneous Maximum (TRS ppb)





PEACE RIVER AREA MONITORING PROGRAM COMMITTEE

Three Creeks 986b Station - December 2018

TOTAL HYDROCARBONS Instantaneous Maximum (THC ppm)

HR START (MST) HR END (MST)	0:00 0:59	1:00 1:59	2:00 2:59	3:00 3:59	4:00 4:59	5:00 5:59	6:00 6:59	7:00 7:59	8:00 8:59	9:00 9:59	10:00 10:59	11:00 11:59	12:00 12:59	13:00 13:59	14:00 14:59	15:00 15:59	16:00 16:59	17:00 17:59	18:00 18:59	19:00 19:59	20:00 20:59	21:00 21:59	22:00 22:59	23:00 23:59	DAILY MIN.	DAILY MAX.	24-HR AVG.	RDGS.
DAY																												
1	2.31	2.35	2.02	2.26	2.22	2.13	S	2.04	2.07	2.06	2.08	2.04	2.06	2.02	2.02	2.08	2.03	2.49	2.44	2.05	2.06	2.07	2.25	2.21	2.02	2.49	2.15	24
2	2.32	2.06	2.17	2.22	2.32	S	2.11	2.13	2.10	2.04	2.08	1.99	1.97	2.04	1.98	2.00	1.97	2.00	2.00	2.00	2.01	2.02	2.02	2.03	1.97	2.32	2.07	24
3	2.06	2.10	2.03	2.07	S	2.05	2.02	2.02	2.05	2.02	2.07	2.27	2.80	2.50	2.29	2.03	2.00	1.99	1.99	1.98	1.98	1.97	1.96	1.96	1.96	2.80	2.11	24
4	1.98	1.99	2.01	S	2.01	2.03	2.01	2.16	S1	S1	2.07	2.05	2.05	2.19	2.09	2.05	2.04	2.09	2.04	2.06	2.10	2.10	2.12	2.22	1.98	2.22	2.07	22
5	2.23	3.18	S	2.00	2.03	2.00	2.13	2.03	2.30	2.16	1.99	1.99	2.07	2.07	1.99	2.01	2.00	2.00	1.98	1.99	2.00	2.65	1.98	1.99	1.98	3.18	2.12	24
6	1.99	S	2.00	2.01	2.02	2.02	2.02	2.01	2.02	2.03	2.03	2.02	2.02	2.02	2.01	2.02	2.02	2.00	2.01	2.02	2.03	2.03	2.03	1.99	2.03	2.02	24	
7	S	2.05	2.06	2.06	2.09	2.07	2.06	2.04	2.03	2.04	2.04	2.02	2.02	2.03	2.03	2.03	2.04	2.03	2.03	2.02	2.01	2.01	2.01	S	2.00	2.09	2.04	24
8	2.05	2.06	2.05	2.06	2.06	2.14	2.07	2.05	2.05	2.04	2.08	2.04	2.05	2.08	2.06	2.06	2.06	2.06	2.06	2.04	2.04	2.01	2.01	2.14	2.06	24		
9	2.07	2.01	2.05	2.05	2.05	2.06	2.06	2.04	2.05	2.06	2.02	2.03	2.03	2.02	2.02	2.05	2.05	2.03	2.04	5.29	S	2.02	2.06	2.01	5.29	2.24	24	
10	2.07	2.08	2.05	2.06	2.06	2.05	2.03	2.02	2.00	1.99	1.98	1.98	1.97	1.97	1.97	1.98	1.97	1.98	1.95	S	1.96	1.96	1.97	1.95	2.08	2.00	24	
11	1.97	1.98	1.97	1.96	1.96	1.96	1.95	1.98	1.96	1.95	1.95	1.95	1.92	1.91	1.94	1.91	1.93	1.92	1.91	S	1.90	1.91	1.89	1.91	1.98	1.94	24	
12	1.90	1.90	1.91	1.91	1.91	1.91	1.91	C	C	C	Y	Y	Y	Y	C	C	C	C	2.01	2.00	2.02	1.99	2.01	1.90	2.02	1.94	20	
13	2.01	1.99	2.01	1.98	2.00	1.99	2.01	1.99	2.00	1.98	2.00	2.01	2.04	2.08	2.10	2.06	2.02	S	2.00	2.02	2.01	2.01	2.02	2.00	1.98	2.10	2.01	24
14	1.99	2.00	2.01	1.99	2.01	1.99	2.01	2.01	2.00	2.03	2.00	1.99	1.98	1.99	1.97	1.99	S	2.00	2.01	2.00	3.20	2.12	2.14	2.34	1.97	3.20	2.08	24
15	2.27	2.46	2.31	2.16	2.21	2.15	2.19	2.16	2.12	2.09	2.13	2.23	2.39	2.39	2.49	S	2.07	2.03	2.05	2.06	2.11	2.14	2.18	2.25	2.03	2.49	2.20	24
16	2.21	2.13	2.09	3.19	2.47	2.06	2.04	2.03	2.01	2.03	2.01	2.01	2.01	2.01	2.01	2.01	2.01	2.01	2.05	2.02	2.05	2.02	2.05	2.01	3.19	2.11	24	
17	2.04	2.06	2.07	2.08	2.07	2.11	2.26	2.06	2.08	2.05	2.08	2.06	2.10	S	2.11	2.09	2.12	2.12	2.09	2.08	2.13	2.19	2.10	2.10	2.04	2.26	2.10	24
18	2.10	2.12	2.11	2.11	3.97	2.07	2.36	2.08	2.05	2.06	2.04	S	2.04	2.02	2.05	2.03	2.07	2.07	2.08	2.07	2.10	2.07	2.09	2.02	3.97	2.17	24	
19	2.10	2.08	2.09	2.07	2.07	2.10	2.13	2.07	2.07	2.04	S	2.02	2.03	2.00	2.03	2.00	2.02	2.03	2.02	2.02	2.01	2.03	2.03	2.02	2.13	2.05	24	
20	2.01	2.02	2.04	2.01	2.04	2.03	2.06	2.09	2.05	2.07	S	2.03	2.02	2.01	1.99	2.01	1.99	2.00	1.99	2.01	2.01	2.25	1.99	2.25	2.03	24		
21	2.37	2.25	2.24	2.24	2.17	2.10	2.20	2.08	2.11	S	2.32	2.73	2.30	2.31	2.36	2.39	2.43	2.98	3.04	2.86	2.64	2.16	2.05	3.61	2.43	24		
22	2.18	2.07	2.03	2.04	2.03	2.10	2.11	2.15	S	2.15	2.15	2.14	2.08	2.05	2.07	2.05	2.05	2.05	2.08	2.09	2.11	2.64	2.15	2.03	2.64	2.12	24	
23	2.17	2.18	2.27	2.66	3.02	2.67	2.09	S	2.10	2.10	2.23	2.18	2.28	2.15	2.15	2.18	2.07	2.10	2.06	2.03	2.06	2.03	2.08	2.03	3.02	2.21	24	
24	2.03	2.06	2.07	2.05	2.07	2.05	S	2.05	2.08	2.07	2.09	2.09	2.11	2.13	2.14	2.15	2.15	2.20	2.19	2.24	2.24	2.26	2.23	2.18	2.03	2.26	2.13	24
25	2.17	2.24	2.17	2.25	2.06	S	2.07	2.18	2.41	2.49	2.62	2.43	2.27	2.42	2.53	2.65	2.76	2.09	2.07	2.08	2.07	2.11	2.16	2.20	2.06	2.76	2.28	24
26	2.23	2.22	2.16	2.14	S	2.15	2.13	2.14	2.17	2.48	2.14	2.09	2.15	2.13	2.06	2.11	2.13	2.15	2.92	2.32	2.31	2.10	2.10	2.06	2.92	2.20	24	
27	3.18	3.18	2.10	S	2.30	2.19	2.37	2.39	2.13	2.38	2.14	2.23	2.11	2.11	2.10	2.09	2.10	2.09	2.16	2.08	2.07	3.18	2.25	24				
28	2.08	2.08	S	2.13	2.11	2.12	2.12	2.14	2.12	2.11	2.13	2.12	3.42	3.11	2.40	2.96	2.80	3.07	2.59	2.54	2.50	2.26	4.27	2.50	24			
29	2.09	S	2.35	2.15	2.14	2.14	2.14	2.13	2.12	2.10	2.11	2.12	2.11	2.13	2.14	2.12	2.12	2.13	2.12	2.15	2.20	2.17	2.08	2.08	2.35	2.14	24	
30	S	2.10	2.17	2.22	2.14	2.30	2.25	2.25	2.42	2.65	2.50	2.77	2.95	2.43	2.22	2.13	2.13	2.13	2.16	2.17	2.14	2.14	S	2.10	2.95	2.30	24	
31	2.14	2.14	2.14	2.14	2.15	2.14	2.14	2.15	2.14	2.13	2.13	2.12	2.11	2.11	2.10	2.10	2.10	2.07	2.06	2.05	S	2.07	2.05	2.15	2.12	24		
HOURLY MAX	3.18	3.18	2.35	3.19	3.02	3.97	2.37	2.39	2.42	2.65	2.62	2.77	3.42	3.11	2.53	2.96	2.80	3.37	3.04	2.86	5.29	2.65	4.27	3.61				
HOURLY AVG	2.15	2.18	2.09	2.15	2.13	2.16	2.10	2.10	2.09	2.12	2.11	2.15	2.19	2.15	2.12	2.12	2.11	2.18	2.14	2.10	2.24	2.10	2.16	2.15	2.12	2.12	24	


STATUS FLAG CODES

C	- MONTHLY CALIBRATION	Q	- QUALITY ASSURANCE
C1	- REPEAT CALIBRATION	R	- RECOVERY
Y	- MAINTENANCE	X	- MACHINE MALFUNCTION
S	- DAILY ZERO/SPAN CHECK	G	- OUT FOR REPAIR
S1	- REPEAT ZERO/SPAN CHECK	P	- POWER FAILURE

MONTHLY SUMMARY

NUMBER OF NON-ZERO READINGS:	700
MAXIMUM INSTANTANEOUS VALUE:	5.29 ppm @ HOUR 20 ON DAY 9
Izs Calibration Time:	32 hrs
Monthly Calibration Time:	6 hrs
Standard Deviation:	0.27

TOTAL HYDROCARBONS Instantaneous Maximum (THC ppm)





PEACE RIVER AREA MONITORING PROGRAM COMMITTEE

Three Creeks 986b Station - December 2018


METHANE MAX Instantaneous Maximum (CH₄ ppm)

HR START (MST) HR END (MST)	0:00 0:59	1:00 1:59	2:00 2:59	3:00 3:59	4:00 4:59	5:00 5:59	6:00 6:59	7:00 7:59	8:00 8:59	9:00 9:59	10:00 10:59	11:00 11:59	12:00 12:59	13:00 13:59	14:00 14:59	15:00 15:59	16:00 16:59	17:00 17:59	18:00 18:59	19:00 19:59	20:00 20:59	21:00 21:59	22:00 22:59	23:00 23:59	DAILY MIN.	DAILY MAX.	24-HR AVG.	RDGS.	
DAY																													
1	2.31	2.35	2.02	2.26	2.22	2.13	S	2.04	2.07	2.06	2.08	2.04	2.06	2.02	2.02	2.08	2.03	2.49	2.44	2.05	2.06	2.07	2.25	2.21	2.02	2.49	2.15	24	
2	2.32	2.06	2.17	2.22	2.32	S	2.11	2.13	2.10	2.04	2.08	1.99	1.97	2.04	1.98	2.00	1.97	2.00	2.00	2.00	2.01	2.02	2.02	2.03	1.97	2.32	2.07	24	
3	2.06	2.10	2.03	2.07	S	2.05	2.02	2.02	2.05	2.02	2.02	2.72	2.80	2.50	2.29	2.03	2.00	1.99	1.99	1.98	1.98	1.97	1.96	1.96	1.96	2.80	2.11	24	
4	1.98	1.99	2.01	S	2.01	2.03	2.16	S1	S1	2.07	2.05	2.05	2.19	2.09	2.05	2.04	2.09	2.04	2.06	2.10	2.10	2.12	2.22	1.98	2.22	2.07	22		
5	2.23	3.18	S	2.00	2.03	2.00	2.13	2.03	2.30	2.16	1.99	1.99	2.07	2.07	1.99	2.01	2.00	2.00	1.98	1.99	2.00	2.65	1.98	1.99	1.98	3.18	2.12	24	
6	1.99	S	2.00	2.01	2.02	2.02	2.01	2.02	2.03	2.03	2.02	2.02	2.02	2.01	2.02	2.02	2.00	2.01	2.00	2.02	2.03	2.03	1.99	2.03	2.02	24			
7	S	2.05	2.06	2.06	2.09	2.07	2.06	2.04	2.03	2.04	2.04	2.02	2.02	2.03	2.03	2.03	2.04	2.03	2.03	2.02	2.00	2.01	S	2.00	2.09	2.04	24		
8	2.05	2.06	2.05	2.06	2.06	2.09	2.07	2.05	2.05	2.04	2.04	2.05	2.04	2.05	2.07	2.06	2.06	2.06	2.06	2.04	2.01	2.01	2.01	2.09	2.05	24			
9	2.07	2.01	2.03	2.05	2.05	2.05	2.06	2.06	2.04	2.05	2.04	2.02	2.03	2.03	2.02	2.02	2.05	2.37	2.03	2.04	3.92	S	2.02	2.06	2.01	3.92	2.18	24	
10	2.07	2.07	2.05	2.06	2.06	2.05	2.03	2.02	2.00	1.99	1.98	1.98	1.97	1.97	1.97	1.97	1.98	1.98	1.95	S	1.96	1.96	1.97	1.95	2.07	2.00	24		
11	1.97	1.98	1.97	1.96	1.96	1.96	1.95	1.98	1.96	1.95	1.95	1.92	1.91	1.91	1.91	1.93	1.92	1.92	S	1.90	1.91	1.89	1.91	1.89	1.98	1.94	24		
12	1.90	1.90	1.91	1.91	1.91	1.90	1.92	1.91	C	C	C	Y	Y	Y	Y	C	C	C	C	2.01	2.00	2.02	1.99	2.01	1.90	2.02	1.94	20	
13	2.01	1.99	2.01	1.98	2.00	1.99	2.01	1.99	2.00	1.98	2.00	2.01	2.04	2.08	2.10	2.06	2.06	2.02	S	2.00	2.02	2.01	2.02	2.00	1.98	2.10	2.01	24	
14	1.99	2.00	2.01	1.99	2.01	1.99	2.01	2.01	2.00	2.03	2.00	1.99	1.98	1.99	1.97	1.99	S	2.00	2.01	2.00	3.20	3.20	2.12	2.14	2.34	1.97	3.20	2.08	24
15	2.27	2.46	2.31	2.16	2.21	2.15	2.19	2.16	2.12	2.09	2.13	2.23	2.39	2.39	2.49	S	2.07	2.03	2.05	2.06	2.11	2.14	2.18	2.25	2.03	2.49	2.20	24	
16	2.21	2.13	2.09	3.19	2.47	2.06	2.04	2.03	2.01	2.03	2.04	2.01	2.03	2.01	2.01	2.03	2.01	2.03	2.02	2.05	2.05	2.02	2.05	2.01	3.19	2.11	24		
17	2.04	2.06	2.07	2.08	2.07	2.11	2.26	2.06	2.08	2.05	2.08	2.06	2.10	S	2.11	2.09	2.12	2.12	2.09	2.08	2.10	2.14	2.10	2.10	2.04	2.26	2.09	24	
18	2.09	2.12	2.11	2.11	3.97	2.07	2.36	2.08	2.05	2.06	2.04	S	2.04	2.02	2.05	2.03	2.07	2.07	2.08	2.07	2.10	2.07	2.09	2.02	3.97	2.17	24		
19	2.10	2.08	2.09	2.07	2.07	2.10	2.13	2.07	2.07	2.04	2.04	S	2.02	2.03	2.00	2.03	2.00	2.02	2.02	2.02	2.03	2.00	2.13	2.05	24				
20	2.01	2.02	2.04	2.01	2.04	2.03	2.06	2.09	2.05	2.07	S	2.03	2.02	2.01	1.99	2.01	2.01	1.99	2.00	1.99	2.01	1.99	2.25	1.99	2.25	2.03	24		
21	2.37	2.25	2.24	2.24	2.17	2.10	2.20	2.08	2.11	S	2.32	2.73	2.30	2.31	2.36	2.39	2.43	2.98	3.04	2.86	2.64	2.16	2.05	3.61	2.43	24			
22	2.18	2.07	2.03	2.04	2.03	2.10	2.11	2.15	S	2.15	2.15	2.14	2.08	2.05	2.07	2.05	2.05	2.05	2.08	2.09	2.11	2.64	2.12	2.64	2.12	24			
23	2.17	2.18	2.27	2.66	3.02	2.67	2.09	S	2.10	2.10	2.23	2.18	2.28	2.15	2.15	2.18	2.07	2.10	2.06	2.06	2.03	2.08	2.03	3.02	2.21	24			
24	2.03	2.06	2.07	2.05	2.07	S	2.05	2.08	2.07	2.09	2.09	2.11	2.13	2.14	2.15	2.15	2.20	2.19	2.21	2.24	2.26	2.23	2.18	2.03	2.26	2.13	24		
25	2.17	2.24	2.17	2.25	2.06	S	2.07	2.18	2.41	2.49	2.62	2.43	2.27	2.42	2.53	2.65	2.65	2.76	2.09	2.07	2.08	2.07	2.11	2.16	2.20	2.06	2.76	2.28	24
26	2.23	2.22	2.16	2.14	S	2.15	2.13	2.14	2.17	2.48	2.14	2.09	2.15	2.13	2.06	2.11	2.13	2.15	2.92	2.32	2.31	2.10	2.10	2.06	2.92	2.20	24		
27	3.18	3.18	2.10	S	2.30	2.19	2.37	2.39	2.13	2.38	2.14	2.23	2.11	2.11	2.10	2.09	2.09	2.16	2.08	2.07	3.18	2.25	2.07	3.18	2.25	24			
28	2.08	2.08	S	2.13	2.11	2.12	2.12	2.14	2.12	2.11	2.13	2.12	3.42	3.11	2.40	2.96	2.80	3.07	2.59	2.54	2.50	2.26	4.27	2.25	2.08	4.27	2.50	24	
29	2.09	S	2.35	2.15	2.14	2.14	2.14	2.13	2.12	2.10	2.11	2.12	2.11	2.13	2.14	2.12	2.13	2.12	2.12	2.15	2.20	2.17	2.08	2.08	2.35	2.14	24		
30	S	2.10	2.17	2.22	2.14	2.30	2.25	2.25	2.42	2.65	2.50	2.77	2.95	2.43	2.22	2.13	2.13	2.13	2.16	2.14	2.14	S	2.10	2.95	2.30	2.95	2.30	24	
31	2.14	2.14	2.14	2.15	2.14	2.14	2.15	2.14	2.14	2.13	2.13	2.12	2.11	2.09	2.10	2.10	2.07	2.06	2.05	S	2.07	2.05	2.05	2.15	2.12	24			
HOURLY MAX	3.18	3.18	2.35	3.19	3.02	3.97	2.37	2.39	2.42	2.65	2.62	2.77	3.42	3.11	2.53	2.96	2.80	3.37	3.04	2.86	3.92	2.65	4.27	3.61					
HOURLY AVG	2.15	2.18	2.09	2.15	2.13	2.16	2.10	2.10	2.09	2.12	2.11	2.15	2.19	2.15	2.12	2.12	2.11	2.18	2.14	2.10	2.20	2.09	2.16	2.15	2.15	2.12	24		

STATUS FLAG CODES	
C	- MONTHLY CALIBRATION
C1	- REPEAT CALIBRATION
Y	- MAINTENANCE
S	- DAILY ZERO/SPAN CHECK
S1	- REPEAT ZERO/SPAN CHECK
Q	- QUALITY ASSURANCE
R	- RECOVERY
X	- MACHINE MALFUNCTION
G	- OUT FOR REPAIR
P	- POWER FAILURE

MONTHLY SUMMARY	
NUMBER OF NON-ZERO READINGS:	700
MAXIMUM INSTANTANEOUS VALUE:	4.27 ppm @ HOUR 22 ON DAY 28
Izs Calibration Time:	32 hrs
Monthly Calibration Time:	6 hrs
Standard Deviation:	0.25
Operational Time:	738 hrs

METHANE MAX Instantaneous Maximum (CH_4 ppm)





PEACE RIVER AREA MONITORING PROGRAM COMMITTEE

Three Creeks 986b Station - December 2018


NON-METHANE HYDROCARBONS Instantaneous Maximum (NMHC ppm)

	HR START (MST)	0:00	1:00	2:00	3:00	4:00	5:00	6:00	7:00	8:00	9:00	10:00	11:00	12:00	13:00	14:00	15:00	16:00	17:00	18:00	19:00	20:00	21:00	22:00	23:00	DAILY MIN.	DAILY MAX.	24-HR AVG.	RDGS.
	HR END (MST)	0:59	1:59	2:59	3:59	4:59	5:59	6:59	7:59	8:59	9:59	10:59	11:59	12:59	13:59	14:59	15:59	16:59	17:59	18:59	19:59	20:59	21:59	22:59	23:59				
DAY																													
1		0.00	0.00	0.00	0.00	0.00	0.00	S	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	24	
2		0.00	0.00	0.00	0.00	0.00	0.00	S	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	24	
3		0.00	0.00	0.00	0.00	0.00	S	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	24	
4		0.00	0.00	0.00	0.00	S	0.00	0.00	0.00	0.00	S1	S1	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	22	
5		0.00	0.00	0.00	S	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	24	
6		0.00	S	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	24	
7		S	0.00	0.00	0.00	0.00	0.01	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.01	0.00	0.00	24		
8		0.00	0.00	0.00	0.01	0.00	0.07	0.00	0.00	0.00	0.04	0.00	0.00	0.01	0.00	0.00	0.00	0.00	0.00	0.00	0.00	S	0.00	0.00	0.00	0.07	0.01	24	
9		0.00	0.00	0.04	0.00	0.00	0.00	0.00	0.00	0.00	0.03	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	1.37	S	0.00	0.00	0.00	1.37	0.06	24	
10		0.00	0.03	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	S	0.00	0.00	0.00	0.00	0.03	0.00	24	
11		0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	S	0.00	0.00	0.00	0.00	0.00	0.00	24		
12		0.00	0.00	0.00	0.00	0.00	0.00	0.00	C	C	C	Y	Y	Y	Y	C	C	C	C	0.00	0.00	0.00	0.00	0.00	0.00	0.00	20		
13		0.00	0.00	0.00	0.00	0.00	0.00	0.00	S	0.00	0.00	0.00	0.00	0.00	0.00	0.00	S	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	24		
14		0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	S	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	24		
15		0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	S	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	24		
16		0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	S	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	24		
17		0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	S	0.00	0.00	0.00	0.00	0.05	0.08	0.00	0.00	0.08	0.01	24	
18		0.02	0.00	0.00	0.00	0.00	0.00	0.00	0.01	0.00	S	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.02	0.00	24
19		0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	S	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	24
20		0.00	0.00	0.00	0.00	0.00	0.00	0.00	S	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	24	
21		0.00	0.00	0.00	0.00	0.00	0.00	0.00	S	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	24	
22		0.00	0.00	0.00	0.00	0.00	0.00	S	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	24	
23		0.00	0.00	0.00	0.00	0.00	S	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	24	
24		0.00	0.00	0.00	0.00	S	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.05	0.00	0.00	0.00	0.00	0.05	0.00	0.00	24	
25		0.00	0.00	0.00	0.00	S	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	24	
26		0.00	0.00	0.00	S	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	24	
27		0.00	0.00	0.00	S	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	24	
28		0.00	0.00	S	0.00	0.00	0.01	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.01	0.00	24	
29		S	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	24	
30		S	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	S	0.00	0.00	0.00	0.00	0.00	0.00	24	
31		0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	S	0.00	0.00	0.00	0.00	0.00	0.00	24	
HOURLY MAX		0.02	0.03	0.04	0.01	0.00	0.07	0.00	0.01	0.00	0.00	0.04	0.00	0.00	0.01	0.00	0.00	0.00	0.00	0.05	1.37	0.08	0.00	0.00	0.00	0.00	0.00	0.00	24
HOURLY AVG		0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.05	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	24

STATUS FLAG CODES	
C	- MONTHLY CALIBRATION
C1	- REPEAT CALIBRATION
Y	- MAINTENANCE
S	- DAILY ZERO/SPAN CHECK
S1	- REPEAT ZERO/SPAN CHECK
Q	- QUALITY ASSURANCE
R	- RECOVERY
X	- MACHINE MALFUNCTION
G	- OUT FOR REPAIR
P	- POWER FAILURE

MONTHLY SUMMARY	
NUMBER OF NON-ZERO READINGS:	16
MAXIMUM INSTANTANEOUS VALUE:	1.37 ppm
@ HOUR	20
ON DAY	9
Izs Calibration Time:	32 hrs
Monthly Calibration Time:	6 hrs
Standard Deviation:	0.05
Operational Time:	738 hrs

NON-METHANE HYDROCARBONS Instantaneous Maximum (NMHC ppm)





PEACE RIVER AREA MONITORING PROGRAM COMMITTEE

Three Creeks 986b Station - December 2018

WIND SPEED Instantaneous Maximum (WS kph)

HR START (MST) HR END (MST)	0:00 0:59	1:00 1:59	2:00 2:59	3:00 3:59	4:00 4:59	5:00 5:59	6:00 6:59	7:00 7:59	8:00 8:59	9:00 9:59	10:00 10:59	11:00 11:59	12:00 12:59	13:00 13:59	14:00 14:59	15:00 15:59	16:00 16:59	17:00 17:59	18:00 18:59	19:00 19:59	20:00 20:59	21:00 21:59	22:00 22:59	23:00 23:59	DAILY MIN.	DAILY MAX.	24-HR AVG.	RDGS.
DAY																												
1	3.1	5.0	5.8	9.6	4.7	5.7	6.6	5.2	8.2	6.5	6.3	7.4	6.8	8.5	8.0	9.0	8.7	4.6	7.3	6.9	8.6	6.2	4.0	4.1	3.1	9.6	6.5	24
2	7.0	5.8	6.5	6.2	4.5	11.3	10.2	8.9	8.5	7.0	7.6	12.7	12.2	12.1	11.1	11.9	12.6	12.3	11.2	14.0	14.6	16.5	12.2	10.8	4.5	16.5	10.3	24
3	12.8	13.8	10.4	11.4	9.9	10.2	7.9	3.8	4.1	2.1	2.3	5.9	8.8	5.3	4.2	0.1	0.1	0.1	9.4	11.6	13.5	12.7	13.5	0.1	13.8	7.3	24	
4	12.8	14.4	15.5	15.2	15.5	9.5	15.0	12.5	13.8	17.5	17.3	17.0	17.0	14.6	40.0	48.2	36.0	37.6	37.5	27.5	19.7	28.4	21.0	16.0	9.5	48.2	21.6	24
5	14.1	4.7	3.2	3.9	5.5	6.8	6.1	6.8	9.1	5.7	8.5	11.3	11.9	14.5	14.8	11.0	9.6	13.3	12.1	10.9	8.0	6.2	7.7	12.6	3.2	14.8	9.1	24
6	10.9	9.2	8.3	7.4	8.6	7.3	6.5	7.0	6.2	6.1	9.7	11.0	14.1	14.4	13.8	16.2	13.9	18.6	14.6	16.4	18.2	14.6	11.7	12.8	6.1	18.6	11.6	24
7	23.6	24.9	24.2	22.3	16.1	16.7	15.5	9.5	8.7	10.8	11.5	13.0	16.1	15.9	14.2	14.5	10.3	7.5	9.4	9.7	9.8	9.5	10.2	15.6	7.5	24.9	14.2	24
8	13.9	7.7	10.7	11.0	15.7	17.1	20.8	19.9	25.2	26.7	29.2	30.7	27.3	28.3	21.6	18.9	12.1	12.4	15.1	17.0	18.2	19.6	24.5	7.7	30.7	19.6	24	
9	19.4	20.9	17.3	14.2	10.8	14.6	13.1	13.8	15.1	10.3	11.4	8.5	15.6	11.2	12.4	9.5	9.3	4.1	7.3	6.2	10.3	13.6	13.0	13.8	4.1	20.9	12.3	24
10	14.9	12.5	14.5	15.4	16.6	20.9	18.8	22.5	20.8	16.2	19.8	24.8	29.0	25.8	29.1	20.4	17.8	13.6	16.8	16.8	22.6	16.8	17.4	17.8	12.5	29.1	19.2	24
11	14.0	13.2	11.8	20.1	23.6	21.1	23.2	19.7	20.3	24.1	24.0	28.8	26.3	25.7	32.8	26.6	27.5	22.9	25.4	26.4	31.9	42.6	36.9	36.2	11.8	42.6	25.2	24
12	26.4	29.0	37.5	33.1	35.9	34.4	22.5	27.1	31.9	29.3	29.8	34.6	38.7	35.3	41.0	33.0	35.5	27.9	40.1	21.5	21.9	20.5	17.2	22.1	17.2	41.0	30.3	24
13	26.1	26.1	26.1	30.7	38.2	30.3	21.6	22.7	26.7	28.8	31.0	35.1	32.7	30.8	33.1	41.0	39.7	37.6	28.1	21.5	16.8	17.9	20.8	22.2	16.8	41.0	28.6	24
14	13.9	19.1	23.8	17.3	13.3	16.5	16.0	10.6	9.9	10.4	13.2	18.7	20.7	20.6	12.9	9.8	12.4	11.8	13.7	14.2	6.3	14.0	11.1	13.1	6.3	23.8	14.3	24
15	14.4	16.9	18.6	37.6	15.5	24.2	25.1	21.3	25.2	37.6	40.1	27.6	34.9	30.5	23.1	9.8	6.0	5.7	8.0	8.6	5.7	4.8	3.8	9.9	3.8	40.1	19.0	24
16	9.9	11.3	6.5	7.6	6.4	13.8	25.6	28.4	27.8	28.2	27.4	27.3	26.0	21.5	18.9	19.7	23.0	15.1	18.0	20.2	14.3	12.1	16.9	15.4	6.4	28.4	18.4	24
17	15.5	13.8	13.5	13.9	19.8	13.0	10.9	14.9	17.8	16.4	23.8	25.2	27.2	23.7	20.2	16.4	14.6	10.5	15.9	17.1	15.4	10.8	6.3	8.1	6.3	27.2	16.0	24
18	14.8	9.0	14.3	10.1	11.2	11.7	15.0	8.4	4.6	7.4	10.1	15.4	15.9	14.6	25.1	15.2	13.1	11.0	15.0	9.1	10.5	13.3	19.9	18.0	4.6	25.1	13.0	24
19	20.6	13.8	19.9	17.6	21.5	16.3	11.1	19.2	23.5	24.6	25.5	30.4	30.4	27.2	30.3	23.9	26.6	25.8	24.8	20.6	19.6	16.3	18.1	18.2	11.1	30.4	21.9	24
20	14.4	16.4	10.1	10.6	7.2	11.4	12.6	10.1	10.8	11.0	10.5	12.8	17.3	21.3	21.6	25.1	36.8	34.6	29.1	25.8	19.7	15.7	8.5	8.8	7.2	36.8	16.8	24
21	13.3	15.1	16.1	22.4	30.1	25.9	25.4	37.1	35.6	25.9	26.4	30.3	36.2	26.5	33.4	19.8	11.0	8.6	6.5	8.5	6.1	4.1	3.5	4.7	3.5	37.1	19.7	24
22	4.9	6.6	3.6	3.6	3.0	4.3	3.9	2.8	3.3	4.8	3.2	2.4	3.8	4.2	4.3	10.3	6.4	5.2	3.1	3.9	2.6	4.1	3.0	3.4	2.4	10.3	4.2	24
23	3.2	3.4	3.6	3.8	3.1	5.0	6.9	8.1	6.2	7.5	5.7	5.8	7.4	8.2	6.4	7.4	5.3	7.2	6.1	5.1	7.6	5.0	5.0	5.8	3.1	8.2	5.8	24
24	5.0	6.0	7.0	8.1	7.5	7.6	7.2	5.5	7.7	9.5	10.8	9.5	8.1	9.7	9.4	9.5	6.0	9.2	10.1	6.7	6.3	8.9	8.0	8.3	5.0	10.8	8.0	24
25	12.8	11.6	13.4	15.2	17.1	17.4	12.5	10.7	6.9	8.8	8.2	7.3	10.8	16.4	14.2	6.8	4.4	4.5	3.8	3.6	4.4	4.7	4.9	5.7	3.6	17.4	9.4	24
26	5.7	3.6	5.1	7.6	9.8	8.1	9.2	8.2	6.4	7.3	9.8	10.5	8.4	9.1	10.2	10.5	10.9	9.8	6.9	9.4	10.5	14.5	13.3	9.3	3.6	14.5	8.9	24
27	6.4	5.9	4.8	6.1	3.9	3.7	5.4	9.1	4.5	7.9	3.9	5.8	6.7	7.7	7.3	6.0	6.2	5.6	5.6	6.6	5.8	6.7	4.7	6.1	3.7	9.1	5.9	24
28	6.5	8.2	6.5	3.5	4.4	9.7	13.5	13.1	12.9	9.3	9.2	8.1	2.4	5.8	6.3	5.3	6.4	5.0	7.8	5.9	7.3	4.6	5.2	5.5	2.4	13.5	7.2	24
29	8.1	9.3	4.9	5.3	4.6	4.9	9.1	9.7	13.4	22.0	20.7	19.5	18.5	20.3	25.6	24.9	22.8	19.5	22.5	19.7	19.2	21.3	25.6	24.9	4.6	25.6	16.5	24
30	24.5	20.2	25.5	21.1	23.7	18.2	17.6	14.6	12.5	9.8	7.7	8.0	10.4	9.7	6.4	8.0	10.3	10.9	14.3	18.9	17.1	13.6	13.1	12.1	6.4	25.5	14.5	24
31	9.4	9.5	9.6	10.9	13.9	11.4	11.7	14.9	18.1	14.2	15.6	17.8	23.0	27.1	21.5	17.3	15.3	16.0	19.4	25.1	28.9	32.3	36.6	26.2	9.4	36.6	18.6	24
HOURLY MAX	26.4	29.0	37.5	37.6	38.2	34.4	25.6	37.1	35.6	37.6	40.1	35.1	38.7	35.3	41.0	48.2	39.7	37.6	40.1	27.5	31.9	42.6	36.9	36.2				


STATUS FLAG CODES

C	- MONTHLY CALIBRATION	Q	- QUALITY ASSURANCE
C1	- REPEAT CALIBRATION	R	- RECOVERY
Y	- MAINTENANCE	X	- MACHINE MALFUNCTION
S	- DAILY ZERO/SPAN CHECK	G	- OUT FOR REPAIR
S1	- REPEAT ZERO/SPAN CHECK	P	- POWER FAILURE

MONTHLY SUMMARY

MAXIMUM INSTANTANEOUS VALUE:	48.2	kph	@ HOUR	15	ON DAY	4
OPERATIONAL TIME:						

WIND SPEED Instantaneous Maximum (WS kph)



APPENDIX IV
REPORT CERTIFICATION FORM

Report Certification Form

Alberta Airshed (if applicable)	EPA Approval or Code of Practice Registration # (if applicable)
YES	N/A
Company Name (if applicable)	Industrial Operation Name (if applicable)
Peace River Area Monitoring Program Committee	Three Creeks 986b Station
Name of the Representative of the Person Responsible	Position / Title of the Representative of the Person Responsible
Mike Bisaga / Lily Lin	Technical Program Managers
Is an External Party Certifying the Report? If 'Yes', fill in the fields below for the external person.	
<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No	
Name of External Person Certifying the Report	Position / Title of External Person Certifying the Report
Cheri Sinclair	Supervisor, Customer Service, Air Services
Company Name for the External Person Certifying the Report	Identification of Qualifications / Professional Designations of the External Person Certifying the Report
Maxxam Analytics, A Bureau Veritas Group Company	B.Sc.

Maxxam Analytics is the designated contractor conducting monitoring and reporting activities. I certify that the submitted data has been (a) reviewed and validated as per the AMD Chapter 6: Ambient Data Quality. I certify that the submitted report (b) accurately reflects the monitoring results and reporting timeframe and (c) meets the specified analysis, summarization and reporting requirements as per the AMD Chapter 9: Reporting.



Signature of the External Person Certifying the Report

23 - Jan - 2019

Report Issued Date (dd-mon-yyyy)

APPENDIX V
DATA VALIDATION CERTIFICATION FORM



Validation Certificate Form

Client: Peace River Area Monitoring Program Committee
Site: Three Creeks 986b Station

Project #: 8449-2018-12-02-C
Contact: Lily Lin

Level 0 Preliminary Verification

Date 18 - Jan - 2019

Level 1 Primary Validation

Date 18 - Jan - 2019

Level 2 Final Validation

Date 23 - Jan - 2019

Level 3 Independent Data Review

Date 23 - Jan - 2019

Post-Final Validation

NA

Date NA

Notes

The Post-Final Validation step serves to re-evaluate the data that errors or omissions are discovered and/or suspected after the initial submittal of data. This validation is performed on an annual basis.