



Peace River Area Monitoring Program

DECEMBER 2023
Monthly Ambient Air Quality Monitoring Report
PRAMP-202312

Operation and Maintenance:

Bureau Veritas Canada

Data Validation and Report:

Peace River Area Monitoring Program

January 12, 2024

Pages may be left blank for double-sided printing

Table of Contents

LIST OF ACRONYMS	4
COVER LETTER	5
NETWORK STATION SUMMARY	6
Listing of Continuous Monitoring Stations	6
Listing of Intermittent Monitoring Stations.....	6
Listing of PRAMP member with EPEA Facility Operating Approval	6
Calibration and Data Submission	6
Monitoring Notes during the Month of December 2023	7
986-C Station	7
842-B Station	7
Reno-B Station	7
PRC Station	7
AQHI – Grimshaw Station	7
VOCs Canister Sampling Program	8
Revisions to Alberta’s Ambient Air Quality Data Warehouse.....	9
Deviations from Authorized Monitoring Methods	9
Disclaimer.....	9
Certification.....	10
Map of PRAMP Continuous Monitoring Network	11
CONTINUOUS NETWORK EQUIPMENT AND MONITORING RESULTS SUMMARY	12
986 -C Station.....	13
842-B Station.....	16
Reno-B Station	21
PRC Station.....	25
AQHI – Grimshaw Station	29
TABLES, CHARTS AND WIND ROSES	34
986-C STATION	35
842-B STATION	55
RENO -B STATION.....	75
PRC STATION	95
AQHI GRIMSHAW STATION.....	116
END OF REPORT	146

LIST OF ACRONYMS

AAAQOs	Alberta Ambient Air Quality Objectives
AEP	Alberta Environment and Parks
AMD	Air Monitoring Directive
AT	Ambient Temperature
BP	Barometric Pressure
CH ₄	Methane
EPEA	Environmental Protection and Enhancement Act
H ₂ S	Hydrogen Sulphide
kph	kilometers per hour
mb	millibar
mm	millimeter
NMHC	Non-Methane Hydrocarbons
ppb	parts per billion
ppm	parts per million
PRAMP	Peace River Area Monitoring Program
RH	Relative Humidity
SO ₂	Sulphur Dioxide
ST	Station Temperature
THC	Total Hydrocarbons
TRS	Total Reduced Sulphur
VWD	Vector Wind Direction
VWS	Vector Wind Speed
WD	Wind Direction
WS	Wind Speed
°C	Degrees Celsius



Peace River Area Monitoring Program
Suite 91, 305 – 4625 Varsity Drive NW
Calgary, AB, T3A 0Z9
Phone #: 780-226-7068 / 587-225-2248
E-mail: pramptech@prampairshed.ca
www.prampairshed.ca

Alberta Environment and Protected Areas (EPA)
11th Floor, Oxbridge Place
9820 106 Street
Edmonton, AB, T5K 2J6

January 12, 2024

RE: PRAMP – December 2023 Monthly Ambient Air Quality Monitoring Report

Enclosed is the December 2023 Monthly Ambient Air Quality Monitoring Report for the continuous ambient air quality monitoring stations of the Peace River Area Monitoring Program (PRAMP) regional air quality monitoring network.

The representative of the Person Responsible for this monitoring program is

PRAMP Airshed
Michael Bisaga / Lily Lin, Technical Program Managers
Suite 91, 305 – 4625 Varsity Drive NW
Calgary, AB, T3A 0Z9
Phone #: 780-226-7068 / 587-225-2248
E-mail: pramptech@prampairshed.ca

This report has been prepared, review and submitted by Michael Bisaga & Lily Lin of the PRAMP Airshed. This report is also submitted on behalf of the industrial member companies to satisfy the requirements of the facility operating approvals.

PRAMP Airshed has retained the services of Bureau Veritas Canada to conduct continuous ambient monitoring on its behalf.

NETWORK STATION SUMMARY

Listing of Continuous Monitoring Stations

The PRAMP continuous ambient air quality monitoring network stations are:

- 986-C Station
- 842-B Station
- Reno-B Station
- AQHI Grimshaw
- Peace River Complex (PRC) Station

Station ID	Station Name	Latitude	Longitude
1562	986-C	56.36980	-116.92500
1561	842-B	56.27406	-116.98129
1563	Reno-B	55.890868	-117.137080
1689	AQHI-Grimshaw	56.18657	-117.604994
1698	PRC	56.38257	-116.769283

Listing of Intermittent Monitoring Stations

- VOC Canister Sampling Station
 - 986-C Station
 - 842-B Station
 - Reno-B Station

Listing of PRAMP member with EPEA Facility Operating Approval

Company	Facility	Approval No.
Canadian Natural Upgrading Limited	Peace River Complex	1642-03-00

Calibration and Data Submission

Hourly data and calibration reports for December 2023 were submitted to Alberta's Ambient Air Data Warehouse through ETS for the 986-C station, 842-B station, Reno-B station, PRC station and AQHI-Grimshaw station.

Monitoring Notes during the Month of December 2023

986-C Station

- Measured parameters were below Alberta Ambient Air Quality Objectives (AAAQOs) where applicable.
- All data collected this month were compliant with the requirements outlined in the AMD 2016.
- All parameters met the 90% operational uptime requirement.
- No major operational issues were recorded this month.

842-B Station

- Measured parameters were below Alberta Ambient Air Quality Objectives (AAAQOs) where applicable.
- All data collected this month were compliant with the requirements outlined in the AMD 2016.
- All parameters met the 90% operational uptime requirement.
- No major operational issues were recorded this month.

Reno-B Station

- Measured parameters were below Alberta Ambient Air Quality Objectives (AAAQOs) where applicable.
- All data collected this month were compliant with the requirements outlined in the AMD 2016.
- All parameters met the 90% operational uptime requirement.
- No major operational issues were recorded this month.

PRC Station

- Measured parameters were below Alberta Ambient Air Quality Objectives (AAAQOs) where applicable.
- All data collected this month were compliant with the requirements outlined in the AMD 2016.
- All parameters met the 90% operational uptime requirement.
- No major operational issues were recorded this month.

AQHI – Grimshaw Station

- All data collected this month were compliant with the requirements outlined in the AMD 2016.
- All parameters met the 90% operational uptime requirement.
- Measured parameters were below Alberta Ambient Air Quality Objectives (AAAQOs) and /or Alberta Ambient Air Quality Guidelines (AAAQGs) where applicable, except PM2.5. One 1-hour PM2.5 exceedance was recorded this month. The recorded exceedances were believed to be the result of local buildup of emissions.

Date	Time (MST)	Parameter	Average Period	Concentration (µg/m ³)	Wind speed (km/hr)	Wind Direction	Reference #
9-Dec	0	PM2.5	1-Hour	200	14.0	278°(W)	EDGE00422947

- PM2.5: A firmware update and an alignment factor was activated on the T640 PM analyzer was completed on December 20 hour 13.

VOCs Canister Sampling Program

- The canister sampling program collects a 1-hour sample of air when the continuously measured non-methane hydrocarbon (NMHC) concentration reaches a specified trigger point. The current trigger point is 0.3 ppm for non-methane hydrocarbons. The trigger point is based on real-time monitoring data that are averaged over a 5-minute period.
- The canister sample collection systems are in place at Station 986-C, 842-B, and the Reno-B Station; a canister sample collection system is not part of the suite of instruments currently deployed at both the PRC station and the AQHI-Grimshaw station.
- Sample analysis and analytical results were prepared and provided by InnoTech Alberta.
- There were no canister events recorded this month. The canister system was triggered at the 986-C station on December 26 at 03:35 due to an analyzer issue. The event is not valid.

Revisions to Alberta's Ambient Air Quality Data Warehouse

No revisions to historical data previously submitted to the Alberta's Ambient Air Quality Data Warehouse were made this month.

Deviations from Authorized Monitoring Methods

No deviations from authorized monitoring methods were recorded this month.

Disclaimer

Baseline corrections were performed on the 1-minute data. 5-minute and hourly data were calculated based on the post-baseline correction 1-minute data set. Data verification/validation were then performed on the 5-minute and hourly data. Hourly data that are included in this report are the post-validation hourly data set.

Equipment calibration / maintenance records were provided by Bureau Veritas.

Certification

This report was prepared and submitted by Lily Lin in accordance with Chapter 9 of the Air Monitoring Directive (AMD 2016).



Lily Lin, Technical Program Manager, PRAMP Airshed


This report was reviewed by Michael Bisaga in accordance with Chapter 9 of the Air Monitoring Directive (AMD 2016).

I certify that I have reviewed and verified this report and that the information is complete, accurate and representative of the monitoring results, reporting timeframe and the specified analysis, summarization and reporting requirements. I also certify that at the time of this report's submission, all air data have been electronically uploaded to Alberta's Ambient Air Quality Data Warehouse as required by the AMD. Uploading of VOC data from the canister sampling program was not required at the time of completing this report.



Michael Bisaga, Technical Program Manager, PRAMP Airshed

Map of PRAMP Continuous Monitoring Network




Legend

- PRAMP Boundary
- Populated Place
- First Nation
- Continuous Monitoring Station
- Passive Monitoring Station

Industrial Facilities

- In-Situ Oil Sands
- Heavy Oil/Bitumen Well or Battery
- Conventional Oil Well or Battery
- Natural Gas Well or Battery
- Gas Plant or Gas Processing
- Compressor Station or Pipeline
- Agricultural Storage and Transfer
- Pulp and Paper
- Well (Not Associated with Batteries)



CONTINUOUS NETWORK EQUIPMENT AND MONITORING RESULTS
SUMMARY

986 -C Station

Equipment Operation Summary

Parameter	Equipment Operational Summary
<p>SO2</p> <p>Thermo 43iQTL #1193585646</p>	<ul style="list-style-type: none"> • A successful monthly calibration was performed on December 7. • No operational issues were identified this month.
<p>TRS</p> <p>Thermo 43iQTL #1191833341</p> <p>TRS convertor CD Nova CDN-101 #530 (BV-supplied)</p>	<ul style="list-style-type: none"> • A successful monthly calibration was performed on December 7. • No operational issues were identified this month.
<p>THC/CH4/NMHC</p> <p>Thermo 55i #12208316589</p> <p>H2 Generator HG300 #191267063</p>	<ul style="list-style-type: none"> • A successful monthly calibration was performed on December 7. • Few NMHC concentration spikes were recorded on December 26. As the spikes were very short duration (2-minutes) and high concentrations, the validity of these events is questionable. Data were invalidated as a result. • The canister system was triggered by a spike event at 03:35 on December 26. As the data were not believed to be real, the canister event is considered invalid.
<p>RH</p> <p>Rotronic HC2-S3 #20626912</p>	<ul style="list-style-type: none"> • The RH probe was checked on December 7. The probe passed the check requirements. • No operational issues were identified this month.
<p>BP</p> <p>MetOne 092 #Y23358</p>	<ul style="list-style-type: none"> • The BP sensor was checked on December 7. The sensor passed the check requirements. • No operational issues were identified this month.
<p>AT</p> <p>Rotronic HC2-S3 #20626912</p>	<ul style="list-style-type: none"> • The AT probe was checked on December 7. The probe passed the check requirements. • No operational issues were identified this month.

Parameter	Equipment Operational Summary
ST COMET #18961918	<ul style="list-style-type: none"> No operational issues were identified this month.
Precipitation RM Young 52202 #TB 16325	<ul style="list-style-type: none"> The precipitation gauge was checked on December 7. The gauge's functionality passed the check requirements. No operational issues were identified this month.
WS/ WD RM Young 05305AQ #180340	<ul style="list-style-type: none"> Wind direction data contained in this report represents where the wind is coming from. The annual wind system calibration was completed on August 3, 2023. The anemometer sensors were check on December 7. The wind system passed the check requirements. No operational issues were identified this month.

Monitored Data Summary

842-B Station

Equipment Operation Summary

Parameter	Equipment Operational Summary
SO2 Thermo 43iQTL #1200736629	<ul style="list-style-type: none"> • A successful monthly calibration was performed on December 5. • Data collected between December 19 hour 13 and December 20 hour 5 were lost due to a power failure. Seventeen hours of downtime were recorded due to this event.
TRS Thermo 43iQTL #1200736630 TRS Convertor CD Nova CDN-101 #583	<ul style="list-style-type: none"> • A successful monthly calibration was performed on December 5. • As marked change in response was noticed on the December 5's calibration, an as-found points check was completed on December 6 to confirm the December 5's calibration result. No issues were identified. Two hours of downtime were recorded due to the additional quality check. • Data collected between December 19 hour 13 and December 20 hour 5 were lost due to a power failure. Seventeen hours of downtime were recorded due to this event.
THC/CH4/NMHC Thermo 55i #1501663728 H2 Generator HG300 #190567058	<ul style="list-style-type: none"> • A successful monthly calibration was performed on December 5. • Multiple flame out events began after the December 5's calibration due to a problem with the H2 generator. A successful additional zero-span check was completed On December 6. The issue improved after the zero-span check. The H2 desiccant was changed following by a successful post-repair calibration on December 6. Data were invalidated back to the time the additional zero-span check was completed. Twenty-two hours of downtime were recorded due to this event. • Data collected between December 19 hour 13 and December 20 hour 5 were lost due to a power failure. Seventeen hours of downtime were recorded due to this event.
RH Rotronic HC2-S3 #20370767	<ul style="list-style-type: none"> • The RH probe was checked on December 5. The probe passed the check requirements. • Data collected between December 19 hour 13 and December 20 hour 5 were lost due to a power failure. Seventeen hours of downtime were recorded due to this event.

Parameter	Equipment Operational Summary
AT Rotronic HC2-S3 #20370767	<ul style="list-style-type: none"> • The AT probe was checked on December 5. The probe passed the check requirements. • Data collected between December 19 hour 13 and December 20 hour 5 were lost due to a power failure. Seventeen hours of downtime were recorded due to this event.
BP MetOne 092 #Y23362	<ul style="list-style-type: none"> • The BP sensor was checked on December 5. The sensor passed the check requirements. • One hour of downtime was recorded on December 17 due to Data collected between December 19 hour 13 and December 20 hour 5 were lost due to a power failure. Seventeen hours of downtime were recorded due to this event.
ST COMET #20790297	<ul style="list-style-type: none"> • Data collected between December 19 hour 13 and December 20 hour 5 were lost due to a power failure. Seventeen hours of downtime were recorded due to this event.
Precipitation RM Young 52202 #TB 15878	<ul style="list-style-type: none"> • The precipitation gauge was checked on December 5. The sensor passed the check requirements. • Data collected between December 19 hour 13 and December 20 hour 5 were lost due to a power failure. Seventeen hours of downtime were recorded due to this event.
WS/ WD RM Young 05305AQ #174802	<ul style="list-style-type: none"> • Wind direction data contained in this report represents where the wind is coming from. • The annual wind system calibration was completed on August 3, 2023. • The anemometer sensors were check on December 5. Both the wind speed sensor and wind direction sensor passed the check requirements. • Data collected between December 19 hour 13 and December 20 hour 5 were lost due to a power failure. Seventeen hours of downtime were recorded due to this event.

Monitored Data Summary

Reno-B Station

Equipment Operation Summary

Parameter	Equipment Operational Summary
SO2 Thermo 43iQTL #12101910505	<ul style="list-style-type: none"> • A successful monthly calibration was performed on December 1. • No operational issues were identified this month.
TRS Thermo 43iQTL #12101910504 TRS Convertor CD Nova CDN-101 #590	<ul style="list-style-type: none"> • A successful monthly calibration was performed on December 1. • No operational issues were identified this month.
THC/CH4/NMHC Thermo 55i #12101910497 H2 Generator HG300 #210467069	<ul style="list-style-type: none"> • A successful monthly calibration was performed on December 1. • No operational issues were identified this month.
RH Rotronic HC2-S3 #20467597	<ul style="list-style-type: none"> • The RH probe was checked on December 1. The probe passed the check requirements. • Due to datalogger polling errors, one hour of downtime was recorded as the hourly data completeness requirement did not meet.
BP MetOne 092 #A17940	<ul style="list-style-type: none"> • The BP sensor was checked on December 1. The sensor passed the check requirements. • Due to datalogger polling errors, one hour of downtime was recorded as the hourly data completeness requirement did not meet.

Parameter	Equipment Operational Summary
AT Rotronic HC2-S3 #20467597	<ul style="list-style-type: none"> • The AT probe was checked on December 1. The probe passed the check requirements. • Due to datalogger polling errors, one hour of downtime was recorded as the hourly data completeness requirement did not meet.
ST COMET #NA	<ul style="list-style-type: none"> • No operational issues were identified this month.
Precipitation RM Young 52202 #TB 15877	<ul style="list-style-type: none"> • The precipitation gauge was checked and tested on December 1. The unit passed the check requirements. • Due to datalogger polling errors, one hour of downtime was recorded as the hourly data completeness requirement did not meet.
WS/ WD RM Young 05305AQ #174795	<ul style="list-style-type: none"> • Wind direction data contained in this report represents where the wind is coming from. • The annual wind system calibration was completed on August 1, 2023. • The anemometer sensors were check on December 1. The wind sensors passed the check requirements. • Due to datalogger polling errors, one hour of downtime was recorded as the hourly data completeness requirement did not meet.

Monitored Data Summary

Equipment Operation Summary

Parameter	Equipment Operational Summary
<p>SO2</p> <p>Thermo 43i #1034746225</p>	<ul style="list-style-type: none"> • A successful monthly calibration was performed on December 6. • The scheduled zero-span check was interrupted by Windows update on December 18. A repeat zero-span check was completed later the day to confirm the analyzer’s functionality. The check results met the calibration check requirements. One hour of downtime was recorded as a result.
<p>H2S</p> <p>Thermo 450i #1308857354</p>	<ul style="list-style-type: none"> • A successful monthly calibration was performed on December 6. • The scheduled zero-span check was interrupted by Windows update on December 18. A repeat zero-span check was completed later the day to confirm the analyzer’s functionality. The check results met the calibration check requirements. One hour of downtime was recorded as a result.
<p>TRS</p> <p>Thermo 450i #1034746224</p> <p>TRS Convertor CD Nova CDN-101 #516</p>	<ul style="list-style-type: none"> • A successful monthly calibration was performed on December 6. • The scheduled zero-span check was interrupted by Windows update on December 18. A repeat zero-span check was completed later the day to confirm the analyzer’s functionality. The check results met the calibration check requirements. One hour of downtime was recorded as a result.
<p>THC/CH4/NMHC</p> <p>Thermo 55i #1034745845</p> <p>H2 Generator HG300 #211067076</p>	<ul style="list-style-type: none"> • The H2 desiccant was changed after the successful as-found points check on December 6. A successful monthly calibration was completed afterwards. • The span gas cylinder was installed on December 6. • The scheduled zero-span check was interrupted by Windows update on December 18. A repeat zero-span check was completed later the day to confirm the analyzer’s functionality. The check results met the calibration check requirements. One hour of downtime was recorded as a result.
<p>RH</p> <p>Rotronic HC2-S3 #20558318</p>	<ul style="list-style-type: none"> • The RH sensor was checked on December 6. The sensor passed the check requirements. • No operational issues were identified.
<p>BP</p> <p>MetOne 092 #B19577</p>	<ul style="list-style-type: none"> • The BP sensor was checked on December 6. The sensor passed the check requirements. • No operational issues were identified.

Parameter	Equipment Operational Summary
AT Rotronic HC2-S3 #20558318	<ul style="list-style-type: none"> • The AT sensor was checked on December 6. The sensor passed the check requirements. • No operational issues were identified.
ST Canadian Natural #NA	<ul style="list-style-type: none"> • No operational issues were identified.
WS/ WD RM Young 05305VK #129612	<ul style="list-style-type: none"> • Wind direction data contained in this report represents where the wind is coming from. • The annual wind system calibration was completed on August 3, 2023. • The anemometer sensors were checked on December 6. The sensors passed the check requirements. • No operational issues were identified.

Monitored Data Summary

AQHI – Grimshaw Station

Equipment Operation Summary

Parameter	Equipment Operational Summary
<p>SO2</p> <p>Teledyne T100 #722</p>	<ul style="list-style-type: none"> • A successful monthly calibration was performed on December 12. • Due to datalogger polling errors, three hours of downtime were recorded as the hourly data completeness requirement did not meet.
<p>TRS</p> <p>Teledyne T100U #132</p> <p>TRS Convertor CD Nova CDN-101 #576</p>	<ul style="list-style-type: none"> • A successful monthly calibration was performed on December 12. • Due to datalogger polling errors, three hours of downtime were recorded as the hourly data completeness requirement did not meet.
<p>NOx/NO/NO2</p> <p>API 200E #594</p> <p>Teledyne T200 #837</p>	<ul style="list-style-type: none"> • A successful monthly calibration was performed on December 12. • Due to datalogger polling errors, five hours of downtime were recorded as the hourly data completeness requirement did not meet.
<p>O3</p> <p>Teledyne T400 #824</p>	<ul style="list-style-type: none"> • A successful monthly calibration was performed on December 12. • Due to datalogger polling errors, five hours of downtime were recorded as the hourly data completeness requirement did not meet.
<p>THC/CH4/NMHC</p> <p>Thermo 55i #1191032505</p> <p>H2 Generator AMA HG300 #190567059</p>	<ul style="list-style-type: none"> • A successful monthly calibration was performed on December 12. • Due to datalogger polling errors, five hours of downtime were recorded as the hourly data completeness requirement did not meet.
<p>PM2.5</p> <p>Teledyne T640 #318</p>	<ul style="list-style-type: none"> • A successful monthly audit was performed on December 12. • A firmware update and an alignment factor activation were completed on the T640 PM analyzer on December 20 hour 13. One hour of downtime was recorded as a result. • Due to datalogger polling errors, five hours of downtime were recorded as the hourly data completeness requirement did not meet.

Parameter	Equipment Operational Summary
RH Vaisala HMP155 #N2910506	<ul style="list-style-type: none"> • The RH probe was checked on December 12. The Probe passed the check requirements. • Due to datalogger polling errors, seven hours of downtime were recorded as the hourly data completeness requirement did not meet.
BP MetOne 092 #A2397	<ul style="list-style-type: none"> • The BP sensor was checked on December 12. The sensor passed the check requirements. • Due to datalogger polling errors, seven hours of downtime were recorded as the hourly data completeness requirement did not meet.
AT Vaisala HMP155 #N2910506	<ul style="list-style-type: none"> • The AT prober was checked on December 12. The probe passed the check requirements. • Due to datalogger polling errors, seven hours of downtime were recorded as the hourly data completeness requirement did not meet.
ST COMET #NA	<ul style="list-style-type: none"> • Due to datalogger polling errors, three hours of downtime were recorded as the hourly data completeness requirement did not meet.
WS/ WD RM Young 05305AQ #174801	<ul style="list-style-type: none"> • Wind direction data contained in this report represents where the wind is coming from. • The last annual wind system calibration was completed on August 2, 2023. • The anemometer sensors were check on December 12. Both the wind speed sensor and wind direction sensor passed the check requirements. • Due to datalogger polling errors, seven hours of downtime were recorded as the hourly data completeness requirement did not meet.

Monitored Data Summary

TABLES, CHARTS AND WIND ROSES

986-C STATION

842-B STATION

RENO -B STATION

PRC STATION

AQHI GRIMSHAW STATION

END OF REPORT

This page, 146 of 146, ends the December 2023 Monthly Ambient Air Quality Monitoring Report.



Peace River Area Monitoring Program

DECEMBER 2023

Ambient Air Monitoring Calibration Report

- 842-B STATION-

CAL-PRAMP-202312-01561

Operation and Maintenance:

Bureau Veritas Canada

Data Validation and Report:

Bureau Veritas Canada

January 6, 2024


DAILY INTERNAL ZERO-SPAN CALIBRATION RECORDS

SO2[ppb] Calibration: PRAMP 842-B Monthly: 12-2023 Type: SpanAndZero - Zero




Legend: Zero (orange square), Zero Ref (red line), Zero Low (blue line), Zero High (purple line)

SO2[ppb] Calibration: PRAMP 842-B Monthly: 12-2023 Type: SpanAndZero - Span




Legend: Span (orange square), Span Ref (red line), Span Low (blue line), Span High (purple line)

TRS[ppb] Calibration: PRAMP 842-B Monthly: 12-2023 Type: SpanAndZero - Zero




Zero Zero Ref Zero Low Zero High

TRS[ppb] Calibration: PRAMP 842-B Monthly: 12-2023 Type: SpanAndZero - Span




Span SpanRef Span Low Span High

THC55[ppm] Calibration: PRAMP 842-B Monthly: 12-2023 Type: SpanAndZero - Zero




Zero Zero Ref Zero Low Zero High

THC55[ppm] Calibration: PRAMP 842-B Monthly: 12-2023 Type: SpanAndZero - Span




Span Span Ref Span Low Span High

CH4[ppm] Calibration: PRAMP 842-B Monthly: 12-2023 Type: SpanAndZero - Zero




Zero Zero Ref Zero Low Zero High

CH4[ppm] Calibration: PRAMP 842-B Monthly: 12-2023 Type: SpanAndZero - Span




Span SpanRef Span Low Span High

NMHC[ppm] Calibration: PRAMP 842-B Monthly: 12-2023 Type: SpanAndZero - Zero



Zero Zero Ref Zero Low Zero High

NMHC[ppm] Calibration: PRAMP 842-B Monthly: 12-2023 Type: SpanAndZero - Span



Span Span Ref Span Low Span High

MULTI-POINT CALIBRATION RECORDS

SO2 Analyzer Calibration by Dilution



DATE:	05-Dec-2023	PREVIOUS CALIBRATION DATE:	08-Nov-2023
PARAMETER:	SO2	PREVIOUS CORRECTION FACTOR:	1.002
CLIENT:	PRAMP	TEMPERATURE (°C):	22.6
LOCATION:	842b	BAROMETRIC (mBar):	927
PURPOSE:	Routine	START TIME (MST):	09:16
PERFORMED BY:	Kevin Sebastian	END TIME (MST):	14:09

ANALYZER:

MAKE/MODEL	Thermo 43iQTL	RANGE	500 ppb
SERIAL #	1200736629	FLOW (mL/min)	421
INITIAL		FINAL	
BKG/OFFSET	8.6	BKG/OFFSET	9.1
COEF/SLOPE	1.126	COEF/SLOPE	1.145
Expected (reference) Value	241.8	Expected (reference) Value	251.9

CALIBRATION SYSTEM:

CALIBRATOR:		ZERO AIR:	
MAKE:	Sabio	MAKE:	Teledyne
MODEL:	2010	MODEL:	T701
ID:	75401122	ID:	4568
MFC CALIBRATION DATE:	08-Sep-2023	OXIDIZER ID:	n/a
CALIBRATION GAS:		FLOWMETERS (if applicable):	
CYLINDER ID:	LL109693	HIGH ID	n/a
CONC (ppm):	25.00	EXPIRY DATE	n/a
CYLINDER (psi):	1500	LOW ID	n/a
EXPIRY DATE	02-Nov-2025	EXPIRY DATE	n/a

CALIBRATION PARAMETERS:

POINT	HIGH	MID	LOW
TARGET	390	190	95
RANGE	300 - 400	150 - 200	50 - 100

SCRUBBER CHECK (15 MINS; TRS/H2S ONLY):

START TIME:	n/a	SO2 Conc (ppb)	n/a
END TIME:	n/a	Analyzer Response (ppb)	n/a

CALIBRATION:


FLOW RATES (mL/min)			CONCENTRATION (ppb)			CORRECTION FACTOR	
DILUENT	GAS	TOTAL	ACTUAL	INDICATED		Initial	Final
				Initial	Final		
3999	60.80	3999	0.00	0.16	0	1.023	0.998
3938	60.80	3999	380.10	371.7	380.7	1.023	0.998
3971	28.80	4000	180.00	n/a	181.8	n/a	0.990
3986	14.40	4000	90.00	n/a	90.8	n/a	0.991

LINEAR REGRESSION ANALYSIS:

	CORRELATION	SLOPE	INTERCEPT
VALUE	1.000	1.001	0.1%

COMMENTS:

Sample filter changed.



TRS Analyzer Calibration by Dilution



DATE:	05-Dec-2023	PREVIOUS CALIBRATION DATE:	09-Nov-2023
PARAMETER:	TRS	PREVIOUS CORRECTION FACTOR:	0.999
CLIENT:	PRAMP	TEMPERATURE (°C):	22.6
LOCATION:	842b	BAROMETRIC (mBar):	927
PURPOSE:	Install/Post-Repair	START TIME (MST):	09:16
PERFORMED BY:	Kevin Sebastian	END TIME (MST):	16:00

ANALYZER:

MAKE/MODEL	Thermo 43iQTL	RANGE	100 ppb
SERIAL #	1200736630	FLOW (mL/min)	373
INITIAL		FINAL	
BKG/OFFSET	14	BKG/OFFSET	15.2
COEF/SLOPE	0.898	COEF/SLOPE	0.946
Expected (reference) Value	50.46	Expected (reference) Value	51.87

CALIBRATION SYSTEM:

CALIBRATOR:		ZERO AIR:	
MAKE:	Sabio	MAKE:	Teledyne
MODEL:	2010	MODEL:	T701
ID:	75401122	ID:	4568
MFC CALIBRATION DATE:	08-Sep-2023	OXIDIZER ID:	n/a
CALIBRATION GAS:		FLOWMETERS (if applicable):	
CYLINDER ID:	LL131374	HIGH ID	n/a
CONC (ppm):	10.09	EXPIRY DATE	n/a
CYLINDER (psi):	1800	LOW ID	n/a
EXPIRY DATE	03-Jan-2026	EXPIRY DATE	n/a

CALIBRATION PARAMETERS:

POINT	HIGH	MID	LOW
TARGET	78	38	19
RANGE	60 - 80	30 - 40	10 - 20

SCRUBBER CHECK (15 MINS; TRS/H2S ONLY):

START TIME:	n/a	SO2 Conc (ppb)	n/a
END TIME:	n/a	Analyzer Response (ppb)	n/a

CALIBRATION:


FLOW RATES (mL/min)			CONCENTRATION (ppb)			CORRECTION FACTOR	
DILUENT	GAS	TOTAL	ACTUAL	INDICATED		Initial	Final
				Initial	Final		
3999	30.90	3999	0.00	0.2	0	1.061	0.991
3968	30.90	3999	77.96	73.65	78.7	1.061	0.991
3984	15.10	3999	38.10	n/a	38.63	n/a	0.986
3993	7.50	4000	18.92	n/a	19.31	n/a	0.980

LINEAR REGRESSION ANALYSIS:

	CORRELATION	SLOPE	INTERCEPT
VALUE	1.000	1.009	0.1%

COMMENTS:

TRS Converter CDNOVA CDN #583. Sample filter changed.
 Initial calibration failed at low point. Repeated from adjusted high.
 14:00 = Daily ZS, adjusted high restarted



TRS Analyzer Calibration by Dilution



DATE:	06-Dec-2023	PREVIOUS CALIBRATION DATE:	05-Dec-2023
PARAMETER:	TRS	PREVIOUS CORRECTION FACTOR:	0.991
CLIENT:	PRAMP	TEMPERATURE (°C):	23.7
LOCATION:	842b	BAROMETRIC (mBar):	928
PURPOSE:	As-Found	START TIME (MST):	17:01
PERFORMED BY:	Kevin Sebastian	END TIME (MST):	18:55

ANALYZER:

MAKE/MODEL	Thermo 43iQTL	RANGE	100 ppb
SERIAL #	1200736630	FLOW (mL/min)	374
INITIAL		FINAL	
BKG/OFFSET	15.2	BKG/OFFSET	n/a
COEF/SLOPE	0.946	COEF/SLOPE	n/a
Expected (reference) Value	51.87	Expected (reference) Value	n/a

CALIBRATION SYSTEM:

CALIBRATOR:		ZERO AIR:	
MAKE:	Sabio	MAKE:	Teledyne
MODEL:	2010	MODEL:	T701
ID:	75401122	ID:	4568
MFC CALIBRATION DATE:	08-Sep-2023	OXIDIZER ID:	n/a
CALIBRATION GAS:		FLOWMETERS (if applicable):	
CYLINDER ID:	LL131374	HIGH ID	n/a
CONC (ppm):	10.09	EXPIRY DATE	n/a
CYLINDER (psi):	1800	LOW ID	n/a
EXPIRY DATE	03-Jan-2026	EXPIRY DATE	n/a

CALIBRATION PARAMETERS:

POINT	HIGH	MID	LOW
TARGET	78	38	19
RANGE	60 - 80	30 - 40	10 - 20

SCRUBBER CHECK (15 MINS; TRS/H2S ONLY):

START TIME:	n/a	SO2 Conc (ppb)	n/a
END TIME:	n/a	Analyzer Response (ppb)	n/a

CALIBRATION:


FLOW RATES (mL/min)			CONCENTRATION (ppb)			CORRECTION FACTOR	
DILUENT	GAS	TOTAL	ACTUAL	INDICATED		Initial	Final
				Initial	Final		
3999	30.90	3999	0.00	-0.35	n/a	0.989	n/a
3968	30.90	3999	77.96	78.51	n/a	0.989	n/a
3984	15.10	3999	38.10	37.42	n/a	1.009	n/a
3993	7.50	4000	18.92	18.5	n/a	1.004	n/a

LINEAR REGRESSION ANALYSIS:

	CORRELATION	SLOPE	INTERCEPT
VALUE	n/a	n/a	n/a

COMMENTS:

TRS Converter CDNOVA CDN #583.
As-found to confirm marked change at monthly calibration



Methane/Non-Methane Analyzer Calibration by Dilution



CALIBRATION:				ANALYZER:			
DATE:	05-Dec-2023	PREVIOUS CALIBRATION DATE:	09-Nov-2023	VALUE	MAKE/MODEL	SERIAL	FLOW (mL/min)
CLIENT:	PRAMP	TEMPERATURE (°C):	22.6		Thermo 55i	1501663728	1094
LOCATION:	842b	BAROMETRIC (mBar):	927	PARAMETER:	CH4	NMHC	THC
PURPOSE	Routine	START TIME (MST):	09:16	RANGE (ppm):	20	20	40
PERFORMED BY:	Kevin Sebastian	END TIME (MST):	13:43	PREVIOUS CF:	1.000	1.001	1.000

CALIBRATION SYSTEM:

CALIBRATOR:		ZERO AIR:		CALIBRATION GAS:		FLOWMETERS (if applicable):	
MAKE:	SABIO	MAKE:	API	CYLINDER ID:	LL68768	HIGH ID:	n/a
MODEL:	2010	MODEL:	T701	CH ₄ /C ₃ H ₈ (ppm):	897.0 301.0	HIGH EXPIRY:	n/a
ID:	26801218	ID:	4568	CYLINDER (psi):	1600	LOW ID:	n/a
MFC CALIBRATION DATE:	12-Sep-2023	OXIDIZER ID:	Internal	EXPIRY DATE	08-Nov-2029	LOW EXPIRY:	n/a

CALIBRATION PARAMETERS:

POINT (CH ₄ /NMHC)	HIGH	MID	LOW	CH ₄ EQUIVILANCE		
TARGET	14	7	3.5	C ₃ H ₈ as CH ₄		827.8
RANGE	12 - 16	6 - 8	2 - 4	THC as CH ₄		1724.8

EXPECTED (REFERENCE) VALUE:

INITIAL	CH ₄	NMHC	THC	FINAL	CH ₄	NMHC	THC
	9.22	10.67	19.89		9.69	10.87	20.56

CALIBRATION:

FLOW RATE			CONCENTRATION (PPM)									CORRECTION FACTOR (CF.)					
(mL/min)			CALCULATED			INITIAL INDICATED			FINAL INDICATED			INITIAL			FINAL		
DILUENT	GAS	TOTAL	CH ₄	NMHC	THC	CH ₄	NMHC	THC	CH ₄	NMHC	THC	CH ₄	NMHC	THC	CH ₄	NMHC	THC
3099	X	3099	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	X	X	X	X	X	X
3049	50.30	3099	14.56	13.44	27.99	13.99	12.88	26.87	14.56	13.43	27.99	1.041	1.043	1.042	1.000	1.000	1.000
3075	25.20	3100	7.29	6.73	14.02	n/a	n/a	n/a	7.17	6.69	13.87	n/a	n/a	n/a	1.017	1.006	1.011
3085	12.60	3098	3.65	3.37	7.01	n/a	n/a	n/a	3.58	3.32	6.90	n/a	n/a	n/a	1.019	1.014	1.017

LINEAR REGRESSION ANALYSIS:


	CORRELATION	SLOPE	INTERCEPT
CH ₄	1.000	1.001	-0.3%
NMHC	1.000	1.000	-0.1%
THC	1.000	1.001	-0.2%

Comments:

H2 = AMA HG300 #190567058
Sample filter changed.

Use Zero Chrom?

No



CAL-PRAMP-202312-01561

Page 16 of 21
CH4 [ppm] NMHC [ppm]

Methane/Non-Methane Analyzer Calibration by Dilution



CALIBRATION:				ANALYZER:			
DATE:	06-Dec-2023	PREVIOUS CALIBRATION DATE:	n/a	VALUE	MAKE/MODEL	SERIAL	FLOW (mL/min)
CLIENT:	PRAMP	TEMPERATURE (°C):	23.7		Thermo 55i	1501663728	1116
LOCATION:	842b	BAROMETRIC (mBar):	929	PARAMETER:	CH4	NMHC	THC
PURPOSE:	Install/Post-Repair	START TIME (MST):	17:01	RANGE (ppm):	20	20	40
PERFORMED BY:	Kevin Sebastian	END TIME (MST):	19:53	PREVIOUS CF:	n/a	n/a	n/a

CALIBRATION SYSTEM:

CALIBRATOR:		ZERO AIR:		CALIBRATION GAS:		FLOWMETERS (if applicable):	
MAKE:	SABIO	MAKE:	API	CYLINDER ID:	LL68768	HIGH ID:	n/a
MODEL:	2010	MODEL:	T701	CH ₄ /C ₃ H ₈ (ppm):	897.0 301.0	HIGH EXPIRY:	n/a
ID:	26801218	ID:	4568	CYLINDER (psi):	1600	LOW ID:	n/a
MFC CALIBRATION DATE:	12-Sep-2023	OXIDIZER ID:	Internal	EXPIRY DATE:	08-Nov-2029	LOW EXPIRY:	n/a

CALIBRATION PARAMETERS:

POINT (CH ₄ /NMHC)	HIGH	MID	LOW	CH ₄ EQUIVILANCE		
TARGET	14	7	3.5	C ₃ H ₈ as CH ₄		827.8
RANGE	12 - 16	6 - 8	2 - 4	THC as CH ₄		1724.8

EXPECTED (REFERENCE) VALUE:

INITIAL	CH ₄	NMHC	THC	FINAL	CH ₄	NMHC	THC
	n/a	n/a	n/a		n/a	9.66	10.85

CALIBRATION:

FLOW RATE			CONCENTRATION (PPM)									CORRECTION FACTOR (CF.)					
(mL/min)			CALCULATED			INITIAL INDICATED			FINAL INDICATED			INITIAL			FINAL		
DILUENT	GAS	TOTAL	CH ₄	NMHC	THC	CH ₄	NMHC	THC	CH ₄	NMHC	THC	CH ₄	NMHC	THC	CH ₄	NMHC	THC
3099	X	3099	0.00	0.00	0.00	n/a	n/a	n/a	0.00	0.00	0.00	X	X	X	X	X	X
3049	50.30	3099	14.56	13.44	27.99	n/a	n/a	n/a	14.52	13.40	27.92	n/a	n/a	n/a	1.003	1.003	1.003
3075	25.20	3100	7.29	6.73	14.02	n/a	n/a	n/a	7.18	6.68	13.86	n/a	n/a	n/a	1.016	1.007	1.012
3085	12.60	3098	3.65	3.37	7.01	n/a	n/a	n/a	3.57	3.34	6.91	n/a	n/a	n/a	1.022	1.008	1.015

LINEAR REGRESSION ANALYSIS:

	CORRELATION	SLOPE	INTERCEPT
CH ₄	1.000	0.998	-0.2%
NMHC	1.000	0.998	-0.1%
THC	1.000	0.998	-0.1%

Comments:

H2 = AMA HG300 #190567058
H2 Dessicant changed prior to thos post-repair

Use Zero Chrom?

No



CAL-PRAMP-202312-01561

Page 18 of 21
CH4 [ppm] NMHC [ppm]

Meteorological System Checklist



Date:	December 5, 2023		
Technician:	Kevin Sebastian		
Station:	PRAMP 842b		
Unit:	Make:	Model:	Serial #:
Precipitation Sampler:	RM Young	52202	TB 15878
Temperature Sensor:	Rotronic	HC2A-S3	20370767
Barometric Pressure Sensor:	MetOne	92	Y23362
Relative Humidity Sensor:	Rotronic	HC2A-S3	20370767
Anemometer:	RM Young	05305AQ	174802

PRECIPITATION SENSOR CHECK

Checklist:	Reply:	Comments:
Previous check date:	November 9, 2023	
Is the sensor Level?	yes	
Is the heater operating properly?	yes	
Are the bucket drain holes clean?	yes	test time: 10:24-10:31
Is the screen on the housing? (screen should be on between July and September)	yes	
Is the housing clean?	yes	
Is the area around the housing clean and free from obstacles?	yes	

TIP TEST - Slowly pour water until 10 tips are heard. (10 tips = 1 mm)

# of Tips	Data Logger Response (mm):	Manual Specification = +/- 0.1 mm
10	1.0	0.00

AMBIENT TEMPERATURE SENSOR CHECK

Previous check date:	November 9, 2023		
Parameter:	Temperature @ 2 metres		
Reference Thermometer ID:	F.S. 181341226 expires July 17, 2024		
Reference Temperature (°C):	5.5		
Station - Ambient Temperature (°C):	5.8		
Temperature Difference (°C):	0.3		

BAROMETRIC PRESSURE SENSOR CHECK

Previous check date:	November 9, 2023		
Reference Barometer ID:	Brunton 05535 Expires July 17, 2024		
Reference Pressure - Units/Reading:	millibar	927	
Station Pressure - Units/Reading:	millibar	927.2	
Pressure Tolerance +/- 15% of error:	788 - 1066	-0.02%	

RELATIVE HUMIDITY (HYGROMETER) SENSOR CHECK

Previous check date:	November 9, 2023		
Reference Hygrometer ID:	F.S. 181341226 expires July 17, 2024		
Reference Hygrometer % RH- Reading:	76.50		
Station Hygrometer % RH- Reading:	80.70		
RH Tolerance +/- 15% of difference:	65.03 - 87.98	-5.5%	

ANEMOMETER - WIND SPEED & WIND DIRECTION SENSOR CHECK

WIND SPEED		WIND DIRECTION	
Previous check date:	November 9, 2023	Previous check date:	November 9, 2023
Wind Speed Observed (kph):	10~20	Wind Direction Observed:	SW
Wind speed on Data Logger (kph):	19	Wind Direction on Data Logger:	SW
		Wind Direction Pass/Fail?:	Pass

Comments



Meteorological Sensor Audit/Calibration

Location Information

Company: PRAMP
Audit Location: 842b
Audit Date: August 3, 2023
Calibration Purpose: routine annual
Performed By: Chris Wesson
Reviewed By: Limin Li
Start/End Time (mst): 15:57 / 17:00
Weather Conditions: Mix of sun and clouds with rain showers

Wind Sensor Information

Sensor ID Data:		Sensor Outputs:	
Sensor Make:	RM Young	Velocity Voltage Output Range:	n/a
Sensor Model:	05305AQ	Velocity Unit Output Range:	0-200
Serial #:	174802	Direction Voltage Output Range:	n/a
Previous Cal/Audit Date:	August 3, 2022	Direction Unit Output Range:	0-360

Wind Calibrator Information

Calibrator I.D. and Expiry Date: RM Young 18802 id# R9133 expires Oct 18, 2024

Wind Speed Audit Data ****+/- 2% of the average correction factor is the limit****

RPM	Wind Speed Generated kph	Clockwise Wind Speed kph	Counter Clockwise Wind Speed kph	Correction Factor
0	0	0.0	0.0	-
1000	18.4	18.3	18.3	1.007
2000	36.9	36.7	36.7	1.004
3000	55.3	55.1	55.1	1.003
4000	73.7	73.6	73.6	1.002
5000	92.2	92.0	92.0	1.002
6000	110.6	110.5	110.5	1.001
7000	129.0	128.9	128.9	1.001
8000	147.4	147.3	147.3	1.001
9000	165.9	165.8	165.8	1.000
10000	184.3	184.2	184.2	1.001
The audit meets AMD requirements.			Average Correction Factor=	1.002

Wind Direction Audit Data ****+/- 3° of the absolute average degrees difference for all points is the limit****

Generated Wind Direction 0-360 (Up)	Generated Wind Direction 360-0 (Down)	Indicated Wind Direction 0-360 (Up)	Indicated Wind Direction 360-0 (Down)	Degrees Difference 0-360 (Up)	Degrees Difference 360-0 (Down)	Average Absolute Degrees Difference
0	355	2	354	2.0	1.0	1.5
30	330	23	332	7.0	-2.0	4.5
60	300	59	303	1.0	-3.0	2.0
90	270	88	274	2.0	-4.0	3.0
120	240	118	242	2.0	-2.0	2.0
150	210	149	209	1.0	1.0	1.0
180	180	179	178	1.0	2.0	1.5
210	150	210	146	0.0	4.0	2.0
240	120	243	117	-3.0	3.0	3.0
270	90	274	89	-4.0	1.0	2.5
300	60	304	57	-4.0	3.0	3.5
330	30	334	29	-4.0	1.0	2.5
355	0	354	2	1.0	2.0	1.5
The audit meets AMD requirements.				Average Absolute Degrees Difference=		2.3

Comments:

Declination = 15 deg East
 Physical inspection completed, no issues
 Potentiometer noisy. Replacement required.

END OF REPORT



Peace River Area Monitoring Program

DECEMBER 2023

Ambient Air Monitoring Calibration Report

- 986-C STATION-

CAL-PRAMP-202312-01562

Operation and Maintenance:

Bureau Veritas Canada


Data Validation and Report:

Bureau Veritas Canada

January 6, 2024


DAILY INTERNAL ZERO-SPAN CALIBRATION RECORDS

SO2[ppb] Calibration: PRAMP 986-C Monthly: 12-2023 Type: SpanAndZero - Zero




Zero Zero Ref Zero Low Zero High

SO2[ppb] Calibration: PRAMP 986-C Monthly: 12-2023 Type: SpanAndZero - Span




Span SpanRef Span Low Span High

TRS[ppb] Calibration: PRAMP 986-C Monthly: 12-2023 Type: SpanAndZero - Zero




Zero Zero Ref Zero Low Zero High

TRS[ppb] Calibration: PRAMP 986-C Monthly: 12-2023 Type: SpanAndZero - Span




Span SpanRef Span Low Span High

THC55[ppm] Calibration: PRAMP 986-C Monthly: 12-2023 Type: SpanAndZero - Zero




Zero Zero Ref Zero Low Zero High

THC55[ppm] Calibration: PRAMP 986-C Monthly: 12-2023 Type: SpanAndZero - Span




Span SpanRef Span Low Span High

CH4[ppm] Calibration: PRAMP 986-C Monthly: 12-2023 Type: SpanAndZero - Zero




Legend: Zero (orange square), Zero Ref (red line), Zero Low (blue line), Zero High (purple line)

CH4[ppm] Calibration: PRAMP 986-C Monthly: 12-2023 Type: SpanAndZero - Span




Legend: Span (orange square), Span Ref (red line), Span Low (blue line), Span High (purple line)

NMHC[ppm] Calibration: PRAMP 986-C Monthly: 12-2023 Type: SpanAndZero - Zero



Zero Zero Ref Zero Low Zero High

NMHC[ppm] Calibration: PRAMP 986-C Monthly: 12-2023 Type: SpanAndZero - Span



Span SpanRef Span Low Span High

MULTI-POINT CALIBRATION RECORDS

SO2 Analyzer Calibration by Dilution



DATE:	07-Dec-2023	PREVIOUS CALIBRATION DATE:	08-Nov-2023
PARAMETER:	SO2	PREVIOUS CORRECTION FACTOR:	1.004
CLIENT:	PRAMP	TEMPERATURE (°C):	21.8
LOCATION:	986c	BAROMETRIC (mBar):	932
PURPOSE:	Routine	START TIME (MST):	10:10
PERFORMED BY:	Kevin Sebastian	END TIME (MST):	15:32

ANALYZER:

MAKE/MODEL	Thermo 43iQTL	RANGE	500 ppb
SERIAL #	1193585646	FLOW (mL/min)	424
INITIAL		FINAL	
BKG/OFFSET	17.9	BKG/OFFSET	18.1
COEF/SLOPE	1.037	COEF/SLOPE	1.052
Expected (reference) Value	251.1	Expected (reference) Value	250.7

CALIBRATION SYSTEM:

CALIBRATOR:		ZERO AIR:	
MAKE:	Sabio	MAKE:	Teledyne
MODEL:	2010	MODEL:	M701
ID:	75401122	ID:	4568
MFC CALIBRATION DATE:	08-Sep-2023	OXIDIZER ID:	n/a
CALIBRATION GAS:		FLOWMETERS (if applicable):	
CYLINDER ID:	LL109693	HIGH ID	n/a
CONC (ppm):	25.00	EXPIRY DATE	n/a
CYLINDER (psi):	1600	LOW ID	n/a
EXPIRY DATE	02-Nov-2025	EXPIRY DATE	n/a

CALIBRATION PARAMETERS:

POINT	HIGH	MID	LOW
TARGET	390	190	95
RANGE	300 - 400	150 - 200	50 - 100

SCRUBBER CHECK (15 MINS; TRS/H2S ONLY):

START TIME:	n/a	SO2 Conc (ppb)	n/a
END TIME:	n/a	Analyzer Response (ppb)	n/a

CALIBRATION:


FLOW RATES (mL/min)			CONCENTRATION (ppb)			CORRECTION FACTOR	
DILUENT	GAS	TOTAL	ACTUAL	INDICATED		Initial	Final
				Initial	Final		
3999	60.80	3999	0.00	0.2	0	1.015	0.998
3939	60.80	4000	380.00	374.7	380.7	1.015	0.998
3971	28.80	4000	180.00	n/a	181.8	n/a	0.990
3987	14.40	4001	89.98	n/a	90.4	n/a	0.995

LINEAR REGRESSION ANALYSIS:

	CORRELATION	SLOPE	INTERCEPT
VALUE	1.000	1.002	0.1%

COMMENTS:

Sample filter changed.



TRS Analyzer Calibration by Dilution



DATE:	07-Dec-2023	PREVIOUS CALIBRATION DATE:	08-Nov-2023
PARAMETER:	TRS	PREVIOUS CORRECTION FACTOR:	0.997
CLIENT:	PRAMP	TEMPERATURE (°C):	21.8
LOCATION:	986C	BAROMETRIC (mBar):	932
PURPOSE:	Routine	START TIME (MST):	10:10
PERFORMED BY:	Kevin Sebastian	END TIME (MST):	15:33

ANALYZER:

MAKE/MODEL	Thermo 43iQTL	RANGE	100 ppb
SERIAL #	1191833341	FLOW (mL/min)	419
INITIAL		FINAL	
BKG/OFFSET	15.9	BKG/OFFSET	15.8
COEF/SLOPE	0.948	COEF/SLOPE	0.963
Expected (reference) Value	57.72	Expected (reference) Value	55.52

CALIBRATION SYSTEM:

CALIBRATOR:		ZERO AIR:	
MAKE:	Sabio	MAKE:	Teledyne
MODEL:	2010	MODEL:	M701
ID:	75401122	ID:	4568
MFC CALIBRATION DATE:	08-Sep-2023	OXIDIZER ID:	n/a
CALIBRATION GAS:		FLOWMETERS (if applicable):	
CYLINDER ID:	LL131374	HIGH ID	n/a
CONC (ppm):	10.09	EXPIRY DATE	n/a
CYLINDER (psi):	1800	LOW ID	n/a
EXPIRY DATE	03-Jan-2026	EXPIRY DATE	n/a

CALIBRATION PARAMETERS:

POINT	HIGH	MID	LOW
TARGET	78	38	19
RANGE	60 - 80	30 - 40	10 - 20

SCRUBBER CHECK (15 MINS; TRS/H2S ONLY):

START TIME:	n/a	SO2 Conc (ppb)	n/a
END TIME:	n/a	Analyzer Response (ppb)	n/a

CALIBRATION:


FLOW RATES (mL/min)			CONCENTRATION (ppb)			CORRECTION FACTOR	
DILUENT	GAS	TOTAL	ACTUAL	INDICATED		Initial	Final
				Initial	Final		
3999	30.90	3999	0.00	0.14	0	1.005	0.987
3970	30.90	4001	77.93	77.67	78.93	1.005	0.987
3987	15.10	4002	38.07	n/a	39.36	n/a	0.967
3992	7.50	3999	18.92	n/a	19.62	n/a	0.964

LINEAR REGRESSION ANALYSIS:

	CORRELATION	SLOPE	INTERCEPT
VALUE	1.000	1.012	0.3%

COMMENTS:

TRS Converter CDNOVA CDN101 #530



Methane/Non-Methane Analyzer Calibration by Dilution



CALIBRATION:				ANALYZER:			
DATE:	07-Dec-2023	PREVIOUS CALIBRATION DATE:	08-Nov-2023	VALUE	MAKE/MODEL	SERIAL	FLOW (mL/min)
CLIENT:	PRAMP	TEMPERATURE (°C):	21.8		Thermo 55i	12208316589	1123
LOCATION:	986C	BAROMETRIC (mBar):	932	PARAMETER:	CH4	NMHC	THC
PURPOSE	Routine	START TIME (MST):	10:10	RANGE (ppm):	20	20	40
PERFORMED BY:	Kevin Sebastian	END TIME (MST):	14:50	PREVIOUS CF:	1.002	0.995	0.999

CALIBRATION SYSTEM:

CALIBRATOR:		ZERO AIR:		CALIBRATION GAS:		FLOWMETERS (if applicable):	
MAKE:	Sabio	MAKE:	Teledyne	CYLINDER ID:	LL68768	HIGH ID:	n/a
MODEL:	2010	MODEL:	M701	CH ₄ /C ₃ H ₈ (ppm):	897.0 301.0	HIGH EXPIRY:	n/a
ID:	26801218	ID:	4568	CYLINDER (psi):	1600	LOW ID:	n/a
MFC CALIBRATION DATE:	12-Sep-2023	OXIDIZER ID:	Internal	EXPIRY DATE	08-Nov-2029	LOW EXPIRY:	n/a

CALIBRATION PARAMETERS:

POINT (CH ₄ /NMHC)	HIGH	MID	LOW	CH ₄ EQUIVILANCE		
TARGET	14	7	3.5	C ₃ H ₈ as CH ₄		827.8
RANGE	12 - 16	6 - 8	2 - 4	THC as CH ₄		1724.8

EXPECTED (REFERENCE) VALUE:


INITIAL	CH ₄	NMHC	THC	FINAL	CH ₄	NMHC	THC
	9.57	11.17	20.74		9.61	11.01	20.62

CALIBRATION:

FLOW RATE			CONCENTRATION (PPM)									CORRECTION FACTOR (CF.)					
(mL/min)			CALCULATED			INITIAL INDICATED			FINAL INDICATED			INITIAL			FINAL		
DILUENT	GAS	TOTAL	CH ₄	NMHC	THC	CH ₄	NMHC	THC	CH ₄	NMHC	THC	CH ₄	NMHC	THC	CH ₄	NMHC	THC
3099	X	3099	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	X	X	X	X	X	X
3048	50.30	3098	14.56	13.44	28.00	14.72	13.33	28.04	14.56	13.32	27.89	0.989	1.008	0.999	1.000	1.009	1.004
3073	25.20	3098	7.30	6.73	14.03	n/a	n/a	n/a	7.35	6.66	14.01	n/a	n/a	n/a	0.993	1.011	1.001
3085	12.60	3098	3.65	3.37	7.01	n/a	n/a	n/a	3.66	3.32	6.98	n/a	n/a	n/a	0.997	1.014	1.005

LINEAR REGRESSION ANALYSIS:

	CORRELATION	SLOPE	INTERCEPT	Comments:	
CH ₄	1.000	1.000	0.1%	H2 = AMA HG300 #191267063 filter changed	
NMHC	1.000	0.991	0.0%		
THC	1.000	0.996	0.0%		
				Use Zero Chrom?	Yes



CAL-PRAMP-202312-01562

Meteorological System Checklist



Date:	December 7, 2023		
Technician:	Kevin Sebastian		
Station:	PRAMP 986c		
Unit:	Make:	Model:	Serial #:
Precipitation Sampler:	RM Young	52202	TB 16325
Temperature Sensor:	Rotronic	HC2-32	20626912
Barometric Pressure Sensor:	MetOne	092	Y23358
Relative Humidity Sensor:	Rotronic	HC2-S3	20626912
Anemometer:	RM Young	05305AQ	180340

PRECIPITATION SENSOR CHECK

Checklist:	Reply:	Comments:
Previous check date:	November 7, 2023	
Is the sensor Level?	yes	
Is the heater operating properly?	yes	
Are the bucket drain holes clean?	yes	Tested 15:28-15:35
Is the screen on the housing? (screen should be on between July and September)	yes	
Is the housing clean?	yes	
Is the area around the housing clean and free from obstacles?	yes	

TIP TEST - Slowly pour water until 10 tips are heard. (10 tips = 1 mm)

# of Tips	Data Logger Response (mm):	Manual Specification = +/- 0.1 mm
10	1.0	0.00

AMBIENT TEMPERATURE SENSOR CHECK

Previous check date:	November 7, 2023	
Parameter:	Temperature @ 2 metres	
Reference Thermometer ID:	FS 181341226 expires July 17, 2024	
Reference Temperature (°C):	-2.6	
Station - Ambient Temperature (°C):	-2.8	
Temperature Difference (°C):	0.2	

BAROMETRIC PRESSURE SENSOR CHECK

Previous check date:	November 7, 2023	
Reference Barometer ID:	Brunton 05535 expires July 17, 2024	
Reference Pressure - Units/Reading:	millibar	934
Station Pressure - Units/Reading:	millibar	933
Pressure Tolerance +/- 15% of error:	794 - 1074	0.11%

RELATIVE HUMIDITY (HYGROMETER) SENSOR CHECK

Previous check date:	November 7, 2023	
Reference Hygrometer ID:	FS 181341226 expires July 17, 2024	
Reference Hygrometer % RH- Reading:	81.30	
Station Hygrometer % RH- Reading:	83.70	
RH Tolerance +/- 15% of difference:	69.11 - 93.50	-3.0%

ANEMOMETER - WIND SPEED & WIND DIRECTION SENSOR CHECK

WIND SPEED		WIND DIRECTION	
Previous check date:	November 7, 2023	Previous check date:	November 7, 2023
Wind Speed Observed (kph):	5~10	Wind Direction Observed:	E
Wind speed on Data Logger (kph):	8.8	Wind Direction on Data Logger:	E
		Wind Direction Pass/Fail?:	Pass

Comments



Meteorological Sensor Audit/Calibration

Location Information

Company: PRAMP
 Audit Location: 986C
 Audit Date: August 3, 2023
 Calibration Purpose: routine annual

Performed By: Chris Wesson
 Reviewed By: Limin Li
 Start/End Time (mst): 14:01 / 15:03
 Weather Conditions: Rain fall heavy at times

Wind Sensor Information

Sensor ID Data:		Sensor Outputs:	
Sensor Make:	RM Young	Velocity Voltage Output Range:	n/a
Sensor Model:	05305AQ	Velocity Unit Output Range:	0-200
Serial #:	180340	Direction Voltage Output Range:	n/a
Previous Cal/Audit Date:	August 5, 2022	Direction Unit Output Range:	0-360

Wind Calibrator Information

Calibrator I.D. and Expiry Date: RM Young 18802 id# R9133 expires Oct 18, 2024

Wind Speed Audit Data ****+/- 2% of the average correction factor is the limit****

RPM	Wind Speed Generated kph	Clockwise Wind Speed kph	Counter Clockwise Wind Speed kph	Correction Factor
0	0	0.0	0.0	-
1000	18.4	18.3	18.3	1.007
2000	36.9	36.7	36.7	1.004
3000	55.3	55.1	55.1	1.003
4000	73.7	73.6	73.6	1.002
5000	92.2	92.0	92.0	1.002
6000	110.6	110.5	110.1	1.003
7000	129.0	128.9	129.3	0.999
8000	147.4	147.3	147.3	1.001
9000	165.9	165.8	165.8	1.000
10000	184.3	184.2	184.2	1.001
The audit meets AMD requirements.			Average Correction Factor=	1.002

Wind Direction Audit Data ****+/- 3° of the absolute average degrees difference for all points is the limit****

Generated Wind Direction 0-360 (Up)	Generated Wind Direction 360-0 (Down)	Indicated Wind Direction 0-360 (Up)	Indicated Wind Direction 360-0 (Down)	Degrees Difference 0-360 (Up)	Degrees Difference 360-0 (Down)	Average Absolute Degrees Difference
0	355	2	354	2.0	1.0	1.5
30	330	31	331	-1.0	-1.0	1.0
60	300	60	301	0.0	-1.0	0.5
90	270	90	271	0.0	-1.0	0.5
120	240	119	240	1.0	0.0	0.5
150	210	148	207	2.0	3.0	2.5
180	180	178	178	2.0	2.0	2.0
210	150	208	148	2.0	2.0	2.0
240	120	240	122	0.0	-2.0	1.0
270	90	272	90	-2.0	0.0	1.0
300	60	301	60	-1.0	0.0	0.5
330	30	331	32	-1.0	-2.0	1.5
355	0	354	2	1.0	2.0	1.5
The audit meets AMD requirements.				Average Absolute Degrees Difference=		1.2

Comments:

Declination = 15 deg East
 Physical inspection completed, no issues

END OF REPORT



Peace River Area Monitoring Program

DECEMBER 2023

Ambient Air Monitoring Calibration Report

- RENO-B STATION-

CAL-PRAMP-202312-01563

Operation and Maintenance:

Bureau Veritas Canada


Data Validation and Report:

Bureau Veritas Canada

January 6, 2024


DAILY INTERNAL ZERO-SPAN CALIBRATION RECORDS

SO2[ppb] Calibration: PRAMP RENO-B Monthly: 12-2023 Type: SpanAndZero - Zero




Zero Zero Ref Zero Low Zero High

SO2[ppb] Calibration: PRAMP RENO-B Monthly: 12-2023 Type: SpanAndZero - Span




Span Span Ref Span Low Span High

TRS[ppb] Calibration: PRAMP RENO-B Monthly: 12-2023 Type: SpanAndZero - Zero




Zero Zero Ref Zero Low Zero High

TRS[ppb] Calibration: PRAMP RENO-B Monthly: 12-2023 Type: SpanAndZero - Span




Span SpanRef Span Low Span High

THC55[ppm] Calibration: PRAMP REND-B Monthly: 12-2023 Type: SpanAndZero - Zero




Zero Zero Ref Zero Low Zero High

THC55[ppm] Calibration: PRAMP REND-B Monthly: 12-2023 Type: SpanAndZero - Span



Span SpanRef Span Low Span High

CH4[ppm] Calibration: PRAMP RENO-B Monthly: 12-2023 Type: SpanAndZero - Zero




Zero Zero Ref Zero Low Zero High

CH4[ppm] Calibration: PRAMP RENO-B Monthly: 12-2023 Type: SpanAndZero - Span




Span Span Ref Span Low Span High

NMHC[ppm] Calibration: PRAMP REND-B Monthly: 12-2023 Type: SpanAndZero - Zero



Zero Zero Ref Zero Low Zero High

NMHC[ppm] Calibration: PRAMP REND-B Monthly: 12-2023 Type: SpanAndZero - Span



Span SpanRef Span Low Span High

MULTI-POINT CALIBRATION RECORDS

SO2 Analyzer Calibration by Dilution



DATE:	01-Dec-2023	PREVIOUS CALIBRATION DATE:	01-Nov-2023
PARAMETER:	SO2	PREVIOUS CORRECTION FACTOR:	0.998
CLIENT:	PRAMP	TEMPERATURE (°C):	23.2
LOCATION:	Reno-B	BAROMETRIC (mBar):	926
PURPOSE:	Routine	START TIME (MST):	09:24
PERFORMED BY:	Kevin Sebastian	END TIME (MST):	14:18

ANALYZER:

MAKE/MODEL	Thermo 43iQTL	RANGE	500 ppb
SERIAL #	12101910505	FLOW (mL/min)	431
INITIAL		FINAL	
BKG/OFFSET	1.4	BKG/OFFSET	1.39
COEF/SLOPE	0.949	COEF/SLOPE	0.955
Expected (reference) Value	217.8	Expected (reference) Value	224.4

CALIBRATION SYSTEM:

CALIBRATOR:		ZERO AIR:	
MAKE:	Sabio	MAKE:	Teledyne
MODEL:	2010	MODEL:	M701
ID:	75401122	ID:	4568
MFC CALIBRATION DATE:	08-Sep-2023	OXIDIZER ID:	n/a
CALIBRATION GAS:		FLOWMETERS (if applicable):	
CYLINDER ID:	LL109693	HIGH ID	n/a
CONC (ppm):	25.00	EXPIRY DATE	n/a
CYLINDER (psi):	1400	LOW ID	n/a
EXPIRY DATE	02-Nov-2025	EXPIRY DATE	n/a

CALIBRATION PARAMETERS:

POINT	HIGH	MID	LOW
TARGET	390	190	95
RANGE	300 - 400	150 - 200	50 - 100

SCRUBBER CHECK (15 MINS; TRS/H2S ONLY):

START TIME:	n/a	SO2 Conc (ppb)	n/a
END TIME:	n/a	Analyzer Response (ppb)	n/a

CALIBRATION:


FLOW RATES (mL/min)			CONCENTRATION (ppb)			CORRECTION FACTOR	
DILUENT	GAS	TOTAL	ACTUAL	INDICATED		Initial	Final
				Initial	Final		
3999	60.80	3999	0.00	-0.2	0	1.014	0.998
3938	60.80	3999	380.10	374.7	380.7	1.014	0.998
3971	28.80	4000	180.00	n/a	180.6	n/a	0.997
3986	14.40	4000	90.00	n/a	89.6	n/a	1.004

LINEAR REGRESSION ANALYSIS:

	CORRELATION	SLOPE	INTERCEPT
VALUE	1.000	1.002	0.0%

COMMENTS:

Sample filter changed



TRS Analyzer Calibration by Dilution



DATE:	01-Dec-2023	PREVIOUS CALIBRATION DATE:	02-Nov-2023
PARAMETER:	TRS	PREVIOUS CORRECTION FACTOR:	1.002
CLIENT:	PRAMP	TEMPERATURE (°C):	23.2
LOCATION:	Reno-B	BAROMETRIC (mBar):	926
PURPOSE:	Routine	START TIME (MST):	09:24
PERFORMED BY:	Kevin Sebastian	END TIME (MST):	14:18

ANALYZER:

MAKE/MODEL	Thermo 43iQTL	RANGE	100 ppb
SERIAL #	12101910504	FLOW (mL/min)	392
INITIAL		FINAL	
BKG/OFFSET	1.05	BKG/OFFSET	1.14
COEF/SLOPE	0.859	COEF/SLOPE	0.932
Expected (reference) Value	35.1	Expected (reference) Value	39.46

CALIBRATION SYSTEM:

CALIBRATOR:		ZERO AIR:	
MAKE:	Sabio	MAKE:	Teledyne
MODEL:	2010	MODEL:	M701
ID:	75401122	ID:	4568
MFC CALIBRATION DATE:	08-Sep-2023	OXIDIZER ID:	n/a
CALIBRATION GAS:		FLOWMETERS (if applicable):	
CYLINDER ID:	LL131374	HIGH ID	n/a
CONC (ppm):	10.09	EXPIRY DATE	n/a
CYLINDER (psi):	1800	LOW ID	n/a
EXPIRY DATE	03-Jan-2026	EXPIRY DATE	n/a

CALIBRATION PARAMETERS:

POINT	HIGH	MID	LOW
TARGET	78	38	19
RANGE	60 - 80	30 - 40	10 - 20

SCRUBBER CHECK (15 MINS; TRS/H2S ONLY):

START TIME:	n/a	SO2 Conc (ppb)	n/a
END TIME:	n/a	Analyzer Response (ppb)	n/a

CALIBRATION:


FLOW RATES (mL/min)			CONCENTRATION (ppb)			CORRECTION FACTOR	
DILUENT	GAS	TOTAL	ACTUAL	INDICATED		Initial	Final
				Initial	Final		
3999	30.90	3999	0.00	0.06	0	1.088	0.998
3968	30.90	3999	77.96	71.75	78.16	1.088	0.998
3985	15.10	4000	38.09	n/a	38.02	n/a	1.002
3993	7.50	4000	18.92	n/a	18.84	n/a	1.004

LINEAR REGRESSION ANALYSIS:

	CORRELATION	SLOPE	INTERCEPT
VALUE	1.000	1.003	-0.1%

COMMENTS:

TRS Converter CDNOVA CDN-101 #590. Sample Filter Changed 10:15- Regulator flushed due to low response	10:12
---	-------



Methane/Non-Methane Analyzer Calibration by Dilution



CALIBRATION:				ANALYZER:			
DATE:	01-Dec-2023	PREVIOUS CALIBRATION DATE:	02-Nov-2023	VALUE	MAKE/MODEL	SERIAL	FLOW (mL/min)
CLIENT:	PRAMP	TEMPERATURE (°C):	23.2		Thermo 55i	12101910497	1058
LOCATION:	Reno-B	BAROMETRIC (mBar):	926	PARAMETER:	CH4	NMHC	THC
PURPOSE:	Routine	START TIME (MST):	09:24	RANGE (ppm):	20	20	40
PERFORMED BY:	Kevin Sebastian	END TIME (MST):	14:18	PREVIOUS CF:	0.998	0.999	0.998

CALIBRATION SYSTEM:

CALIBRATOR:		ZERO AIR:		CALIBRATION GAS:		FLOWMETERS (if applicable):	
MAKE:	Sabio	MAKE:	Teledyne	CYLINDER ID:	LL68768	HIGH ID:	n/a
MODEL:	2010	MODEL:	M701	CH ₄ /C ₃ H ₈ (ppm):	897.0 301.0	HIGH EXPIRY:	n/a
ID:	26701218	ID:	4568	CYLINDER (psi):	1600	LOW ID:	n/a
MFC CALIBRATION DATE:	08-Sep-2023	OXIDIZER ID:	Internal	EXPIRY DATE	11-Aug-2029	LOW EXPIRY:	n/a

CALIBRATION PARAMETERS:

POINT (CH ₄ /NMHC)	HIGH	MID	LOW	CH ₄ EQUIVILANCE		
TARGET	14	7	3.5	C ₃ H ₈ as CH ₄		827.8
RANGE	12 - 16	6 - 8	2 - 4	THC as CH ₄		1724.8

EXPECTED (REFERENCE) VALUE:


INITIAL	CH ₄	NMHC	THC	FINAL	CH ₄	NMHC	THC
	9.71	11.11	20.83		9.83	11.01	20.84

CALIBRATION:

FLOW RATE			CONCENTRATION (PPM)									CORRECTION FACTOR (CF.)					
(mL/min)			CALCULATED			INITIAL INDICATED			FINAL INDICATED			INITIAL			FINAL		
DILUENT	GAS	TOTAL	CH ₄	NMHC	THC	CH ₄	NMHC	THC	CH ₄	NMHC	THC	CH ₄	NMHC	THC	CH ₄	NMHC	THC
3099	X	3099	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	X	X	X	X	X	X
3049	50.30	3099	14.56	13.44	27.99	14.12	12.83	26.95	14.55	13.43	27.98	1.031	1.047	1.039	1.001	1.000	1.001
3075	25.20	3100	7.29	6.73	14.02	n/a	n/a	n/a	7.26	6.73	14.00	n/a	n/a	n/a	1.004	1.000	1.001
3085	12.60	3098	3.65	3.37	7.01	n/a	n/a	n/a	3.65	3.41	7.06	n/a	n/a	n/a	1.000	0.987	0.994

LINEAR REGRESSION ANALYSIS:

	CORRELATION	SLOPE	INTERCEPT	Comments:	
CH ₄	1.000	0.999	0.0%	Sample filter changed	
NMHC	1.000	0.999	0.1%		
THC	1.000	0.999	0.0%		
				Use Zero Chrom?	No



CAL-PRAMP-202312-01563

Page 14 of 17
CH4 [ppm] NMHC [ppm]

Meteorological System Checklist



Date:		December 1, 2023	
Technician:		Kevin Sebastian	
Station:		PRAMP Reno	
Unit:	Make:	Model:	Serial #:
Precipitation Sampler:	RM Young	52202	TB 15877
Temperature Sensor:	Rotronic	HC2-S3	20467597
Barometric Pressure Sensor:	MetOne	92	A17940
Relative Humidity Sensor:	Rotronic	HC2-S3	20467597
Anemometer:	RM Young	05305AQ	174795
PRECIPITATION SENSOR CHECK			
Checklist:	Reply:	Comments:	
Previous check date:	October 5, 2023	Audit: 12:53-13:02	
Is the sensor Level?	yes		
Is the heater operating properly?	yes		
Are the bucket drain holes clean?	yes		
Is the screen on the housing? (screen should be on between July and September)	yes		
Is the housing clean?	yes		
Is the area around the housing clean and free from obstacles?	yes		
TIP TEST - Slowly pour water until 10 tip are heard. (10 tips = 1 ml)			
# of Tips	Data Logger Response (ml):	Manual Specification = +/- 0.1 ml	
10	1.00	0.00	
AMBIENT TEMPERATURE SENSOR CHECK			
Previous check date:	November 2, 2023		
Parameter:	Temperature @ 2 metres		
Reference Thermometer ID:	FS 181341226 expires July 17, 2024		
Reference Temperature (°C):	-11.4		
Station - Ambient Temperature (°C):	-10.8		
Temperature Difference (°C):	-0.6		
BAROMETRIC PRESSURE SENSOR CHECK			
Previous check date:	November 2, 2023		
Reference Barometer ID:	Equipment ID - 05535 Brunton Expiry - July 17 2024		
Reference Pressure - Units/Reading:	millibar	924.3	
Station Pressure - Units/Reading:	millibar	924.1	
Pressure Tolerance +/- 15% of error:	786 - 1063	0.02%	
RELATIVE HUMIDITY (HYGROMETER) SENSOR CHECK			
Previous check date:	November 2, 2023		
Reference Hygrometer ID:	FS 181341226 expires July 17, 2024		
Reference Hygrometer % RH- Reading:	82.70		
Station Hygrometer % RH- Reading:	84.20		
RH Tolerance +/- 15% of difference:	70.30 - 95.11	-1.8%	
ANEMOMETER - WIND SPEED & WIND DIRECTION SENSOR CHECK			
WIND SPEED		WIND DIRECTION	
Previous check date:	November 2, 2023	Previous check date:	November 2, 2023
Wind Speed Observed (kph):	5~10	Wind Direction Observed:	N
Wind speed on Data Logger (kph):	9.8	Wind Direction on Data Logger:	N
		Wind Direction Pass/Fail?:	Pass
Comments			



Meteorological Sensor Audit/Calibration

Location Information

Company: PRAMP
Audit Location: Reno-B
Audit Date: August 1, 2023
Calibration Purpose: routine annual

Performed By: Chris Wesson
Reviewed By: Limin Li
Start/End Time (mst): 09:54 / 11:18
Weather Conditions: Mainly cloudy with drizzle

Wind Sensor Information

Sensor ID Data:		Sensor Outputs:	
Sensor Make:	RM Young	Velocity Voltage Output Range:	n/a
Sensor Model:	05305AQ	Velocity Unit Output Range:	0-200
Serial #:	174795	Direction Voltage Output Range:	n/a
Previous Cal/Audit Date:	November 23, 2022	Direction Unit Output Range:	0-360

Wind Calibrator Information

Calibrator I.D. and Expiry Date: RM Young 18802 id# R9133 expires Oct 18, 2024

Wind Speed Audit Data ****+/- 2% of the average correction factor is the limit****

RPM	Wind Speed Generated kph	Clockwise Wind Speed kph	Counter Clockwise Wind Speed kph	Correction Factor
0	0	0.0	0.0	-
1000	18.4	18.3	18.3	1.007
2000	36.9	36.7	36.7	1.004
3000	55.3	55.1	55.1	1.003
4000	73.7	73.6	73.6	1.002
5000	92.2	92.0	92.0	1.002
6000	110.6	110.5	110.5	1.001
7000	129.0	128.9	128.9	1.001
8000	147.4	147.3	147.3	1.001
9000	165.9	165.8	165.8	1.000
10000	184.3	184.2	184.6	0.999
The audit meets AMD requirements.			Average Correction Factor=	1.002

Wind Direction Audit Data ****+/- 3° of the absolute average degrees difference for all points is the limit****

Generated Wind Direction 0-360 (Up)	Generated Wind Direction 360-0 (Down)	Indicated Wind Direction 0-360 (Up)	Indicated Wind Direction 360-0 (Down)	Degrees Difference 0-360 (Up)	Degrees Difference 360-0 (Down)	Average Absolute Degrees Difference
0	355	1	354	1.0	1.0	1.0
30	330	30	332	0.0	-2.0	1.0
60	300	60	302	0.0	-2.0	1.0
90	270	89	272	1.0	-2.0	1.5
120	240	120	241	0.0	-1.0	0.5
150	210	149	211	1.0	-1.0	1.0
180	180	181	182	-1.0	-2.0	1.5
210	150	211	150	-1.0	0.0	0.5
240	120	241	120	-1.0	0.0	0.5
270	90	271	90	-1.0	0.0	0.5
300	60	301	60	-1.0	0.0	0.5
330	30	331	29	-1.0	1.0	1.0
355	0	354	1	1.0	1.0	1.0
The audit meets AMD requirements.				Average Absolute Degrees Difference=		0.9

Comments:

Declination = 15 deg East
Physical inspection completed, no issues

END OF REPORT



Peace River Area Monitoring Program

DECEMBER 2023

Ambient Air Monitoring Calibration Report

- AQHI - GRIMSHAW STATION-

CAL-PRAMP-202312-01689

Operation and Maintenance:

Bureau Veritas Canada


Data Validation and Report:

Bureau Veritas Canada

January 6, 2024


DAILY INTERNAL ZERO-SPAN CALIBRATION RECORDS

SO2[ppb] Calibration: AQHI Grimshaw Monthly: 12-2023 Type: SpanAndZero - Zero




Zero Zero Ref Zero Low Zero High

SO2[ppb] Calibration: AQHI Grimshaw Monthly: 12-2023 Type: SpanAndZero - Span




Span SpanRef Span Low Span High

TRS[ppb] Calibration: AQHI Grimshaw Monthly: 12-2023 Type: SpanAndZero - Zero




Zero Zero Ref Zero Low Zero High

TRS[ppb] Calibration: AQHI Grimshaw Monthly: 12-2023 Type: SpanAndZero - Span




Span SpanRef Span Low Span High

NOX[ppb] Calibration: AQHI Grimshaw Monthly: 12-2023 Type: SpanAndZero - Zero




Zero Zero Ref Zero Low Zero High

NOX[ppb] Calibration: AQHI Grimshaw Monthly: 12-2023 Type: SpanAndZero - Span



Span SpanRef Span Low Span High

NO2[ppb] Calibration: AQHI Grimshaw Monthly: 12-2023 Type: SpanAndZero - Zero




Zero Zero Ref Zero Low Zero High

NO2[ppb] Calibration: AQHI Grimshaw Monthly: 12-2023 Type: SpanAndZero - Span




Span SpanRef Span Low Span High

O3[ppb] Calibration: AQHI Grimshaw Monthly: 12-2023 Type: SpanAndZero - Zero




Zero Zero Ref Zero Low Zero High

O3[ppb] Calibration: AQHI Grimshaw Monthly: 12-2023 Type: SpanAndZero - Span




Span SpanRef Span Low Span High

THC55[ppm] Calibration: AQHI Grimshaw Monthly: 12-2023 Type: SpanAndZero - Zero




Zero Zero Ref Zero Low Zero High

THC55[ppm] Calibration: AQHI Grimshaw Monthly: 12-2023 Type: SpanAndZero - Span




Span SpanRef Span Low Span High

CH4[ppm] Calibration: AQHI Grimshaw Monthly: 12-2023 Type: SpanAndZero - Zero




Zero Zero Ref Zero Low Zero High

CH4[ppm] Calibration: AQHI Grimshaw Monthly: 12-2023 Type: SpanAndZero - Span




Span SpanRef Span Low Span High

NMHC[ppm] Calibration: AQHI Grimshaw Monthly: 12-2023 Type: SpanAndZero - Zero



Zero Zero Ref Zero Low Zero High

NMHC[ppm] Calibration: AQHI Grimshaw Monthly: 12-2023 Type: SpanAndZero - Span



Span SpanRef Span Low Span High

MULTI-POINT CALIBRATION RECORDS

SO2 Analyzer Calibration by Dilution



DATE:	12-Dec-2023	PREVIOUS CALIBRATION DATE:	15-Nov-2023
PARAMETER:	SO2	PREVIOUS CORRECTION FACTOR:	1.000
CLIENT:	PRAMP	TEMPERATURE (°C):	22.2
LOCATION:	Grimshaw	BAROMETRIC (mBar):	940
PURPOSE:	Routine	START TIME (MST):	07:46
PERFORMED BY:	Kevin Sebastian	END TIME (MST):	13:17

ANALYZER:

MAKE/MODEL	Teledyne T100	RANGE	500 ppb
SERIAL #	722	FLOW (mL/min)	483
INITIAL		FINAL	
BKG/OFFSET	32.2	BKG/OFFSET	32.8
COEF/SLOPE	0.894	COEF/SLOPE	0.94
Expected (reference) Value	244.4	Expected (reference) Value	255.6

CALIBRATION SYSTEM:

CALIBRATOR:		ZERO AIR:	
MAKE:	Sabio	MAKE:	Teledyne
MODEL:	2010	MODEL:	M701
ID:	75401122	ID:	5004
MFC CALIBRATION DATE:	09-Sep-2023	OXIDIZER ID:	n/a
CALIBRATION GAS:		FLOWMETERS (if applicable):	
CYLINDER ID:	LL109693	HIGH ID	n/a
CONC (ppm):	25.00	EXPIRY DATE	n/a
CYLINDER (psi):	1400	LOW ID	n/a
EXPIRY DATE	02-Nov-2025	EXPIRY DATE	n/a

CALIBRATION PARAMETERS:

POINT	HIGH	MID	LOW
TARGET	390	190	95
RANGE	300 - 400	150 - 200	50 - 100

SCRUBBER CHECK (15 MINS; TRS/H2S ONLY):

START TIME:	n/a	SO2 Conc (ppb)	n/a
END TIME:	n/a	Analyzer Response (ppb)	n/a

CALIBRATION:


FLOW RATES (mL/min)			CONCENTRATION (ppb)			CORRECTION FACTOR	
DILUENT	GAS	TOTAL	ACTUAL	INDICATED		Initial	Final
				Initial	Final		
4000	60.80	4000	0.00	0.1	0	1.053	1.000
3939	60.80	4000	380.00	361	379.9	1.053	1.000
3971	28.80	4000	180.00	n/a	179.7	n/a	1.002
3985	14.40	3999	90.02	n/a	90.1	n/a	0.999

LINEAR REGRESSION ANALYSIS:

	CORRELATION	SLOPE	INTERCEPT
VALUE	1.000	1.000	0.0%

COMMENTS:

Sample filter changed.
08:46-08:48 Flushed out regulator to check response



TRS Analyzer Calibration by Dilution



DATE:	12-Dec-2023	PREVIOUS CALIBRATION DATE:	15-Nov-2023
PARAMETER:	TRS	PREVIOUS CORRECTION FACTOR:	1.005
CLIENT:	PRAMP	TEMPERATURE (°C):	22.2
LOCATION:	Grimshaw	BAROMETRIC (mBar):	940
PURPOSE:	Routine	START TIME (MST):	07:46
PERFORMED BY:	Kevin Sebastian	END TIME (MST):	13:17

ANALYZER:

MAKE/MODEL	Teledyne T100U	RANGE	100 ppb
SERIAL #	132	FLOW (mL/min)	535
INITIAL		FINAL	
BKG/OFFSET	58.6	BKG/OFFSET	60.1
COEF/SLOPE	0.708	COEF/SLOPE	0.726
Expected (reference) Value	46.48	Expected (reference) Value	48.54

CALIBRATION SYSTEM:

CALIBRATOR:		ZERO AIR:	
MAKE:	Sabio	MAKE:	Teledyne
MODEL:	2010	MODEL:	M701
ID:	75401122	ID:	4568
MFC CALIBRATION DATE:	08-Sep-2023	OXIDIZER ID:	n/a
CALIBRATION GAS:		FLOWMETERS (if applicable):	
CYLINDER ID:	LL131374	HIGH ID	LOW
CONC (ppm):	10.09	EXPIRY DATE	n/a
CYLINDER (psi):	1800	LOW ID	n/a
EXPIRY DATE	03-Jan-2026	EXPIRY DATE	n/a

CALIBRATION PARAMETERS:

POINT	HIGH	MID	LOW
TARGET	78	38	19
RANGE	60 - 80	30 - 40	10 - 20

SCRUBBER CHECK (15 MINS; TRS/H2S ONLY):

START TIME:	n/a	SO2 Conc (ppb)	n/a
END TIME:	n/a	Analyzer Response (ppb)	n/a

CALIBRATION:


FLOW RATES (mL/min)			CONCENTRATION (ppb)			CORRECTION FACTOR	
DILUENT	GAS	TOTAL	ACTUAL	INDICATED		Initial	Final
				Initial	Final		
4000	30.90	4000	0.00	0.57	0	1.029	0.998
3969	30.90	4000	77.95	76.3	78.12	1.029	0.998
3985	15.10	4000	38.09	n/a	37.32	n/a	1.021
3992	7.50	3999	18.92	n/a	18.23	n/a	1.038

LINEAR REGRESSION ANALYSIS:

	CORRELATION	SLOPE	INTERCEPT
VALUE	1.000	1.004	-0.5%

COMMENTS:

Converter, CDNova CDN-101 #576.



NOx Calibration by Dilution/Gas-Phase Titration



CALIBRATION:				ANALYZER:			
DATE:	12-Dec-2023	PREVIOUS CALIBRATION DATE:	16-Nov-2023	MAKE/MODEL:	Teledyne T200	PREVIOUS CF.	
CLIENT:	PRAMP	TEMPERATURE (°C):	22.2	SERIAL #:	837	NOx	0.997
LOCATION:	Grimshaw	BAROMETRIC (mBar):	940	FLOW (mL/min)	443	NO	0.998
PURPOSE:	Routine	START TIME (MST):	07:46	RANGE (ppb)	500	NO2	1.001
PERFORMED BY:	Kevin Sebastian	END TIME (MST):	14:59	GPT FOR O3?		Yes	

CALIBRATOR:		ZERO AIR:		CALIBRATION GAS:		FLOWMETERS (if applicable):	
MAKE:	Sabio	MAKE:	Teledyne	CYLINDER ID:	EY0001716	HIGH ID:	n/a
MODEL:	2010	MODEL:	M701	NO/NOx (PPM):	49.2 49.4	HIGH EXPIRY:	n/a
ID:	26801218	ID:	4568	CYLINDER (psi):	1000	LOW ID:	n/a
MFC CALIBRATION DATE:	12-Sep-2023	OXIDIZER ID:	n/a	EXPIRY DATE	11-Nov-2029	LOW EXPIRY:	n/a

CALIBRATION SETTINGS:							
INITIAL	NOx	NO	NO2	FINAL	NOx	NO	NO2
BKG/OFFSET:	2	-0.5	n/a	BKG/OFFSET:	0.6	-0.4	n/a
SLOPE/COEF/CE:	1	1	0.996	SLOPE/COEF/CE:	0.996	0.994	0.996

EXPECTED (REFERENCE) VALUE:							
INITIAL	NOx	NO	NO2	FINAL	NOx	NO	NO2
	308.4	2.3	306.1		309.1	2.8	306.3


CALIBRATION PARAMETERS:				
POINT	NO TARGET (PPB)	NO2 TARGET (PPB)	NO2 RANGE	O3 POINT
HIGH	395	250	240-275	n/a
MID	180	154	150-157	Mid
LOW	90	54	50-58	Low
EXTRA 1	n/a	340	300-370	High

FLOW RATE			CONCENTRATION (ppb)									CORRECTION FACTOR (CF.)					
(mL/min)			CALCULATED			INITIAL INDICATED			FINAL INDICATED			INITIAL			FINAL		
DILUENT	GAS	TOTAL	NO	NOx	NO2	NO	NOx	NO2	NO	NOx	NO2	NO	NOx	NO2	NO	NOx	NO2
5000	5000	5000	0.0	0.0	0.0	0.2	-0.7	-0.9	0.0	0.0	0.0	0.999	1.002	1.001	0.997	1.001	0.997
4960	40.20	5000	395.6	397.2	1.6	396.0	395.5	-0.6	395.3	398.2	2.9	0.999	1.002	1.001	0.997	1.001	0.997
4982	18.30	5000	180.1	180.8	0.7	n/a	n/a	n/a	180.8	180.6	-0.2	n/a	n/a	0.996	1.001	1.001	0.996
4991	9.20	5000	90.5	90.9	0.4	n/a	n/a	n/a	89.6	90.1	0.4	n/a	n/a	1.010	1.009	1.009	1.009

GPT CALIBRATION:										
Point	CALIBRATOR			INDICATED (ppb)			NO DROP / O3 Conc (ppb)	NO2 GAIN (ppb)	NO2 Corr. FACTOR	CONV. EFFICIENCY
	GAS	TOTAL	O3 SETPOINT	NO	NOx	NO2				
REFERENCE	40.10	5000	0	395.8	399.1	3.3	264.7	264.4	1.001	99.89%
AS-FOUND HIGH	40.10	5000	260	131.1	398.9	267.7	264.7	264.4	1.001	99.89%
ADJUSTED HIGH	n/a	n/a	n/a	n/a	n/a	n/a	n/a	n/a	n/a	n/a
MID	40.10	5000	150	244.5	397.7	153.3	151.3	150	1.009	99.14%
LOW	40.10	5000	55	343.3	400.1	56.8	52.5	53.5	0.981	101.90%
NO2 adjustment not required.									AVERAGE:	100.31%

LINEAR REGRESSION ANALYSIS:				COMMENTS:
	CORRELATION	SLOPE	INTERCEPT	
NO	1.000	1.000	-0.03%	
NOx	1.000	1.003	-0.11%	
NO2	1.000	0.994	0.14%	

Sample filter changed
Additional point for O3: Setpoint = 350, O3 conc = 362.3



CAL-PRAMP-202312-01689

Ozone Calibration by Direct GPT



DATE:	12-Dec-2023	PREVIOUS CALIBRATION DATE:	16-Nov-2023
PARAMETER:	O3	PREVIOUS CORRECTION FACTOR:	0.998
CLIENT:	PRAMP	TEMPERATURE (°C):	22.5
LOCATION:	Grimshaw	BAROMETRIC (mBar):	938
PURPOSE:	Routine	START TIME (MST):	14:15
PERFORMED BY:	Kevin Sebastian	END TIME (MST):	18:36

ANALYZER:

MAKE/MODEL	Teledyne T400	RANGE	500 ppb
SERIAL #	824	FLOW (mL/min)	758
INITIAL		FINAL	
BKG/OFFSET	-1.8	BKG/OFFSET	-2.1
COEF/SLOPE	0.987	COEF/SLOPE	0.988
Expected (reference) Value	262	Expected (reference) Value	268

CALIBRATION SYSTEM:

CALIBRATOR:		ZERO AIR:	
MAKE:	Sabio	MAKE:	Teledyne
MODEL:	2010	MODEL:	M701
ID:	26801218	ID:	4568
MFC CALIBRATION DATE:	12-Sep-2023	OXIDIZER ID:	n/a
CALIBRATION METHOD:		Direct GPT	
GPT DATE:	12-Dec-2023	GPT END TIME:	14:00

CALIBRATION PARAMETERS:

POINT	HIGH	MID	LOW
RANGE	300 - 400	150 - 200	50 - 100

CALIBRATION:


FLOW RATES (mL/min)			CONCENTRATION (ppb)			CORRECTION FACTOR	
DILUENT	GAS	TOTAL	ACTUAL	INDICATED		Initial	Final
				Initial	Final		
5000	 	5000	0.0	-1.4	0.0	 	
5000	 	5000	362.3	360.9	361.2	1.000	1.003
5000	 	5000	151.3	n/a	152.2	n/a	0.994
5000	 	5000	52.5	n/a	52.2	n/a	1.006

LINEAR REGRESSION ANALYSIS:

	CORRELATION	SLOPE	INTERCEPT
VALUE	1.000	0.997	0.1%

COMMENTS:

Sample filter changed



Methane/Non-Methane Analyzer Calibration by Dilution



CALIBRATION:				ANALYZER:			
DATE:	12-Dec-2023	PREVIOUS CALIBRATION DATE:	15-Nov-2023	VALUE	MAKE/MODEL	SERIAL	FLOW (mL/min)
CLIENT:	PRAMP	TEMPERATURE (°C):	22.2		Thermo 55i	1191032505	1071
LOCATION:	Grimshaw	BAROMETRIC (mBar):	939	PARAMETER:	CH4	NMHC	THC
PURPOSE	Routine	START TIME (MST):	12:47	RANGE (ppm):	20	20	40
PERFORMED BY:	Kevin Sebastian	END TIME (MST):	18:25	PREVIOUS CF:	1.001	1.001	1.001

CALIBRATION SYSTEM:

CALIBRATOR:		ZERO AIR:		CALIBRATION GAS:		FLOWMETERS (if applicable):	
MAKE:	Sabio	MAKE:	Teledyne	CYLINDER ID:	LL68768	HIGH ID:	n/a
MODEL:	2010	MODEL:	M701	CH ₄ /C ₃ H ₈ (ppm):	897.0 301.0	HIGH EXPIRY:	n/a
ID:	75401122	ID:	4568	CYLINDER (psi):	1900	LOW ID:	n/a
MFC CALIBRATION DATE:	08-Sep-2023	OXIDIZER ID:	Internal	EXPIRY DATE	08-Nov-2029	LOW EXPIRY:	n/a

CALIBRATION PARAMETERS:

POINT (CH ₄ /NMHC)	HIGH	MID	LOW	CH ₄ EQUIVILANCE		
TARGET	14	7	3.5	C ₃ H ₈ as CH ₄		827.8
RANGE	12 - 16	6 - 8	2 - 4	THC as CH ₄		1724.8

EXPECTED (REFERENCE) VALUE:


INITIAL	CH ₄	NMHC	THC	FINAL	CH ₄	NMHC	THC
	9.10	10.57	19.68		9.59	11.16	20.75

CALIBRATION:

FLOW RATE			CONCENTRATION (PPM)									CORRECTION FACTOR (CF.)					
(mL/min)			CALCULATED			INITIAL INDICATED			FINAL INDICATED			INITIAL			FINAL		
DILUENT	GAS	TOTAL	CH ₄	NMHC	THC	CH ₄	NMHC	THC	CH ₄	NMHC	THC	CH ₄	NMHC	THC	CH ₄	NMHC	THC
3250	X	3250	0.00	0.00	0.00	0.01	0.00	0.01	0.00	0.00	0.00	X	X	X	X	X	X
3196	52.80	3249	14.58	13.45	28.03	14.83	13.77	28.61	14.58	13.45	28.03	0.983	0.977	0.980	1.000	1.000	1.000
3223	26.40	3249	7.29	6.73	14.01	n/a	n/a	n/a	7.28	6.75	14.03	n/a	n/a	n/a	1.001	0.996	0.999
3237	13.20	3250	3.64	3.36	7.01	n/a	n/a	n/a	3.67	3.40	7.07	n/a	n/a	n/a	0.993	0.989	0.991

LINEAR REGRESSION ANALYSIS:

	CORRELATION	SLOPE	INTERCEPT	Comments:	
CH ₄	1.000	0.999	0.0%	H2 = AMA HG300 #190567059 Sample Filter Changed	
NMHC	1.000	0.999	0.1%		
THC	1.000	0.999	0.1%		
				Use Zero Chrom?	Yes



CAL-PRAMP-202312-01689

Page 21 of 25
CH4 [ppm] NMHC [ppm]



Teledyne T640 Audit/Calibration

Date/Previous Audit Date:	December 12, 2023	November 16, 2023	Weather Conditions:		Mix of sun and clouds
Company:	PRAMP		Start Time (mst):		15:09
Station:	Grimshaw		End Time (mst):		15:32
Parameter:	PM 2.5		Performed By/Reviewer:		Kevin Sebastian Limin Li
Instrument Data:					
Make/Model:	Teledyne T640		Serial Number:		318
Owner:	PRAMP		Alarms (detail in comments):		Yes
Reference Standards/I.D./Expiry Date:					
Flow Standard: DeltaCal DC1 #206578, expires Nov 27, 2024			Temperature: DeltaCal DC1 #206578, expires Nov 27, 2024		
Digital Manometer: DeltaCal DC1 #206578, expires Nov 27, 2024			Pressure: DeltaCal DC1 #206578, expires Nov 27, 2024		
DIAGNOSTICS:					
Ambient Pressure (mmHg)	703.4	Ambient Temp (°C)	-2.3	ASC Heater Duty (%)	0.0
Box Temp (°C)	26.5	Current PMT HV (V)	1527	LED Temp (°C)	34.88
P3 Value	47	PMT Setting (V)	1532	Pump PWM (%)	66
Sample Flow (L/min)	4.91	Sample RH (%RH)	11.8	Sample Temp (°C)	24.3
Item:	As-found		As-left		Tolerance
	Reference	T640x	Reference	T640x	
Zero Test (Leak Check)	PM10	0.0	0	0.0	0.0 to 0.2
	PM2.5	0.0	0	0.0	
Ambient Pressure (mmHg)	702.0	708.4	n/a	n/a	+/- 10 mm Hg
Ambient Temperature (°C)	-2.30	-2.3	n/a		+/- 2°C
Sample Flow (L/min)	5.03	4.98	n/a	n/a	+/- 5% of T640x (e.g., 4.75 – 5.25 lpm)
Additional Monthly Maintenance :					Completed
Inlet cleaned?					Yes
Sample tubing inspected (inner and outer)?					Yes
Quarterly Audit/Calibration:					
SpanDust™ Standard	Peak at Channel		Lot No:		Expiry:
	10.9		100128-050-046		1-31-2025
Item:	Verification:		Calibration (if needed):		Tolerance
	Reference	T640x	Reference	T640x	
Peak Channel	10.9	11	10.9	n/a	± 0.5
PMT Setting (V)	n/a	1546	n/a	n/a	n/a
Peak Channel Counts:	n/a	1014	n/a	n/a	n/a
Comments:					
Alert: 11/16/2023: System Reset No issues					

Meteorological System Checklist



Date:	December 12, 2023		
Technician:	Kevin Sebastian		
Station:	PRAMP Grimshaw		
Unit:	Make:	Model:	Serial #:
Temperature Sensor:	Vaisala	HMP155	N2910506
Barometric Pressure Sensor:	MetOne	92	A2397
Relative Humidity Sensor:	Vaisala	HMP155	N2910506
Anemometer:	RM Young	05305AQ	174801
AMBIENT TEMPERATURE SENSOR CHECK			
Previous check date:	November 15, 2023		
Parameter:	Temperature @ 2 metres		
Reference Thermometer ID:	F.S. 181341226 expires July 17, 2024		
Reference Temperature (°C):	-2.5		
Station - Ambient Temperature (°C):	-2.3		
Temperature Difference (°C):	-0.2		
BAROMETRIC PRESSURE SENSOR CHECK			
Previous check date:	November 15, 2023		
Reference Barometer ID:	Brunton 05535 Expires July 17, 2024		
Reference Pressure - Units/Reading:	millibar	937.7	
Station Pressure - Units/Reading:	millibar	936	
Pressure Tolerance +/- 15% of error:	797 - 1078	0.18%	
RELATIVE HUMIDITY (HYGROMETER) SENSOR CHECK			
Previous check date:	November 15, 2023		
Reference Hygrometer ID:	F.S. 181341226 expires July 17, 2024		
Reference Hygrometer % RH- Reading:	66.70		
Station Hygrometer % RH- Reading:	64.80		
RH Tolerance +/- 15% of difference:	56.70 - 76.71	2.8%	
ANEMOMETER - WIND SPEED & WIND DIRECTION SENSOR CHECK			
WIND SPEED		WIND DIRECTION	
Previous check date:	November 15, 2023	Previous check date:	November 15, 2023
Wind Speed Observed (kph):	5~10	Wind Direction Observed:	W
Wind speed on Data Logger (kph):	9.6	Wind Direction on Data Logger:	W
		Wind Direction Pass/Fail?:	Pass
Comments			
No issues			



Meteorological Sensor Audit/Calibration

Location Information

Company: PRAMP
 Audit Location: Grimshaw
 Audit Date: August 2, 2023
 Calibration Purpose: routine annual

Performed By: Chris Wesson
 Reviewed By: Limin Li
 Start/End Time (mst): 14:55 / 16:16
 Weather Conditions: Mainly cloudy with sunny breaks

Wind Sensor Information

Sensor ID Data:		Sensor Outputs:	
Sensor Make:	RM Young	Velocity Voltage Output Range:	n/a
Sensor Model:	05305AQ	Velocity Unit Output Range:	0-200
Serial #:	174801	Direction Voltage Output Range:	n/a
Previous Cal/Audit Date:	July 12, 2022	Direction Unit Output Range:	0-360

Wind Calibrator Information

Calibrator I.D. and Expiry Date: RM Young 18802 id# R9133 expires Oct 18, 2024

Wind Speed Audit Data ****+/- 2% of the average correction factor is the limit****

RPM	Wind Speed Generated kph	Clockwise Wind Speed kph	Counter Clockwise Wind Speed kph	Correction Factor
0	0	0.0	0.0	-
1000	18.4	18.3	18.3	1.007
2000	36.9	36.7	36.7	1.004
3000	55.3	55.1	55.1	1.003
4000	73.7	73.6	73.6	1.002
5000	92.2	92.0	92.0	1.002
6000	110.6	110.5	110.5	1.001
7000	129.0	128.9	128.9	1.001
8000	147.4	147.3	147.3	1.001
9000	165.9	165.8	165.8	1.000
10000	184.3	184.2	184.2	1.001
The audit meets AMD requirements.			Average Correction Factor=	1.002

Wind Direction Audit Data ****+/- 3° of the absolute average degrees difference for all points is the limit****

Generated Wind Direction 0-360 (Up)	Generated Wind Direction 360-0 (Down)	Indicated Wind Direction 0-360 (Up)	Indicated Wind Direction 360-0 (Down)	Degrees Difference 0-360 (Up)	Degrees Difference 360-0 (Down)	Average Absolute Degrees Difference
0	355	2	354	2.0	1.0	1.5
30	330	29	333	1.0	-3.0	2.0
60	300	58	300	2.0	0.0	1.0
90	270	88	271	2.0	-1.0	1.5
120	240	119	237	1.0	3.0	2.0
150	210	149	206	1.0	4.0	2.5
180	180	178	177	2.0	3.0	2.5
210	150	206	148	4.0	2.0	3.0
240	120	238	119	2.0	1.0	1.5
270	90	272	87	-2.0	3.0	2.5
300	60	304	57	-4.0	3.0	3.5
330	30	333	30	-3.0	0.0	1.5
355	0	354	1	1.0	1.0	1.0
The audit meets AMD requirements.				Average Absolute Degrees Difference=		2.0

Comments:

Declination = 15 deg East
 Horizontal bearings replaced.
 Potentiometer noisy. Replacement required.

END OF REPORT



Peace River Area Monitoring Program

DECEMBER 2023

Ambient Air Monitoring Calibration Report

- PEACE RIVER COMPLEX (PRC) STATION-

CAL-PRAMP-202312-01698

Operation and Maintenance:

Bureau Veritas Canada


Data Validation and Report:

Bureau Veritas Canada

January 6, 2024


DAILY INTERNAL ZERO-SPAN CALIBRATION RECORDS

SO2[ppb] Calibration: Peace River Complex [PRC] Monthly: 12-2023 Type: SpanAndZero - Zero




Zero Zero Ref Zero Low Zero High

SO2[ppb] Calibration: Peace River Complex [PRC] Monthly: 12-2023 Type: SpanAndZero - Span




Span SpanRef Span Low Span High

H2S[ppb] Calibration: Peace River Complex (PRC) Monthly: 12-2023 Type: SpanAndZero - Zero




Zero Zero Ref Zero Low Zero High

H2S[ppb] Calibration: Peace River Complex (PRC) Monthly: 12-2023 Type: SpanAndZero - Span




Span SpanRef Span Low Span High

TRS[ppb] Calibration: Peace River Complex (PRC) Monthly: 12-2023 Type: SpanAndZero - Zero




Zero Zero Ref Zero Low Zero High

TRS[ppb] Calibration: Peace River Complex (PRC) Monthly: 12-2023 Type: SpanAndZero - Span




Span SpanRef Span Low Span High

THC55[ppm] Calibration: Peace River Complex (PRC) Monthly: 12-2023 Type: SpanAndZero - Zero




Zero Zero Ref Zero Low Zero High

THC55[ppm] Calibration: Peace River Complex (PRC) Monthly: 12-2023 Type: SpanAndZero - Span




Span Span Ref Span Low Span High

CH4[ppm] Calibration: Peace River Complex (PRC) Monthly: 12-2023 Type: SpanAndZero - Zero




Zero Zero Ref Zero Low Zero High

CH4[ppm] Calibration: Peace River Complex (PRC) Monthly: 12-2023 Type: SpanAndZero - Span




Span SpanRef Span Low Span High

NMHC[ppm] Calibration: Peace River Complex (PRC) Monthly: 12-2023 Type: SpanAndZero - Zero



Zero Zero Ref Zero Low Zero High

NMHC[ppm] Calibration: Peace River Complex (PRC) Monthly: 12-2023 Type: SpanAndZero - Span



Span Span Ref Span Low Span High

MULTI-POINT CALIBRATION RECORDS

SO2 Analyzer Calibration by Dilution



DATE:	06-Dec-2023	PREVIOUS CALIBRATION DATE:	01-Nov-2023
PARAMETER:	SO2	PREVIOUS CORRECTION FACTOR:	0.999
CLIENT:	PRAMP	TEMPERATURE (°C):	21.4
LOCATION:	Peace River Compliance	BAROMETRIC (mBar):	932
PURPOSE:	Routine	START TIME (MST):	10:36
PERFORMED BY:	Kevin Sebastian	END TIME (MST):	15:44

ANALYZER:

MAKE/MODEL	Thermo 43i	RANGE	500 ppb
SERIAL #	1034746225	FLOW (mL/min)	436
INITIAL		FINAL	
BKG/OFFSET	19.9	BKG/OFFSET	20.2
COEF/SLOPE	1.147	COEF/SLOPE	1.151
Expected (reference) Value	269.6	Expected (reference) Value	272

CALIBRATION SYSTEM:

CALIBRATOR:		ZERO AIR:	
MAKE:	Sabio	MAKE:	Teledyne
MODEL:	2010	MODEL:	M701
ID:	75401122	ID:	5004
MFC CALIBRATION DATE:	08-Sep-2023	OXIDIZER ID:	n/a
CALIBRATION GAS:		FLOWMETERS (if applicable):	
CYLINDER ID:	LL109693	HIGH ID	n/a
CONC (ppm):	25.00	EXPIRY DATE	n/a
CYLINDER (psi):	1600	LOW ID	n/a
EXPIRY DATE	02-Nov-2025	EXPIRY DATE	n/a

CALIBRATION PARAMETERS:

POINT	HIGH	MID	LOW
TARGET	390	190	95
RANGE	300 - 400	150 - 200	50 - 100

SCRUBBER CHECK (15 MINS; TRS/H2S ONLY):

START TIME:	n/a	SO2 Conc (ppb)	n/a
END TIME:	n/a	Analyzer Response (ppb)	n/a

CALIBRATION:


FLOW RATES (mL/min)			CONCENTRATION (ppb)			CORRECTION FACTOR	
DILUENT	GAS	TOTAL	ACTUAL	INDICATED		Initial	Final
				Initial	Final		
3999	60.80	3999	0.00	0	0	0.996	0.995
3939	60.80	4000	380.00	381.7	382.1	0.996	0.995
3970	28.80	3999	180.05	n/a	181.5	n/a	0.992
3985	14.40	3999	90.02	n/a	90.4	n/a	0.996

LINEAR REGRESSION ANALYSIS:

	CORRELATION	SLOPE	INTERCEPT
VALUE	1.000	1.006	0.0%

COMMENTS:

Sample filter changed.



H2S Analyzer Calibration by Dilution



DATE:	06-Dec-2023	PREVIOUS CALIBRATION DATE:	01-Nov-2023
PARAMETER:	H2S	PREVIOUS CORRECTION FACTOR:	0.999
CLIENT:	PRAMP	TEMPERATURE (°C):	21.4
LOCATION:	Peace River Compliance	BAROMETRIC (mBar):	932
PURPOSE:	Routine	START TIME (MST):	10:36
PERFORMED BY:	Kevin Sebastian	END TIME (MST):	15:44

ANALYZER:

MAKE/MODEL	Thermo 450i	RANGE	100 ppb
SERIAL #	1308857354	FLOW (mL/min)	931
INITIAL		FINAL	
BKG/OFFSET	14.9	BKG/OFFSET	15.7
COEF/SLOPE	1.045	COEF/SLOPE	1.091
Expected (reference) Value	35.4	Expected (reference) Value	38.2

CALIBRATION SYSTEM:

CALIBRATOR:		ZERO AIR:	
MAKE:	Sabio	MAKE:	Teledyne
MODEL:	2010	MODEL:	M701
ID:	75401122	ID:	4568
MFC CALIBRATION DATE:	08-Sep-2023	OXIDIZER ID:	n/a
CALIBRATION GAS:		FLOWMETERS (if applicable):	
CYLINDER ID:	LL131374	HIGH ID	n/a
CONC (ppm):	10.09	EXPIRY DATE	n/a
CYLINDER (psi):	1800	LOW ID	n/a
EXPIRY DATE	03-Jan-2026	EXPIRY DATE	n/a

CALIBRATION PARAMETERS:

POINT	HIGH	MID	LOW
TARGET	78	38	19
RANGE	60 - 80	30 - 40	10 - 20

SCRUBBER CHECK (15 MINS; TRS/H2S ONLY):

START TIME:	n/a	SO2 Conc (ppb)	n/a
END TIME:	n/a	Analyzer Response (ppb)	n/a

CALIBRATION:


FLOW RATES (mL/min)			CONCENTRATION (ppb)			CORRECTION FACTOR	
DILUENT	GAS	TOTAL	ACTUAL	INDICATED		Initial	Final
				Initial	Final		
3999	30.90	3999	0.00	0.2	0	1.032	0.998
3969	30.90	4000	77.95	75.7	78.1	1.032	0.998
3984	15.10	3999	38.10	n/a	38.1	n/a	1.000
3992	7.50	3999	18.92	n/a	19.1	n/a	0.991

LINEAR REGRESSION ANALYSIS:

	CORRELATION	SLOPE	INTERCEPT
VALUE	1.000	1.001	0.0%

COMMENTS:

Sample filter changed



TRS Analyzer Calibration by Dilution



DATE:	06-Dec-2023	PREVIOUS CALIBRATION DATE:	02-Nov-2023
PARAMETER:	TRS	PREVIOUS CORRECTION FACTOR:	0.996
CLIENT:	PRAMP	TEMPERATURE (°C):	21.4
LOCATION:	Peace River Compliance	BAROMETRIC (mBar):	932
PURPOSE:	Routine	START TIME (MST):	10:36
PERFORMED BY:	Kevin Sebastian	END TIME (MST):	15:44

ANALYZER:

MAKE/MODEL	Thermo 450i	RANGE	100 ppb
SERIAL #	1034746224	FLOW (mL/min)	700
INITIAL		FINAL	
BKG/OFFSET	26.4	BKG/OFFSET	28.2
COEF/SLOPE	1.023	COEF/SLOPE	1.083
Expected (reference) Value	51.25	Expected (reference) Value	55.14

CALIBRATION SYSTEM:

CALIBRATOR:		ZERO AIR:	
MAKE:	Sabio	MAKE:	Teledyne
MODEL:	2010	MODEL:	M701
ID:	75401122	ID:	4568
MFC CALIBRATION DATE:	08-Sep-2023	OXIDIZER ID:	n/a
CALIBRATION GAS:		FLOWMETERS (if applicable):	
CYLINDER ID:	LL131374	HIGH ID	n/a
CONC (ppm):	10.09	EXPIRY DATE	n/a
CYLINDER (psi):	1800	LOW ID	n/a
EXPIRY DATE	03-Jan-2026	EXPIRY DATE	n/a

CALIBRATION PARAMETERS:

POINT	HIGH	MID	LOW
TARGET	78	38	19
RANGE	60 - 80	30 - 40	10 - 20

SCRUBBER CHECK (15 MINS; TRS/H2S ONLY):

START TIME:	n/a	SO2 Conc (ppb)	n/a
END TIME:	n/a	Analyzer Response (ppb)	n/a

CALIBRATION:


FLOW RATES (mL/min)			CONCENTRATION (ppb)			CORRECTION FACTOR	
DILUENT	GAS	TOTAL	ACTUAL	INDICATED		Initial	Final
				Initial	Final		
3999	30.90	3999	0.00	0.02	0	1.053	1.000
3969	30.90	4000	77.95	74.06	77.97	1.053	1.000
3984	15.10	3999	38.10	n/a	38.1	n/a	1.000
3992	7.50	3999	18.92	n/a	18.82	n/a	1.005

LINEAR REGRESSION ANALYSIS:

	CORRELATION	SLOPE	INTERCEPT
VALUE	1.000	1.001	0.0%

COMMENTS:

TRS Converter CDNOVA CDN-101 #516



Methane/Non-Methane Analyzer Calibration by Dilution



CALIBRATION:				ANALYZER:			
DATE:	06-Dec-2023	PREVIOUS CALIBRATION DATE:	01-Nov-2023	VALUE	MAKE/MODEL	SERIAL	FLOW (mL/min)
CLIENT:	PRAMP	TEMPERATURE (°C):	21.4		Thermo 55i	1034745845	1088
LOCATION:	Peace River Compliance	BAROMETRIC (mBar):	932	PARAMETER:	CH4	NMHC	THC
PURPOSE	Routine	START TIME (MST):	10:36	RANGE (ppm):	20	20	40
PERFORMED BY:	Kevin Sebastian	END TIME (MST):	15:44	PREVIOUS CF:	1.001	1.001	1.001

CALIBRATION SYSTEM:

CALIBRATOR:		ZERO AIR:		CALIBRATION GAS:		FLOWMETERS (if applicable):	
MAKE:	Sabio	MAKE:	Teledyne	CYLINDER ID:	LL68768	HIGH ID:	n/a
MODEL:	2010	MODEL:	M701	CH ₄ /C ₃ H ₈ (ppm):	897.0 301.0	HIGH EXPIRY:	n/a
ID:	26801218	ID:	4568	CYLINDER (psi):	1600	LOW ID:	n/a
MFC CALIBRATION DATE:	12-Sep-2023	OXIDIZER ID:	Internal	EXPIRY DATE	08-Nov-2029	LOW EXPIRY:	n/a

CALIBRATION PARAMETERS:

POINT (CH ₄ /NMHC)	HIGH	MID	LOW	CH ₄ EQUIVILANCE		
TARGET	14	7	3.5	C ₃ H ₈ as CH ₄		827.8
RANGE	12 - 16	6 - 8	2 - 4	THC as CH ₄		1724.8

EXPECTED (REFERENCE) VALUE:


INITIAL	CH ₄	NMHC	THC	FINAL	CH ₄	NMHC	THC
	9.62	11.08	20.70		9.85	11.03	20.88

CALIBRATION:

FLOW RATE			CONCENTRATION (PPM)									CORRECTION FACTOR (CF.)					
(mL/min)			CALCULATED			INITIAL INDICATED			FINAL INDICATED			INITIAL			FINAL		
DILUENT	GAS	TOTAL	CH ₄	NMHC	THC	CH ₄	NMHC	THC	CH ₄	NMHC	THC	CH ₄	NMHC	THC	CH ₄	NMHC	THC
3099	X	3099	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	X	X	X	X	X	X
3049	50.30	3099	14.56	13.44	27.99	14.27	13.44	27.72	14.61	13.47	28.08	1.020	1.000	1.010	0.997	0.997	0.997
3074	25.10	3099	7.27	6.70	13.97	n/a	n/a	n/a	7.31	6.74	14.05	n/a	n/a	n/a	0.994	0.995	0.994
3085	12.60	3098	3.65	3.37	7.01	n/a	n/a	n/a	3.67	3.39	7.06	n/a	n/a	n/a	0.994	0.993	0.994

LINEAR REGRESSION ANALYSIS:

	CORRELATION	SLOPE	INTERCEPT	Comments:	
CH ₄	1.000	1.003	0.0%	H2 Dessicant changed after AF high. H2 = AMA HG300 #211067076	
NMHC	1.000	1.002	0.0%		
THC	1.000	1.003	0.0%		
				Use Zero Chrom?	No



CAL-PRAMP-202312-01698

Meteorological System Checklist



Date:	December 6, 2023		
Technician:	Kevin Sebastian		
Station:	Peace River Compliance		
Unit:	Make:	Model:	Serial #:
Precipitation Sampler:	RM Young	52202	TB 16325
Temperature Sensor:	Rotronic	HC2-S3	20558318
Barometric Pressure Sensor:	MetOne	092	B19577
Relative Humidity Sensor:	Rotronic	HC2-S3	20558318
Anemometer:	RM Young	05305VK	129612
AMBIENT TEMPERATURE SENSOR CHECK			
Previous check date:	November 1, 2023		
Parameter:	Temperature @ 2 metres		
Reference Thermometer ID:	F.S. 181341226 expires July 17, 2024		
Reference Temperature (°C):	3.2		
Station - Ambient Temperature (°C):	3.0		
Temperature Difference (°C):	0.2		
BAROMETRIC PRESSURE SENSOR CHECK			
Previous check date:	November 1, 2023		
Reference Barometer ID:	Brunton 05535 Expires July 17, 2024		
Reference Pressure - Units/Reading:	millibar	928.3	
Station Pressure - Units/Reading:	millibar	931	
Pressure Tolerance +/- 15% of error:	789 - 1068	-0.29%	
RELATIVE HUMIDITY (HYGROMETER) SENSOR CHECK			
Previous check date:	November 1, 2023		
Reference Hygrometer ID:	F.S. 181341226 expires July 17, 2024		
Reference Hygrometer % RH- Reading:	60.70		
Station Hygrometer % RH- Reading:	65.70		
RH Tolerance +/- 15% of difference:	51.60 - 69.81	-8.2%	
ANEMOMETER - WIND SPEED & WIND DIRECTION SENSOR CHECK			
WIND SPEED		WIND DIRECTION	
Previous check date:	November 1, 2023	Previous check date:	November 1, 2023
Wind Speed Observed (kph):	10~20	Wind Direction Observed:	W
Wind speed on Data Logger (kph):	12	Wind Direction on Data Logger:	W
		Wind Direction Pass/Fail?:	Pass
Comments			
No issues.			



Meteorological Sensor Audit/Calibration

Location Information

Company: PRAMP
Audit Location: PRC Compliance
Audit Date: August 3, 2023
Calibration Purpose: routine annual

Performed By: Chris Wesson
Reviewed By: Limin Li
Start/End Time (mst): 10:40 / 11:32
Weather Conditions: Mainly cloudy with drizzle

Wind Sensor Information

Sensor ID Data:		Sensor Outputs:	
Sensor Make:	RM Young	Velocity Voltage Output Range:	0-1
Sensor Model:	05305VK	Velocity Unit Output Range:	0-200
Serial #:	129612	Direction Voltage Output Range:	0-1
Previous Cal/Audit Date:	August 17, 2022	Direction Unit Output Range:	0-360

Wind Calibrator Information

Calibrator I.D. and Expiry Date: RM Young 18802 id# R9133 expires Oct 18, 2024

Wind Speed Audit Data ****+/- 2% of the average correction factor is the limit****

RPM	Wind Speed Generated kph	Clockwise Wind Speed kph	Counter Clockwise Wind Speed kph	Correction Factor
0	0	0.1	0.1	-
1000	18.4	18.6	18.5	0.994
2000	36.9	37.1	37.0	0.995
3000	55.3	55.5	55.4	0.997
4000	73.7	74.0	73.8	0.998
5000	92.2	92.5	92.4	0.997
6000	110.6	111.0	111.0	0.996
7000	129.0	129.6	129.6	0.995
8000	147.4	148.0	148.3	0.995
9000	165.9	166.7	166.6	0.995
10000	184.3	185.3	185.3	0.995
The audit meets AMD requirements.			Average Correction Factor=	0.996

Wind Direction Audit Data ****+/- 3° of the absolute average degrees difference for all points is the limit****

Generated Wind Direction 0-360 (Up)	Generated Wind Direction 360-0 (Down)	Indicated Wind Direction 0-360 (Up)	Indicated Wind Direction 360-0 (Down)	Degrees Difference 0-360 (Up)	Degrees Difference 360-0 (Down)	Average Absolute Degrees Difference
0	355	0	352	0.0	3.0	1.5
30	330	31	328	-1.0	2.0	1.5
60	300	59	298	1.0	2.0	1.5
90	270	90	270	0.0	0.0	0.0
120	240	120	240	0.0	0.0	0.0
150	210	151	210	-1.0	0.0	0.5
180	180	180	180	0.0	0.0	0.0
210	150	211	150	-1.0	0.0	0.5
240	120	241	120	-1.0	0.0	0.5
270	90	271	91	-1.0	-1.0	1.0
300	60	299	59	1.0	1.0	1.0
330	30	328	29	2.0	1.0	1.5
355	0	353	1	2.0	1.0	1.5
The audit meets AMD requirements.				Average Absolute Degrees Difference=		0.8

Comments:

Declination = 15 deg East
Physical inspection completed, bearings replaced

END OF REPORT

List of SOPs

MONITOR	SOP
SULPHUR DIOXIDE (SO ₂)	Bureau Veritas EMS SOP-00209: Ambient Sulphur Monitoring
HYDROGEN SULPHIDE (H ₂ S)	Bureau Veritas EMS SOP-00209: Ambient Sulphur Monitoring
TOTAL REDUCED SULPHUR (TRS)	Bureau Veritas EMS SOP-00209: Ambient Sulphur Monitoring
TOTAL HYDROCARBONS (THC), METHANE (CH ₄), NON-METHANE(NMHC)	Bureau Veritas EMS SOP-00001: Methane, Non-Methane Hydrocarbon Analyzer Monitoring
OXIDES OF NITROGEN (NO _x), NITRIC OXIDE (NO) & NITROGEN DIOXIDE (NO ₂)	Bureau Veritas EMS SOP-00213: Ambient NO/NO ₂ /NO _x Monitoring
OZONE (O ₃)	Bureau Veritas EMS SOP-00212: Ambient O ₃ Monitoring
PARTICULATE MATTER < 2.5 MICRONS (PM _{2.5})	Bureau Veritas EMS SOP-00015: Teledyne API PM Monitor Model T640
WIND SPEED (WS) & WIND DIRECTION (WD)	Bureau Veritas EMS SOP-00013: RM Young Wind Monitor Calibration



Peace River Area Monitoring Program

DECEMBER 2023

**Monthly Ambient Air Quality Monitoring Integrated
Sampling Report**

PRAMP-202312-INTEGRATED

January 12, 2024

Pages may be left blank for double-sided printing

Table of Contents

NETWORK STATION SUMMARY	5
Listing of Integrated Sampling Stations	5
Listing of Passives: 1-Month Exposure Sampling Sites	5
List of Contractors who performed the air monitoring activities	5
Monitoring Notes during the Month of December 2023	6
Revisions to Alberta’s Ambient Air Quality Data Warehouse.....	6
Deviations from Authorized Monitoring Methods	6
Certification.....	7
Map of PRAMP Continuous Monitoring and Integrated Sampling Network.....	8
INTEGRATED SAMPLING RESULTS SUMMARY	9
ANALYTICAL SAMPLING RESULTS	10
NMHC Canisters – VOCs.....	11
Passives	20
End of Report	22



Peace River Area Monitoring Program
Suite 91, 305 – 4625 Varsity Drive NW
Calgary, AB, T3A 0Z9
Phone #: 780-226-7068 / 587-225-2248
E-mail: pramptech@prampairshed.ca
www.prampairshed.ca

January 12, 2024

Alberta Environment and Protected Areas (EPA)
11th Floor, Oxbridge Place
9820 106 Street
Edmonton, AB, T5K 2J6

RE: PRAMP –December 2023 Monthly Ambient Air Quality Monitoring Integrated Sampling Report

Enclosed is the December 2023 Monthly Ambient Air Quality Monitoring Integrated Sampling Report for the Peace River Area Monitoring Program's (PRAMP) regional air quality monitoring network. This report summarizes monitoring data for samples collected using integrated methods, including volatile organic compounds (NMHC canister sampling program), hydrogen sulphide, and sulphur dioxide (passive sampling program).

The representative of the Person Responsible for this monitoring program is

PRAMP Airshed
Michael Bisaga / Lily Lin, Technical Program Managers
Suite 91, 305 – 4625 Varsity Drive NW
Calgary, AB, T3A 0Z9
Phone #: 780-226-7068 / 587-225-2248
E-mail: pramptech@prampairshed.ca

This report has been prepared, reviewed and submitted by Michael Bisaga & Lily Lin of the PRAMP Airshed. This report is also submitted on behalf of the industrial member companies to satisfy the requirements of the facility Operating Approvals.

NETWORK STATION SUMMARY

Listing of Integrated Sampling Stations

- 986-C Station
- 842-B Station
- Reno-B Station
- Peace River Complex (PRC) Station

Station Name	986-C	842-B	Reno-B	PRC
Station ID	1562	1561	1563	1698
Coordinates	56.36980, -116.92500	56.27406, -116.98129	55.86936, -117.05739	56.38257, -116.769283
NMHC Canister (VOCs)	√	√	√	
Passives: 2-Month exposure (PACs)	√			
Passives: 1-Month Exposure (H ₂ S, SO ₂)				√

Listing of Passives: 1-Month Exposure Sampling Sites

Site ID	Latitude	Longitude
1	56.377841	-116.787142
2	56.378638	-116.780496
3	56.382958	-116.783813
4	56.377044	-116.794220
7	56.384796	-116.780488
8	56.388710	-116.771234
9	56.388943	-116.756205
10	56.388642	-116.797817
11	56.383771	-116.841165
12	56.388962	-116.885263
13	56.390972	-116.822083
14	56.424825	-116.853181

List of Contractors who performed the air monitoring activities

Sampling Program	Monitoring Activities Conducted By	Sample Analysis Conducted By	Data/Report Prepared By	Electronic Submission Conducted By
NMHC Canister (VOCs)	Bureau Veritas	InnoTech Alberta Inc	PRAMP	PRAMP
Passives: PACs	PRAMP	ECCC	AEP	AEP
Passives: H ₂ S, SO ₂	PRAMP	Bureau Veritas	PRAMP	PRAMP

Monitoring Notes during the Month of December 2023

- **NMHC Canister Sampling Program - Volatile Organic Compounds (VOCs)**
 - The canister sampling program collects a 1-hour sample of air when the continuously measured non-methane hydrocarbon (NMHC) concentration reaches a specified trigger point. The current trigger point is 0.3 ppm for non-methane hydrocarbons. The trigger point is based on real-time monitoring data that are averaged over a 5-minute period.
 - There were no valid canister events recorded this month. The canister system was triggered at the 986-C station on December 26 at 03:35 due to an analyzer issue. The event is not valid.

- **Passive Polycyclic Aromatic Compounds (PACs) Sampling Program**
 - The PAC sampling program began in December 2019, and is designed to collect a 2-month integrated sample.
 - The sample media for sampling period of November and December were installed on October 31. They were removed on December 31.
 - The sample media for sampling period of January and February 2024 were installed on December 31. They are scheduled to be removed by the end February.

- **Passives H₂S and SO₂ Sampling Program**
 - The passive sample filters were installed at the stations on December 1 and were removed on December 29.

Revisions to Alberta's Ambient Air Quality Data Warehouse

No revisions to historical data previously submitted to the Alberta's Ambient Air Quality Data Warehouse were made this month.

Deviations from Authorized Monitoring Methods

There were no deviations from authorized monitoring methods.

Certification

The report was prepared and submitted by Lily Lin in accordance with Chapter 9 of the Air Monitoring Directive (AMD 2016).

A handwritten signature in blue ink, appearing to read 'Lily Lin', with a stylized flourish at the end.

Lily Lin, Technical Program Manager, PRAMP Airshed

The report was reviewed by Mike Bisaga in accordance with Chapter 9 of the Air Monitoring Directive (AMD 2016).


I certify that I have reviewed and verified this report and that the information is complete, accurate and representative of the monitoring results, reporting timeframe and the specified analysis, summarization and reporting requirements. I also certify that at the time of this report's submission, all air data have been electronically uploaded to Alberta ETS as required by the AMD.

A handwritten signature in blue ink, appearing to read 'Michael Bisaga', with a stylized flourish at the end.

Michael Bisaga, Technical Program Manager, PRAMP Airshed


January 12, 2024

Map of PRAMP Continuous Monitoring and Integrated Sampling Network



Legend

- PRAMP Boundary
 - Populated Place
 - First Nation
 - ✱ Continuous Monitoring Station
 - ▲ Passive Monitoring Station
- ### Industrial Facilities
- In-Situ Oil Sands
 - Heavy Oil/Bitumen Well or Battery
 - Conventional Oil Well or Battery
 - Natural Gas Well or Battery
 - Gas Plant or Gas Processing
 - Compressor Station or Pipeline
 - Agricultural Storage and Transfer
 - Pulp and Paper
 - Well (Not Associated with Batteries)



INTEGRATED SAMPLING RESULTS SUMMARY

- **NMHC analytical results**

No analytical results are attached as there were canister events recorded this month.

- **Passive analytical results**

	H ₂ S		SO ₂	
Minimum (ppb)	0.10	#1	0.2	#1
Maximum (ppb)	0.23	#14	0.5	#2
Average (ppb)	0.15	-	0.31	-

ANALYTICAL SAMPLING RESULTS

NMHC Canisters – VOCs

Passives

End of Report



Peace River Area Monitoring Program

DECEMBER 2023

Ambient Air Monitoring

Certified Laboratory Analysis Report

LAB-PRAMP-202312

Operation and Maintenance:

Bureau Veritas Canada

Data Validation and Report:

Peace River Area Monitoring Program

January 6, 2024

Table of Contents

NMHC Canister Analytical Results	3
Passive Sampling Analytical Results.....	48
End of Report	55

NMHC Canister Analytical Results

Passive Sampling Analytical Results

End of Report